

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120122\
 Data File : BM037818.D
 Acq On : 01 Dec 2022 20:06
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV004

Quant Time: Dec 02 01:18:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.098	152	189911	20.000	ng/u1	0.00
20) Naphthalene-d8	10.916	136	870662	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.721	164	588244	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.462	188	1296968	20.000	ng/u1	0.00
79) Chrysene-d12	21.627	240	1371732	20.000	ng/u1	0.00
88) Perylene-d12	24.109	264	1432722	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	28992	7.162	ng/uL	0.00
4) Pyridine-d5	3.869	84	223846	17.098	ng/u1	0.00
7) Phenol-d5	7.245	99	326242	19.067	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	218133	21.852	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	262286	20.275	ng/u1	0.00
15) 4-Methylphenol-d8	8.804	113	269642	19.817	ng/u1	0.00
21) Nitrobenzene-d5	9.269	128	146546	21.189	ng/u1	0.00
24) 2-Nitrophenol-d4	9.992	143	153563	21.457	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	283029	21.324	ng/u1	0.00
31) 4-Chloroaniline-d4	11.051	131	413756	20.324	ng/u1	0.00
46) Dimethylphthalate-d6	14.133	166	858637	21.257	ng/u1	0.00
49) Acenaphthylene-d8	14.421	160	1041365	21.198	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	177136	21.127	ng/u1	0.00
60) Fluorene-d10	15.710	176	782264	21.346	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.821	200	137613	20.142	ng/u1	0.00
73) Anthracene-d10	17.562	188	1257301	21.080	ng/u1	0.00
81) Pyrene-d10	19.839	212	1463289	20.594	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.950	264	1512054	21.067	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.463	88	30183	6.958	ng/uL	96
5) Pyridine	3.887	79	214038	16.674	ng/u1	97
6) Benzaldehyde	7.228	77	209217	25.390	ng/u1	99
8) Phenol	7.275	94	320020	18.302	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.510	93	262093	18.789	ng/u1	99
12) 2-Chlorophenol	7.657	128	269292	19.340	ng/u1	98
13) 2-Methylphenol	8.540	108	255070	18.292	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.628	45	501435	19.346	ng/u1	100
16) Acetophenone	8.928	105	422968	19.707	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.910	70	216181	19.996	ng/u1	96
18) 4-Methylphenol	8.869	108	283127	18.818	ng/u1	99
19) Hexachloroethane	9.187	117	105811	19.895	ng/u1	100
22) Nitrobenzene	9.310	77	290548	20.046	ng/u1	98
23) Isophorone	9.839	82	617071	19.788	ng/u1	100
25) 2-Nitrophenol	10.028	139	160213	20.658	ng/u1	97
26) 2,4-Dimethylphenol	10.081	107	299424	20.085	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.322	93	376166	19.976	ng/u1	98
29) 2,4-Dichlorophenol	10.557	162	275392	20.513	ng/u1	95
30) Naphthalene	10.969	128	932678	19.738	ng/u1	100
32) 4-Chloroaniline	11.075	127	412508	20.023	ng/u1	98
33) Hexachlorobutadiene	11.251	225	167453	20.137	ng/u1	97
34) Caprolactam	11.857	113	95278	19.544	ng/u1	81
35) 4-Chloro-3-methylphenol	12.186	107	289859	20.690	ng/u1	98
36) 2-Methylnaphthalene	12.563	142	671282	20.534	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120122\
 Data File : BM037818.D
 Acq On : 01 Dec 2022 20:06
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SICV004


Quant Time: Dec 02 01:18:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.780	142	644077	19.685	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.927	216	336922	19.999	ng/ul	99
40) Hexachlorocyclopentadiene	12.910	237	202352	21.842	ng/ul	97
41) 2,4,6-Trichlorophenol	13.163	196	229202	20.840	ng/ul	99
42) 2,4,5-Trichlorophenol	13.233	196	245841	20.988	ng/ul	99
43) 1,1'-Biphenyl	13.563	154	885625	20.712	ng/ul	99
44) 2-Chloronaphthalene	13.610	162	670676	19.988	ng/ul	98
45) 2-Nitroaniline	13.810	65	198034	20.859	ng/ul	98
47) Dimethylphthalate	14.180	163	843964	20.060	ng/ul	99
48) 2,6-Dinitrotoluene	14.298	165	181903	21.176	ng/ul	96
50) Acenaphthylene	14.451	152	1112522	20.426	ng/ul	100
51) 3-Nitroaniline	14.627	138	215186	20.988	ng/ul	97
52) Acenaphthene	14.786	153	755063	20.006	ng/ul	99
53) 2,4-Dinitrophenol	14.827	184	91067	18.152	ng/ul	96
55) 4-Nitrophenol	14.921	109	103435	20.015	ng/ul	95
56) Dibenzofuran	15.121	168	1061516	20.890	ng/ul	98
57) 2,4-Dinitrotoluene	15.080	165	248793	20.738	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.339	232	219239	21.279	ng/ul	97
59) Diethylphthalate	15.533	149	866469	20.082	ng/ul	99
61) Fluorene	15.768	166	885822	20.707	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.757	204	422161	20.700	ng/ul	99
63) 4-Nitroaniline	15.786	138	233014	22.590	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.839	198	136493	19.380	ng/ul	98
67) N-Nitrosodiphenylamine	15.968	169	751602	20.109	ng/ul	99
68) 4-Bromophenyl-phenylether	16.651	248	263596	19.805	ng/ul	100
69) Hexachlorobenzene	16.768	284	315481	20.298	ng/ul	98
70) Atrazine	16.915	200	254773	19.793	ng/ul	99
71) Pentachlorophenol	17.110	266	192916	20.181	ng/ul	98
72) Phenanthrene	17.504	178	1406474	20.009	ng/ul	100
74) Anthracene	17.598	178	1426982	19.900	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.533	216	322900	18.465	ng/uL	99
76) Pentachlorobenzene	15.039	250	340789	17.913	ng/uL	100
77) Carbazole	17.862	167	1371480	20.783	ng/ul	100
78) Di-n-butylphthalate	18.415	149	1628277	19.626	ng/ul	99
80) Fluoranthene	19.509	202	1692188	19.497	ng/ul	97
82) Pyrene	19.868	202	1759948	19.639	ng/ul	100
83) Butylbenzylphthalate	20.750	149	746779	19.113	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.539	252	689002	23.976	ng/ul	99
85) Benzo(a)anthracene	21.609	228	1777609	20.254	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.521	149	1149686	19.473	ng/ul	99
87) Chrysene	21.668	228	1662803	20.354	ng/ul	99
89) Di-n-octyl phthalate	22.474	149	1977265	19.177	ng/ul	100
90) Benzo(b)fluoranthene	23.350	252	1838581	19.983	ng/ul	99
91) Benzo(k)fluoranthene	23.403	252	1788853	20.154	ng/ul	99
93) Benzo(a)pyrene	24.003	252	1613514	20.222	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.715	276	1906382	21.252	ng/ul	98
95) Dibenzo(a,h)anthracene	26.727	278	1688356	20.680	ng/ul	99
96) Benzo(g,h,i)perylene	27.515	276	1610865	20.578	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 02 01:18:46 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 02 01:14:59 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.958 ng/uL
RT: 3.463 min Scan# 11
Delta R.T. -0.001 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion: 88 Resp: 30183
Ion Ratio Lower Upper
88 100
43 58.3 51.3 76.9
58 100.5 79.0 118.6



#5
Pyridine
Concen: 16.674 ng/uL
RT: 3.887 min Scan# 85
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion: 79 Resp: 214038
Ion Ratio Lower Upper
79 100
52 96.8 75.2 112.8
51 44.0 33.9 50.9









#6
Benzaldehyde
Concen: 25.390 ng/ul
RT: 7.228 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion: 77 Resp: 209217
Ion Ratio Lower Upper
77 100
105 103.4 81.2 121.8
106 98.8 79.0 118.6



#8
Phenol
Concen: 18.302 ng/ul
RT: 7.275 min Scan# 661
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion: 94 Resp: 320020
Ion Ratio Lower Upper
94 100
65 26.8 21.1 31.7
66 31.6 25.2 37.8









#10
Bis(2-Chloroethyl)ether
Concen: 18.789 ng/ul
RT: 7.510 min Scan# 701
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion: 93 Resp: 262093
Ion Ratio Lower Upper
93 100
63 82.5 65.7 98.5
95 33.9 26.8 40.2



#12
2-Chlorophenol
Concen: 19.340 ng/ul
RT: 7.657 min Scan# 726
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:128 Resp: 269292
Ion Ratio Lower Upper
128 100
64 45.1 37.3 55.9
130 32.7 27.1 40.7









#13
2-Methylphenol
Concen: 18.292 ng/ul
RT: 8.540 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:108 Resp: 255070
Ion Ratio Lower Upper
108 100
107 87.8 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.346 ng/ul
RT: 8.628 min Scan# 891
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion: 45 Resp: 501435
Ion Ratio Lower Upper
45 100
77 10.2 8.2 12.4
79 8.1 6.7 10.1









#16
Acetophenone
Concen: 19.707 ng/ul
RT: 8.928 min Scan# 942
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:105 Resp: 422968
Ion Ratio Lower Upper
105 100
77 72.4 59.0 88.4
51 30.1 24.4 36.6



#17
N-Nitroso-di-n-propylamine
Concen: 19.996 ng/ul
RT: 8.910 min Scan# 939
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion: 70 Resp: 216181
Ion Ratio Lower Upper
70 100
42 73.3 55.8 83.6
101 9.2 8.0 12.0
130 24.0 18.6 27.8









#18
4-Methylphenol
Concen: 18.818 ng/ul
RT: 8.869 min Scan# 911
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:108 Resp: 283127
Ion Ratio Lower Upper
108 100
107 110.7 89.3 133.9



#19
Hexachloroethane
Concen: 19.895 ng/ul
RT: 9.187 min Scan# 986
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:117 Resp: 105811
Ion Ratio Lower Upper
117 100
201 102.2 82.3 123.5
199 62.1 49.8 74.6









#22
Nitrobenzene
Concen: 20.046 ng/ul
RT: 9.310 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion: 77 Resp: 290548
Ion Ratio Lower Upper
77 100
123 50.0 41.4 62.0
65 13.5 10.9 16.3



#23
Isophorone
Concen: 19.788 ng/ul
RT: 9.839 min Scan# 1097
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion: 82 Resp: 617071
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 19.0 15.3 22.9









#25
2-Nitrophenol
Concen: 20.658 ng/ul
RT: 10.028 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:139 Resp: 160213
Ion Ratio Lower Upper
139 100
65 44.5 36.6 54.8
109 23.3 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 20.085 ng/ul
RT: 10.081 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:107 Resp: 299424
Ion Ratio Lower Upper
107 100
121 53.2 41.8 62.8
122 93.1 73.1 109.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.976 ng/ul
 RT: 10.322 min Scan# 1179
 Delta R.T. -0.000 min
 Lab File: BM037818.D
 Acq: 01 Dec 2022 20:06




Instrument :
 BNA_M
 ClientSampleId :
 SICV004


Tgt Ion: 93	Resp: 376166
Ion Ratio	Lower Upper
93	100
95	33.4 27.9 41.9
123	12.3 9.7 14.5



#29
 2,4-Dichlorophenol
 Concen: 20.513 ng/ul
 RT: 10.557 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BM037818.D
 Acq: 01 Dec 2022 20:06

Tgt Ion: 162	Resp: 275392
Ion Ratio	Lower Upper
162	100
164	67.4 50.7 76.1
98	36.0 27.3 40.9









#30
Naphthalene
Concen: 19.738 ng/ul
RT: 10.969 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:128 Resp: 932678
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 20.023 ng/ul
RT: 11.075 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:127 Resp: 412508
Ion Ratio Lower Upper
127 100
129 32.5 26.9 40.3









#33
Hexachlorobutadiene
Concen: 20.137 ng/ul
RT: 11.251 min Scan# 1337
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:225 Resp: 167453
Ion Ratio Lower Upper
225 100
223 62.5 49.3 73.9
227 65.4 49.4 74.0



#34
Caprolactam
Concen: 19.544 ng/ul
RT: 11.857 min Scan# 1440
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:113 Resp: 95278
Ion Ratio Lower Upper
113 100
55 195.0 186.2 279.2
56 133.9 119.8 179.8









#35
4-Chloro-3-methylphenol
Concen: 20.690 ng/ul
RT: 12.186 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:107 Resp: 289859
Ion Ratio Lower Upper
107 100
144 27.1 23.0 34.6
142 84.8 69.1 103.7



#36
2-Methylnaphthalene
Concen: 20.534 ng/ul
RT: 12.563 min Scan# 1560
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:142 Resp: 671282
Ion Ratio Lower Upper
142 100
141 87.1 70.1 105.1





#37

1-Methylnaphthalene

Concen: 19.685 ng/ul

RT: 12.780 min Scan# 1598

Delta R.T. -0.006 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Instrument :

BNA_M

ClientSampleId :

SICV004





Tgt Ion:142 Resp: 644077

Ion Ratio Lower Upper

142 100

141 91.9 73.4 110.2

116 3.4 2.6 4.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.999 ng/ul

RT: 12.927 min Scan# 1622

Delta R.T. -0.000 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Tgt Ion:216 Resp: 336922




Ion Ratio Lower Upper


216 100

214 77.7 62.7 94.1

179 19.4 15.9 23.9

108 15.7 13.0 19.4









#40
Hexachlorocyclopentadiene
Concen: 21.842 ng/ul
RT: 12.910 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:237 Resp: 202352
Ion Ratio Lower Upper
237 100
235 63.6 49.2 73.8
272 14.9 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 20.840 ng/ul
RT: 13.163 min Scan# 1662
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:196 Resp: 229202
Ion Ratio Lower Upper
196 100
198 93.8 75.6 113.4
200 30.7 24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 20.988 ng/ul
RT: 13.233 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:196 Resp: 245841
Ion Ratio Lower Upper
196 100
198 96.5 76.1 114.1
200 30.2 24.8 37.2



#43
1,1'-Biphenyl
Concen: 20.712 ng/ul
RT: 13.563 min Scan# 1730
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:154 Resp: 885625
Ion Ratio Lower Upper
154 100
153 40.2 31.4 47.0
76 13.0 10.2 15.2









#44
2-Chloronaphthalene
Concen: 19.988 ng/ul
RT: 13.610 min Scan# 1738
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:162 Resp: 670676
Ion Ratio Lower Upper
162 100
164 32.2 26.5 39.7
127 37.1 30.5 45.7



#45
2-Nitroaniline
Concen: 20.859 ng/ul
RT: 13.810 min Scan# 1772
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion: 65 Resp: 198034
Ion Ratio Lower Upper
65 100
92 71.5 54.7 82.1
138 122.8 96.8 145.2









#47
Dimethylphthalate
Concen: 20.060 ng/ul
RT: 14.180 min Scan# 1835
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:163 Resp: 843964
Ion Ratio Lower Upper
163 100
194 4.4 3.6 5.4
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 21.176 ng/ul
RT: 14.298 min Scan# 1855
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:165 Resp: 181903
Ion Ratio Lower Upper
165 100
89 47.1 40.6 60.8
121 20.2 16.4 24.6









#50
Acenaphthylene
Concen: 20.426 ng/ul
RT: 14.451 min Scan# 1881
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:152 Resp: 1112522
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 13.2 10.6 15.8



#51
3-Nitroaniline
Concen: 20.988 ng/ul
RT: 14.627 min Scan# 1911
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:138 Resp: 215186
Ion Ratio Lower Upper
138 100
108 9.5 7.8 11.8
92 101.0 83.8 125.6









#52
Acenaphthene
Concen: 20.006 ng/ul
RT: 14.786 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:153 Resp: 755063
Ion Ratio Lower Upper
153 100
152 46.6 37.0 55.4
154 89.0 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 18.152 ng/ul
RT: 14.827 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:184 Resp: 91067
Ion Ratio Lower Upper
184 100
63 56.7 43.4 65.0
154 53.1 39.7 59.5









#55
4-Nitrophenol
Concen: 20.015 ng/ul
RT: 14.921 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:109 Resp: 103435
Ion Ratio Lower Upper
109 100
139 167.2 131.6 197.4
65 147.8 109.4 164.0



#56
Dibenzofuran
Concen: 20.890 ng/ul
RT: 15.121 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:168 Resp: 1061516
Ion Ratio Lower Upper
168 100
139 33.5 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 20.738 ng/ul

RT: 15.080 min Scan# 1988

Delta R.T. -0.000 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Instrument :

BNA_M

ClientSampleId :

SICV004





Tgt Ion:165 Resp: 248793

Ion Ratio Lower Upper

165 100

63 39.6 30.2 45.2

182 2.5 1.8 2.8



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.279 ng/ul

RT: 15.339 min Scan# 2032

Delta R.T. -0.006 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Tgt Ion:232 Resp: 219239




Ion Ratio Lower Upper


232 100

131 42.6 32.7 49.1

130 2.3 1.8 2.6

166 30.0 22.2 33.4









#59
Diethylphthalate
Concen: 20.082 ng/ul
RT: 15.533 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:149 Resp: 866469
Ion Ratio Lower Upper
149 100
177 22.2 18.0 27.0
150 12.1 9.8 14.8



#61
Fluorene
Concen: 20.707 ng/ul
RT: 15.768 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:166 Resp: 885822
Ion Ratio Lower Upper
166 100
165 98.5 77.0 115.6
167 13.1 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.700 ng/ul
RT: 15.757 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:204 Resp: 422161
Ion Ratio Lower Upper
204 100
206 32.8 26.2 39.4
141 54.6 44.7 67.1



#63
4-Nitroaniline
Concen: 22.590 ng/ul
RT: 15.786 min Scan# 2108
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:138 Resp: 233014
Ion Ratio Lower Upper
138 100
92 45.0 36.6 54.8
108 61.6 48.2 72.2









#66
4,6-Dinitro-2-methylphenol
Concen: 19.380 ng/ul
RT: 15.839 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:198 Resp: 136493
Ion Ratio Lower Upper
198 100
182 3.6 2.9 4.3
77 20.6 17.5 26.3



#67
N-Nitrosodiphenylamine
Concen: 20.109 ng/ul
RT: 15.968 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:169 Resp: 751602
Ion Ratio Lower Upper
169 100
168 66.8 53.8 80.6
167 36.5 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 19.805 ng/ul
RT: 16.651 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:248 Resp: 263596
Ion Ratio Lower Upper
248 100
250 98.0 78.1 117.1
141 64.1 51.6 77.4



#69
Hexachlorobenzene
Concen: 20.298 ng/ul
RT: 16.768 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:284 Resp: 315481
Ion Ratio Lower Upper
284 100
142 29.1 24.5 36.7
249 28.7 22.9 34.3









#70
Atrazine
Concen: 19.793 ng/ul
RT: 16.915 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:200 Resp: 254773
Ion Ratio Lower Upper
200 100
173 26.7 22.0 33.0
215 49.4 40.0 60.0



#71
Pentachlorophenol
Concen: 20.181 ng/ul
RT: 17.110 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:266 Resp: 192916
Ion Ratio Lower Upper
266 100
264 60.4 50.3 75.5
268 65.3 52.2 78.2





#72

Phenanthrene

Concen: 20.009 ng/ul

RT: 17.504 min Scan# 2400

Delta R.T. -0.000 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Instrument :

BNA_M

ClientSampleId :

SICV004





Tgt Ion:178 Resp: 1406474

Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.9 15.3 22.9



#74

Anthracene

Concen: 19.900 ng/ul

RT: 17.598 min Scan# 2416

Delta R.T. -0.000 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06




Tgt Ion:178 Resp: 1426982


Ion Ratio Lower Upper

178 100

179 15.7 12.2 18.4

176 18.8 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.465 ng/uL
RT: 13.533 min Scan# 1725
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:216 Resp: 322900
Ion Ratio Lower Upper
216 100
214 78.1 63.0 94.6
218 49.1 40.0 60.0



#76
Pentachlorobenzene
Concen: 17.913 ng/uL
RT: 15.039 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:250 Resp: 340789
Ion Ratio Lower Upper
250 100
248 62.6 50.5 75.7
252 63.8 51.1 76.7









#77
 Carbazole
 Concen: 20.783 ng/ul
 RT: 17.862 min Scan# 2461
 Delta R.T. -0.000 min
 Lab File: BM037818.D
 Acq: 01 Dec 2022 20:06




Instrument :
 BNA_M
 ClientSampleId :
 SICV004


Tgt Ion:167 Resp: 1371480
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.2 25.8
 139 12.2 10.0 15.0



#78
 Di-n-butylphthalate
 Concen: 19.626 ng/ul
 RT: 18.415 min Scan# 2555
 Delta R.T. -0.000 min
 Lab File: BM037818.D
 Acq: 01 Dec 2022 20:06

Tgt Ion:149 Resp: 1628277
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.2
 104 4.2 3.4 5.0





#80

Fluoranthene

Concen: 19.497 ng/ul

RT: 19.509 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Instrument :

BNA_M

ClientSampleId :

SICV004





Tgt Ion:202 Resp: 1692188

Ion Ratio Lower Upper

202 100

101 10.8 9.7 14.5

100 8.7 7.7 11.5



#82

Pyrene

Concen: 19.639 ng/ul

RT: 19.868 min Scan# 2802

Delta R.T. -0.006 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06




Tgt Ion:202 Resp: 1759948


Ion Ratio Lower Upper

202 100

101 13.8 11.0 16.4

100 11.5 9.0 13.6









#83
Butylbenzylphthalate
Concen: 19.113 ng/ul
RT: 20.750 min Scan# 2953
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:149 Resp: 746779
Ion Ratio Lower Upper
149 100
91 65.9 51.7 77.5
206 20.5 17.1 25.7



#84
3,3'-Dichlorobenzidine
Concen: 23.976 ng/ul
RT: 21.539 min Scan# 3086
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:252 Resp: 689002
Ion Ratio Lower Upper
252 100
254 64.3 52.2 78.4
126 11.1 8.8 13.2









#85
Benzo(a)anthracene
Concen: 20.254 ng/ul
RT: 21.609 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:228 Resp: 1777609
Ion Ratio Lower Upper
228 100
229 19.8 16.2 24.2
226 26.6 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.473 ng/ul
RT: 21.521 min Scan# 3083
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:149 Resp: 1149686
Ion Ratio Lower Upper
149 100
167 28.4 23.3 34.9
279 4.2 3.6 5.4





#87

Chrysene

Concen: 20.354 ng/ul

RT: 21.668 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Instrument :

BNA_M

ClientSampleId :

SICV004





Tgt Ion:228 Resp: 1662803

Ion Ratio Lower Upper

228 100

226 30.0 24.2 36.2

229 19.2 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 19.177 ng/ul




RT: 22.474 min Scan# 3245


Delta R.T. -0.006 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Tgt Ion:149 Resp: 1977265









#90
Benzo(b)fluoranthene
Concen: 19.983 ng/ul
RT: 23.350 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06




Instrument :
BNA_M
ClientSampleId :
SICV004


Tgt Ion:252 Resp: 1838581
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 11.0 9.3 13.9



#91
Benzo(k)fluoranthene
Concen: 20.154 ng/ul
RT: 23.403 min Scan# 3403
Delta R.T. -0.006 min
Lab File: BM037818.D
Acq: 01 Dec 2022 20:06

Tgt Ion:252 Resp: 1788853
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.8
125 10.6 9.3 13.9





#93

Benzo(a)pyrene

Concen: 20.222 ng/ul

RT: 24.003 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06

Instrument :

BNA_M

ClientSampleId :

SICV004





Tgt Ion:252 Resp: 1613514

Ion Ratio Lower Upper

252 100

253 21.4 17.4 26.2

125 11.8 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.252 ng/ul

RT: 26.715 min Scan# 3966

Delta R.T. -0.006 min

Lab File: BM037818.D

Acq: 01 Dec 2022 20:06




Tgt Ion:276 Resp: 1906382


Ion Ratio Lower Upper

276 100

138 27.5 22.7 34.1

277 24.2 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 20.680 ng/ul
 RT: 26.727 min Scan# 3970
 Delta R.T. -0.012 min
 Lab File: BM037818.D
 Acq: 01 Dec 2022 20:06

Instrument :
 BNA_M
 ClientSampleId :
 SICV004

Tgt Ion:278 Resp: 1688356
 Ion Ratio Lower Upper
 278 100
 139 16.9 13.9 20.9
 279 24.3 19.0 28.6



#96
 Benzo(g,h,i)perylene
 Concen: 20.578 ng/ul
 RT: 27.515 min Scan# 4102
 Delta R.T. -0.012 min
 Lab File: BM037818.D
 Acq: 01 Dec 2022 20:06

Tgt Ion:276 Resp: 1610865
 Ion Ratio Lower Upper
 276 100
 138 23.5 19.2 28.8
 277 23.6 19.1 28.7

