

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120215\
 Data File : BM003363.D
 Acq On : 02 Dec 2015 19:22
 Operator : UM/IZ
 Sample : G3707-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 LOQ-WATER2015-02

Manual Integrations
 APPROVED

Mmdadoda
 12/3/2015 7:07:57 PM

Quant Time: Dec 03 00:39:55 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 16:19:00 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	53743	5.00	ng	0.00
7) Naphthalene-d8	10.51	136	224549	5.00	ng	0.00
13) Acenaphthene-d10	14.37	164	126541	5.00	ng	0.00
19) Phenanthrene-d10	17.11	188	296051	5.00	ng	0.00
26) Chrysene-d12	21.31	240	158562	5.00	ng	-0.02
35) Perylene-d12	23.58	264	130795	5.00	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.31	112	74326	6.93	ng	0.00
5) Phenol-d6	6.88	99	94439	7.16	ng	0.00
8) Nitrobenzene-d5	8.88	82	35933	3.43	ng	0.00
14) 2,4,6-Tribromophenol	15.86	330	17196	5.30	ng	0.00
15) 2-Fluorobiphenyl	13.00	172	141996	3.73	ng	0.00
29) Terphenyl-d14	19.76	244	106531	4.14	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.23	88	863	0.18	ng	# 84
3) n-Nitrosodimethylamine	3.54	42	596	0.12	ng	# 98
6) bis(2-Chloroethyl)ether	7.16	93	1800	0.14	ng	98
9) Nitrobenzene	8.92	77	1585m	0.14	ng	
10) Naphthalene	10.56	128	7128	0.15	ng	98
11) Hexachlorobutadiene	10.85	225	1057m	0.14	ng	
12) 2-Methylnaphthalene	12.18	142	4595	0.14	ng	99
16) Acenaphthylene	14.08	152	5589	0.11	ng	99
17) Acenaphthene	14.44	154	4939	0.15	ng	# 100
18) Fluorene	15.42	166	5067	0.13	ng	# 95
20) 4-Bromophenyl-phenylether	16.32	248	1248	0.15	ng	# 37
21) Hexachlorobenzene	16.42	284	1460	0.12	ng	# 62
22) Pentachlorophenol	16.78	266	166	0.36	ng	# 83
23) Phenanthrene	17.16	178	8205m	0.14	ng	
24) Anthracene	17.24	178	7351	0.12	ng	97
25) Fluoranthene	19.19	202	7718	0.13	ng	99
27) Benzidine	19.38	184	1260	0.43	ng	# 81
28) Pyrene	19.55	202	8110	0.14	ng	99
30) Benzo(a)anthracene	21.29	228	6224m	0.15	ng	
31) 3,3'-Dichlorobenzidine	21.25	252	1407	0.20	ng	# 93
32) Chrysene	21.35	228	7735m	0.16	ng	
33) Bis(2-ethylhexyl)phthalate	21.23	149	2683	0.34	ng	# 95
34) Indeno(1,2,3-cd)pyrene	25.85	276	5198	0.14	ng	# 80
36) Benzo(b)fluoranthene	22.89	252	5370	0.13	ng	# 90
37) Benzo(k)fluoranthene	22.94	252	4695	0.13	ng	# 87
38) Benzo(a)pyrene	23.47	252	4351	0.13	ng	# 84
39) Dibenzo(a,h)anthracene	25.87	278	4197	0.14	ng	# 88
40) Benzo(g,h,i)perylene	26.55	276	4913	0.16	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120215\
Data File : BM003363.D
Acq On : 02 Dec 2015 19:22
Operator : UM/IZ
Sample : G3707-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02

Manual Integrations
APPROVED

Mmdadoda
12/3/2015 7:07:57 PM

Quant Time: Dec 03 00:39:55 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 02 16:19:00 2015
Response via : Initial Calibration

