

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120222\
 Data File : BM037820.D
 Acq On : 02 Dec 2022 09:41
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020060

Quant Time: Dec 02 21:45:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.092	152	144381	20.000	ng/u1	0.00
20) Naphthalene-d8	10.916	136	701299	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.721	164	490566	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.456	188	1110367	20.000	ng/u1	0.00
79) Chrysene-d12	21.627	240	1235162	20.000	ng/u1	0.00
88) Perylene-d12	24.109	264	1275847	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	24948	8.107	ng/uL	0.00
4) Pyridine-d5	3.863	84	198746	19.968	ng/u1	0.00
7) Phenol-d5	7.245	99	257085	19.764	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	161550	21.287	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	204520	20.795	ng/u1	0.00
15) 4-Methylphenol-d8	8.804	113	206638	19.976	ng/u1	0.00
21) Nitrobenzene-d5	9.269	128	108340	19.448	ng/u1	0.00
24) 2-Nitrophenol-d4	9.992	143	107061	18.572	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.528	165	215750	20.181	ng/u1	0.00
31) 4-Chloroaniline-d4	11.051	131	327273	19.958	ng/u1	0.00
46) Dimethylphthalate-d6	14.127	166	680302	20.195	ng/u1	0.00
49) Acenaphthylene-d8	14.415	160	826092	20.164	ng/u1	0.00
54) 4-Nitrophenol-d4	14.904	143	134043	19.171	ng/u1	0.00
60) Fluorene-d10	15.710	176	647920	21.201	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.821	200	103680	17.726	ng/u1	0.00
73) Anthracene-d10	17.556	188	1059773	20.754	ng/u1	0.00
81) Pyrene-d10	19.839	212	1271105	19.867	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.950	264	1315188	20.577	ng/u1	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.463	88	26084	7.909	ng/uL#	90
5) Pyridine	3.881	79	198014	20.290	ng/u1	91
6) Benzaldehyde	7.228	77	139109	22.205	ng/u1	99
8) Phenol	7.269	94	265771	19.993	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.510	93	218948	20.646	ng/u1	95
12) 2-Chlorophenol	7.651	128	221198	20.896	ng/u1	98
13) 2-Methylphenol	8.534	108	210718	19.876	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.628	45	455431	23.112	ng/u1	97
16) Acetophenone	8.928	105	341182	20.909	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.910	70	175513	21.354	ng/u1	94
18) 4-Methylphenol	8.863	108	231239	20.216	ng/u1	98
19) Hexachloroethane	9.186	117	86489	21.390	ng/u1	94
22) Nitrobenzene	9.310	77	240822	20.627	ng/u1	98
23) Isophorone	9.839	82	498626	19.851	ng/u1	99
25) 2-Nitrophenol	10.022	139	122176	19.558	ng/u1	95
26) 2,4-Dimethylphenol	10.080	107	246671	20.542	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.316	93	305028	20.110	ng/u1	98
29) 2,4-Dichlorophenol	10.557	162	224074	20.722	ng/u1	98
30) Naphthalene	10.969	128	787891	20.701	ng/u1	99
32) 4-Chloroaniline	11.075	127	337109	20.315	ng/u1	97
33) Hexachlorobutadiene	11.251	225	144026	21.502	ng/u1	98
34) Caprolactam	11.851	113	70639	17.989	ng/u1	91
35) 4-Chloro-3-methylphenol	12.180	107	234603	20.789	ng/u1	97
36) 2-Methylnaphthalene	12.563	142	549371	20.863	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120222\
 Data File : BM037820.D
 Acq On : 02 Dec 2022 09:41
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020060


Quant Time: Dec 02 21:45:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.780	142	554770	21.050	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.927	216	290336	20.665	ng/ul	99
40) Hexachlorocyclopentadiene	12.904	237	155408	20.115	ng/ul	100
41) 2,4,6-Trichlorophenol	13.163	196	184843	20.153	ng/ul	98
42) 2,4,5-Trichlorophenol	13.233	196	202695	20.750	ng/ul	99
43) 1,1'-Biphenyl	13.563	154	730714	20.492	ng/ul	99
44) 2-Chloronaphthalene	13.604	162	574767	20.540	ng/ul	99
45) 2-Nitroaniline	13.804	65	160589	20.283	ng/ul	93
47) Dimethylphthalate	14.174	163	712322	20.302	ng/ul	99
48) 2,6-Dinitrotoluene	14.298	165	136736	19.087	ng/ul	98
50) Acenaphthylene	14.445	152	932042	20.519	ng/ul	99
51) 3-Nitroaniline	14.621	138	152406	17.824	ng/ul	98
52) Acenaphthene	14.786	153	654914	20.808	ng/ul	99
53) 2,4-Dinitrophenol	14.821	184	65255	15.597	ng/ul	92
55) 4-Nitrophenol	14.921	109	90795	21.067	ng/ul	95
56) Dibenzofuran	15.115	168	900681	21.254	ng/ul	99
57) 2,4-Dinitrotoluene	15.074	165	205494	20.540	ng/ul	88
58) 2,3,4,6-Tetrachlorophenol	15.339	232	179432	20.883	ng/ul	98
59) Diethylphthalate	15.527	149	735688	20.446	ng/ul	100
61) Fluorene	15.762	166	773187	21.673	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.757	204	373536	21.963	ng/ul	97
63) 4-Nitroaniline	15.780	138	169371	19.690	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.833	198	112693	18.690	ng/ul	100
67) N-Nitrosodiphenylamine	15.968	169	648958	20.281	ng/ul	98
68) 4-Bromophenyl-phenylether	16.645	248	225048	19.751	ng/ul	93
69) Hexachlorobenzene	16.762	284	279598	21.013	ng/ul	99
70) Atrazine	16.909	200	215994	19.601	ng/ul	95
71) Pentachlorophenol	17.104	266	156919	19.174	ng/ul	97
72) Phenanthrene	17.504	178	1249566	20.764	ng/ul	99
74) Anthracene	17.592	178	1291365	21.035	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.527	216	296545	19.808	ng/uL	98
76) Pentachlorobenzene	15.039	250	333609	20.482	ng/uL	98
77) Carbazole	17.862	167	1191470	21.090	ng/ul	100
78) Di-n-butylphthalate	18.415	149	1378604	19.409	ng/ul	99
80) Fluoranthene	19.503	202	1523685	19.497	ng/ul	98
82) Pyrene	19.868	202	1620982	20.089	ng/ul	99
83) Butylbenzylphthalate	20.750	149	633578	18.009	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.539	252	546458	21.119	ng/ul	100
85) Benzo(a)anthracene	21.609	228	1637251	20.718	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.521	149	983850	18.507	ng/ul	99
87) Chrysene	21.668	228	1541961	20.961	ng/ul	100
89) Di-n-octyl phthalate	22.474	149	1758967	19.157	ng/ul	100
90) Benzo(b)fluoranthene	23.350	252	1721241	21.007	ng/ul	98
91) Benzo(k)fluoranthene	23.397	252	1640362	20.753	ng/ul	98
93) Benzo(a)pyrene	23.997	252	1483403	20.878	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.709	276	1698323	21.261	ng/ul	98
95) Dibenzo(a,h)anthracene	26.721	278	1549851	21.317	ng/ul	99
96) Benzo(g,h,i)perylene	27.509	276	1495021	21.446	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 02 21:45:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 02 01:14:59 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.909 ng/uL
RT: 3.463 min Scan# 11
Delta R.T. -0.001 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion: 88 Resp: 26084
Ion Ratio Lower Upper
88 100
43 46.2 51.3 76.9#
58 100.6 79.0 118.6



#5
Pyridine
Concen: 20.290 ng/uL
RT: 3.881 min Scan# 84
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion: 79 Resp: 198014
Ion Ratio Lower Upper
79 100
52 104.1 75.2 112.8
51 46.7 33.9 50.9









#6
Benzaldehyde
Concen: 22.205 ng/ul
RT: 7.228 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion: 77 Resp: 139109
Ion Ratio Lower Upper
77 100
105 101.8 81.2 121.8
106 96.1 79.0 118.6



#8
Phenol
Concen: 19.993 ng/ul
RT: 7.269 min Scan# 660
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion: 94 Resp: 265771
Ion Ratio Lower Upper
94 100
65 27.5 21.1 31.7
66 33.5 25.2 37.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.646 ng/ul
RT: 7.510 min Scan# 701
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion: 93 Resp: 218948
Ion Ratio Lower Upper
93 100
63 87.8 65.7 98.5
95 33.4 26.8 40.2



#12
2-Chlorophenol
Concen: 20.896 ng/ul
RT: 7.651 min Scan# 725
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:128 Resp: 221198
Ion Ratio Lower Upper
128 100
64 48.1 37.3 55.9
130 33.3 27.1 40.7





#13

2-Methylphenol

Concen: 19.876 ng/ul

RT: 8.534 min Scan# 875

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :





SSTD020060

Tgt Ion:108 Resp: 210718

Ion Ratio Lower Upper

108 100

107 87.9 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 23.112 ng/ul

RT: 8.628 min Scan# 891

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41




Tgt Ion: 45 Resp: 455431


Ion Ratio Lower Upper

45 100

77 9.3 8.2 12.4

79 7.5 6.7 10.1









#16
 Acetophenone
 Concen: 20.909 ng/ul
 RT: 8.928 min Scan# 942
 Delta R.T. -0.000 min
 Lab File: BM037820.D
 Acq: 02 Dec 2022 09:41




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020060


Tgt Ion:105 Resp: 341182
 Ion Ratio Lower Upper
 105 100
 77 74.1 59.0 88.4
 51 33.8 24.4 36.6



#17
 N-Nitroso-di-n-propylamine
 Concen: 21.354 ng/ul
 RT: 8.910 min Scan# 939
 Delta R.T. -0.006 min
 Lab File: BM037820.D
 Acq: 02 Dec 2022 09:41

Tgt Ion: 70 Resp: 175513
 Ion Ratio Lower Upper
 70 100
 42 75.9 55.8 83.6
 101 9.6 8.0 12.0
 130 21.2 18.6 27.8









#18
4-Methylphenol
Concen: 20.216 ng/ul
RT: 8.863 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:108 Resp: 231239
Ion Ratio Lower Upper
108 100
107 113.9 89.3 133.9



#19
Hexachloroethane
Concen: 21.390 ng/ul
RT: 9.186 min Scan# 986
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:117 Resp: 86489
Ion Ratio Lower Upper
117 100
201 108.1 82.3 123.5
199 67.5 49.8 74.6









#22
Nitrobenzene
Concen: 20.627 ng/ul
RT: 9.310 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion: 77 Resp: 240822
Ion Ratio Lower Upper
77 100
123 49.8 41.4 62.0
65 13.5 10.9 16.3



#23
Isophorone
Concen: 19.851 ng/ul
RT: 9.839 min Scan# 1097
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion: 82 Resp: 498626
Ion Ratio Lower Upper
82 100
95 6.4 5.5 8.3
138 18.9 15.3 22.9









#25
2-Nitrophenol
Concen: 19.558 ng/ul
RT: 10.022 min Scan# 1128
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:139 Resp: 122176
Ion Ratio Lower Upper
139 100
65 49.4 36.6 54.8
109 24.8 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 20.542 ng/ul
RT: 10.080 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:107 Resp: 246671
Ion Ratio Lower Upper
107 100
121 54.2 41.8 62.8
122 92.3 73.1 109.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.110 ng/ul
 RT: 10.316 min Scan# 1178
 Delta R.T. -0.006 min
 Lab File: BM037820.D
 Acq: 02 Dec 2022 09:41




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020060


Tgt Ion: 93 Resp: 305028
 Ion Ratio Lower Upper
 93 100
 95 33.4 27.9 41.9
 123 11.4 9.7 14.5



#29
 2,4-Dichlorophenol
 Concen: 20.722 ng/ul
 RT: 10.557 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BM037820.D
 Acq: 02 Dec 2022 09:41

Tgt Ion: 162 Resp: 224074
 Ion Ratio Lower Upper
 162 100
 164 64.8 50.7 76.1
 98 32.9 27.3 40.9









#30
Naphthalene
Concen: 20.701 ng/ul
RT: 10.969 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:128 Resp: 787891
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 12.7 10.4 15.6



#32
4-Chloroaniline
Concen: 20.315 ng/ul
RT: 11.075 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:127 Resp: 337109
Ion Ratio Lower Upper
127 100
129 31.9 26.9 40.3









#33
Hexachlorobutadiene
Concen: 21.502 ng/ul
RT: 11.251 min Scan# 1337
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:225 Resp: 144026
Ion Ratio Lower Upper
225 100
223 63.4 49.3 73.9
227 62.6 49.4 74.0



#34
Caprolactam
Concen: 17.989 ng/ul
RT: 11.851 min Scan# 1439
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:113 Resp: 70639
Ion Ratio Lower Upper
113 100
55 210.0 186.2 279.2
56 146.9 119.8 179.8









#35
4-Chloro-3-methylphenol
Concen: 20.789 ng/ul
RT: 12.180 min Scan# 1495
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:107 Resp: 234603
Ion Ratio Lower Upper
107 100
144 26.9 23.0 34.6
142 84.0 69.1 103.7



#36
2-Methylnaphthalene
Concen: 20.863 ng/ul
RT: 12.563 min Scan# 1560
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:142 Resp: 549371
Ion Ratio Lower Upper
142 100
141 89.4 70.1 105.1





#37

1-Methylnaphthalene

Concen: 21.050 ng/ul

RT: 12.780 min Scan# 1597

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :

SSTD020060





Tgt Ion:142 Resp: 554770

Ion Ratio Lower Upper

142 100

141 91.7 73.4 110.2

116 3.4 2.6 4.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.665 ng/ul

RT: 12.927 min Scan# 1622

Delta R.T. -0.000 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Tgt Ion:216 Resp: 290336




Ion Ratio Lower Upper


216 100

214 77.0 62.7 94.1

179 19.5 15.9 23.9

108 15.7 13.0 19.4









#40
Hexachlorocyclopentadiene
Concen: 20.115 ng/ul
RT: 12.904 min Scan# 1618
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:237 Resp: 155408
Ion Ratio Lower Upper
237 100
235 61.1 49.2 73.8
272 13.1 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 20.153 ng/ul
RT: 13.163 min Scan# 1662
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:196 Resp: 184843
Ion Ratio Lower Upper
196 100
198 97.0 75.6 113.4
200 31.5 24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 20.750 ng/ul
RT: 13.233 min Scan# 1
Delta R.T. -0.001 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:196 Resp: 202695
Ion Ratio Lower Upper
196 100
198 96.4 76.1 114.1
200 30.9 24.8 37.2



#43
1,1'-Biphenyl
Concen: 20.492 ng/ul
RT: 13.563 min Scan# 1730
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:154 Resp: 730714
Ion Ratio Lower Upper
154 100
153 38.7 31.4 47.0
76 11.7 10.2 15.2









#44
2-Chloronaphthalene
Concen: 20.540 ng/ul
RT: 13.604 min Scan# 1737
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:162 Resp: 574767
Ion Ratio Lower Upper
162 100
164 32.5 26.5 39.7
127 37.2 30.5 45.7



#45
2-Nitroaniline
Concen: 20.283 ng/ul
RT: 13.804 min Scan# 1771
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion: 65 Resp: 160589
Ion Ratio Lower Upper
65 100
92 68.8 54.7 82.1
138 108.7 96.8 145.2





#47

Dimethylphthalate

Concen: 20.302 ng/ul

RT: 14.174 min Scan# 1834

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :

SSTD020060





Tgt Ion:163 Resp: 712322

Ion Ratio Lower Upper

163 100

194 3.9 3.6 5.4

164 10.1 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 19.087 ng/ul

RT: 14.298 min Scan# 1855

Delta R.T. -0.000 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41




Tgt Ion:165 Resp: 136736


Ion Ratio Lower Upper

165 100

89 51.3 40.6 60.8

121 22.2 16.4 24.6









#50
Acenaphthylene
Concen: 20.519 ng/ul
RT: 14.445 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:152 Resp: 932042
Ion Ratio Lower Upper
152 100
151 19.0 15.8 23.6
153 13.2 10.6 15.8



#51
3-Nitroaniline
Concen: 17.824 ng/ul
RT: 14.621 min Scan# 1910
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:138 Resp: 152406
Ion Ratio Lower Upper
138 100
108 10.2 7.8 11.8
92 107.0 83.8 125.6









#52
Acenaphthene
Concen: 20.808 ng/ul
RT: 14.786 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:153 Resp: 654914
Ion Ratio Lower Upper
153 100
152 47.9 37.0 55.4
154 87.5 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 15.597 ng/ul
RT: 14.821 min Scan# 1944
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:184 Resp: 65255
Ion Ratio Lower Upper
184 100
63 60.9 43.4 65.0
154 54.4 39.7 59.5









#55
4-Nitrophenol
Concen: 21.067 ng/ul
RT: 14.921 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:109 Resp: 90795
Ion Ratio Lower Upper
109 100
139 165.2 131.6 197.4
65 150.3 109.4 164.0



#56
Dibenzofuran
Concen: 21.254 ng/ul
RT: 15.115 min Scan# 1994
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:168 Resp: 900681
Ion Ratio Lower Upper
168 100
139 35.3 27.6 41.4









#57
2,4-Dinitrotoluene
Concen: 20.540 ng/ul
RT: 15.074 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:165 Resp: 205494
Ion Ratio Lower Upper
165 100
63 45.0 30.2 45.2
182 2.7 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.883 ng/ul
RT: 15.339 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:232 Resp: 179432
Ion Ratio Lower Upper
232 100
131 42.3 32.7 49.1
130 2.0 1.8 2.6
166 29.2 22.2 33.4









#59
Diethylphthalate
Concen: 20.446 ng/ul
RT: 15.527 min Scan# 2064
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:149 Resp: 735688
Ion Ratio Lower Upper
149 100
177 22.9 18.0 27.0
150 12.3 9.8 14.8



#61
Fluorene
Concen: 21.673 ng/ul
RT: 15.762 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:166 Resp: 773187
Ion Ratio Lower Upper
166 100
165 98.1 77.0 115.6
167 13.9 10.5 15.7





#62

4-Chlorophenyl-phenylether

Concen: 21.963 ng/ul

RT: 15.757 min Scan# 2103

Delta R.T. -0.000 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :

SSTD020060





Tgt Ion:204 Resp: 373536

Ion Ratio Lower Upper

204 100

206 32.6 26.2 39.4

141 52.8 44.7 67.1



#63

4-Nitroaniline

Concen: 19.690 ng/ul

RT: 15.780 min Scan# 2107

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41




Tgt Ion:138 Resp: 169371


Ion Ratio Lower Upper

138 100

92 44.6 36.6 54.8

108 61.0 48.2 72.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.690 ng/ul
 RT: 15.833 min Scan# 2116
 Delta R.T. -0.006 min
 Lab File: BM037820.D
 Acq: 02 Dec 2022 09:41




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020060


Tgt Ion:198 Resp: 112693
 Ion Ratio Lower Upper
 198 100
 182 4.1 2.9 4.3
 77 21.8 17.5 26.3



#67
 N-Nitrosodiphenylamine
 Concen: 20.281 ng/ul
 RT: 15.968 min Scan# 2139
 Delta R.T. -0.000 min
 Lab File: BM037820.D
 Acq: 02 Dec 2022 09:41

Tgt Ion:169 Resp: 648958
 Ion Ratio Lower Upper
 169 100
 168 65.2 53.8 80.6
 167 36.3 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 19.751 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:248 Resp: 225048
Ion Ratio Lower Upper
248 100
250 103.7 78.1 117.1
141 70.9 51.6 77.4



#69
Hexachlorobenzene
Concen: 21.013 ng/ul
RT: 16.762 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:284 Resp: 279598
Ion Ratio Lower Upper
284 100
142 31.6 24.5 36.7
249 28.5 22.9 34.3





#70

Atrazine

Concen: 19.601 ng/ul

RT: 16.909 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :

SSTD020060





Tgt Ion:200 Resp: 215994

Ion Ratio Lower Upper

200 100

173 26.0 22.0 33.0

215 45.3 40.0 60.0



#71

Pentachlorophenol

Concen: 19.174 ng/ul

RT: 17.104 min Scan# 2332

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41




Tgt Ion:266 Resp: 156919


Ion Ratio Lower Upper

266 100

264 61.8 50.3 75.5

268 62.0 52.2 78.2









#72
Phenanthrene
Concen: 20.764 ng/ul
RT: 17.504 min Scan# 2400
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:178 Resp: 1249566
Ion Ratio Lower Upper
178 100
179 14.9 12.5 18.7
176 18.7 15.3 22.9



#74
Anthracene
Concen: 21.035 ng/ul
RT: 17.592 min Scan# 2415
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:178 Resp: 1291365
Ion Ratio Lower Upper
178 100
179 14.8 12.2 18.4
176 18.4 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.808 ng/uL
RT: 13.527 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:216 Resp: 296545
Ion Ratio Lower Upper
216 100
214 77.6 63.0 94.6
218 47.6 40.0 60.0



#76
Pentachlorobenzene
Concen: 20.482 ng/uL
RT: 15.039 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:250 Resp: 333609
Ion Ratio Lower Upper
250 100
248 62.0 50.5 75.7
252 65.4 51.1 76.7





#77

Carbazole

Concen: 21.090 ng/ul

RT: 17.862 min Scan# 2461

Delta R.T. -0.000 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :





BNA_M

ClientSampleId :

SSTD020060

Tgt Ion:167 Resp: 1191470

Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.2	25.8
139	12.5	10.0	15.0



#78

Di-n-butylphthalate

Concen: 19.409 ng/ul

RT: 18.415 min Scan# 2555




Delta R.T. -0.000 min


Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Tgt Ion:149 Resp: 1378604

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.4	11.2
104	4.4	3.4	5.0









#80
Fluoranthene
Concen: 19.497 ng/ul
RT: 19.503 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:202 Resp: 1523685
Ion Ratio Lower Upper
202 100
101 11.5 9.7 14.5
100 8.6 7.7 11.5



#82
Pyrene
Concen: 20.089 ng/ul
RT: 19.868 min Scan# 2802
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:202 Resp: 1620982
Ion Ratio Lower Upper
202 100
101 12.9 11.0 16.4
100 11.2 9.0 13.6









#83
Butylbenzylphthalate
Concen: 18.009 ng/ul
RT: 20.750 min Scan# 2952
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:149 Resp: 633578
Ion Ratio Lower Upper
149 100
91 67.6 51.7 77.5
206 20.9 17.1 25.7



#84
3,3'-Dichlorobenzidine
Concen: 21.119 ng/ul
RT: 21.539 min Scan# 3086
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:252 Resp: 546458
Ion Ratio Lower Upper
252 100
254 65.4 52.2 78.4
126 10.8 8.8 13.2









#85
Benzo(a)anthracene
Concen: 20.718 ng/ul
RT: 21.609 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41




Instrument :
BNA_M
ClientSampleId :
SSTD020060


Tgt Ion:228 Resp: 1637251
Ion Ratio Lower Upper
228 100
229 19.9 16.2 24.2
226 27.3 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.507 ng/ul
RT: 21.521 min Scan# 3083
Delta R.T. -0.006 min
Lab File: BM037820.D
Acq: 02 Dec 2022 09:41

Tgt Ion:149 Resp: 983850
Ion Ratio Lower Upper
149 100
167 28.3 23.3 34.9
279 4.1 3.6 5.4





#87

Chrysene

Concen: 20.961 ng/ul

RT: 21.668 min Scan# 3108

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :

SSTD020060





Tgt Ion:228 Resp: 1541961

Ion Ratio Lower Upper

228 100

226 30.3 24.2 36.2

229 19.9 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 19.157 ng/ul




RT: 22.474 min Scan# 3245


Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Tgt Ion:149 Resp: 1758967





#90

Benzo(b)fluoranthene

Concen: 21.007 ng/ul

RT: 23.350 min Scan# 3394

Delta R.T. -0.006 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :

SSTD020060





Tgt Ion:252 Resp: 1721241

Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 10.2 9.3 13.9



#91

Benzo(k)fluoranthene

Concen: 20.753 ng/ul

RT: 23.397 min Scan# 3402

Delta R.T. -0.012 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41




Tgt Ion:252 Resp: 1640362


Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.8

125 10.7 9.3 13.9





#93

Benzo(a)pyrene

Concen: 20.878 ng/ul

RT: 23.997 min Scan# 3

Delta R.T. -0.012 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :

SSTD020060





Tgt Ion:252 Resp: 1483403

Ion Ratio Lower Upper

252 100

253 22.2 17.4 26.2

125 12.3 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.261 ng/ul

RT: 26.709 min Scan# 3965

Delta R.T. -0.012 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41




Tgt Ion:276 Resp: 1698323


Ion Ratio Lower Upper

276 100

138 26.6 22.7 34.1

277 25.6 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 21.317 ng/ul

RT: 26.721 min Scan# 3967

Delta R.T. -0.018 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Instrument :

BNA_M

ClientSampleId :

SSTD020060





Tgt Ion:278 Resp: 1549851

Ion Ratio Lower Upper

278 100

139 17.1 13.9 20.9

279 24.0 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 21.446 ng/ul

RT: 27.509 min Scan# 4101

Delta R.T. -0.018 min

Lab File: BM037820.D

Acq: 02 Dec 2022 09:41

Tgt Ion:276 Resp: 1495021

Ion Ratio Lower Upper

276 100

138 22.5 19.2 28.8

277 23.4 19.1 28.7

