

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120222\
 Data File : BM037837.D
 Acq On : 02 Dec 2022 20:08
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020061

Quant Time: Dec 02 21:50:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.092	152	235103	20.000	ng/ul	0.00
20) Naphthalene-d8	10.916	136	1106365	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.727	164	750966	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.462	188	1551222	20.000	ng/ul	0.00
79) Chrysene-d12	21.633	240	1182338	20.000	ng/ul	0.00
88) Perylene-d12	24.115	264	909765	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	38962	7.775	ng/uL	0.00
4) Pyridine-d5	3.863	84	323220	19.943	ng/ul	0.00
7) Phenol-d5	7.251	99	418054	19.737	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	255337	20.662	ng/ul	0.00
11) 2-Chlorophenol-d4	7.622	132	338886	21.161	ng/ul	0.00
15) 4-Methylphenol-d8	8.810	113	341102	20.250	ng/ul	0.00
21) Nitrobenzene-d5	9.269	128	181566	20.659	ng/ul	0.00
24) 2-Nitrophenol-d4	9.992	143	200582	22.056	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.533	165	363606	21.559	ng/ul	0.00
31) 4-Chloroaniline-d4	11.051	131	525450	20.312	ng/ul	0.00
46) Dimethylphthalate-d6	14.133	166	1062019	20.595	ng/ul	0.00
49) Acenaphthylene-d8	14.421	160	1271055	20.267	ng/ul	0.00
54) 4-Nitrophenol-d4	14.915	143	198686	18.563	ng/ul	0.00
60) Fluorene-d10	15.709	176	956271	20.440	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.827	200	171107	20.940	ng/ul	0.00
73) Anthracene-d10	17.562	188	1475145	20.678	ng/ul	0.00
81) Pyrene-d10	19.844	212	1546878	25.258	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.956	264	951330	20.873	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.463	88	42168	7.852	ng/uL#	83
5) Pyridine	3.887	79	319397	20.099	ng/ul	96
6) Benzaldehyde	7.228	77	222469	21.808	ng/ul	99
8) Phenol	7.275	94	430440	19.885	ng/ul	93
10) Bis(2-Chloroethyl)ether	7.510	93	348813	20.199	ng/ul	98
12) 2-Chlorophenol	7.657	128	359740	20.870	ng/ul	99
13) 2-Methylphenol	8.539	108	343997	19.927	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.628	45	692614	21.585	ng/ul	99
16) Acetophenone	8.928	105	560734	21.103	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.916	70	289435	21.626	ng/ul	94
18) 4-Methylphenol	8.875	108	379830	20.393	ng/ul	96
19) Hexachloroethane	9.186	117	146865	22.306	ng/ul	95
22) Nitrobenzene	9.310	77	404026	21.936	ng/ul	96
23) Isophorone	9.839	82	823084	20.771	ng/ul	99
25) 2-Nitrophenol	10.027	139	219828	22.306	ng/ul	99
26) 2,4-Dimethylphenol	10.080	107	415320	21.923	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.322	93	491686	20.548	ng/ul	98
29) 2,4-Dichlorophenol	10.563	162	373622	21.901	ng/ul	99
30) Naphthalene	10.969	128	1254424	20.891	ng/ul	100
32) 4-Chloroaniline	11.080	127	530487	20.264	ng/ul	98
33) Hexachlorobutadiene	11.251	225	238776	22.596	ng/ul	98
34) Caprolactam	11.869	113	94105	15.191	ng/ul	89
35) 4-Chloro-3-methylphenol	12.192	107	385217	21.638	ng/ul	99
36) 2-Methylnaphthalene	12.563	142	864671	20.815	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120222\
 Data File : BM037837.D
 Acq On : 02 Dec 2022 20:08
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020061

Quant Time: Dec 02 21:50:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.786	142	870540	20.938	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.927	216	461719	21.468	ng/ul	100
40) Hexachlorocyclopentadiene	12.904	237	264814	22.391	ng/ul	99
41) 2,4,6-Trichlorophenol	13.168	196	312120	22.230	ng/ul	100
42) 2,4,5-Trichlorophenol	13.239	196	335012	22.403	ng/ul	100
43) 1,1'-Biphenyl	13.563	154	1140924	20.901	ng/ul	99
44) 2-Chloronaphthalene	13.610	162	899704	21.004	ng/ul	98
45) 2-Nitroaniline	13.810	65	265701	21.922	ng/ul	91
47) Dimethylphthalate	14.180	163	1118631	20.827	ng/ul	99
48) 2,6-Dinitrotoluene	14.304	165	228287	20.817	ng/ul	97
50) Acenaphthylene	14.451	152	1422685	20.460	ng/ul	99
51) 3-Nitroaniline	14.627	138	241656	18.462	ng/ul	96
52) Acenaphthene	14.786	153	990092	20.549	ng/ul	98
53) 2,4-Dinitrophenol	14.833	184	122810	19.175	ng/ul	90
55) 4-Nitrophenol	14.933	109	141714	21.480	ng/ul	93
56) Dibenzofuran	15.121	168	1350957	20.825	ng/ul	97
57) 2,4-Dinitrotoluene	15.080	165	326161	21.296	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.345	232	295734	22.484	ng/ul	97
59) Diethylphthalate	15.533	149	1141645	20.726	ng/ul	99
61) Fluorene	15.768	166	1142480	20.920	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.757	204	563658	21.649	ng/ul	100
63) 4-Nitroaniline	15.786	138	242668	18.429	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.839	198	180987	21.486	ng/ul	97
67) N-Nitrosodiphenylamine	15.968	169	949284	21.235	ng/ul	100
68) 4-Bromophenyl-phenylether	16.651	248	342991	21.547	ng/ul	98
69) Hexachlorobenzene	16.768	284	417584	22.464	ng/ul	99
70) Atrazine	16.915	200	320069	20.790	ng/ul	99
71) Pentachlorophenol	17.109	266	247876	21.681	ng/ul	99
72) Phenanthrene	17.503	178	1735355	20.641	ng/ul	99
74) Anthracene	17.598	178	1801174	21.001	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.533	216	466469	22.303	ng/ul	95
76) Pentachlorobenzene	15.039	250	519045	22.810	ng/ul	98
77) Carbazole	17.862	167	1595565	20.216	ng/ul	99
78) Di-n-butylphthalate	18.415	149	2008407	20.240	ng/ul	99
80) Fluoranthene	19.509	202	1905186	25.468	ng/ul	96
82) Pyrene	19.874	202	1973242	25.547	ng/ul	96
83) Butylbenzylphthalate	20.750	149	806073	23.935	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.538	252	467171	18.861	ng/ul	98
85) Benzo(a)anthracene	21.615	228	1596451	21.104	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.521	149	1114910	21.909	ng/ul	98
87) Chrysene	21.668	228	1468753	20.858	ng/ul	99
89) Di-n-octyl phthalate	22.474	149	1653167	25.250	ng/ul	100
90) Benzo(b)fluoranthene	23.356	252	1297377	22.206	ng/ul	98
91) Benzo(k)fluoranthene	23.403	252	1237733	21.961	ng/ul	98
93) Benzo(a)pyrene	24.009	252	1065893	21.038	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.721	276	1226940	21.540	ng/ul	95
95) Dibenzo(a,h)anthracene	26.738	278	1111105	21.432	ng/ul	98
96) Benzo(g,h,i)perylene	27.526	276	1081962	21.766	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.852 ng/uL
RT: 3.463 min Scan# 11
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion: 88 Resp: 42168
Ion Ratio Lower Upper
88 100
43 40.3 51.3 76.9#
58 91.0 79.0 118.6



#5
Pyridine
Concen: 20.099 ng/uL
RT: 3.887 min Scan# 85
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion: 79 Resp: 319397
Ion Ratio Lower Upper
79 100
52 97.6 75.2 112.8
51 46.2 33.9 50.9









#6
Benzaldehyde
Concen: 21.808 ng/ul
RT: 7.228 min Scan# 61
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion: 77 Resp: 222469
Ion Ratio Lower Upper
77 100
105 100.7 81.2 121.8
106 98.5 79.0 118.6



#8
Phenol
Concen: 19.885 ng/ul
RT: 7.275 min Scan# 661
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion: 94 Resp: 430440
Ion Ratio Lower Upper
94 100
65 29.2 21.1 31.7
66 35.8 25.2 37.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.199 ng/ul
RT: 7.510 min Scan# 701
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion: 93 Resp: 348813
Ion Ratio Lower Upper
93 100
63 84.5 65.7 98.5
95 32.6 26.8 40.2



#12
2-Chlorophenol
Concen: 20.870 ng/ul
RT: 7.657 min Scan# 726
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:128 Resp: 359740
Ion Ratio Lower Upper
128 100
64 47.4 37.3 55.9
130 34.5 27.1 40.7









#13
2-Methylphenol
Concen: 19.927 ng/ul
RT: 8.539 min Scan# 81
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:108 Resp: 343997
Ion Ratio Lower Upper
108 100
107 86.7 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.585 ng/ul
RT: 8.628 min Scan# 891
Delta R.T. -0.006 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion: 45 Resp: 692614
Ion Ratio Lower Upper
45 100
77 10.4 8.2 12.4
79 8.0 6.7 10.1









#16
Acetophenone
Concen: 21.103 ng/ul
RT: 8.928 min Scan# 942
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:105 Resp: 560734
Ion Ratio Lower Upper
105 100
77 75.2 59.0 88.4
51 32.6 24.4 36.6



#17
N-Nitroso-di-n-propylamine
Concen: 21.626 ng/ul
RT: 8.916 min Scan# 940
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion: 70 Resp: 289435
Ion Ratio Lower Upper
70 100
42 76.5 55.8 83.6
101 9.9 8.0 12.0
130 21.8 18.6 27.8









#18
4-Methylphenol
Concen: 20.393 ng/ul
RT: 8.875 min Scan# 911
Delta R.T. 0.005 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:108 Resp: 379830
Ion Ratio Lower Upper
108 100
107 107.7 89.3 133.9



#19
Hexachloroethane
Concen: 22.306 ng/ul
RT: 9.186 min Scan# 986
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:117 Resp: 146865
Ion Ratio Lower Upper
117 100
201 106.7 82.3 123.5
199 68.1 49.8 74.6









#22
Nitrobenzene
Concen: 21.936 ng/ul
RT: 9.310 min Scan# 1007
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion: 77 Resp: 404026
Ion Ratio Lower Upper
77 100
123 48.2 41.4 62.0
65 13.7 10.9 16.3



#23
Isophorone
Concen: 20.771 ng/ul
RT: 9.839 min Scan# 1097
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion: 82 Resp: 823084
Ion Ratio Lower Upper
82 100
95 6.6 5.5 8.3
138 18.8 15.3 22.9









#25
2-Nitrophenol
Concen: 22.306 ng/ul
RT: 10.027 min Scan# 1128
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:139 Resp: 219828
Ion Ratio Lower Upper
139 100
65 46.6 36.6 54.8
109 26.0 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 21.923 ng/ul
RT: 10.080 min Scan# 1138
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:107 Resp: 415320
Ion Ratio Lower Upper
107 100
121 53.7 41.8 62.8
122 90.5 73.1 109.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.548 ng/ul
 RT: 10.322 min Scan# 1178
 Delta R.T. -0.001 min
 Lab File: BM037837.D
 Acq: 02 Dec 2022 20:08




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020061


Tgt Ion: 93	Resp: 491686
Ion Ratio	Lower Upper
93 100	
95 33.6	27.9 41.9
123 12.1	9.7 14.5



#29
 2,4-Dichlorophenol
 Concen: 21.901 ng/ul
 RT: 10.563 min Scan# 1220
 Delta R.T. -0.001 min
 Lab File: BM037837.D
 Acq: 02 Dec 2022 20:08

Tgt Ion:162	Resp: 373622		
Ion Ratio	Lower	Upper	
162 100			
164 64.4	50.7	76.1	
98 34.7	27.3	40.9	









#30
Naphthalene
Concen: 20.891 ng/ul
RT: 10.969 min Scan# 1289
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:128 Resp: 1254424
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 20.264 ng/ul
RT: 11.080 min Scan# 1308
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:127 Resp: 530487
Ion Ratio Lower Upper
127 100
129 32.2 26.9 40.3





#33

Hexachlorobutadiene

Concen: 22.596 ng/ul

RT: 11.251 min Scan# 11

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061





Tgt Ion:225 Resp: 238776

Ion Ratio Lower Upper

225 100

223 62.4 49.3 73.9

227 64.3 49.4 74.0



#34

Caprolactam

Concen: 15.191 ng/ul

RT: 11.869 min Scan# 1442

Delta R.T. 0.011 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08




Tgt Ion:113 Resp: 94105


Ion Ratio Lower Upper

113 100

55 207.8 186.2 279.2

56 155.4 119.8 179.8









#35
4-Chloro-3-methylphenol
Concen: 21.638 ng/ul
RT: 12.192 min Scan# 1497
Delta R.T. 0.005 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:107 Resp: 385217
Ion Ratio Lower Upper
107 100
144 26.9 23.0 34.6
142 86.2 69.1 103.7



#36
2-Methylnaphthalene
Concen: 20.815 ng/ul
RT: 12.563 min Scan# 1560
Delta R.T. -0.006 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:142 Resp: 864671
Ion Ratio Lower Upper
142 100
141 88.7 70.1 105.1





#37

1-Methylnaphthalene

Concen: 20.938 ng/ul

RT: 12.786 min Scan# 1

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061





Tgt Ion:142 Resp: 870540

Ion Ratio Lower Upper

142 100

141 90.9 73.4 110.2

116 3.5 2.6 4.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.468 ng/ul

RT: 12.927 min Scan# 1622

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Tgt Ion:216 Resp: 461719




Ion Ratio Lower Upper


216 100

214 78.5 62.7 94.1

179 20.0 15.9 23.9

108 17.0 13.0 19.4









#40
Hexachlorocyclopentadiene
Concen: 22.391 ng/ul
RT: 12.904 min Scan# 1618
Delta R.T. -0.006 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:237 Resp: 264814
Ion Ratio Lower Upper
237 100
235 62.3 49.2 73.8
272 13.1 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 22.230 ng/ul
RT: 13.168 min Scan# 1663
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:196 Resp: 312120
Ion Ratio Lower Upper
196 100
198 94.6 75.6 113.4
200 31.3 24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 22.403 ng/ul
RT: 13.239 min Scan# 1
Delta R.T. 0.005 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:196 Resp: 335012
Ion Ratio Lower Upper
196 100
198 95.0 76.1 114.1
200 31.2 24.8 37.2



#43
1,1'-Biphenyl
Concen: 20.901 ng/ul
RT: 13.563 min Scan# 1730
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:154 Resp: 1140924
Ion Ratio Lower Upper
154 100
153 39.0 31.4 47.0
76 12.0 10.2 15.2





#44

2-Chloronaphthalene

Concen: 21.004 ng/ul

RT: 13.610 min Scan# 1

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061





Tgt Ion:162 Resp: 899704

Ion Ratio Lower Upper

162 100

164 32.3 26.5 39.7

127 36.8 30.5 45.7



#45

2-Nitroaniline

Concen: 21.922 ng/ul

RT: 13.810 min Scan# 1772

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08




Tgt Ion: 65 Resp: 265701


Ion Ratio Lower Upper

65 100

92 65.9 54.7 82.1

138 107.8 96.8 145.2









#47
Dimethylphthalate
Concen: 20.827 ng/ul
RT: 14.180 min Scan# 1835
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:163 Resp: 1118631
Ion Ratio Lower Upper
163 100
194 4.1 3.6 5.4
164 9.8 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.817 ng/ul
RT: 14.304 min Scan# 1856
Delta R.T. 0.005 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:165 Resp: 228287
Ion Ratio Lower Upper
165 100
89 53.3 40.6 60.8
121 20.6 16.4 24.6









#50
Acenaphthylene
Concen: 20.460 ng/ul
RT: 14.451 min Scan# 1881
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:152 Resp: 1422685
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.6
153 13.2 10.6 15.8



#51
3-Nitroaniline
Concen: 18.462 ng/ul
RT: 14.627 min Scan# 1911
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:138 Resp: 241656
Ion Ratio Lower Upper
138 100
108 9.5 7.8 11.8
92 108.6 83.8 125.6





#52

Acenaphthene

Concen: 20.549 ng/ul

RT: 14.786 min Scan# 1938

Delta R.T. -0.006 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061





Tgt Ion:153 Resp: 990092

Ion Ratio Lower Upper

153 100

152 48.1 37.0 55.4

154 88.2 69.8 104.6



#53

2,4-Dinitrophenol

Concen: 19.175 ng/ul

RT: 14.833 min Scan# 1946

Delta R.T. 0.005 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08




Tgt Ion:184 Resp: 122810


Ion Ratio Lower Upper

184 100

63 60.3 43.4 65.0

154 57.0 39.7 59.5









#55
4-Nitrophenol
Concen: 21.480 ng/ul
RT: 14.933 min Scan# 1963
Delta R.T. 0.005 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:109 Resp: 141714
Ion Ratio Lower Upper
109 100
139 151.1 131.6 197.4
65 142.0 109.4 164.0



#56
Dibenzofuran
Concen: 20.825 ng/ul
RT: 15.121 min Scan# 1995
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:168 Resp: 1350957
Ion Ratio Lower Upper
168 100
139 36.2 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 21.296 ng/ul

RT: 15.080 min Scan# 1988

Delta R.T. -0.001 min

Lab File: BM037837.D


Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061






Tgt Ion:165 Resp: 326161

Ion Ratio Lower Upper

165 100

63 42.8 30.2 45.2

182 2.4 1.8 2.8



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.484 ng/ul

RT: 15.345 min Scan# 2033

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Tgt Ion:232 Resp: 295734




Ion Ratio Lower Upper


232 100

131 43.1 32.7 49.1

130 2.0 1.8 2.6

166 28.9 22.2 33.4









#59
Diethylphthalate
Concen: 20.726 ng/ul
RT: 15.533 min Scan# 2065
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:149 Resp: 1141645
Ion Ratio Lower Upper
149 100
177 22.2 18.0 27.0
150 11.9 9.8 14.8



#61
Fluorene
Concen: 20.920 ng/ul
RT: 15.768 min Scan# 2105
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:166 Resp: 1142480
Ion Ratio Lower Upper
166 100
165 97.9 77.0 115.6
167 13.1 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 21.649 ng/ul
RT: 15.757 min Scan# 2103
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:204 Resp: 563658
Ion Ratio Lower Upper
204 100
206 32.6 26.2 39.4
141 55.8 44.7 67.1



#63
4-Nitroaniline
Concen: 18.429 ng/ul
RT: 15.786 min Scan# 2108
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:138 Resp: 242668
Ion Ratio Lower Upper
138 100
92 46.0 36.6 54.8
108 68.9 48.2 72.2









#66
4,6-Dinitro-2-methylphenol
Concen: 21.486 ng/ul
RT: 15.839 min Scan# 2116
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:198 Resp: 180987
Ion Ratio Lower Upper
198 100
182 3.9 2.9 4.3
77 23.6 17.5 26.3



#67
N-Nitrosodiphenylamine
Concen: 21.235 ng/ul
RT: 15.968 min Scan# 2139
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:169 Resp: 949284
Ion Ratio Lower Upper
169 100
168 66.9 53.8 80.6
167 36.8 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 21.547 ng/ul
RT: 16.651 min Scan# 21
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:248 Resp: 342991
Ion Ratio Lower Upper
248 100
250 100.0 78.1 117.1
141 65.7 51.6 77.4



#69
Hexachlorobenzene
Concen: 22.464 ng/ul
RT: 16.768 min Scan# 2275
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:284 Resp: 417584
Ion Ratio Lower Upper
284 100
142 30.9 24.5 36.7
249 29.0 22.9 34.3





#70

Atrazine

Concen: 20.790 ng/ul

RT: 16.915 min Scan# 21

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061





Tgt Ion:200 Resp: 320069

Ion Ratio Lower Upper

200 100

173 27.2 22.0 33.0

215 49.1 40.0 60.0



#71

Pentachlorophenol

Concen: 21.681 ng/ul

RT: 17.109 min Scan# 2333

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08




Tgt Ion:266 Resp: 247876


Ion Ratio Lower Upper

266 100

264 63.6 50.3 75.5

268 66.1 52.2 78.2









#72
Phenanthrene
Concen: 20.641 ng/ul
RT: 17.503 min Scan# 2400
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:178 Resp: 1735355
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 19.4 15.3 22.9



#74
Anthracene
Concen: 21.001 ng/ul
RT: 17.598 min Scan# 2416
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:178 Resp: 1801174
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.303 ng/uL
RT: 13.533 min Scan# 1
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:216 Resp: 466469
Ion Ratio Lower Upper
216 100
214 74.5 63.0 94.6
218 45.9 40.0 60.0



#76
Pentachlorobenzene
Concen: 22.810 ng/uL
RT: 15.039 min Scan# 1981
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:250 Resp: 519045
Ion Ratio Lower Upper
250 100
248 64.0 50.5 75.7
252 66.1 51.1 76.7





#77

Carbazole

Concen: 20.216 ng/ul

RT: 17.862 min Scan# 2461

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061





Tgt Ion:167 Resp: 1595565

Ion Ratio Lower Upper

167 100

166 20.9 17.2 25.8

139 12.4 10.0 15.0



#78

Di-n-butylphthalate

Concen: 20.240 ng/ul

RT: 18.415 min Scan# 2555

Delta R.T. -0.001 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08




Tgt Ion:149 Resp: 2008407


Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.2

104 4.4 3.4 5.0









#80
Fluoranthene
Concen: 25.468 ng/ul
RT: 19.509 min Scan# 21
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:202 Resp: 1905186
Ion Ratio Lower Upper
202 100
101 10.7 9.7 14.5
100 8.2 7.7 11.5



#82
Pyrene
Concen: 25.547 ng/ul
RT: 19.874 min Scan# 2803
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:202 Resp: 1973242
Ion Ratio Lower Upper
202 100
101 12.1 11.0 16.4
100 10.0 9.0 13.6





#83

Butylbenzylphthalate

Concen: 23.935 ng/ul

RT: 20.750 min Scan# 2952

Delta R.T. -0.006 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061





Tgt Ion:149 Resp: 806073

Ion Ratio Lower Upper

149 100

91 68.5 51.7 77.5

206 20.5 17.1 25.7



#84

3,3'-Dichlorobenzidine

Concen: 18.861 ng/ul

RT: 21.538 min Scan# 3086

Delta R.T. -0.006 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08




Tgt Ion:252 Resp: 467171


Ion Ratio Lower Upper

252 100

254 63.7 52.2 78.4

126 11.1 8.8 13.2









#85
Benzo(a)anthracene
Concen: 21.104 ng/ul
RT: 21.615 min Scan# 3099
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:228 Resp: 1596451
Ion Ratio Lower Upper
228 100
229 19.8 16.2 24.2
226 26.7 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.909 ng/ul
RT: 21.521 min Scan# 3083
Delta R.T. -0.006 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:149 Resp: 1114910
Ion Ratio Lower Upper
149 100
167 27.8 23.3 34.9
279 4.2 3.6 5.4





#87

Chrysene

Concen: 20.858 ng/ul

RT: 21.668 min Scan# 3108

Delta R.T. -0.006 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Instrument :

BNA_M

ClientSampleId :

SSTD020061





Tgt Ion:228 Resp: 1468753

Ion Ratio Lower Upper

228 100

226 29.6 24.2 36.2

229 20.0 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 25.250 ng/ul




RT: 22.474 min Scan# 3245


Delta R.T. -0.006 min

Lab File: BM037837.D

Acq: 02 Dec 2022 20:08

Tgt Ion:149 Resp: 1653167









#90
Benzo(b)fluoranthene
Concen: 22.206 ng/ul
RT: 23.356 min Scan# 31
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:252 Resp: 1297377
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 9.7 9.3 13.9



#91
Benzo(k)fluoranthene
Concen: 21.961 ng/ul
RT: 23.403 min Scan# 3403
Delta R.T. -0.006 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:252 Resp: 1237733
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 9.8 9.3 13.9









#93
Benzo(a)pyrene
Concen: 21.038 ng/ul
RT: 24.009 min Scan# 3
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08




Instrument :
BNA_M
ClientSampleId :
SSTD020061


Tgt Ion:252 Resp: 1065893
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 10.5 10.2 15.4



#94
Indeno(1,2,3-cd)pyrene
Concen: 21.540 ng/ul
RT: 26.721 min Scan# 3967
Delta R.T. -0.001 min
Lab File: BM037837.D
Acq: 02 Dec 2022 20:08

Tgt Ion:276 Resp: 1226940
Ion Ratio Lower Upper
276 100
138 23.5 22.7 34.1
277 25.5 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 21.432 ng/ul
 RT: 26.738 min Scan# 3967
 Delta R.T. -0.001 min
 Lab File: BM037837.D
 Acq: 02 Dec 2022 20:08

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020061

Tgt Ion:278 Resp: 1111105
 Ion Ratio Lower Upper
 278 100
 139 15.8 13.9 20.9
 279 24.0 19.0 28.6



#96
 Benzo(g,h,i)perylene
 Concen: 21.766 ng/ul
 RT: 27.526 min Scan# 4104
 Delta R.T. -0.001 min
 Lab File: BM037837.D
 Acq: 02 Dec 2022 20:08

Tgt Ion:276 Resp: 1081962
 Ion Ratio Lower Upper
 276 100
 138 19.8 19.2 28.8
 277 24.1 19.1 28.7

