


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120324\
Data File : BM048808.D
Acq On : 02 Dec 2024 16:53
Operator : RC/JU
Sample : DFTPP068
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP068

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM120324.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Dec 03 00:22:16 2024



AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.9	234187	PASS
68	69	0.00	2	1.5	3823	PASS
69	198	0.00	100	40.5	249877	PASS
70	69	0.00	2	0.5	1257	PASS
127	198	10	80	46.7	288171	PASS
197	198	0.00	2	0.4	2701	PASS
198	198	100	100	100.0	617323	PASS
199	198	5	9	6.9	42469	PASS
275	198	10	60	23.7	146219	PASS
365	198	1	100	3.4	21202	PASS
441	198	0.01	100	10.5	64688	PASS
442	198	50	100	64.5	398400	PASS
443	442	15	24	19.2	76499	PASS

m/z	Abundance
40.00	3377.0
44.00	4257.0
49.00	502.0
51.00	416.0
72.90	212.0
83.90	479.0
86.00	203.0
167.00	230.0
190.80	200.0
207.00	1247.0
208.00	239.0
208.90	269.0
280.90	250.0

Instrument :

BNA_M

ClientSampleId :

DFTPP068

m/z	Abundance
35.90	174.0
37.00	952.0
38.00	2857.0
39.00	18224.0
40.00	5733.0
41.10	653.0
42.90	190.0
44.00	5686.0
45.00	620.0
46.90	204.0
47.90	308.0
49.10	2106.0
50.00	60432.0
51.00	242176.0
52.00	12581.0
53.00	755.0
54.90	1046.0
56.00	6332.0
57.00	16214.0
58.00	882.0
59.10	199.0
61.00	3080.0
62.00	3288.0
63.00	8882.0
64.00	1289.0
65.00	4790.0
66.00	375.0
67.10	379.0
68.00	4157.0
69.00	250496.0
70.00	1233.0
70.90	261.0
73.00	2030.0
74.00	23560.0
75.00	40368.0
76.10	14004.0
77.00	291264.0
78.00	18776.0
79.00	16165.0
80.00	12919.0
81.00	19144.0
82.00	5342.0
83.00	4777.0
84.00	1004.0
85.00	2458.0
86.00	4858.0
87.00	2808.0
88.00	978.0
88.90	489.0
91.00	3386.0
92.00	4813.0
93.00	28976.0
94.00	2470.0
94.90	392.0
96.10	1387.0
97.00	412.0
98.00	21528.0
99.00	19104.0
100.00	1631.0
101.00	10622.0
102.00	709.0
103.00	2479.0
104.00	5853.0
105.00	5929.0
106.00	1985.0
107.00	84208.0
108.00	12405.0
109.00	1726.0
110.00	144960.0
111.00	21664.0
112.00	3110.0
113.00	810.0
113.80	206.0
114.30	173.0
114.90	217.0
116.00	3459.0
117.00	70736.0
118.00	5070.0
119.00	493.0
120.10	796.0
121.00	345.0
121.90	4899.0
123.00	7115.0
124.00	3957.0
125.00	3191.0
126.10	782.0
127.00	270656.0
128.00	19928.0
129.00	101320.0
130.00	8436.0
130.90	1584.0
132.00	769.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

132.90	478.0
133.90	2602.0
135.00	8499.0
136.00	3685.0
137.00	4800.0
137.90	873.0
138.80	440.0
139.90	1202.0
141.00	11004.0
141.90	4281.0
143.00	2815.0
144.00	802.0
145.00	900.0
146.00	2106.0
147.00	6980.0
148.00	18216.0
149.00	3572.0
150.00	956.0
151.00	1426.0
151.90	629.0
153.00	4730.0
153.90	3265.0
155.00	7125.0
156.00	11565.0
157.00	2006.0
157.90	2197.0
158.90	1925.0
160.00	4402.0
161.00	6336.0
162.00	1942.0
163.00	512.0
164.00	497.0
165.00	5582.0
166.00	4430.0
167.00	32024.0
168.00	19792.0
169.00	3210.0
170.00	809.0
171.10	1434.0
171.90	2089.0
172.90	3241.0
173.90	6527.0
175.00	11578.0
176.00	2858.0
176.90	4717.0
177.90	1797.0
179.00	19832.0
180.00	14700.0
181.00	6392.0
181.90	1051.0
183.10	341.0
184.00	1683.0
185.00	10314.0
186.00	77744.0
187.00	23368.0
188.00	2299.0
189.00	4536.0
189.90	892.0
191.00	2177.0
192.00	6962.0
193.00	7737.0
194.00	1630.0
195.00	853.0
196.00	12621.0
196.90	2148.0
197.90	525632.0
198.90	37968.0
199.90	2742.0
201.40	1934.0
202.00	603.0
203.00	3978.0
204.00	21600.0
205.00	35336.0
206.00	142976.0
207.00	19688.0
208.00	5270.0
209.00	1771.0
209.90	2014.0
211.00	5290.0
212.00	968.0
212.90	268.0
213.90	252.0
215.00	1755.0
216.00	3561.0
216.90	40832.0
218.00	5297.0
219.00	559.0
220.20	683.0
221.00	17768.0
221.90	5238.0
222.90	8596.0
224.00	80936.0
225.00	20872.0
226.00	1651.0
227.00	36192.0
228.00	4867.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

229.00	6886.0
229.90	1068.0
231.00	2703.0
232.00	364.0
232.90	575.0
233.90	2180.0
234.90	2405.0
236.00	1634.0
237.00	2517.0
237.90	451.0
238.90	1247.0
239.90	1037.0
240.90	1917.0
242.00	4303.0
243.00	4145.0
244.00	60824.0
245.00	7779.0
245.90	12545.0
247.00	2588.0
247.90	337.0
248.90	2532.0
249.90	328.0
251.00	507.0
252.00	288.0
253.00	900.0
254.00	1277.0
255.00	286208.0
256.00	42496.0
257.00	3159.0
258.00	18504.0
259.00	2768.0
259.90	529.0
261.10	347.0
263.00	167.0
264.00	361.0
264.90	6837.0
265.90	1175.0
266.90	153.0
270.00	342.0
270.90	658.0
271.90	921.0
273.00	7883.0
274.00	20472.0
275.00	106624.0
276.00	14514.0
277.00	10776.0
278.00	2140.0
278.90	516.0
280.90	449.0
281.90	413.0
282.90	1226.0
283.90	738.0
285.00	1897.0
285.90	406.0
288.90	280.0
290.00	449.0
291.00	432.0
292.10	557.0
292.90	2311.0
293.90	602.0
295.00	836.0
296.00	34920.0
296.90	5640.0
297.90	431.0
300.80	239.0
302.00	520.0
303.00	3833.0
304.00	946.0
304.90	157.0
307.90	342.0
308.90	383.0
309.90	532.0
313.00	303.0
314.00	1105.0
314.90	3745.0
316.00	1919.0
316.90	546.0
320.90	751.0
321.90	532.0
323.00	9391.0
324.00	1924.0
325.90	157.0
326.90	2366.0
328.00	969.0
329.00	160.0
332.00	725.0
332.90	768.0
334.00	6146.0
335.10	1688.0
338.90	235.0
341.00	1008.0
342.10	282.0
345.90	2189.0
346.90	395.0
352.10	2446.0
353.00	1776.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

354.00	2514.0
355.00	693.0
359.10	250.0
364.90	13431.0
366.00	2009.0
366.90	166.0
370.00	355.0
371.00	894.0
372.00	5062.0
372.90	1235.0
383.00	1254.0
383.90	398.0
390.00	353.0
390.90	336.0
392.10	314.0
400.90	288.0
401.90	1527.0
403.00	2172.0
404.00	800.0
404.90	159.0
421.00	1581.0
422.00	2266.0
423.10	11360.0
424.00	2149.0
441.10	36880.0
442.00	207680.0
443.00	42536.0
444.10	4186.0
445.10	349.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

m/z	Abundance
36.00	449.0
37.00	812.0
38.10	3267.0
39.00	20352.0
40.00	4982.0
41.10	898.0
42.00	207.0
43.10	170.0
44.00	6268.0
45.00	666.0
47.00	237.0
47.90	243.0
49.00	2121.0
50.00	71872.0
51.00	277184.0
52.10	13792.0
53.00	640.0
55.00	1314.0
56.00	8072.0
57.00	19728.0
58.10	670.0
59.10	233.0
60.00	245.0
61.00	3054.0
62.00	4002.0
63.00	10919.0
63.90	1394.0
65.00	5953.0
66.00	448.0
67.10	266.0
68.00	4333.0
69.00	300864.0
70.00	1464.0
72.10	200.0
73.10	2389.0
74.00	29128.0
75.00	49472.0
76.00	16848.0
77.00	355584.0
78.10	23944.0
79.00	18880.0
80.00	16253.0
81.00	23472.0
82.00	5222.0
83.00	5867.0
83.90	922.0
85.00	3439.0
86.00	6001.0
87.00	2975.0
88.00	1342.0
89.00	816.0
91.00	4918.0
92.00	4874.0
93.00	36720.0
94.00	2708.0
95.00	730.0
96.00	1600.0
96.90	411.0
98.00	28848.0
99.00	23816.0
100.00	1917.0
101.00	14486.0
102.10	672.0
103.00	3790.0
104.00	7733.0
105.00	7516.0
106.00	2351.0
107.00	105960.0
108.00	14818.0
109.00	2566.0
110.00	186112.0
111.00	28768.0
112.00	3394.0
113.00	1184.0
114.20	339.0
114.90	355.0
116.00	4936.0
117.00	92368.0
118.00	6988.0
118.90	763.0
120.00	1246.0
120.90	336.0
122.00	5937.0
123.00	9305.0
124.00	3891.0
125.00	4360.0
126.10	1457.0
127.00	352064.0
128.00	26080.0
129.00	134464.0
130.00	11226.0
131.00	2103.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

132.00	895.0
132.90	570.0
134.00	4149.0
135.00	11256.0
136.00	4928.0
137.00	6800.0
137.90	1309.0
138.90	494.0
140.00	1152.0
141.00	15988.0
142.00	5098.0
143.00	4307.0
144.00	993.0
145.00	674.0
146.00	3115.0
147.00	9898.0
148.00	24072.0
149.00	5408.0
149.90	1464.0
151.00	2403.0
151.70	1233.0
153.00	5911.0
154.00	4405.0
155.00	10323.0
156.00	16064.0
157.00	2698.0
158.00	3261.0
159.00	2631.0
160.00	5815.0
161.00	8810.0
162.00	2378.0
163.00	735.0
164.10	891.0
165.00	7448.0
166.00	5381.0
167.00	47328.0
168.00	27608.0
169.00	3882.0
170.10	1105.0
170.90	1677.0
172.00	3450.0
172.90	4291.0
174.00	8256.0
175.00	15811.0
176.00	3912.0
177.00	7080.0
178.00	2651.0
178.90	29448.0
180.00	19904.0
181.00	9972.0
182.00	1605.0
182.90	570.0
184.00	2631.0
185.00	14773.0
186.00	112264.0
187.00	31216.0
188.00	3483.0
189.00	7444.0
190.00	1175.0
190.90	3408.0
192.00	10152.0
193.00	11776.0
194.00	2419.0
194.90	774.0
196.00	20536.0
196.80	3496.0
197.90	778560.0
198.90	54944.0
200.00	3936.0
201.50	2510.0
203.00	6523.0
204.00	31568.0
205.00	53304.0
206.00	215168.0
207.00	29376.0
208.00	7609.0
209.00	2191.0
210.00	3187.0
211.00	8969.0
211.90	1154.0
212.90	531.0
214.10	394.0
215.00	2728.0
216.00	4438.0
217.00	62616.0
218.00	8506.0
219.00	701.0
221.00	28536.0
221.80	7323.0
223.00	14545.0
224.00	129792.0
225.00	31328.0
226.00	3499.0
227.00	54088.0
228.00	6916.0
229.00	11011.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

230.00	1574.0
231.00	3919.0
231.80	625.0
232.90	866.0
234.00	2950.0
234.90	4093.0
235.90	3374.0
237.00	4112.0
238.00	625.0
238.90	1974.0
239.90	1331.0
241.00	3014.0
242.00	7517.0
243.00	7823.0
244.00	101168.0
245.00	13497.0
246.00	19632.0
247.00	3470.0
248.00	1033.0
249.00	3250.0
250.00	426.0
251.00	905.0
251.90	866.0
252.90	1593.0
254.00	2281.0
255.00	468672.0
256.00	69288.0
257.00	5626.0
258.00	28944.0
259.00	4686.0
260.00	861.0
260.90	749.0
262.00	212.0
262.90	425.0
264.00	603.0
265.00	12345.0
266.00	1972.0
266.90	226.0
269.90	457.0
271.00	675.0
271.90	1113.0
273.00	14838.0
274.00	37784.0
275.00	187200.0
276.00	25880.0
277.00	17856.0
278.00	2553.0
279.10	939.0
281.00	560.0
281.90	524.0
282.90	1985.0
284.00	1138.0
285.00	3037.0
286.00	688.0
287.90	215.0
289.00	695.0
290.00	633.0
290.90	237.0
292.00	685.0
293.00	4323.0
294.00	1026.0
295.00	1034.0
296.00	62384.0
297.00	8748.0
298.00	802.0
301.00	864.0
301.90	753.0
303.00	6758.0
304.00	1877.0
304.90	308.0
308.00	786.0
308.90	472.0
310.00	757.0
310.90	200.0
311.90	247.0
313.10	433.0
314.00	2716.0
315.00	6280.0
316.00	3525.0
317.00	718.0
320.00	336.0
321.00	2074.0
322.00	778.0
323.00	18040.0
324.10	3558.0
325.10	296.0
326.00	489.0
326.90	3251.0
328.00	2248.0
328.90	307.0
332.00	1281.0
333.00	1535.0
334.00	11684.0
335.00	3499.0
336.00	269.0
338.90	215.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

339.90	336.0
341.10	2128.0
342.00	414.0
346.00	4489.0
346.90	695.0
350.90	239.0
352.00	5322.0
353.00	3584.0
354.10	5071.0
355.10	988.0
359.00	676.0
361.00	157.0
365.00	26264.0
366.00	4072.0
367.00	250.0
370.00	566.0
371.00	1370.0
372.00	9392.0
373.00	2329.0
373.80	163.0
377.00	214.0
383.00	2390.0
384.10	712.0
385.00	201.0
390.00	1339.0
390.90	730.0
392.10	731.0
400.90	362.0
402.00	3574.0
403.00	4639.0
404.10	1889.0
405.00	212.0
414.90	231.0
421.00	3695.0
422.10	3717.0
423.10	27584.0
424.00	5245.0
425.00	566.0
441.00	79264.0
442.10	499584.0
443.10	95672.0
444.10	9176.0
445.00	591.0

Instrument :

BNA_M

ClientSampleId :

DFTPP068

m/z	Abundance
36.00	321.0
37.00	680.0
38.00	2248.0
39.00	13719.0
40.00	5086.0
41.10	466.0
41.90	250.0
42.80	158.0
43.20	175.0
44.00	6070.0
44.90	556.0
47.00	215.0
48.00	361.0
49.00	1907.0
50.00	44024.0
51.00	184448.0
52.10	8872.0
53.10	413.0
55.00	1093.0
56.00	5435.0
57.00	11720.0
58.10	563.0
60.10	257.0
61.00	2481.0
62.00	2599.0
63.00	7663.0
64.00	1098.0
65.00	3298.0
66.00	221.0
68.00	2978.0
69.00	198272.0
70.00	1075.0
71.00	259.0
72.10	210.0
73.10	1339.0
74.00	18592.0
75.00	32296.0
76.10	11861.0
77.00	234560.0
78.10	15125.0
79.00	12907.0
80.00	10016.0
81.00	15127.0
82.00	4273.0
83.00	3531.0
84.10	873.0
85.00	2178.0
86.00	4948.0
87.00	1919.0
88.00	1280.0
88.90	344.0
90.10	223.0
91.00	3164.0
92.00	3691.0
93.00	23344.0
94.00	2078.0
95.00	451.0
96.10	1133.0
97.10	353.0
98.00	19408.0
99.00	16640.0
100.00	1412.0
101.00	8468.0
101.90	384.0
103.00	2325.0
104.00	4599.0
105.00	5298.0
105.90	1777.0
107.00	68000.0
108.00	9245.0
109.00	1658.0
110.00	126368.0
111.00	17704.0
112.00	2165.0
113.10	674.0
114.00	177.0
114.80	279.0
116.00	3010.0
117.00	58688.0
118.00	4406.0
119.00	622.0
120.00	918.0
120.90	322.0
122.00	3996.0
123.00	6154.0
123.90	3117.0
125.00	2934.0
125.90	769.0
127.00	241792.0
128.00	18648.0
129.00	92312.0
130.00	7446.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

130.90	1686.0
131.90	774.0
133.00	728.0
134.00	2532.0
135.00	7817.0
136.00	2882.0
137.00	4644.0
138.00	895.0
138.90	337.0
140.00	917.0
141.00	11503.0
142.00	4404.0
143.00	2861.0
143.90	463.0
144.90	569.0
145.90	2496.0
147.00	6599.0
148.00	16371.0
149.00	3137.0
149.90	1144.0
151.00	1495.0
151.90	809.0
152.90	4155.0
153.90	3538.0
155.00	8031.0
156.00	10866.0
157.00	2093.0
158.00	1993.0
159.00	1487.0
160.00	3708.0
160.90	6468.0
161.90	1819.0
162.90	525.0
164.00	544.0
165.00	5114.0
166.00	4157.0
167.00	32040.0
168.00	19040.0
169.00	2650.0
170.00	561.0
170.80	906.0
172.00	2758.0
173.00	3222.0
174.00	7097.0
175.00	11907.0
176.00	2699.0
177.00	4553.0
178.00	1614.0
178.90	20784.0
180.00	14736.0
181.00	7196.0
181.90	982.0
182.90	369.0
183.90	1683.0
185.00	10169.0
186.00	81456.0
187.00	22128.0
188.00	1962.0
188.90	5714.0
190.00	1071.0
190.90	2741.0
192.00	7740.0
193.00	8730.0
193.90	1658.0
195.00	884.0
196.00	13257.0
196.90	2460.0
197.90	547776.0
198.90	34496.0
200.00	2691.0
200.80	559.0
201.40	2238.0
202.90	4484.0
204.00	21672.0
205.00	37544.0
206.00	148288.0
207.00	20864.0
208.00	5736.0
209.00	1877.0
210.00	2509.0
211.00	6878.0
211.80	1695.0
213.00	509.0
213.90	324.0
214.90	1705.0
216.00	3311.0
216.90	44656.0
218.00	6250.0
218.90	649.0
220.20	802.0
221.00	19432.0
221.90	5876.0
223.00	9921.0
224.00	95264.0
225.00	22720.0
226.00	2418.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

227.00	40976.0
228.00	5378.0
229.00	8284.0
230.00	996.0
231.10	3276.0
232.00	803.0
232.90	630.0
233.90	2702.0
234.90	2892.0
236.00	1865.0
237.00	3110.0
237.90	423.0
239.00	1802.0
240.00	1128.0
240.90	1925.0
242.00	5675.0
243.00	5339.0
244.00	73264.0
245.00	9012.0
246.00	13992.0
247.00	2556.0
247.90	592.0
248.90	2372.0
250.00	452.0
250.80	379.0
251.10	415.0
252.10	705.0
253.00	1404.0
253.90	1728.0
255.00	354560.0
256.00	52304.0
257.00	4042.0
258.00	22848.0
259.00	3215.0
260.00	516.0
260.90	840.0
263.00	279.0
263.90	652.0
264.90	9439.0
266.10	1356.0
266.90	269.0
269.90	403.0
271.00	872.0
272.00	791.0
273.00	11099.0
274.00	28920.0
275.00	144832.0
276.00	19352.0
277.00	13998.0
278.00	2487.0
278.90	608.0
280.80	256.0
281.90	407.0
283.00	1538.0
284.00	1138.0
285.00	2702.0
286.00	569.0
288.90	359.0
289.90	460.0
290.90	284.0
292.00	645.0
292.90	3288.0
294.10	840.0
294.90	914.0
296.00	51592.0
297.00	7275.0
297.90	501.0
300.90	684.0
302.00	1051.0
303.00	4907.0
304.00	1500.0
305.00	210.0
307.90	569.0
308.90	484.0
310.00	521.0
312.00	190.0
312.80	367.0
314.00	2328.0
315.00	5787.0
316.00	2571.0
317.00	632.0
318.90	160.0
319.80	266.0
321.00	1738.0
322.00	710.0
323.00	15292.0
324.00	2610.0
325.90	333.0
326.90	3362.0
328.00	1426.0
328.80	252.0
331.90	1489.0
333.00	1221.0
334.00	10069.0
335.00	2446.0
335.90	198.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068

338.90	287.0
339.90	211.0
341.00	1837.0
342.00	391.0
346.00	3659.0
347.00	498.0
351.00	305.0
352.00	4427.0
353.00	2287.0
354.00	5651.0
355.00	774.0
358.00	178.0
358.90	344.0
365.00	23912.0
366.00	3094.0
366.80	453.0
370.10	400.0
371.00	1034.0
372.00	8028.0
373.00	1899.0
374.00	286.0
377.00	180.0
383.00	1986.0
384.00	593.0
384.90	189.0
390.00	1078.0
391.10	794.0
392.00	382.0
400.80	545.0
402.00	3198.0
403.00	4507.0
404.00	1665.0
405.00	239.0
414.80	178.0
421.00	4031.0
422.00	3953.0
423.10	26312.0
424.00	5513.0
425.10	501.0
441.10	77920.0
442.10	487936.0
443.10	91288.0
444.10	8562.0
445.00	557.0

Instrument :
BNA_M
ClientSampleId :
DFTPP068