

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120415\
 Data File : BM003377.D
 Acq On : 03 Dec 2015 19:03
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 04 00:19:33 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 16:19:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	63	0.00
2	1,4-Dioxane	0.448	0.409	8.7	65	0.00
3	n-Nitrosodimethylamine	0.461	0.453	1.7	64	0.00
4 S	2-Fluorophenol	0.997	0.937	6.0	67	0.00
5 S	Phenol-d6	1.227	1.183	3.6	65	0.00
6	bis(2-Chloroethyl)ether	1.193	1.145	4.0	63	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	62	0.00
8 S	Nitrobenzene-d5	0.233	0.225	3.4	65	0.00
9	Nitrobenzene	0.252	0.244	3.2	64	0.00
10	Naphthalene	1.077	1.024	4.9	64	0.00
11	Hexachlorobutadiene	0.164	0.163	0.6	67	0.00
12	2-Methylnaphthalene	0.738	0.709	3.9	63	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	61	0.00
14 S	2,4,6-Tribromophenol	0.104	0.114	-9.6	65	0.00
15 S	2-Fluorobiphenyl	1.505	1.398	7.1	63	0.00
16	Acenaphthylene	1.976	1.923	2.7	61	0.00
17	Acenaphthene	1.321	1.200	9.2	62	0.00
18	Fluorene	1.490	1.463	1.8	63	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	62	0.00
20	4-Bromophenyl-phenylether	0.144	0.130	9.7	64	-0.03
21	Hexachlorobenzene	0.173	0.161	6.9	66	0.00
22	Pentachlorophenol	0.055	0.053	3.6	64	0.00
23	Phenanthrene	0.964	0.994	-3.1	72	0.08
24	Anthracene	1.016	1.016	0.0	63	0.00
25	Fluoranthene	0.999	1.026	-2.7	63	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	64	0.00
27	Benzydine	0.409	0.147	64.1#	26#	0.00
28	Pvrene	1.812	1.601	11.6	62	0.00
29 S	Terphenyl-d14	0.811	0.769	5.2	65	0.00
30	Benzo(a)anthracene	1.350	1.287	4.7	64	0.00
31	3,3'-Dichlorobenzidine	0.349	0.327	6.3	62	0.00
32	Chrysene	1.563	1.544	1.2	67	0.00
33	Bis(2-ethylhexyl)phthalate	0.684	0.723	-5.7	68	0.00
34	Indeno(1,2,3-cd)pyrene	1.163	0.927	20.3	66	0.00
35 I	Perylene-d12	1.000	1.000	0.0	66	0.00
36	Benzo(b)fluoranthene	1.522	1.498	1.6	68	0.00
37	Benzo(k)fluoranthene	1.366	1.295	5.2	66	0.00
38 C	Benzo(a)pyrene	1.249	1.150	7.9	66	0.00
39	Dibenzo(a,h)anthracene	1.136	0.907	20.2	66	0.00
40	Benzo(a,h,i)perylene	1.199	0.954	20.4	68	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120415\
Data File : BM003377.D
Acq On : 03 Dec 2015 19:03
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 04 00:19:33 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 02 16:19:00 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				