

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120415\
 Data File : BM003377.D
 Acq On : 03 Dec 2015 19:03
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 04 00:19:33 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 16:19:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	63	0.00
2	1,4-Dioxane	1.000	0.914	8.6	65	0.00
3	n-Nitrosodimethylamine	1.000	0.982	1.8	64	0.00
4 S	2-Fluorophenol	1.000	0.939	6.1	67	0.00
5 S	Phenol-d6	1.000	0.964	3.6	65	0.00
6	bis(2-Chloroethyl)ether	1.000	0.960	4.0	63	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	62	0.00
8 S	Nitrobenzene-d5	1.000	0.967	3.3	65	0.00
9	Nitrobenzene	1.000	0.965	3.5	64	0.00
10	Naphthalene	1.000	0.951	4.9	64	0.00
11	Hexachlorobutadiene	1.000	0.988	1.2	67	0.00
12	2-Methylnaphthalene	1.000	0.960	4.0	63	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	61	0.00
14 S	2,4,6-Tribromophenol	1.000	1.065	-6.5	65	0.00
15 S	2-Fluorobiphenyl	1.000	0.929	7.1	63	0.00
16	Acenaphthylene	1.000	0.973	2.7	61	0.00
17	Acenaphthene	1.000	0.908	9.2	62	0.00
18	Fluorene	1.000	0.981	1.9	63	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	62	0.00
20	4-Bromophenyl-phenylether	1.000	0.901	9.9	64	-0.03
21	Hexachlorobenzene	1.000	0.980	2.0	66	0.00
22	Pentachlorophenol	1.000	0.990	1.0	64	0.00
23	Phenanthrene	1.000	1.031	-3.1	72	0.08
24	Anthracene	1.000	1.000	0.0	63	0.00
25	Fluoranthene	1.000	1.027	-2.7	63	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	64	0.00
27	Benzydine	1.000	0.595	40.5#	26	0.00
28	Pvrene	1.000	0.884	11.6	62	0.00
29 S	Terphenyl-d14	1.000	0.949	5.1	65	0.00
30	Benzo(a)anthracene	1.000	0.954	4.6	64	0.00
31	3,3'-Dichlorobenzidine	1.000	0.890	11.0	62	0.00
32	Chrysene	1.000	0.988	1.2	67	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	1.039	-3.9	68	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	0.797	20.3	66	0.00
35 I	Perylene-d12	5.000	5.000	0.0	66	0.00
36	Benzo(b)fluoranthene	1.000	0.984	1.6	68	0.00
37	Benzo(k)fluoranthene	1.000	0.948	5.2	66	0.00
38 C	Benzo(a)pyrene	1.000	0.921	7.9	66	0.00
39	Dibenzo(a,h)anthracene	1.000	0.799	20.1	66	0.00
40	Benzo(a,h,i)perylene	1.000	0.795	20.5	68	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120415\
Data File : BM003377.D
Acq On : 03 Dec 2015 19:03
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 04 00:19:33 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 02 16:19:00 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				