


Data Path : Z:\HPCHEM1\BNA_M\Data\BM120417\
Data File : BM013158.D
Acq On : 04 Dec 2017 10:06
Operator : SJ/JU
Sample : DFTPP55
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP55

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM113017.M
Title : SVOA CALIBRATION
Last Update : Tue Dec 05 04:53:13 2017



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2451

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.8	47008	PASS
68	69	0.00	2	1.4	702	PASS
69	198	0.00	100	40.3	48906	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.7	55453	PASS
197	198	0.00	2	0.4	492	PASS
198	198	100	100	100.0	121221	PASS
199	198	5	9	6.5	7859	PASS
275	198	10	60	29.8	36138	PASS
365	198	1	100	4.1	5030	PASS
441	443	0.01	100	78.9	16959	PASS
442	198	50	100	90.4	109592	PASS
443	442	15	24	19.6	21501	PASS

Scan 2451 (17.504 min): BM013158.D
DFTPP55

m/z	Abundance
44.00	2580.0
48.90	432.0
146.90	371.0
207.00	1179.0
221.00	328.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

m/z	Abundance
37.10	325.0
38.00	724.0
39.10	4282.0
40.00	430.0
43.90	2358.0
48.90	481.0
50.00	16240.0
51.00	58720.0
52.00	2675.0
55.10	447.0
56.00	1636.0
57.00	3032.0
61.00	622.0
61.90	570.0
63.00	2040.0
64.00	381.0
64.90	1205.0
68.10	1034.0
69.00	58656.0
73.00	874.0
74.00	6219.0
75.00	9746.0
76.10	3280.0
77.00	61328.0
78.00	3908.0
78.90	4723.0
79.90	3843.0
81.00	4421.0
82.00	1533.0
83.00	731.0
84.00	314.0
85.00	1084.0
86.00	1489.0
87.10	335.0
88.00	464.0
91.00	1460.0
91.90	2614.0
93.00	10101.0
93.90	867.0
98.10	7524.0
99.00	6408.0
100.20	462.0
101.00	1990.0
102.80	641.0
103.90	1626.0
105.00	2443.0
106.20	608.0
107.00	20328.0
108.00	3192.0
109.00	933.0
110.00	30648.0
111.00	5468.0
111.80	327.0
112.20	457.0
115.90	1605.0
117.00	21816.0
118.00	1850.0
120.10	376.0
122.00	951.0
123.00	2436.0
124.00	644.0
124.70	514.0
125.20	326.0
127.00	64488.0
128.00	4585.0
129.00	26984.0
130.00	2426.0
130.90	424.0
134.10	1052.0
134.90	2501.0
135.90	571.0
137.00	1779.0
140.00	602.0
141.00	4178.0
141.90	719.0
142.90	765.0
144.10	365.0
145.90	707.0
146.90	1913.0
147.90	4069.0
148.90	937.0
150.10	542.0
151.10	962.0
153.00	875.0
154.00	1246.0
155.00	2418.0
155.90	3070.0
156.90	650.0
157.90	924.0
159.10	680.0
160.00	1511.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

160.90	1254.0
162.10	387.0
164.90	1318.0
166.10	691.0
167.00	7783.0
168.00	3788.0
169.00	321.0
171.90	1055.0
173.00	1880.0
174.00	1297.0
175.00	2608.0
176.20	498.0
177.00	1935.0
178.00	485.0
179.00	5352.0
180.00	3688.0
181.00	1617.0
183.90	632.0
185.00	2085.0
186.00	19576.0
187.00	5804.0
188.10	322.0
188.90	1342.0
190.90	502.0
192.00	1715.0
193.00	1974.0
193.90	331.0
194.90	301.0
195.90	4294.0
196.80	867.0
198.00	123832.0
198.90	9257.0
199.90	927.0
201.40	308.0
203.10	969.0
204.00	5461.0
205.10	7769.0
206.00	30272.0
207.00	6415.0
208.10	1267.0
209.00	866.0
209.90	412.0
210.90	1732.0
212.00	428.0
215.90	886.0
217.00	8340.0
217.90	850.0
220.90	4264.0
221.90	1283.0
223.00	2436.0
224.00	20808.0
225.10	5371.0
225.90	494.0
227.00	10920.0
227.90	1195.0
228.90	1631.0
229.90	354.0
230.80	711.0
231.20	502.0
234.10	831.0
234.90	1340.0
236.00	952.0
236.90	805.0
238.90	362.0
241.00	356.0
241.70	844.0
242.10	623.0
243.10	758.0
244.00	14839.0
245.10	2369.0
246.00	4249.0
247.10	865.0
249.00	313.0
253.10	680.0
254.10	583.0
255.00	81352.0
256.00	12793.0
257.00	1764.0
258.00	5457.0
259.10	723.0
265.00	2268.0
272.90	2302.0
274.00	7320.0
275.00	36016.0
276.00	4965.0
276.90	4137.0
278.10	505.0
281.10	471.0
281.90	421.0
283.10	587.0
284.90	504.0
292.90	450.0
296.00	10366.0
297.00	1742.0
303.10	1751.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

303.90	361.0
313.80	624.0
315.00	1106.0
316.00	787.0
321.00	374.0
322.00	317.0
323.10	2787.0
324.00	1024.0
327.00	1156.0
328.00	470.0
333.10	386.0
334.00	2370.0
335.10	725.0
341.00	506.0
345.80	1055.0
351.90	584.0
353.00	804.0
354.20	698.0
365.00	4208.0
365.90	819.0
372.00	1730.0
383.00	863.0
402.10	411.0
402.90	650.0
404.20	374.0
421.10	624.0
422.10	677.0
423.10	4065.0
424.10	1340.0
441.10	13925.0
442.00	86512.0
443.10	18344.0
444.10	2138.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

m/z	Abundance
37.00	311.0
38.00	581.0
39.10	4524.0
39.90	617.0
44.00	1244.0
49.00	784.0
50.00	15746.0
51.10	56888.0
52.00	2375.0
55.00	931.0
55.90	1666.0
57.00	4213.0
61.10	889.0
62.00	841.0
63.00	1993.0
64.10	369.0
65.00	918.0
68.00	1072.0
69.00	59064.0
73.00	1206.0
74.10	7706.0
75.00	10280.0
76.00	3555.0
77.00	62464.0
78.00	3960.0
79.00	6079.0
80.00	4284.0
81.00	5093.0
82.10	509.0
82.70	508.0
83.90	366.0
85.00	803.0
86.00	1605.0
87.00	815.0
91.00	1320.0
92.00	1375.0
93.00	9634.0
93.90	817.0
95.90	475.0
98.00	8362.0
99.00	4374.0
99.90	407.0
100.20	366.0
101.00	3030.0
102.90	964.0
104.00	2106.0
105.00	2144.0
106.00	582.0
107.00	22096.0
108.00	3626.0
108.90	727.0
110.00	34000.0
111.00	6253.0
112.10	1153.0
112.80	303.0
116.00	1704.0
117.00	23888.0
118.00	1222.0
119.10	338.0
120.10	303.0
121.80	1725.0
123.00	2562.0
123.90	1048.0
125.00	978.0
126.10	332.0
127.00	62856.0
128.00	5129.0
129.00	29552.0
130.00	2878.0
131.00	540.0
134.00	1131.0
134.90	2444.0
135.80	640.0
137.10	1297.0
137.80	591.0
139.90	352.0
141.00	4399.0
142.00	1209.0
142.80	558.0
143.90	351.0
145.90	805.0
146.90	1759.0
148.00	5084.0
148.90	918.0
150.80	420.0
151.80	669.0
152.90	1717.0
154.00	677.0
155.00	3721.0
156.00	3337.0
158.10	814.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

158.90	813.0
160.10	1534.0
161.00	2050.0
162.10	526.0
163.10	358.0
165.00	1631.0
165.90	1669.0
167.00	9041.0
168.00	5790.0
169.00	533.0
170.80	380.0
171.80	432.0
172.10	527.0
173.00	771.0
174.00	2147.0
175.10	2965.0
176.10	874.0
176.90	1911.0
177.90	723.0
178.90	6137.0
180.00	4985.0
180.90	2169.0
182.10	340.0
183.80	734.0
185.10	2929.0
186.00	24376.0
187.00	7378.0
188.00	611.0
189.00	2770.0
189.80	480.0
190.90	1003.0
192.00	1836.0
193.10	2193.0
194.00	921.0
196.00	4319.0
197.10	1346.0
197.90	147840.0
199.00	10224.0
200.00	917.0
203.00	1936.0
204.10	5892.0
205.00	11346.0
206.00	40456.0
207.00	7636.0
208.10	2039.0
209.00	733.0
210.20	686.0
211.00	1628.0
215.00	760.0
216.00	775.0
216.90	14403.0
218.00	1491.0
221.00	5271.0
222.00	1166.0
223.00	2374.0
224.00	22176.0
225.10	6563.0
225.90	1097.0
227.00	11414.0
228.10	2141.0
229.00	1995.0
231.00	705.0
234.00	758.0
235.10	1145.0
236.00	485.0
237.00	1157.0
239.00	388.0
240.00	479.0
241.00	749.0
242.00	1654.0
243.20	1580.0
244.10	16880.0
245.00	2865.0
245.90	4745.0
246.90	1310.0
248.90	414.0
249.20	403.0
250.10	321.0
253.00	890.0
254.10	503.0
255.00	97872.0
256.00	14342.0
257.10	1365.0
258.00	7652.0
258.90	897.0
264.90	3144.0
266.00	715.0
273.10	3109.0
274.00	7245.0
275.00	43048.0
276.00	5850.0
277.00	4035.0
278.00	785.0
280.90	566.0
283.00	810.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

284.90	514.0
292.90	1152.0
296.00	16026.0
297.00	2299.0
303.00	1746.0
303.90	647.0
309.90	402.0
314.10	722.0
314.90	1801.0
316.00	661.0
320.90	478.0
322.00	392.0
323.10	3893.0
324.10	1321.0
326.90	1153.0
328.00	419.0
331.80	318.0
333.00	701.0
334.00	3112.0
335.10	546.0
341.00	595.0
345.90	787.0
352.00	1033.0
353.00	1008.0
354.10	1764.0
365.00	5883.0
365.90	901.0
371.10	385.0
372.00	2459.0
382.90	380.0
390.00	567.0
402.10	628.0
402.90	734.0
403.80	368.0
421.10	1142.0
422.00	1423.0
423.10	6945.0
424.00	2182.0
441.10	19296.0
442.10	131712.0
443.10	24520.0
444.10	2237.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

m/z	Abundance
37.90	547.0
39.10	2307.0
39.80	467.0
44.00	2566.0
49.00	431.0
50.00	8423.0
51.00	25416.0
52.00	1871.0
56.00	1110.0
57.00	1984.0
60.90	709.0
62.00	365.0
63.10	1039.0
65.10	432.0
67.80	541.0
69.00	29000.0
73.00	600.0
73.90	3647.0
75.00	5723.0
76.00	1961.0
77.10	36648.0
78.10	2619.0
79.00	3682.0
80.00	2028.0
81.10	2431.0
82.00	703.0
83.10	756.0
83.80	363.0
86.00	1016.0
87.00	482.0
91.10	728.0
92.00	698.0
93.00	6328.0
98.00	3971.0
98.90	1813.0
100.10	419.0
101.00	2220.0
103.00	738.0
103.90	794.0
105.00	1048.0
107.00	10655.0
108.00	1825.0
109.00	353.0
110.00	19392.0
111.00	3187.0
116.00	1195.0
117.00	15194.0
118.10	1229.0
121.80	1170.0
122.90	1326.0
123.80	1034.0
124.90	484.0
126.10	353.0
127.00	39016.0
127.90	2752.0
129.00	17384.0
130.00	1467.0
131.00	369.0
132.80	468.0
133.90	571.0
135.00	1058.0
136.00	612.0
136.80	425.0
137.00	421.0
140.90	2228.0
142.00	963.0
142.80	569.0
146.00	430.0
147.10	1494.0
148.00	3338.0
149.10	417.0
151.20	395.0
152.90	803.0
154.00	398.0
155.00	1353.0
156.10	1748.0
157.00	508.0
158.00	423.0
159.10	347.0
159.90	649.0
160.90	1816.0
161.90	528.0
165.10	1158.0
166.10	474.0
167.00	6796.0
168.00	2905.0
168.90	667.0
173.00	821.0
173.80	775.0
174.20	615.0
175.00	1515.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

176.10	685.0
176.90	700.0
177.50	379.0
178.90	2988.0
179.90	2610.0
181.00	1430.0
184.90	2315.0
186.10	12798.0
187.10	3253.0
188.00	574.0
189.00	1483.0
191.10	683.0
192.00	1073.0
192.90	2353.0
194.20	314.0
196.00	2559.0
196.80	610.0
198.00	91992.0
198.90	4097.0
199.80	442.0
201.30	607.0
202.90	468.0
204.00	3132.0
205.00	5830.0
206.00	28232.0
207.00	5313.0
208.00	1058.0
209.00	597.0
211.00	1267.0
215.10	717.0
216.00	721.0
217.00	8851.0
218.00	1703.0
221.00	3651.0
221.80	925.0
223.00	1959.0
224.00	15589.0
225.00	4389.0
225.90	373.0
227.00	9255.0
228.00	1144.0
228.90	1653.0
230.90	732.0
234.00	394.0
234.70	333.0
235.90	437.0
237.00	400.0
238.90	373.0
241.10	417.0
242.00	913.0
242.80	565.0
243.30	616.0
244.00	12448.0
245.00	1402.0
245.90	3547.0
246.90	474.0
248.90	347.0
250.00	460.0
253.00	464.0
254.10	708.0
255.00	64944.0
256.00	9542.0
257.00	1211.0
258.00	5804.0
259.00	640.0
265.00	1655.0
273.00	2293.0
274.00	4495.0
275.00	29352.0
276.00	3829.0
277.00	4175.0
278.00	808.0
280.90	544.0
284.00	541.0
285.00	501.0
292.90	857.0
296.00	11209.0
297.10	1722.0
302.80	1177.0
304.10	311.0
314.00	754.0
314.90	1662.0
316.00	893.0
322.10	347.0
323.00	3333.0
324.00	487.0
327.00	553.0
334.00	2508.0
335.00	839.0
340.90	595.0
345.90	920.0
347.00	356.0
352.00	1212.0
353.20	993.0
354.00	1353.0

Instrument :

BNA_M

ClientSampleId :

DFTPP55

355.00	468.0
365.00	4999.0
365.90	745.0
372.00	2278.0
402.00	1009.0
402.90	1302.0
403.70	361.0
420.90	788.0
422.10	864.0
423.00	6048.0
424.20	1700.0
425.10	469.0
441.10	17656.0
442.10	110552.0
443.10	21640.0
444.10	2113.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55