

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM120420\
 Data File : BM028039.D
 Acq On : 04 Dec 2020 20:20
 Operator : JU/CG
 Sample : PB133360BL
 Misc : ME-SOIL-BL-03
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK60

Quant Time: Dec 05 00:22:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM120320.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 04 23:57:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.29	152	230355	20.00	ng/ul	0.00
20) Naphthalene-d8	11.12	136	962053	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.91	164	573993	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.63	188	1197100	20.00	ng/ul	0.00
79) Chrysene-d12	21.73	240	1154385	20.00	ng/ul	0.00
88) Perylene-d12	24.14	264	1129071	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.48	96	36125	5.85	ng/uL	0.00
4) Pyridine-d5	3.93	84	498544	29.96	ng/ul	0.00
7) Phenol-d5	7.42	99	605418	31.57	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	7.60	67	378518	31.76	ng/ul	0.00
11) 2-Chlorophenol-d4	7.81	132	507415	31.77	ng/ul	0.00
15) 4-Methylphenol-d8	8.99	113	485187	32.08	ng/ul	0.00
21) Nitrobenzene-d5	9.47	128	240715	32.23	ng/ul	0.00
24) 2-Nitrophenol-d4	10.19	143	251168	33.65	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.73	165	461725	30.50	ng/ul	0.00
31) 4-Chloroaniline-d4	11.25	131	694120	33.52	ng/ul	0.00
46) Dimethylphthalate-d6	14.30	166	1488376	35.05	ng/ul	0.00
49) Acenaphthylene-d8	14.61	160	1674969	30.73	ng/ul	0.00
54) 4-Nitrophenol-d4	15.07	143	259739	31.71	ng/ul	0.00
60) Fluorene-d10	15.89	176	1226187	32.22	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.99	200	196741	28.32	ng/ul	0.00
73) Anthracene-d10	17.73	188	1907823	32.51	ng/ul	0.00
81) Pyrene-d10	19.98	212	2106051	34.75	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.99	264	2071757	34.47	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM120420\
Data File : BM028039.D
Acq On : 04 Dec 2020 20:20
Operator : JU/CG
Sample : PB133360BL
Misc : ME-SOIL-BL-03
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK60

Quant Time: Dec 05 00:22:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM120320.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 04 23:57:08 2020
Response via : Initial Calibration

