


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM120420\
 Data File : BM028020.D
 Acq On : 04 Dec 2020 08:50
 Operator : JU/CG
 Sample : DFTPP01
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP011

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM120320.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 07 08:15:07 2020



AutoFind: Scans 2523, 2524, 2525; Background Corrected with Scan 2517

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.8	119960	PASS
68	69	0.00	2	1.9	2257	PASS
69	198	0.00	100	43.9	117637	PASS
70	69	0.00	2	0.5	547	PASS
127	198	10	80	51.5	138056	PASS
197	198	0.00	2	0.7	1922	PASS
198	198	100	100	100.0	267818	PASS
199	198	5	9	6.7	18034	PASS
275	198	10	60	27.6	73842	PASS
365	198	1	100	3.3	8933	PASS
441	198	0.01	100	13.5	36080	PASS
442	198	50	100	91.0	243712	PASS
443	442	15	24	19.5	47434	PASS

m/z	Abundance
40.00	866.0
40.90	197.0
43.00	188.0
44.00	731.0
45.00	167.0
48.80	298.0
73.00	1045.0
78.10	215.0
95.90	237.0
132.90	323.0
147.10	263.0
166.00	181.0
167.00	720.0
190.80	319.0
192.90	322.0
207.00	2982.0
208.00	686.0
208.90	351.0
221.10	169.0
281.00	992.0
355.10	220.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

m/z	Abundance
37.00	536.0
38.10	1576.0
39.10	8629.0
39.90	915.0
41.00	354.0
44.00	868.0
44.90	301.0
49.00	1122.0
50.00	29576.0
51.10	111928.0
52.00	5644.0
53.00	285.0
55.00	769.0
56.00	3670.0
57.00	7767.0
57.90	374.0
59.00	217.0
61.00	1397.0
62.00	1495.0
63.00	4010.0
64.10	497.0
65.00	1988.0
66.00	191.0
68.10	2239.0
69.00	108920.0
70.00	488.0
73.10	1942.0
74.00	10659.0
75.00	16382.0
76.00	5563.0
77.10	113536.0
78.10	8432.0
79.00	7661.0
80.00	5983.0
81.00	8358.0
82.00	2119.0
83.00	1683.0
83.90	369.0
85.00	1311.0
86.00	2424.0
87.00	1178.0
88.00	451.0
88.90	252.0
91.00	1545.0
92.00	2674.0
93.00	14075.0
94.00	863.0
95.00	201.0
96.00	939.0
97.10	243.0
98.00	10766.0
99.00	9202.0
100.10	603.0
101.00	4702.0
102.10	251.0
103.00	1759.0
104.00	2789.0
105.00	2408.0
106.10	727.0
107.00	32040.0
108.00	5630.0
109.10	1125.0
110.00	61400.0
111.00	9831.0
112.00	1189.0
113.00	339.0
114.90	229.0
116.00	2081.0
117.00	28392.0
118.00	2384.0
119.00	397.0
119.90	541.0
122.00	2138.0
123.00	3234.0
123.90	1777.0
125.00	1422.0
127.00	122712.0
128.00	9966.0
129.00	46792.0
130.00	4292.0
131.00	667.0
132.00	348.0
132.90	565.0
134.00	1341.0
135.00	3759.0
136.00	1817.0
137.10	2285.0
137.90	430.0
139.00	229.0
140.00	419.0
141.00	5776.0
141.90	1916.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

143.00	1340.0
144.00	520.0
145.10	379.0
145.90	988.0
147.00	3821.0
148.00	7548.0
149.00	1289.0
150.10	397.0
151.10	937.0
152.00	388.0
153.00	2005.0
154.00	1301.0
155.00	2676.0
156.00	4689.0
157.00	837.0
157.90	1345.0
159.10	799.0
160.00	2001.0
161.00	2318.0
162.00	1139.0
162.90	251.0
163.90	254.0
165.00	2263.0
166.00	1508.0
167.00	11499.0
168.00	7532.0
169.00	1244.0
170.00	407.0
170.90	506.0
171.90	1092.0
173.00	1317.0
174.00	2086.0
175.00	3881.0
176.00	1980.0
177.00	2132.0
178.00	901.0
179.00	8816.0
180.00	6700.0
181.00	2868.0
182.10	284.0
183.00	373.0
184.00	632.0
185.00	4268.0
186.00	28920.0
187.00	8471.0
188.00	1145.0
189.00	2014.0
190.10	342.0
191.00	1391.0
192.00	2674.0
193.00	2943.0
194.00	786.0
195.10	388.0
196.00	6957.0
197.10	1736.0
198.00	216640.0
199.00	15487.0
200.00	1170.0
201.50	1014.0
202.90	1542.0
204.00	7424.0
205.00	12450.0
206.10	56440.0
207.10	11095.0
208.00	2415.0
209.00	1031.0
210.00	1342.0
211.10	2128.0
211.90	304.0
213.00	214.0
214.90	673.0
216.00	1427.0
217.00	14128.0
218.00	2182.0
219.00	293.0
220.20	290.0
221.00	9374.0
223.00	3389.0
224.00	30528.0
225.00	8143.0
226.00	1212.0
227.00	12271.0
228.00	1852.0
229.00	2859.0
230.00	378.0
231.00	1399.0
231.90	182.0
232.80	285.0
234.00	837.0
235.00	960.0
235.90	752.0
237.00	1826.0
238.10	287.0
239.00	618.0
240.00	311.0
241.00	864.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

242.00	1832.0
243.00	1683.0
244.10	23184.0
245.10	2855.0
246.00	4212.0
247.00	938.0
248.00	152.0
249.00	738.0
251.00	262.0
252.00	257.0
252.80	279.0
254.20	1027.0
255.00	116040.0
256.00	17560.0
257.00	1688.0
258.00	7614.0
259.00	923.0
260.00	189.0
260.90	216.0
263.90	210.0
265.00	2983.0
265.90	499.0
267.00	254.0
271.00	318.0
271.80	259.0
272.20	215.0
273.00	3388.0
274.00	8839.0
275.00	56144.0
276.00	7374.0
277.00	4112.0
277.90	1046.0
278.90	219.0
281.00	983.0
281.90	386.0
283.10	838.0
284.10	314.0
285.00	859.0
291.90	183.0
293.00	1115.0
294.10	442.0
294.90	339.0
296.00	14131.0
297.00	2012.0
297.90	185.0
301.00	293.0
302.00	262.0
303.00	1681.0
304.00	450.0
307.80	258.0
308.90	192.0
309.90	248.0
314.00	721.0
314.90	1833.0
315.90	970.0
316.90	190.0
321.00	413.0
321.90	370.0
323.00	4519.0
324.10	813.0
324.90	218.0
326.90	938.0
328.10	466.0
331.90	411.0
333.00	630.0
334.00	2898.0
335.10	616.0
341.10	618.0
342.00	306.0
346.00	950.0
347.00	206.0
351.90	1528.0
353.00	1036.0
354.10	1523.0
354.80	392.0
359.10	187.0
365.00	5555.0
365.90	930.0
371.00	420.0
372.00	2461.0
372.90	756.0
383.00	772.0
384.00	211.0
390.00	353.0
391.10	363.0
392.10	162.0
402.10	980.0
403.10	1335.0
404.00	739.0
405.10	234.0
421.10	1219.0
422.00	1200.0
423.00	8302.0
424.00	3728.0
424.90	735.0
441.00	22152.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

442.00	149568.0
443.00	30840.0
444.00	3255.0
444.90	208.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

m/z	Abundance
36.00	158.0
37.00	560.0
38.10	1499.0
39.00	11100.0
39.90	1335.0
41.10	296.0
43.00	151.0
44.00	846.0
44.90	361.0
49.10	1146.0
50.10	36904.0
51.10	141248.0
52.00	7168.0
53.00	362.0
55.00	726.0
56.00	4521.0
57.00	9572.0
58.00	451.0
58.90	153.0
61.00	1720.0
62.10	2143.0
63.00	5670.0
64.00	683.0
65.00	2657.0
66.00	231.0
68.10	2597.0
69.00	137984.0
70.00	644.0
73.10	2099.0
74.10	14086.0
75.00	22296.0
76.10	7266.0
77.10	148288.0
78.10	10364.0
79.00	10440.0
80.00	7269.0
81.00	10454.0
82.00	2612.0
83.00	2426.0
84.10	248.0
85.00	1524.0
86.00	2838.0
87.00	1466.0
87.90	663.0
89.00	246.0
91.00	2458.0
92.00	2919.0
93.00	17816.0
94.00	1076.0
95.00	301.0
96.10	867.0
96.80	390.0
98.00	12884.0
99.00	11705.0
100.00	1178.0
101.00	5654.0
102.00	343.0
103.00	2273.0
104.00	4428.0
105.00	3945.0
106.10	1283.0
107.00	44096.0
108.00	6458.0
109.10	1473.0
110.00	80360.0
111.00	11876.0
112.00	1211.0
112.90	447.0
115.00	325.0
116.00	2452.0
117.00	37920.0
118.00	3159.0
118.90	315.0
120.10	572.0
120.90	192.0
122.00	3160.0
123.00	4653.0
124.00	2139.0
125.00	1877.0
126.20	1120.0
127.00	159360.0
128.00	12138.0
129.00	60480.0
130.00	5313.0
130.90	1144.0
132.00	477.0
133.00	610.0
134.00	1781.0
135.00	5033.0
136.00	2114.0
137.00	2435.0
138.00	564.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

138.90	477.0
140.00	761.0
141.00	8377.0
142.00	2256.0
142.90	1607.0
143.90	448.0
145.00	493.0
146.00	1426.0
147.00	4641.0
148.00	9736.0
149.00	1974.0
150.00	543.0
151.00	968.0
151.90	596.0
153.00	2617.0
154.00	1839.0
155.00	4431.0
156.00	5575.0
157.00	1254.0
158.00	1542.0
159.00	956.0
160.00	2462.0
161.00	3409.0
162.00	1064.0
162.90	275.0
163.90	439.0
165.00	3195.0
166.00	2584.0
167.00	16086.0
168.00	9576.0
169.00	1374.0
170.00	848.0
171.10	783.0
172.00	1374.0
173.00	2288.0
174.00	3219.0
175.10	6426.0
176.00	2045.0
177.00	3329.0
177.90	996.0
179.00	11979.0
180.00	8927.0
181.00	3905.0
182.00	499.0
182.90	279.0
183.90	736.0
185.00	5021.0
186.00	43184.0
187.00	11898.0
188.00	1225.0
189.00	2819.0
189.90	418.0
191.00	1738.0
192.00	3703.0
193.00	4339.0
194.00	1075.0
195.00	400.0
196.00	10315.0
197.10	2081.0
198.00	311808.0
199.00	21320.0
200.00	1639.0
201.50	1481.0
203.00	2486.0
204.00	11198.0
205.00	18432.0
206.10	80776.0
207.10	13341.0
208.00	3590.0
209.10	1277.0
210.00	1379.0
211.10	3018.0
211.90	668.0
212.90	179.0
215.00	982.0
216.00	1900.0
217.00	20272.0
218.00	2581.0
219.00	274.0
219.80	298.0
221.10	15008.0
222.00	2903.0
223.00	5109.0
224.00	43592.0
225.10	11620.0
226.10	1450.0
227.00	16664.0
228.00	2722.0
229.00	3773.0
230.00	529.0
231.00	1660.0
232.00	206.0
232.90	332.0
234.00	1142.0
235.00	1496.0
236.00	945.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

237.00	2820.0
237.90	282.0
239.00	864.0
240.00	645.0
241.00	1183.0
242.00	1995.0
243.10	2492.0
244.10	33840.0
245.10	4744.0
246.00	7235.0
247.00	1667.0
248.00	224.0
249.00	1380.0
250.00	372.0
250.90	279.0
251.90	349.0
252.90	890.0
254.20	1932.0
255.00	178752.0
256.00	27728.0
257.00	1869.0
258.00	9892.0
259.00	1466.0
260.10	247.0
260.90	322.0
264.00	383.0
265.00	3637.0
265.90	572.0
266.90	189.0
270.00	197.0
270.90	405.0
272.10	430.0
273.00	5462.0
274.00	14467.0
275.00	85424.0
276.00	10826.0
277.00	6572.0
278.00	1372.0
278.90	398.0
281.10	1070.0
282.10	316.0
283.00	1087.0
284.00	576.0
285.10	1027.0
289.00	291.0
290.00	296.0
291.00	180.0
291.90	358.0
293.00	1337.0
294.00	484.0
294.90	525.0
296.00	21096.0
297.00	3013.0
297.90	273.0
300.90	298.0
302.00	350.0
303.00	2404.0
303.90	735.0
305.00	209.0
308.00	261.0
308.90	276.0
310.10	311.0
310.90	164.0
313.00	298.0
314.00	1508.0
315.00	2355.0
316.00	1365.0
317.00	273.0
321.00	737.0
322.00	382.0
323.00	7183.0
324.00	1370.0
325.10	195.0
325.80	163.0
326.90	1221.0
328.00	835.0
329.00	212.0
331.90	372.0
333.00	635.0
334.00	4508.0
335.00	989.0
341.10	1012.0
342.10	215.0
346.00	1857.0
347.00	358.0
352.00	2199.0
353.00	1581.0
354.00	2548.0
355.00	626.0
359.00	166.0
365.00	10241.0
366.00	1294.0
371.00	609.0
372.10	3811.0
373.00	906.0
383.00	1133.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

383.90	180.0
384.90	175.0
390.00	472.0
391.10	262.0
392.00	196.0
401.00	192.0
402.00	1674.0
403.10	2382.0
403.90	977.0
404.90	189.0
415.10	191.0
421.00	2402.0
422.00	2296.0
423.00	17416.0
424.10	5462.0
425.00	747.0
440.20	241.0
441.00	40168.0
442.00	277440.0
443.00	53048.0
444.00	4946.0
445.00	436.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

m/z	Abundance
37.00	414.0
38.10	1672.0
39.10	7985.0
39.90	826.0
40.90	212.0
44.00	987.0
44.90	326.0
49.10	814.0
50.00	27224.0
51.10	106704.0
52.00	5332.0
53.00	252.0
55.00	691.0
56.00	3356.0
57.00	7798.0
58.00	336.0
59.00	207.0
61.00	1469.0
62.00	1491.0
63.00	4084.0
64.00	600.0
65.00	1988.0
66.90	181.0
68.10	1936.0
69.00	106008.0
69.90	511.0
71.00	181.0
73.00	1839.0
74.00	10060.0
75.00	16704.0
76.10	5669.0
77.10	112568.0
78.10	7698.0
79.00	8146.0
80.00	5580.0
81.00	8732.0
82.00	1954.0
83.00	2031.0
83.90	219.0
85.00	1535.0
86.00	2378.0
86.90	1106.0
88.00	416.0
89.00	205.0
91.00	2383.0
92.00	2231.0
93.00	13783.0
94.00	1115.0
95.10	242.0
96.00	773.0
98.00	10900.0
99.00	9283.0
100.00	840.0
101.00	4882.0
101.90	218.0
103.00	1772.0
104.00	3056.0
105.00	2924.0
106.00	1048.0
107.00	35368.0
108.00	5215.0
109.20	1030.0
110.00	65008.0
111.00	9758.0
112.00	1467.0
113.00	412.0
115.20	169.0
116.00	1985.0
117.00	29680.0
118.00	2247.0
118.90	302.0
119.90	371.0
121.10	164.0
122.00	2524.0
123.00	3799.0
124.10	1668.0
125.00	1385.0
127.00	132096.0
128.00	9812.0
129.00	50528.0
130.00	3886.0
131.00	846.0
132.00	336.0
133.00	563.0
134.00	1434.0
135.00	3942.0
136.00	1742.0
137.00	2615.0
138.10	308.0
138.80	290.0
140.00	748.0
141.00	7585.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

142.00	1901.0
143.00	1384.0
144.10	376.0
144.90	395.0
146.00	1121.0
147.00	3401.0
148.00	7584.0
149.00	1780.0
149.90	455.0
151.00	950.0
151.70	582.0
152.90	2059.0
154.00	1516.0
155.00	3942.0
156.10	4859.0
157.00	964.0
157.90	1524.0
158.90	845.0
160.00	2217.0
161.00	2694.0
162.00	953.0
162.90	255.0
164.00	533.0
165.00	2486.0
166.00	2509.0
167.00	13262.0
168.00	7741.0
169.00	1179.0
170.00	472.0
171.00	527.0
172.00	1131.0
173.00	1455.0
174.00	2855.0
175.00	5072.0
176.00	1767.0
177.00	2716.0
178.10	776.0
179.00	9884.0
180.00	7380.0
181.00	2621.0
182.00	478.0
182.90	209.0
184.00	824.0
185.00	4210.0
186.00	36088.0
187.00	10382.0
188.00	1044.0
189.00	2532.0
189.90	436.0
191.00	1363.0
192.00	3361.0
193.00	3847.0
194.00	738.0
194.90	685.0
196.00	8283.0
197.10	1949.0
198.00	275008.0
199.00	17296.0
200.00	1472.0
201.50	1441.0
203.00	1781.0
204.00	9130.0
205.10	16632.0
206.10	71488.0
207.10	12097.0
208.10	3074.0
209.10	1312.0
210.00	1555.0
211.10	2991.0
212.90	257.0
214.90	860.0
216.10	1289.0
217.00	20272.0
218.00	2376.0
219.00	331.0
221.00	12620.0
221.80	2836.0
223.00	3967.0
224.10	38712.0
225.00	9859.0
226.00	1133.0
227.00	15332.0
228.00	2418.0
229.00	3885.0
229.90	485.0
231.00	1549.0
232.00	191.0
232.90	259.0
234.00	1231.0
235.00	1187.0
236.00	641.0
237.00	1906.0
237.90	253.0
239.00	850.0
239.90	585.0
241.00	738.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

242.00	1971.0
243.00	2352.0
244.10	32200.0
245.00	4696.0
246.00	6323.0
247.00	1008.0
248.10	256.0
249.00	1247.0
250.10	344.0
251.00	328.0
251.80	232.0
253.10	641.0
254.20	1807.0
255.00	166336.0
256.00	23024.0
257.00	1917.0
258.00	9158.0
259.00	1474.0
261.00	255.0
263.00	178.0
264.00	346.0
265.00	3504.0
265.90	595.0
267.10	234.0
270.00	297.0
271.00	350.0
272.00	456.0
273.00	4566.0
274.00	13365.0
275.00	79960.0
276.00	10482.0
277.00	5853.0
278.00	1183.0
278.90	190.0
281.00	963.0
282.00	336.0
283.10	711.0
283.90	716.0
284.90	1132.0
286.00	278.0
289.00	210.0
290.00	179.0
291.00	198.0
292.10	242.0
292.90	1320.0
294.00	376.0
295.10	528.0
296.00	19992.0
297.10	2491.0
297.90	208.0
300.90	243.0
302.00	494.0
303.00	1980.0
304.00	848.0
308.00	259.0
308.90	325.0
310.00	349.0
312.90	197.0
314.00	889.0
315.00	2243.0
316.00	1326.0
317.00	158.0
321.10	722.0
322.10	373.0
323.00	6991.0
324.00	1531.0
325.00	191.0
327.00	1246.0
328.00	673.0
328.90	150.0
331.90	452.0
333.00	686.0
334.00	4602.0
335.10	1146.0
336.00	243.0
341.10	937.0
342.00	246.0
345.90	1425.0
347.00	514.0
352.10	2373.0
353.00	1624.0
354.00	2673.0
355.00	471.0
365.00	11005.0
366.00	1853.0
370.00	355.0
371.10	650.0
372.00	4280.0
373.00	1198.0
373.90	184.0
383.00	1165.0
384.10	284.0
390.10	510.0
391.00	426.0
392.00	374.0
401.00	417.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011

402.00	1685.0
403.00	2373.0
404.00	978.0
405.10	270.0
421.00	2147.0
422.00	2148.0
423.00	16656.0
424.00	4812.0
425.10	528.0
438.90	166.0
440.10	421.0
441.00	45920.0
442.00	304128.0
443.00	58416.0
444.00	5354.0
445.00	156.0

Instrument :
BNA_M
ClientSampleId :
DFTPP011