


Data Path : Z:\HPCHEM1\BNA_M\Data\BM120517\
Data File : BM013185.D
Acq On : 05 Dec 2017 11:19
Operator : SJ/JU
Sample : DFTPP56
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP56

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM113017.M
Title : SVOA CALIBRATION
Last Update : Fri Dec 08 03:46:15 2017



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2451

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.5	96616	PASS
68	69	0.00	2	1.5	1561	PASS
69	198	0.00	100	40.8	104994	PASS
70	69	0.00	2	0.4	433	PASS
127	198	10	80	45.4	116893	PASS
197	198	0.00	2	0.5	1392	PASS
198	198	100	100	100.0	257450	PASS
199	198	5	9	6.5	16748	PASS
275	198	10	60	27.9	71746	PASS
365	198	1	100	3.8	9854	PASS
441	443	0.01	100	75.3	30402	PASS
442	198	50	100	82.7	212885	PASS
443	442	15	24	19.0	40354	PASS

Scan 2451 (17.504 min): BM013185.D
DFTPP56

m/z	Abundance
40.10	366.0
44.00	2779.0
73.20	337.0
207.10	1174.0
208.00	422.0
280.90	373.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
37.20	569.0
38.00	2006.0
39.10	6350.0
40.00	895.0
41.00	426.0
44.00	1655.0
45.10	431.0
49.00	968.0
50.00	26416.0
51.00	95224.0
52.00	4542.0
53.00	479.0
54.90	876.0
55.90	3495.0
57.00	5704.0
61.10	1153.0
62.00	1248.0
63.00	3711.0
63.90	790.0
65.00	1868.0
67.90	1708.0
68.10	1741.0
69.00	101936.0
69.90	439.0
73.10	1162.0
74.00	9040.0
75.10	14439.0
76.10	5688.0
77.10	105104.0
78.10	6051.0
79.00	8025.0
80.00	5876.0
81.00	10015.0
82.00	2910.0
83.00	2664.0
85.10	1243.0
86.10	3393.0
87.00	1039.0
87.90	463.0
91.10	2422.0
92.00	2072.0
93.10	13701.0
93.90	895.0
95.90	534.0
98.00	10934.0
99.00	8341.0
100.00	500.0
101.00	4465.0
101.90	305.0
103.00	1082.0
104.00	2686.0
105.00	3450.0
106.00	1107.0
107.00	35352.0
108.00	4617.0
108.90	1377.0
110.00	56808.0
111.00	9234.0
111.80	1035.0
113.00	907.0
115.00	692.0
116.00	2100.0
117.00	31920.0
118.00	2239.0
119.00	347.0
120.00	711.0
121.90	2149.0
123.00	3918.0
124.00	1563.0
124.90	1467.0
126.00	321.0
127.00	102112.0
128.00	8414.0
129.00	44088.0
129.90	3814.0
131.00	1153.0
132.00	308.0
132.80	359.0
133.90	1335.0
135.00	3296.0
135.90	2176.0
137.10	2321.0
138.20	680.0
140.00	484.0
141.00	5711.0
142.00	2141.0
142.90	1225.0
144.00	335.0
144.90	376.0
146.20	857.0
147.00	3430.0
147.90	6256.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

149.00	1556.0
150.00	321.0
150.90	1045.0
152.20	548.0
153.00	1487.0
154.10	1249.0
155.10	3487.0
156.00	4568.0
157.00	660.0
157.80	1338.0
159.00	1162.0
160.00	1846.0
160.90	2899.0
162.10	565.0
163.90	570.0
165.00	2895.0
166.00	1040.0
167.00	12751.0
168.00	5663.0
168.90	675.0
169.90	319.0
171.00	786.0
171.90	1163.0
172.90	1442.0
174.00	2563.0
175.10	5573.0
176.10	1127.0
176.90	1751.0
177.90	686.0
179.00	8392.0
179.90	5155.0
181.00	2387.0
181.90	649.0
183.80	698.0
185.00	4530.0
186.00	31024.0
187.00	9454.0
187.80	484.0
188.10	511.0
188.90	2483.0
190.80	927.0
192.00	3165.0
193.00	3602.0
194.00	429.0
194.90	423.0
196.00	5798.0
197.10	546.0
197.90	215232.0
199.00	15334.0
200.00	1415.0
201.40	928.0
203.00	2339.0
204.00	8928.0
205.00	14949.0
206.00	57768.0
207.00	10344.0
208.00	2241.0
208.90	1099.0
210.00	912.0
210.90	1923.0
211.90	323.0
215.00	946.0
216.10	1426.0
217.00	16342.0
217.90	2264.0
221.00	8044.0
222.00	2334.0
223.10	3077.0
224.10	30944.0
225.10	9726.0
226.00	974.0
227.00	16600.0
228.00	2229.0
229.00	3143.0
231.00	1146.0
233.00	506.0
233.90	1263.0
235.00	584.0
235.80	567.0
237.10	1696.0
237.90	375.0
238.90	418.0
240.10	524.0
240.80	1177.0
242.10	1277.0
243.00	1524.0
244.00	25360.0
245.00	3497.0
246.00	5805.0
247.10	968.0
249.00	727.0
251.10	339.0
251.90	377.0
252.90	609.0
254.10	1401.0
255.00	130648.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

256.00	19488.0
257.10	1642.0
258.00	7084.0
259.00	1507.0
265.00	4414.0
266.20	500.0
270.80	410.0
272.10	485.0
273.10	3737.0
274.10	7539.0
275.00	50296.0
276.00	7526.0
277.00	4426.0
278.00	784.0
281.00	622.0
282.80	310.0
284.00	417.0
285.00	791.0
293.00	858.0
294.00	448.0
295.10	377.0
296.00	15053.0
296.90	3080.0
297.80	463.0
302.00	712.0
303.00	1980.0
304.20	338.0
314.00	703.0
314.90	1694.0
316.00	836.0
320.90	849.0
321.80	309.0
323.00	4318.0
324.10	708.0
327.00	1249.0
328.00	795.0
331.90	388.0
332.90	350.0
334.00	2963.0
334.90	1081.0
340.90	508.0
346.10	687.0
352.10	2246.0
352.90	896.0
353.80	1117.0
355.20	457.0
365.00	5180.0
366.00	1172.0
372.00	2554.0
372.90	490.0
383.00	676.0
400.70	322.0
402.00	1024.0
403.00	1007.0
403.80	333.0
421.00	871.0
422.00	744.0
423.00	7380.0
424.00	1739.0
425.00	301.0
441.00	19232.0
442.10	131904.0
443.00	26928.0
443.90	1904.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
36.90	792.0
38.00	1933.0
39.00	9777.0
39.90	739.0
40.90	376.0
41.20	336.0
43.90	1987.0
45.00	335.0
49.00	1310.0
50.00	30640.0
51.10	116816.0
52.00	6543.0
53.00	399.0
54.90	952.0
55.90	3742.0
57.00	7195.0
58.00	547.0
61.00	1788.0
62.00	2124.0
63.10	4994.0
64.00	443.0
65.10	2893.0
66.00	319.0
68.00	2943.0
69.00	127792.0
70.10	1082.0
73.00	953.0
74.00	15666.0
75.00	20080.0
76.10	7256.0
77.00	129112.0
78.10	9342.0
79.00	11202.0
80.00	7782.0
81.00	9697.0
82.00	2424.0
83.10	2728.0
84.00	409.0
84.90	1603.0
86.10	3757.0
87.10	1963.0
87.90	749.0
91.10	2377.0
92.00	2868.0
93.00	19192.0
94.10	831.0
95.10	322.0
95.90	739.0
97.10	312.0
98.00	12298.0
99.00	11667.0
100.00	1569.0
101.00	4945.0
102.90	2477.0
104.00	4391.0
104.90	3366.0
106.10	1181.0
107.00	45280.0
108.00	7680.0
109.00	1506.0
110.00	75304.0
111.00	13183.0
112.00	1101.0
113.10	390.0
114.90	353.0
116.10	2778.0
117.00	39952.0
117.90	2095.0
119.10	493.0
120.00	458.0
121.90	2801.0
123.10	4399.0
123.90	1414.0
125.00	1838.0
126.10	775.0
127.00	143680.0
128.00	10411.0
129.00	58872.0
130.00	5320.0
131.00	851.0
132.00	812.0
133.30	473.0
134.00	1387.0
135.10	4396.0
136.10	2055.0
137.10	1796.0
139.80	1062.0
141.00	7649.0
142.00	2611.0
143.00	1191.0
144.10	401.0
145.00	327.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

146.00	1774.0
147.00	3589.0
148.00	10912.0
149.00	2871.0
151.10	1003.0
152.00	604.0
152.90	2598.0
154.00	2340.0
155.00	6570.0
156.10	4856.0
157.20	1184.0
157.80	1250.0
159.00	1434.0
159.90	2271.0
160.90	3333.0
162.10	923.0
163.20	537.0
163.90	316.0
165.00	3576.0
165.90	2403.0
167.00	15874.0
168.00	7805.0
169.00	1239.0
170.10	922.0
171.00	1000.0
171.90	1588.0
173.00	1668.0
174.00	2665.0
175.00	6660.0
176.10	1904.0
177.10	2804.0
178.00	813.0
178.90	12243.0
180.00	8775.0
181.00	3729.0
182.30	710.0
183.10	383.0
184.00	658.0
185.00	4681.0
186.00	47464.0
187.00	11941.0
188.00	968.0
189.00	2660.0
189.90	542.0
190.90	2120.0
191.90	3135.0
193.00	5457.0
194.10	1317.0
195.10	416.0
196.00	8016.0
197.10	1913.0
198.00	310528.0
199.00	18112.0
200.00	2008.0
201.30	1299.0
203.10	2267.0
204.00	11291.0
205.00	20648.0
206.00	91208.0
207.00	12725.0
208.00	3143.0
209.10	1244.0
210.10	1767.0
211.10	2579.0
215.00	956.0
216.00	2117.0
217.00	23912.0
217.90	2761.0
219.90	397.0
221.10	11947.0
222.00	1999.0
223.00	4612.0
224.00	46520.0
225.10	12674.0
226.20	1453.0
227.00	25064.0
228.00	3807.0
228.90	4556.0
230.00	1003.0
231.00	1568.0
233.00	607.0
234.00	1593.0
235.00	1585.0
236.00	1223.0
237.00	1182.0
237.90	314.0
239.00	707.0
240.10	517.0
241.10	1256.0
242.00	3211.0
243.10	3223.0
244.00	35328.0
245.00	5284.0
245.90	8248.0
247.10	1435.0
248.90	1338.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

250.00	437.0
251.00	503.0
251.90	756.0
252.80	673.0
253.20	483.0
254.10	1335.0
255.00	201024.0
256.00	27160.0
257.00	2592.0
258.00	12737.0
259.00	2319.0
261.00	526.0
264.00	421.0
265.00	5374.0
265.90	458.0
270.70	445.0
272.00	385.0
272.90	5132.0
274.00	14781.0
275.00	86112.0
276.00	11724.0
277.00	7709.0
278.00	837.0
280.90	771.0
283.10	1156.0
284.10	722.0
285.00	1588.0
291.80	373.0
293.00	2583.0
294.00	551.0
294.80	587.0
296.00	28152.0
297.00	3689.0
298.10	426.0
300.90	391.0
302.10	303.0
303.00	3296.0
304.00	768.0
307.70	323.0
309.90	365.0
313.80	518.0
315.10	1890.0
316.00	1634.0
321.10	559.0
323.10	8378.0
324.00	1725.0
327.00	1835.0
328.10	1015.0
331.90	304.0
333.10	1058.0
334.00	4998.0
335.00	1228.0
341.10	972.0
341.90	545.0
345.90	1138.0
346.90	456.0
352.00	1919.0
352.90	1893.0
354.10	2113.0
355.10	839.0
365.00	12415.0
366.00	1362.0
370.80	446.0
372.10	4281.0
373.00	1094.0
383.00	1257.0
384.20	444.0
390.10	633.0
391.20	414.0
402.10	1305.0
403.00	2794.0
403.80	831.0
421.10	1861.0
422.00	1685.0
423.10	13855.0
424.10	2794.0
425.10	440.0
441.10	35072.0
442.10	252480.0
443.00	45368.0
444.00	3864.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
38.00	1616.0
39.00	7133.0
39.90	872.0
43.10	366.0
44.00	2142.0
48.90	606.0
50.00	21184.0
51.00	77808.0
52.10	3740.0
55.00	469.0
56.00	2362.0
56.90	5288.0
60.90	1160.0
62.00	1206.0
62.90	3215.0
65.10	1819.0
67.90	1120.0
69.00	85256.0
69.90	860.0
73.00	1658.0
74.00	9882.0
75.00	14107.0
76.10	4881.0
77.00	91424.0
78.00	6050.0
79.00	8394.0
80.00	5798.0
81.00	8623.0
82.00	1940.0
83.00	1438.0
85.00	1556.0
85.90	2166.0
86.80	882.0
91.00	1695.0
92.10	2817.0
93.00	13412.0
94.10	601.0
96.10	899.0
98.00	11045.0
99.00	8630.0
100.00	343.0
101.00	4117.0
102.90	1159.0
104.00	2921.0
105.10	2260.0
106.00	1340.0
107.00	32128.0
108.00	3614.0
109.10	877.0
110.00	54568.0
110.90	9515.0
112.00	915.0
113.10	442.0
116.00	2502.0
117.00	32336.0
118.10	2069.0
119.00	401.0
119.90	607.0
121.00	538.0
121.90	2790.0
123.00	3605.0
123.90	1647.0
125.10	1467.0
126.10	849.0
127.00	104888.0
128.00	7711.0
129.00	42832.0
130.00	3820.0
130.90	805.0
132.20	480.0
134.00	1993.0
135.00	4140.0
136.00	1150.0
137.00	2206.0
138.00	711.0
139.80	675.0
141.00	5466.0
142.00	2525.0
142.90	1053.0
146.10	1344.0
146.90	3134.0
148.00	7513.0
148.90	1448.0
149.80	504.0
151.00	956.0
151.60	646.0
153.00	1971.0
153.90	1328.0
155.00	3865.0
156.00	6243.0
157.10	881.0
158.00	1221.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

159.00	962.0
160.00	2370.0
161.00	2525.0
161.90	987.0
164.00	531.0
165.00	3076.0
165.90	2145.0
167.00	14769.0
168.00	6735.0
168.90	1219.0
170.00	437.0
170.90	1000.0
172.00	1223.0
172.90	1772.0
174.10	2837.0
175.00	4198.0
175.90	1896.0
176.80	2775.0
178.00	820.0
179.00	9943.0
180.00	6383.0
181.00	3178.0
182.20	803.0
183.10	550.0
183.80	1146.0
185.00	5254.0
186.00	32096.0
187.00	10424.0
188.00	1084.0
189.10	2182.0
189.90	310.0
191.10	1453.0
192.00	3559.0
193.00	3396.0
193.90	1063.0
195.90	6598.0
197.10	1719.0
197.90	246592.0
198.90	16800.0
199.90	1562.0
200.90	533.0
201.50	1415.0
202.10	499.0
203.00	2721.0
204.10	9451.0
205.00	16840.0
206.00	65360.0
207.00	12040.0
208.00	3113.0
209.00	831.0
210.10	1075.0
211.00	3442.0
211.90	370.0
212.90	312.0
215.00	943.0
216.00	1472.0
217.00	18224.0
218.00	2567.0
220.10	419.0
221.00	8990.0
222.10	2639.0
223.00	4600.0
224.00	37248.0
225.10	7926.0
226.00	1654.0
227.00	18464.0
228.00	1978.0
229.00	2744.0
230.10	635.0
231.00	1114.0
231.90	692.0
233.30	321.0
234.00	1134.0
234.90	1841.0
236.00	1341.0
237.00	1312.0
239.00	343.0
240.00	971.0
241.10	1452.0
242.00	2547.0
243.00	2733.0
244.00	31624.0
245.00	4320.0
246.00	7476.0
247.00	1405.0
248.90	1266.0
251.00	436.0
253.10	768.0
254.10	921.0
255.00	167040.0
256.00	24056.0
257.00	2124.0
258.00	10743.0
259.00	2285.0
260.10	452.0
265.00	4163.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

266.00	798.0
271.00	323.0
271.90	694.0
273.00	4643.0
274.00	13339.0
275.00	78832.0
276.00	9352.0
277.00	6713.0
278.10	1534.0
279.00	486.0
280.80	315.0
281.20	474.0
282.00	352.0
282.80	858.0
284.00	519.0
285.10	1286.0
289.00	335.0
290.00	696.0
292.10	491.0
293.00	1571.0
294.00	517.0
294.90	690.0
296.00	25008.0
297.00	2854.0
301.90	814.0
303.10	2362.0
304.00	845.0
309.10	473.0
310.10	677.0
312.80	326.0
314.10	1743.0
315.00	2429.0
316.00	1629.0
317.00	334.0
321.00	1042.0
321.80	483.0
323.00	6404.0
324.00	1272.0
327.00	1551.0
328.00	582.0
332.20	814.0
333.10	949.0
334.10	3908.0
335.00	766.0
340.90	582.0
342.00	305.0
346.00	1855.0
352.00	2292.0
353.00	1759.0
354.00	2574.0
355.20	445.0
358.90	360.0
365.00	11968.0
366.00	1961.0
371.10	581.0
372.00	4310.0
373.10	832.0
383.00	1190.0
384.10	616.0
390.00	1018.0
390.90	329.0
392.10	385.0
402.00	1908.0
403.10	2892.0
403.70	502.0
405.10	306.0
415.00	347.0
421.20	2058.0
422.10	2312.0
423.00	15774.0
424.00	3097.0
441.00	36904.0
442.10	254272.0
443.10	48768.0
444.00	4529.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56