

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120615\  
 Data File : BM003394.D  
 Acq On : 06 Dec 2015 18:09  
 Operator : UM/IZ  
 Sample : SSTDC0020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 LabSampleId :  
 SSTD02038

Quant Time: Dec 07 04:23:27 2015  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM120115.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Mon Dec 07 04:19:45 2015  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound                    | AvqRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|-----------------------------|-------|-------|-------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0   | 94    | 0.00     |
| 2    | 1,4-Dioxane                 | 0.479 | 0.496 | -3.5  | 95    | 0.00     |
| 3 S  | 1,4-Dioxane-d8              | 0.402 | 0.422 | -5.0  | 95    | 0.00     |
| 4    | Benzaldehyde                | 0.982 | 0.974 | 0.8   | 91    | 0.00     |
| 5 S  | Phenol-d5                   | 1.597 | 1.653 | -3.5  | 96    | 0.00     |
| 6    | Phenol                      | 1.686 | 1.776 | -5.3  | 97    | 0.00     |
| 7 S  | Bis-(2-Chloroethyl)ether-d8 | 0.881 | 0.932 | -5.8  | 96    | 0.00     |
| 8    | Bis(2-Chloroethyl)ether     | 1.300 | 1.368 | -5.2  | 96    | 0.00     |
| 9 S  | 2-Chlorophenol-d4           | 1.320 | 1.380 | -4.5  | 95    | 0.00     |
| 10   | 2-Chlorophenol              | 1.398 | 1.461 | -4.5  | 95    | 0.00     |
| 11   | 2-Methylphenol              | 1.312 | 1.351 | -3.0  | 95    | 0.00     |
| 12   | 2,2'-oxybis(1-Chloropropane | 1.298 | 1.397 | -7.6  | 97    | 0.00     |
| 13 S | 4-Methylphenol-d8           | 1.339 | 1.383 | -3.3  | 95    | 0.00     |
| 14   | Acetophenone                | 2.054 | 2.219 | -8.0  | 95    | 0.00     |
| 15 P | N-Nitroso-di-n-propylamine  | 0.969 | 1.046 | -7.9  | 95    | 0.00     |
| 16   | 4-Methylphenol              | 1.453 | 1.532 | -5.4  | 96    | 0.00     |
| 17   | Hexachloroethane            | 0.514 | 0.542 | -5.4  | 96    | 0.00     |
| 18 I | Naphthalene-d8              | 1.000 | 1.000 | 0.0   | 93    | 0.00     |
| 19 S | Nitrobenzene-d5             | 0.136 | 0.146 | -7.4  | 95    | 0.00     |
| 20   | Nitrobenzene                | 0.315 | 0.339 | -7.6  | 96    | 0.00     |
| 21   | Isophorone                  | 0.589 | 0.629 | -6.8  | 95    | 0.00     |
| 22 S | 2-Nitrophenol-d4            | 0.143 | 0.153 | -7.0  | 96    | 0.00     |
| 23 C | 2-Nitrophenol               | 0.164 | 0.180 | -9.8  | 97    | 0.00     |
| 24   | 2,4-Dimethylphenol          | 0.344 | 0.371 | -7.8  | 96    | 0.00     |
| 25   | Bis(2-Chloroethoxy)methane  | 0.394 | 0.414 | -5.1  | 96    | 0.00     |
| 26 S | 2,4-Dichlorophenol-d3       | 0.283 | 0.300 | -6.0  | 94    | 0.00     |
| 27 C | 2,4-Dichlorophenol          | 0.298 | 0.316 | -6.0  | 94    | 0.00     |
| 28   | Naphthalene                 | 1.010 | 1.060 | -5.0  | 96    | 0.00     |
| 29 S | 4-Chloroaniline-d4          | 0.320 | 0.355 | -10.9 | 104   | 0.00     |
| 30   | 4-Chloroaniline             | 0.329 | 0.369 | -12.2 | 104   | 0.00     |
| 31 C | Hexachlorobutadiene         | 0.180 | 0.189 | -5.0  | 96    | 0.00     |
| 32   | Caprolactam                 | 0.087 | 0.082 | 5.7   | 88    | 0.00     |
| 33 C | 4-Chloro-3-methylphenol     | 0.313 | 0.333 | -6.4  | 94    | 0.00     |
| 34   | 2-Methylnaphthalene         | 0.737 | 0.774 | -5.0  | 94    | 0.00     |
| 35   | 1-Methylnaphthalene         | 0.697 | 0.734 | -5.3  | 95    | 0.00     |
| 36 I | Acenaphthene-d10            | 1.000 | 1.000 | 0.0   | 92    | 0.00     |
| 37   | 1,2,4,5-Tetrachlorobenzene  | 0.629 | 0.662 | -5.2  | 95    | 0.00     |
| 38   | Hexachlorocyclopentadiene   | 0.362 | 0.356 | 1.7   | 98    | 0.00     |
| 39 C | 2,4,6-Trichlorophenol       | 0.392 | 0.409 | -4.3  | 92    | 0.00     |
| 40   | 2,4,5-Trichlorophenol       | 0.434 | 0.457 | -5.3  | 93    | 0.00     |
| 41   | 1,1'-Biphenyl               | 1.667 | 1.763 | -5.8  | 94    | 0.00     |
| 42   | 2-Chloronaphthalene         | 1.244 | 1.300 | -4.5  | 94    | 0.00     |
| 43   | 2-Nitroaniline              | 0.278 | 0.299 | -7.6  | 93    | 0.00     |
| 44 S | Dimethylphthalate-d6        | 1.504 | 1.560 | -3.7  | 92    | 0.00     |

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120615\  
 Data File : BM003394.D  
 Acq On : 06 Dec 2015 18:09  
 Operator : UM/IZ  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 LabSampleId :  
 SSTD02038

Quant Time: Dec 07 04:23:27 2015  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM120115.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Mon Dec 07 04:19:45 2015  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound                    | AvqRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 45   | Dimethylphthalate           | 1.547 | 1.607 | -3.9 | 92    | 0.00     |
| 46   | 2,6-Dinitrotoluene          | 0.284 | 0.301 | -6.0 | 91    | 0.00     |
| 47 S | Acenaphthylene-d8           | 1.848 | 1.953 | -5.7 | 93    | 0.00     |
| 48   | Acenaphthylene              | 2.019 | 2.150 | -6.5 | 94    | 0.00     |
| 49   | 3-Nitroaniline              | 0.299 | 0.301 | -0.7 | 92    | 0.00     |
| 50 C | Acenaphthene                | 1.349 | 1.417 | -5.0 | 94    | 0.00     |
| 51   | 2,4-Dinitrophenol           | 0.148 | 0.126 | 14.9 | 94    | 0.00     |
| 52 S | 4-Nitrophenol-d4            | 0.270 | 0.253 | 6.3  | 88    | 0.00     |
| 53   | 4-Nitrophenol               | 0.211 | 0.205 | 2.8  | 91    | 0.00     |
| 54   | Dibenzofuran                | 1.911 | 1.991 | -4.2 | 93    | 0.00     |
| 55   | 2,4-Dinitrotoluene          | 0.413 | 0.436 | -5.6 | 91    | 0.00     |
| 56   | 2,3,4,6-Tetrachlorophenol   | 0.339 | 0.351 | -3.5 | 90    | 0.00     |
| 57   | Diethylphthalate            | 1.499 | 1.550 | -3.4 | 91    | 0.00     |
| 58 S | Fluorene-d10                | 1.294 | 1.337 | -3.3 | 92    | 0.00     |
| 59   | Fluorene                    | 1.481 | 1.547 | -4.5 | 93    | 0.00     |
| 60   | 4-Chlorophenyl-phenylether  | 0.723 | 0.752 | -4.0 | 93    | 0.00     |
| 61   | 4-Nitroaniline              | 0.335 | 0.316 | 5.7  | 87    | 0.00     |
| 62 I | Phenanthrene-d10            | 1.000 | 1.000 | 0.0  | 91    | 0.00     |
| 63 S | 4,6-Dinitro-2-methylphenol- | 0.102 | 0.099 | 2.9  | 95    | 0.00     |
| 64   | 4,6-Dinitro-2-methylphenol  | 0.111 | 0.109 | 1.8  | 93    | 0.00     |
| 65   | N-Nitrosodiphenylamine      | 0.610 | 0.645 | -5.7 | 92    | 0.00     |
| 66   | 4-Bromophenyl-phenylether   | 0.201 | 0.211 | -5.0 | 91    | 0.00     |
| 67   | Hexachlorobenzene           | 0.223 | 0.228 | -2.2 | 89    | 0.00     |
| 68   | Atrazine                    | 0.203 | 0.204 | -0.5 | 90    | 0.00     |
| 69 C | Pentachlorophenol           | 0.121 | 0.110 | 9.1  | 87    | 0.00     |
| 70   | Phenanthrene                | 1.101 | 1.138 | -3.4 | 92    | 0.00     |
| 71 S | Anthracene-d10              | 0.889 | 0.926 | -4.2 | 91    | 0.00     |
| 72   | Anthracene                  | 1.110 | 1.159 | -4.4 | 91    | 0.00     |
| 73   | 1,2,3,4-Tetrachlorobenzene  | 0.322 | 0.345 | -7.1 | 94    | 0.00     |
| 74   | Pentachlorobenzene          | 0.286 | 0.301 | -5.2 | 93    | 0.00     |
| 75   | Carbazole                   | 0.959 | 1.003 | -4.6 | 91    | 0.00     |
| 76   | Di-n-butylphthalate         | 1.033 | 1.079 | -4.5 | 90    | 0.00     |
| 77 C | Fluoranthene                | 1.112 | 1.155 | -3.9 | 90    | 0.00     |
| 78 I | Chrysene-d12                | 1.000 | 1.000 | 0.0  | 90    | 0.00     |
| 79 S | Pyrene-d10                  | 0.979 | 1.014 | -3.6 | 91    | 0.00     |
| 80   | Pyrene                      | 1.294 | 1.336 | -3.2 | 90    | 0.00     |
| 81   | Butylbenzylphthalate        | 0.422 | 0.447 | -5.9 | 91    | 0.00     |
| 82   | 3,3'-Dichlorobenzidine      | 0.326 | 0.318 | 2.5  | 88    | 0.00     |
| 83   | Benzo(a)anthracene          | 1.143 | 1.209 | -5.8 | 93    | 0.00     |
| 84   | Bis(2-ethylhexyl)phthalate  | 0.587 | 0.644 | -9.7 | 92    | 0.00     |
| 85   | Chrysene                    | 1.102 | 1.141 | -3.5 | 93    | 0.00     |
| 86 I | Perylene-d12                | 1.000 | 1.000 | 0.0  | 94    | 0.00     |
| 87   | Di-n-octyl phthalate        | 1.186 | 1.194 | -0.7 | 95    | 0.00     |

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120615\  
Data File : BM003394.D  
Acq On : 06 Dec 2015 18:09  
Operator : UM/IZ  
Sample : SSTDCCC020  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_M  
LabSampleId :  
SSTD02038

Quant Time: Dec 07 04:23:27 2015  
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM120115.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Mon Dec 07 04:19:45 2015  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 88   | Benzo(b)fluoranthene   | 1.169 | 1.223 | -4.6 | 94    | 0.00     |
| 89   | Benzo(k)fluoranthene   | 1.147 | 1.181 | -3.0 | 93    | 0.00     |
| 90 S | Benzo(a)pyrene-d12     | 0.965 | 1.007 | -4.4 | 94    | 0.00     |
| 91 C | Benzo(a)pyrene         | 1.134 | 1.178 | -3.9 | 93    | 0.00     |
| 92   | Indeno(1,2,3-cd)pyrene | 1.271 | 1.309 | -3.0 | 92    | 0.00     |
| 93   | Dibenzo(a,h)anthracene | 1.076 | 1.118 | -3.9 | 93    | 0.00     |
| 94   | Benzo(a,h,i)perylene   | 1.070 | 1.116 | -4.3 | 94    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0