

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120616\
 Data File : BM008181.D
 Acq On : 06 Dec 2016 12:52
 Operator : UM/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.465

Quant Time: Dec 06 13:48:46 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM112116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 03 04:54:05 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	80	0.00
3	Naphthalene	1.119	1.073	4.1	76	0.00
4 SURR	2-Methylnaphthalene-d10	0.566	0.541	4.4	77	0.00
5	2-Methylnaphthalene	0.721	0.664	7.9	74	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	73	0.00
7	Acenaphthylene	1.810	1.768	2.3	71	0.00
8 C	Acenaphthene	1.377	1.312	4.7	71	0.00
9	Fluorene	1.583	1.407	11.1	66	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	65	0.00
11	Pentachlorophenol	0.140	0.113	19.3	54	0.00
12	Phenanthrene	1.246	1.178	5.5	62	0.00
13	Anthracene	1.170	1.187	-1.5	66	0.00
14 SURR	Fluoranthene-d10	1.132	1.033	8.7	60	0.00
15 C	Fluoranthene	1.608	1.375	14.5	58	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	53	0.00
17	Pyrene	1.389	1.500	-8.0	58	0.00
18	Benzo(a)anthracene	1.228	1.249	-1.7	56	0.00
19	Chrysene	1.527	1.410	7.7	51	0.00
20 I	Perylene-d12	1.000	1.000	0.0	58	0.00
21	Benzo(b)fluoranthene	1.596	1.473	7.7	54	0.00
22	Benzo(k)fluoranthene	1.647	1.462	11.2	53	0.00
23 C	Benzo(a)pyrene	1.495	1.420	5.0	56	0.00
24	Indeno(1,2,3-cd)pyrene	1.848	1.740	5.8	54	0.00
25	Dibenzo(a,h)anthracene	1.481	1.376	7.1	54	0.00
26	Benzo(a,h,i)perylene	1.640	1.564	4.6	55	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0