

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120616\
 Data File : BM008194.D
 Acq On : 06 Dec 2016 20:40
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.466

Quant Time: Dec 07 06:15:18 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM112116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 07 05:56:15 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
3	Naphthalene	1.119	1.087	2.9	82	0.00
4 SURR	2-Methylnaphthalene-d10	0.566	0.574	-1.4	86	0.01
5	2-Methylnaphthalene	0.721	0.708	1.8	83	0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
7	Acenaphthylene	1.810	1.732	4.3	86	0.00
8 C	Acenaphthene	1.377	1.327	3.6	88	0.00
9	Fluorene	1.583	1.501	5.2	86	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
11	Pentachlorophenol	0.140	0.099	29.3#	60	0.00
12	Phenanthrene	1.246	1.190	4.5	79	0.00
13	Anthracene	1.170	1.287	-10.0	90	0.00
14 SURR	Fluoranthene-d10	1.132	1.018	10.1	74	0.00
15 C	Fluoranthene	1.608	1.316	18.2	69	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	54	0.00
17	Pyrene	1.389	1.763	-26.9#	68	0.00
18	Benzo(a)anthracene	1.228	1.148	6.5	52	0.01
19	Chrysene	1.527	1.530	-0.2	55	0.01
20 I	Perylene-d12	1.000	1.000	0.0	55	0.00
21	Benzo(b)fluoranthene	1.596	1.344	15.8	47	0.00
22	Benzo(k)fluoranthene	1.647	1.753	-6.4	61	0.01
23 C	Benzo(a)pyrene	1.495	1.391	7.0	52	0.01
24	Indeno(1,2,3-cd)pyrene	1.848	1.717	7.1	51	0.01
25	Dibenzo(a,h)anthracene	1.481	1.338	9.7	50	0.02
26	Benzo(a,h,i)perylene	1.640	1.539	6.2	51	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0