


Data Path : Z:\HPCHEM1\BNA M\DATA\BM120616\
 Data File : BM008180.D
 Acq On : 06 Dec 2016 12:16
 Operator : UM/SJ
 Sample : DFTPP61
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP61

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-TEST.M
 Title : SVOA CALIBRATION
 Last Update : Wed Nov 30 00:32:18 2016



AutoFind: Scans 2537, 2538, 2539; Background Corrected with Scan 2532

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.1	42626	PASS
68	69	0.00	2	1.9	868	PASS
69	198	0.00	100	37.4	45408	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.2	52394	PASS
197	198	0.00	2	0.5	641	PASS
198	198	100	100	100.0	121413	PASS
199	198	5	9	6.8	8240	PASS
275	198	10	60	29.9	36304	PASS
365	198	1	100	4.7	5763	PASS
441	443	0.01	100	79.8	14669	PASS
442	198	50	100	80.5	97696	PASS
443	442	15	24	18.8	18392	PASS

Scan 2532 (17.480 min): BM008180.D
DFTPP61

m/z	Abundance
40.00	2897.0
44.10	652.0
207.00	1259.0
281.00	849.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

m/z	Abundance
38.20	499.0
39.00	2475.0
40.00	2497.0
44.10	803.0
49.20	334.0
50.10	9741.0
51.10	42048.0
52.00	1965.0
56.00	1246.0
57.00	2886.0
61.00	495.0
62.00	533.0
63.00	1261.0
64.00	333.0
65.10	867.0
68.00	780.0
69.00	43232.0
73.10	769.0
74.10	4588.0
75.00	7288.0
76.10	2664.0
77.10	48344.0
78.10	3221.0
79.00	3937.0
80.00	2683.0
81.00	4618.0
82.00	958.0
82.90	758.0
85.00	979.0
85.90	1200.0
87.00	679.0
91.00	937.0
91.90	963.0
93.10	6624.0
94.00	457.0
98.00	5236.0
99.00	4912.0
100.90	1829.0
103.00	580.0
104.00	1294.0
105.00	1452.0
106.10	479.0
107.00	15720.0
108.00	2478.0
109.10	498.0
110.00	25408.0
111.00	4363.0
112.00	713.0
115.90	937.0
117.00	15312.0
118.00	1399.0
122.10	675.0
123.00	1670.0
124.00	770.0
125.10	722.0
127.00	47448.0
128.00	3777.0
129.00	18512.0
130.00	1785.0
130.90	440.0
134.00	639.0
135.00	1817.0
136.00	818.0
137.00	1333.0
139.00	307.0
141.00	2508.0
142.10	659.0
142.90	696.0
146.00	387.0
147.00	1608.0
148.00	3508.0
149.10	812.0
151.10	379.0
153.00	784.0
153.90	705.0
155.00	1661.0
156.00	2114.0
157.10	519.0
157.80	483.0
158.20	405.0
160.00	889.0
161.00	1036.0
162.10	382.0
165.00	1171.0
166.10	646.0
167.00	5650.0
168.00	4307.0
169.00	533.0
172.00	471.0
172.90	584.0
174.00	1013.0
175.10	2205.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

175.90	428.0
176.90	1311.0
177.60	315.0
178.00	368.0
179.00	3875.0
180.00	3345.0
181.00	1320.0
183.90	399.0
185.10	1914.0
186.00	14519.0
187.00	5158.0
188.00	706.0
189.00	1044.0
191.00	415.0
191.90	1356.0
193.00	1536.0
196.00	2535.0
197.10	481.0
198.00	99720.0
199.00	7294.0
199.90	770.0
201.60	412.0
202.90	892.0
204.00	5050.0
205.00	6525.0
206.10	27816.0
207.10	5105.0
208.00	1525.0
208.90	555.0
210.20	339.0
211.00	1288.0
211.70	350.0
215.00	329.0
216.10	893.0
217.00	7507.0
218.10	762.0
221.10	4168.0
223.00	1818.0
224.00	15584.0
225.00	3585.0
226.00	595.0
227.00	7721.0
227.90	911.0
228.90	1326.0
229.90	303.0
230.90	657.0
234.00	458.0
235.00	470.0
235.90	408.0
237.00	1190.0
239.10	349.0
239.80	315.0
240.90	410.0
242.00	776.0
243.10	720.0
244.10	11155.0
245.10	1830.0
246.00	2379.0
247.10	689.0
249.00	541.0
254.10	540.0
255.00	61552.0
256.00	9376.0
256.90	820.0
258.00	4021.0
258.90	582.0
265.00	1997.0
266.00	314.0
273.00	1785.0
274.00	4838.0
275.00	26088.0
276.00	3578.0
277.00	2420.0
278.00	605.0
280.90	530.0
283.10	355.0
285.10	430.0
293.00	615.0
296.00	8460.0
297.00	1592.0
302.90	965.0
313.90	616.0
315.10	1010.0
316.10	593.0
321.00	308.0
323.00	2427.0
323.90	616.0
327.00	719.0
328.00	408.0
334.00	1425.0
335.00	509.0
345.90	648.0
352.00	736.0
353.00	435.0
354.00	688.0
365.00	3236.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

365.90	572.0
372.00	1204.0
383.00	357.0
402.10	678.0
403.00	506.0
421.10	440.0
422.00	526.0
423.00	3404.0
424.10	1574.0
441.10	8785.0
442.10	61064.0
443.00	11360.0
444.00	1031.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

m/z	Abundance
38.00	684.0
39.10	3351.0
40.00	2755.0
44.00	558.0
49.00	449.0
50.10	13636.0
51.10	50344.0
52.00	2658.0
55.00	387.0
56.00	1694.0
57.10	3786.0
61.00	702.0
62.10	893.0
63.10	2281.0
65.10	818.0
68.00	1061.0
69.00	53360.0
73.00	825.0
74.10	6702.0
75.00	9764.0
76.00	3278.0
77.10	60776.0
78.10	4294.0
79.10	4190.0
80.00	3347.0
81.00	4347.0
82.00	1222.0
83.00	1135.0
85.00	644.0
86.00	1742.0
87.10	639.0
87.90	324.0
91.00	1061.0
92.00	1278.0
93.00	8710.0
94.00	529.0
96.00	458.0
98.00	6846.0
99.00	5593.0
100.00	673.0
101.00	2639.0
102.90	766.0
104.00	2075.0
105.00	1888.0
105.90	708.0
107.00	19696.0
108.00	3457.0
109.20	421.0
110.00	36008.0
111.00	5821.0
112.00	724.0
116.00	945.0
117.00	19720.0
118.10	1688.0
119.00	431.0
120.00	305.0
122.00	1098.0
123.00	1895.0
124.00	1085.0
124.90	981.0
126.10	541.0
127.00	62304.0
128.00	4725.0
129.00	26680.0
130.00	2141.0
131.10	449.0
133.90	626.0
135.00	2610.0
136.00	1085.0
137.00	1628.0
139.90	372.0
141.00	3521.0
142.00	1272.0
143.00	974.0
146.00	784.0
147.10	1890.0
148.00	4688.0
149.00	1275.0
151.20	613.0
153.00	1399.0
154.00	909.0
155.00	2490.0
156.00	2780.0
157.00	463.0
157.90	793.0
158.90	475.0
160.00	1258.0
161.00	1905.0
161.90	383.0
165.00	1664.0
165.90	1074.0
167.00	8586.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

168.00	4971.0
169.00	1144.0
172.00	868.0
172.90	994.0
174.00	1590.0
175.00	3043.0
176.10	703.0
177.00	1856.0
177.80	357.0
179.00	6058.0
180.00	4189.0
181.00	1909.0
182.10	380.0
183.90	606.0
185.00	3432.0
186.00	21464.0
187.00	6871.0
188.00	835.0
189.00	1444.0
191.00	785.0
192.00	2092.0
193.00	1944.0
193.90	401.0
196.00	4102.0
196.80	472.0
197.10	676.0
198.00	146496.0
199.00	9887.0
200.00	856.0
201.20	442.0
201.60	391.0
203.10	1097.0
204.00	5821.0
205.10	9377.0
206.10	42384.0
207.00	7039.0
208.00	1907.0
209.00	927.0
210.00	635.0
211.00	1930.0
211.90	333.0
215.00	600.0
216.00	922.0
217.00	11714.0
218.00	1391.0
221.00	5106.0
221.70	1658.0
223.00	2402.0
224.10	23664.0
225.00	6837.0
226.10	590.0
227.00	11398.0
228.00	1316.0
228.90	2104.0
230.00	377.0
231.00	983.0
233.10	376.0
233.90	870.0
234.90	788.0
235.80	451.0
237.00	1387.0
238.90	403.0
240.00	403.0
241.00	778.0
242.10	1514.0
243.00	1139.0
244.00	18088.0
245.10	2560.0
246.00	4194.0
247.00	1023.0
249.00	500.0
253.10	455.0
254.10	748.0
255.00	98432.0
256.00	13383.0
257.00	1226.0
258.00	6918.0
259.00	1118.0
265.00	2860.0
266.10	494.0
271.00	334.0
272.00	461.0
273.00	3036.0
274.00	7393.0
275.00	42648.0
276.00	6363.0
277.00	3870.0
278.00	704.0
281.10	428.0
281.80	348.0
283.00	516.0
285.10	658.0
288.90	342.0
292.90	965.0
295.00	342.0
296.00	14159.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

297.00	1757.0
303.10	1637.0
303.80	415.0
314.00	678.0
315.00	1481.0
316.00	704.0
321.00	498.0
323.00	3895.0
324.00	868.0
327.00	761.0
328.00	682.0
333.10	376.0
334.00	2675.0
335.00	492.0
341.00	548.0
346.00	852.0
346.90	365.0
351.90	908.0
353.00	819.0
354.00	1393.0
355.10	457.0
365.00	6758.0
366.00	705.0
371.10	341.0
372.10	2013.0
372.90	377.0
383.00	731.0
401.90	884.0
402.90	1232.0
404.00	470.0
421.00	632.0
422.00	876.0
423.10	6318.0
424.10	1950.0
441.10	17752.0
442.10	112872.0
443.10	21680.0
444.00	1724.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

m/z	Abundance
38.00	446.0
39.10	2644.0
40.00	2619.0
43.90	468.0
48.90	312.0
50.10	9979.0
51.10	35488.0
52.00	1836.0
54.90	308.0
56.00	976.0
57.00	2342.0
60.90	532.0
61.90	702.0
63.10	1578.0
64.10	328.0
65.00	743.0
68.00	763.0
69.00	39632.0
73.00	793.0
74.00	4706.0
75.10	7400.0
76.10	2843.0
77.10	45016.0
78.10	3248.0
79.00	3436.0
80.00	2560.0
81.00	3376.0
82.00	951.0
83.00	798.0
85.00	799.0
86.00	1143.0
87.10	331.0
90.90	830.0
92.00	1304.0
93.00	6136.0
98.00	5252.0
99.00	3966.0
100.00	338.0
101.00	1680.0
102.90	893.0
104.00	1507.0
104.90	1225.0
106.00	482.0
107.00	14867.0
108.00	2234.0
109.10	466.0
110.00	25608.0
111.00	4071.0
112.00	774.0
113.10	333.0
115.90	734.0
117.00	15706.0
118.00	978.0
122.00	1334.0
123.10	1333.0
124.00	787.0
124.90	529.0
127.00	47432.0
128.10	3668.0
129.00	20088.0
130.00	1589.0
134.00	591.0
135.00	2204.0
136.00	701.0
137.00	1169.0
140.00	383.0
141.00	2649.0
142.10	1049.0
143.00	526.0
145.80	578.0
147.00	1466.0
148.00	4188.0
149.00	1050.0
150.10	411.0
153.10	1217.0
154.10	609.0
155.10	1650.0
156.00	2484.0
157.00	800.0
158.00	685.0
159.10	527.0
159.90	721.0
161.00	1445.0
162.00	464.0
165.00	1506.0
166.00	1149.0
167.00	6547.0
168.00	4572.0
169.00	857.0
172.00	549.0
173.10	667.0
174.10	1025.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

175.00	2247.0
176.00	910.0
176.90	1221.0
178.00	499.0
179.00	4684.0
180.00	2989.0
181.00	1437.0
184.10	402.0
185.00	2189.0
186.00	18440.0
187.00	5313.0
187.90	660.0
189.00	1429.0
190.00	359.0
191.00	898.0
192.00	1678.0
193.00	2009.0
196.00	3419.0
196.70	719.0
197.10	768.0
198.00	118024.0
199.00	7539.0
200.00	781.0
201.50	430.0
203.00	969.0
204.00	4836.0
205.00	8131.0
206.10	33928.0
207.10	7027.0
208.00	1572.0
209.00	860.0
210.00	406.0
211.00	1568.0
215.00	459.0
216.00	976.0
217.00	9311.0
218.00	1343.0
221.00	3920.0
221.90	1275.0
223.00	1953.0
224.00	21064.0
225.10	5256.0
226.10	693.0
227.00	8657.0
227.90	1134.0
229.00	1996.0
231.10	676.0
233.90	512.0
234.90	604.0
235.90	342.0
236.90	1024.0
240.10	377.0
241.10	449.0
242.00	991.0
243.00	1127.0
244.10	14644.0
245.00	1533.0
246.00	3329.0
246.90	598.0
249.00	578.0
253.00	344.0
254.20	543.0
255.00	83448.0
256.00	13358.0
256.90	1444.0
258.00	4920.0
259.00	966.0
265.00	2769.0
265.90	395.0
271.00	373.0
272.10	325.0
273.00	2837.0
274.00	6196.0
275.00	40176.0
276.00	5222.0
277.00	3742.0
278.00	867.0
281.10	585.0
282.90	472.0
285.00	749.0
293.00	790.0
296.00	12926.0
297.00	1746.0
302.00	307.0
303.00	1255.0
303.90	379.0
314.10	574.0
315.00	1572.0
316.10	770.0
321.10	327.0
323.10	3249.0
324.10	636.0
326.90	737.0
327.90	513.0
331.80	304.0
333.00	478.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

334.00	2580.0
335.10	591.0
340.90	436.0
346.00	1046.0
352.10	914.0
353.10	637.0
354.00	1478.0
355.00	431.0
365.00	7296.0
366.10	881.0
372.00	2148.0
373.10	525.0
383.10	610.0
402.00	839.0
403.00	1397.0
404.10	316.0
421.00	1042.0
422.10	1140.0
423.10	6642.0
424.10	1940.0
425.00	388.0
441.10	17472.0
442.00	119152.0
443.10	22136.0
444.00	2061.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61