

Data Path : Z:\HPCHEM1\BNA M\Data\BM120715\
 Data File : BM003407.D
 Acq On : 07 Dec 2015 14:48
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM120715

Quant Time: Dec 07 15:30:40 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM120715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 15:08:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.449	0.391	12.9	109	0.00
3	n-Nitrosodimethylamine	0.394	0.385	2.3	114	0.00
4 S	2-Fluorophenol	0.857	0.832	2.9	118	-0.02
5 S	Phenol-d6	1.203	1.204	-0.1	120	0.00
6	2-Chlorophenol	1.120	1.127	-0.6	119	0.00
7 C	Phenol	1.258	1.240	1.4	118	0.00
8	bis(2-Chloroethyl)ether	0.896	0.903	-0.8	118	0.00
9	2-Methylphenol	0.708	0.710	-0.3	121	0.00
10	3+4-Methylphenols	1.029	1.034	-0.5	119	0.00
11 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
12 S	Nitrobenzene-d5	0.264	0.270	-2.3	122	0.00
13	Nitrobenzene	0.275	0.279	-1.5	123	0.00
14 C	2-Nitrophenol	0.112	0.110	1.8	124	0.00
15	2,4-Dimethylphenol	0.264	0.269	-1.9	122	0.00
16 C	2,4-Dichlorophenol	0.246	0.251	-2.0	119	0.00
17	Hexachlorobutadiene	0.183	0.175	4.4	113	0.00
18 C	4-Chloro-3-methylphenol	0.214	0.211	1.4	119	0.03
19	2-Methylnaphthalene	0.773	0.764	1.2	118	0.00
20 I	Acenaphthene-d10	1.000	1.000	0.0	145	0.00
21 S	2,4,6-Tribromophenol	0.123	0.096	22.0	101	0.00
22 C	2,4,6-Trichlorophenol	0.335	0.280	16.4	127	0.00
23	2,4,5-Trichlorophenol	0.368	0.308	16.3	116	0.00
24 S	2-Fluorobiphenyl	1.760	1.534	12.8	124	0.00
25 P	2,4-Dinitrophenol	0.038	0.037#	2.6	137	0.00
26 P	4-Nitrophenol	0.153	0.141	7.8	120	0.00
27 I	Phenanthrene-d10	1.000	1.000	0.0	133	0.00
28	4,6-Dinitro-2-methylphenol	0.061	0.064	-4.9	147	0.00
29	Hexachlorobenzene	0.212	0.182	14.2	107	0.00
30 C	Pentachlorophenol	0.065	0.049	24.6#	120	0.00
31 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
32	Benzidine	0.323	0.379	-17.3	126	0.00
33 S	Terphenyl-d14	0.676	0.794	-17.5	104	0.00
34	3,3'-Dichlorobenzidine	0.304	0.348	-14.5	134	0.00
35 I	Perylene-d12	1.000	1.000	0.0	114	0.01

(#) = Out of Range

SPCC's out = 1 CCC's out = 1