

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120715\
 Data File : BM003404.D
 Acq On : 07 Dec 2015 12:46
 Operator : UM/IZ
 Sample : SSTDICC002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002

Manual Integrations
 APPROVED

Mmdadoda
 12/8/2015 6:14:42 PM

Quant Time: Dec 07 14:53:05 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM120715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 13:56:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	66083	5.00	ng	0.00
11) Naphthalene-d8	10.52	136	227218	5.00	ng	0.00
20) Acenaphthene-d10	14.37	164	105696	5.00	ng	0.00
27) Phenanthrene-d10	17.12	188	165373	5.00	ng	0.00
31) Chrysene-d12	21.31	240	206205	5.00	ng	0.00
35) Perylene-d12	23.57	264	159000	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.31	112	21923	1.96	ng	0.00
5) Phenol-d6	6.88	99	31806	2.11	ng	0.00
12) Nitrobenzene-d5	8.86	82	24084	2.09	ng	0.00
21) 2,4,6-Tribromophenol	15.85	330	5956	2.63	ng	0.00
24) 2-Fluorobiphenyl	12.99	172	73243	1.91	ng	0.00
33) Terphenyl-d14	19.75	244	52326	1.84	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.22	88	10672	1.69	ng	# 56
3) n-Nitrosodimethylamine	3.53	42	9825	1.91	ng	82
6) 2-Chlorophenol	7.29	128	29702	2.08	ng	89
7) Phenol	6.91	94	33144	2.08	ng	99
8) bis(2-Chloroethyl)ether	7.12	93	22887	1.93	ng	# 64
9) 2-Methylphenol	8.17	107	18418	2.07	ng	95
10) 3+4-Methylphenols	8.48	107	26710	2.11	ng	# 36
13) Nitrobenzene	8.91	77	25249	2.15	ng	# 44
14) 2-Nitrophenol	9.65	139	9861	2.08	ng	# 71
15) 2,4-Dimethylphenol	9.69	122	23763m	2.04	ng	
16) 2,4-Dichlorophenol	10.16	162	21967m	2.07	ng	
17) Hexachlorobutadiene	10.84	225	15939	1.88	ng	97
18) 4-Chloro-3-methylphenol	11.77	107	18818	2.11	ng	# 66
19) 2-Methylnaphthalene	12.18	142	64362	1.81	ng	95
22) 2,4,6-Trichlorophenol	12.80	196	14293	2.14	ng	# 51
23) 2,4,5-Trichlorophenol	12.86	196	17163m	2.50	ng	
25) 2,4-Dinitrophenol	14.49	184	1915	3.09	ng	# 63
26) 4-Nitrophenol	14.57	139	8704m	3.36	ng	
28) 4,6-Dinitro-2-methylphenol	15.50	198	4616	2.66	ng	# 29
29) Hexachlorobenzene	16.42	284	14223	2.01	ng	# 12
30) Pentachlorophenol	16.75	266	4119	2.18	ng	# 46
32) Benzidine	19.38	184	24114	2.08	ng	96
34) 3,3'-Dichlorobenzidine	21.24	252	29014	2.54	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA M\DATA\BM120715\
Data File : BM003404.D
Acq On : 07 Dec 2015 12:46
Operator : UM/IZ
Sample : SSTDIC002
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDIC002

Manual Integrations
APPROVED

Mmdadoda
12/8/2015 6:14:42 PM

Quant Time: Dec 07 14:53:05 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM120715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 13:56:34 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.72 min Scan# 226

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Instrument :

BNA_M





ClientSampleId :

SSTDICC002

Tot Ion:	152	Resp:	66083
Ion Ratio	Lower	Upper	
152	100		
150	161.8	116.3	174.5
115	56.4	31.9	47.9#

Manual Integrations
APPROVED

Mmdadoda
12/8/2015 6:14:42 PM



#2

1,4-Dioxane

Concen: 1.69 ng




RT: 3.22 min Scan# 27


Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Tgt Ion:	88	Resp:	10672
Ion Ratio	Lower	Upper	
88	100		
58	60.9	75.5	113.3#
43	0.0	32.7	49.1#





#3





n-Nitrosodimethylamine
Concen: 1.91 ng
RT: 3.53 min Scan# 42
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDICC002

Tot Ion	Ratio	Lower	Upper
42	100		
74	147.9	101.0	151.4
44	7.6	6.3	9.5

Manual Integrations
APPROVED




Mmdadoda
12/8/2015 6:14:42 PM






#4

2-Fluorophenol
Concen: 1.96 ng
RT: 5.31 min Scan# 129
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Tgt Ion	Ratio	Lower	Upper
112	100		
64	29.3	49.3	73.9#
63	12.7	26.2	39.4#





#5

Phenol-d6

Concen: 2.11 ng

RT: 6.88 min Scan# 195

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Tot Ion:	99	Resp:	31806
Ion Ratio	Lower	Upper	
99	100		
42	7.7	12.1	18.1#
71	24.3	26.4	39.6#

Instrument :





BNA_M

ClientSampleId :

SSTDICC002

Manual Integrations
APPROVED

Mmdadoda
12/8/2015 6:14:42 PM



#6

2-Chlorophenol

Concen: 2.08 ng


RT: 7.29 min Scan# 210




Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Tgt Ion:	128	Resp:	29702
Ion Ratio	Lower	Upper	
128	100		
130	34.2	22.2	62.2
64	7.8	0.0	28.7





#7

Phenol

Concen: 2.08 ng

RT: 6.91 min Scan# 196

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Tot Ion:	94	Resp:	33144
Ion Ratio	Lower	Upper	
94	100		
65	17.3	0.0	37.7
66	25.3	6.3	46.3

Instrument :





BNA_M

ClientSampleId :

SSTDICC002

Manual Integrations
APPROVED

Mmdadoda
12/8/2015 6:14:42 PM



#8

bis(2-Chloroethyl)ether

Concen: 1.93 ng


RT: 7.12 min Scan# 204


Delta R.T. -0.03 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Tgt Ion:	93	Resp:	22887
Ion Ratio	Lower	Upper	
93	100		
63	94.7	58.2	87.4#
95	0.0	26.2	39.4#





#9

2-Methylphenol

Concen: 2.07 ng

RT: 8.17 min Scan# 243

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Instrument :

BNA_M


ClientSampled :

SSTDICC002

Tot Ion:	107	Resp:	18418
Ion Ratio	Lower	Upper	
107	100		
108	133.1	100.9	151.3
77	24.5	20.3	30.5
79	28.6	22.8	34.2

Manual Integrations
APPROVED

Mmdadoda
12/8/2015 6:14:42 PM






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BMC

Ion 79.00 (78.70 to 79.70): BMC



#10

3+4-Methylphenols

Concen: 2.11 ng

RT: 8.48 min Scan# 256

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46




Tgt Ion:	107	Resp:	26710
Ion Ratio	Lower	Upper	
107	100		
108	84.8	86.3	126.3#
77	32.1	151.7	191.7#
79	27.4	7.4	47.4


Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BMC

Ion 79.00 (78.70 to 79.70): BMC









#11
Naphthalene-d8
Concen: 5.00 ng
RT: 10.52 min Scan# 339
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDICC002

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	12.2	9.1	13.7	
54	1.0	2.6	3.8	
68	1.4	2.2	3.2	




Manual Integrations
APPROVED


Mmdadoda
12/8/2015 6:14:42 PM



#12
Nitrobenzene-d5
Concen: 2.09 ng
RT: 8.86 min Scan# 272
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	0.0	0.0	0.0	
54	48.9	59.0	88.6	









#13
Nitrobenzene
Concen: 2.15 ng
RT: 8.91 min Scan# 274
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDIC002

Tot Ion	Ratio	Lower	Upper
77	100		
123	0.0	37.7	56.5#
65	12.3	9.8	14.6




Manual Integrations
APPROVED


Mmdadoda
12/8/2015 6:14:42 PM



#14
2-Nitrophenol
Concen: 2.08 ng
RT: 9.65 min Scan# 305
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Tgt Ion	Ratio	Lower	Upper
139	100		
109	12.7	19.5	29.3#
65	6.7	18.6	28.0#





#15

2,4-Dimethylphenol

Concen: 2.04 ng/mL

RT: 9.69 min Scan# 307

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Instrument :

BNA_M


ClientSampleId :

SSTDICC002




Tot Ion:	122	Resp:	23763
Ion Ratio	Lower	Upper	
122	100		
107	74.7	68.6	102.8
121	0.0	44.8	67.2#

Manual Integrations
APPROVED

Mmdadoda
12/8/2015 6:14:42 PM



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E



#16

2,4-Dichlorophenol

Concen: 2.07 ng/mL


RT: 10.16 min Scan# 326

Delta R.T. -0.00 min



Lab File: BM003404.D


Acq: 07 Dec 2015 12:46

Tgt Ion:	162	Resp:	21967
Ion Ratio	Lower	Upper	
162	100		
164	68.2	49.2	89.2
98	10.7	0.0	37.4



Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): BM









#17
Hexachlorobutadiene
Concen: 1.88 ng
RT: 10.84 min Scan# 351
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDICC002

Tot Ion	Ratio	Lower	Upper
225	100		
223	65.2	53.0	79.4
227	60.9	52.2	78.4




Manual Integrations
APPROVED


Mmdadoda
12/8/2015 6:14:42 PM



#18
4-Chloro-3-methylphenol
Concen: 2.11 ng
RT: 11.77 min Scan# 385
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Tgt Ion	Ratio	Lower	Upper
107	100		
144	8.6	16.2	24.2#
142	34.0	49.8	74.6#









#19
2-Methylnaphthalene
Concen: 1.81 ng
RT: 12.18 min Scan# 400
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDIC002

Tot Ion	Ratio	Lower	Upper
142	100		
141	85.5	71.4	107.0
115	31.4	29.3	43.9




Manual Integrations
APPROVED


Mmdadoda
12/8/2015 6:14:42 PM



#20
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.37 min Scan# 471
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Tgt Ion	Ratio	Lower	Upper
164	100		
162	88.7	79.2	118.8
160	35.5	33.7	50.5









#21
 2,4,6-Tribromophenol
 Concen: 2.63 ng
 RT: 15.85 min Scan# 509
 Delta R.T. -0.00 min
 Lab File: BM003404.D
 Acq: 07 Dec 2015 12:46

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002

Tot Ion	Ratio	Lower	Upper
330	100		
332	70.4	74.0	111.0#
141	42.2	24.1	36.1#




Manual Integrations
 APPROVED


Mmdadoda
 12/8/2015 6:14:42 PM



#22
 2,4,6-Trichlorophenol
 Concen: 2.14 ng
 RT: 12.80 min Scan# 423
 Delta R.T. -0.00 min
 Lab File: BM003404.D
 Acq: 07 Dec 2015 12:46

Tgt Ion	Ratio	Lower	Upper
196	100		
198	109.5	54.2	81.2#
200	39.4	17.1	25.7#









#23
 2,4,5-Trichlorophenol
 Concen: 2.50 ng/ml
 RT: 12.86 min Scan# 425
 Delta R.T. -0.00 min
 Lab File: BM003404.D
 Acq: 07 Dec 2015 12:46

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002

Manual Integrations
 APPROVED




Mmdadoda
 12/8/2015 6:14:42 PM


Tot Ion	Ion	Ratio	Resp	Lower	Upper
196	100				
198	84.8	73.6	110.4		
97	23.3	0.0	0.0#		
132	26.5	0.0	0.0#		



#24
 2-Fluorobiphenyl
 Concen: 1.91 ng/ml
 RT: 12.99 min Scan# 430
 Delta R.T. -0.00 min
 Lab File: BM003404.D
 Acq: 07 Dec 2015 12:46

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
172	100				
171	31.6	28.0	42.0		
170	20.5	18.9	28.3		









#25
2,4-Dinitrophenol
Concen: 3.09 ng
RT: 14.49 min Scan# 474
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDIC002

Tot Ion	Ratio	Lower	Upper
184	100		
63	23.7	52.0	78.0#
154	47.5	51.4	77.2#




Manual Integrations
APPROVED


Mmdadoda
12/8/2015 6:14:42 PM



#26
4-Nitrophenol
Concen: 3.36 ng m
RT: 14.57 min Scan# 476
Delta R.T. -0.00 min
Lab File: BM003404.D
Acq: 07 Dec 2015 12:46

Tgt Ion	Ratio	Lower	Upper
139	100		
109	63.8	60.0	100.0
65	63.3	85.7	125.7#





#27

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.12 min Scan# 549

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Instrument :

BNA_M





ClientSampleId :

SSTDICC002

Tot Ion:	188	Resp:	165373
Ion Ratio	Lower	Upper	
188	100		
94	2.2	2.9	4.3#
80	1.7	2.5	3.7#

Manual Integrations
APPROVED

Mmdadoda
12/8/2015 6:14:42 PM



#28

4,6-Dinitro-2-methylphenol

Concen: 2.66 ng




RT: 15.50 min Scan# 500


Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Tgt Ion:	198	Resp:	4616
Ion Ratio	Lower	Upper	
198	100		
51	10.6	60.8	100.8#
105	10.4	30.0	70.0#





#29

Hexachlorobenzene

Concen: 2.01 ng

RT: 16.42 min Scan# 526

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Instrument :

BNA_M





ClientSampleId :

SSTDICC002

Tot Ion:	284	Resp:	14223
Ion Ratio	Lower	Upper	
284	100		
142	86.0	23.3	34.9#
249	65.9	23.3	34.9#

Manual Integrations
APPROVED

Mmdadoda
12/8/2015 6:14:42 PM



#30

Pentachlorophenol

Concen: 2.18 ng




RT: 16.75 min Scan# 537


Delta R.T. -0.03 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Tgt Ion:	266	Resp:	4119
Ion Ratio	Lower	Upper	
266	100		
268	36.3	62.6	94.0#
264	107.3	47.9	71.9#





#31

Chrysene-d12

Concen: 5.00 ng

RT: 21.31 min Scan# 757

Delta R.T. -0.01 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46

Instrument :

BNA_M


ClientSampleId :

SSTDICC002

Tot Ion:	240	Resp:	206205
Ion Ratio	Lower	Upper	
240	100		
120	6.6	10.4	15.6#
236	30.8	22.0	33.0

Manual Integrations
APPROVED




Mmdadoda
12/8/2015 6:14:42 PM



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#32

Benzidine

Concen: 2.08 ng


RT: 19.38 min Scan# 657

Delta R.T. -0.00 min

Lab File: BM003404.D

Acq: 07 Dec 2015 12:46



Tgt Ion:	184	Resp:	24114
Ion Ratio	Lower	Upper	
184	100		
185	16.5	11.6	17.4
183	9.6	8.6	13.0




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#33
 Terphenyl-d14
 Concen: 1.84 ng
 RT: 19.75 min Scan# 679
 Delta R.T. -0.00 min
 Lab File: BM003404.D
 Acq: 07 Dec 2015 12:46

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002

Tot Ion	Ratio	Lower	Upper
244	100		
212	9.6	7.3	10.9
122	19.3	14.8	22.2




Manual Integrations
 APPROVED




Mmdadoda
 12/8/2015 6:14:42 PM



#34
 3,3'-Dichlorobenzidine
 Concen: 2.54 ng
 RT: 21.24 min Scan# 754
 Delta R.T. -0.01 min
 Lab File: BM003404.D
 Acq: 07 Dec 2015 12:46

Tgt Ion	Ratio	Lower	Upper
252	100		
254	63.3	53.7	80.5
126	2.3	2.8	4.2#





#35
 Pervlene-d12
 Concen: 5.00 ng
 RT: 23.57 min Scan# 898
 Delta R.T. 0.00 min
 Lab File: BM003404.D
 Acq: 07 Dec 2015 12:46

Tot Ion	Ratio	Lower	Upper
264	100		
260	26.2	19.7	29.5
265	21.0	17.4	26.2

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002

Manual Integrations
 APPROVED

Mmdadoda
 12/8/2015 6:14:42 PM

