

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM120717\
 Data File : BM013242.D
 Acq On : 07 Dec 2017 14:29
 Operator : SJ/JU
 Sample : PB104759BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK59

Quant Time: Dec 08 01:55:05 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 08 01:54:15 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	222952	20.00	ng/ul	0.00
18) Naphthalene-d8	10.49	136	865332	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.35	164	456383	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.10	188	910758	20.00	ng/ul	0.01
75) Chrysene-d12	21.29	240	873911	20.00	ng/ul	0.01
83) Perylene-d12	23.52	264	953987	20.00	ng/ul	0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.25	96	23624	5.25	ng/uL	0.00
5) Phenol-d5	6.89	99	503802	26.11	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.05	67	287282	26.11	ng/ul	0.00
9) 2-Chlorophenol-d4	7.24	132	430309	28.55	ng/ul	0.00
13) 4-Methylphenol-d8	8.42	113	414311	25.14	ng/ul	0.00
19) Nitrobenzene-d5	8.86	128	205658	29.97	ng/ul	0.00
22) 2-Nitrophenol-d4	9.57	143	230948	31.44	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.11	165	412547	27.26	ng/ul	0.00
29) 4-Chloroaniline-d4	10.63	131	528396	37.99	ng/ul	0.01
43) Dimethylphthalate-d6	13.76	166	1149969	28.32	ng/ul	0.00
46) Acenaphthylene-d8	14.04	160	1478045	29.49	ng/ul	0.00
51) 4-Nitrophenol-d4	14.56	143	195290	27.68	ng/ul	0.00
57) Fluorene-d10	15.35	176	949585	27.19	ng/ul	0.01
62) 4,6-Dinitro-2-methylphenol	15.47	200	123403	21.54	ng/ul	0.01
70) Anthracene-d10	17.19	188	1373929	30.12	ng/ul	0.00
76) Pyrene-d10	19.49	212	1310570	30.10	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.38	264	1393460	28.68	ng/ul	0.02

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA M\DATA\BM120717\
Data File : BM013242.D
Acq On : 07 Dec 2017 14:29
Operator : SJ/JU
Sample : PB104759BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK59

Quant Time: Dec 08 01:55:05 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 08 01:54:15 2017
Response via : Initial Calibration

