

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM120717\
 Data File : BM013249.D
 Acq On : 07 Dec 2017 18:40
 Operator : SJ/JU
 Sample : I6719-19
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 JHPG1

Manual Integrations
 APPROVED

Sohil
 12/8/2017 11:15:24 AM

Quant Time: Dec 08 04:23:34 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 08 03:46:15 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	193377	20.00	ng/ul	0.00
18) Naphthalene-d8	10.48	136	775257	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.34	164	415465	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.09	188	876118	20.00	ng/ul	0.00
75) Chrysene-d12	21.28	240	859258	20.00	ng/ul	0.00
83) Perylene-d12	23.51	264	888485	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	11995	3.07	ng/uL	0.00
5) Phenol-d5	6.88	99	240262	14.36	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.04	67	159847	16.75	ng/ul	0.00
9) 2-Chlorophenol-d4	7.23	132	226397	17.32	ng/ul	0.00
13) 4-Methylphenol-d8	8.40	113	222735	15.58	ng/ul	0.00
19) Nitrobenzene-d5	8.85	128	108331	17.62	ng/ul	0.00
22) 2-Nitrophenol-d4	9.57	143	119035	18.09	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.10	165	207303	15.29	ng/ul	0.00
29) 4-Chloroaniline-d4	10.62	131	276829	22.21	ng/ul	0.00
43) Dimethylphthalate-d6	13.76	166	617786	16.71	ng/ul	0.00
46) Acenaphthylene-d8	14.03	160	804761	17.64	ng/ul	0.00
51) 4-Nitrophenol-d4	14.56	143	9913m	1.54	ng/ul	0.00
57) Fluorene-d10	15.34	176	519232	16.33	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.46	200	5141	0.93	ng/ul	0.00
70) Anthracene-d10	17.19	188	782581	17.84	ng/ul	0.00
76) Pyrene-d10	19.49	212	830178	19.39	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.37	264	811258	17.93	ng/ul	0.00
Target Compounds						
6) Phenol	6.90	94	32880	1.988	ng/ul#	85
44) Dimethylphthalate	13.80	163	163841	4.633	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA M\DATA\BM120717\
Data File : BM013249.D
Acq On : 07 Dec 2017 18:40
Operator : SJ/JU
Sample : I6719-19
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
JHPG1

Quant Time: Dec 08 04:23:34 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 08 03:46:15 2017
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
12/8/2017 11:15:24 AM





#6

Phenol

Concen: 1.988 ng/ul

RT: 6.90 min Scan# 649

Delta R.T. -0.01 min

Lab File: BM013249.D

Acq: 07 Dec 2017 18:40

Instrument :

BNA_M

ClientSampleId :





JHPG1

Tot Ion: 94 Resp: 32880

Ion	Ratio	Lower	Upper
94	100		
65	30.9	28.2	42.4
66	44.1	47.2	70.8#

Manual Integrations
APPROVED

Sohil
12/8/2017 11:15:24 AM



#44

Dimethylphthalate

Concen: 4.633 ng/ul

RT: 13.80 min Scan# 1822

Delta R.T. 0.00 min

Lab File: BM013249.D

Acq: 07 Dec 2017 18:40

Tgt Ion: 163 Resp: 163841

Ion	Ratio	Lower	Upper
163	100		
194	4.6	3.5	5.3
164	9.4	8.2	12.4

