

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120722\
 Data File : BM037869.D
 Acq On : 07 Dec 2022 13:28
 Operator : CG/JU
 Sample : SSTD01004
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010006

Quant Time: Dec 08 01:01:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 08 00:59:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.086	152	186542	20.000	ng/u1	0.00
20) Naphthalene-d8	10.910	136	745375	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.715	164	414779	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.456	188	817554	20.000	ng/u1	0.00
79) Chrysene-d12	21.621	240	779862	20.000	ng/u1	0.00
88) Perylene-d12	24.109	264	876621	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	16827	4.232	ng/uL	0.00
4) Pyridine-d5	3.869	84	113804	8.850	ng/u1	0.00
7) Phenol-d5	7.239	99	146482	8.716	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.410	67	92517	9.436	ng/u1	0.00
11) 2-Chlorophenol-d4	7.616	132	123418	9.713	ng/u1	0.00
15) 4-Methylphenol-d8	8.798	113	109850	8.219	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	58424	9.867	ng/u1	0.00
24) 2-Nitrophenol-d4	9.986	143	58158	9.492	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.527	165	114746	10.099	ng/u1	0.00
31) 4-Chloroaniline-d4	11.045	131	142855	8.197	ng/u1	0.00
46) Dimethylphthalate-d6	14.121	166	294894	10.354	ng/u1	0.00
49) Acenaphthylene-d8	14.415	160	361226	10.428	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	41143	6.959	ng/u1	0.00
60) Fluorene-d10	15.704	176	262439	10.156	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.821	200	31610	7.340	ng/u1	0.00
73) Anthracene-d10	17.551	188	387365	10.303	ng/u1	0.00
81) Pyrene-d10	19.833	212	430632	10.660	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.944	264	440457	10.030	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.463	88	17999	4.224	ng/uL	95
5) Pyridine	3.887	79	120815	9.582	ng/u1	97
6) Benzaldehyde	7.222	77	56293	6.955	ng/u1	97
8) Phenol	7.269	94	148257	8.632	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.504	93	123310	9.000	ng/u1	99
12) 2-Chlorophenol	7.651	128	125995	9.212	ng/u1	99
13) 2-Methylphenol	8.533	108	109698	8.009	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.622	45	228540	8.976	ng/u1	98
16) Acetophenone	8.922	105	178494	8.466	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.904	70	83998	7.910	ng/u1	97
18) 4-Methylphenol	8.863	108	118142	7.994	ng/u1	99
19) Hexachloroethane	9.180	117	51026	9.768	ng/u1	97
22) Nitrobenzene	9.304	77	127911	10.308	ng/u1	98
23) Isophorone	9.833	82	228664	8.565	ng/u1	99
25) 2-Nitrophenol	10.022	139	64322	9.688	ng/u1	98
26) 2,4-Dimethylphenol	10.075	107	124381	9.745	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.310	93	153819	9.542	ng/u1	99
29) 2,4-Dichlorophenol	10.551	162	113571	9.882	ng/u1	99
30) Naphthalene	10.963	128	405065	10.013	ng/u1	98
32) 4-Chloroaniline	11.074	127	144954	8.219	ng/u1	97
33) Hexachlorobutadiene	11.239	225	77634	10.905	ng/u1	98
34) Caprolactam	11.851	113	22793	5.461	ng/u1	91
35) 4-Chloro-3-methylphenol	12.180	107	99522	8.298	ng/u1	100
36) 2-Methylnaphthalene	12.557	142	258074	9.221	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120722\
 Data File : BM037869.D
 Acq On : 07 Dec 2022 13:28
 Operator : CG/JU
 Sample : SSTD01004
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD010006


Quant Time: Dec 08 01:01:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 08 00:59:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.774	142	261589	9.339	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.921	216	143117	12.048	ng/ul	96
40) Hexachlorocyclopentadiene	12.898	237	66281	10.147	ng/ul	98
41) 2,4,6-Trichlorophenol	13.157	196	80926	10.435	ng/ul	93
42) 2,4,5-Trichlorophenol	13.227	196	84652	10.249	ng/ul	95
43) 1,1'-Biphenyl	13.557	154	346134	11.481	ng/ul	98
44) 2-Chloronaphthalene	13.598	162	261052	11.034	ng/ul	99
45) 2-Nitroaniline	13.804	65	57831	8.639	ng/ul	98
47) Dimethylphthalate	14.168	163	293098	9.880	ng/ul	98
48) 2,6-Dinitrotoluene	14.292	165	52858	8.727	ng/ul	94
50) Acenaphthylene	14.439	152	397690	10.355	ng/ul	99
51) 3-Nitroaniline	14.621	138	46214	6.392	ng/ul	92
52) Acenaphthene	14.780	153	278558	10.467	ng/ul	98
53) 2,4-Dinitrophenol	14.827	184	17475	4.940	ng/ul	96
55) 4-Nitrophenol	14.921	109	26751	7.341	ng/ul	86
56) Dibenzofuran	15.115	168	383464	10.702	ng/ul	99
57) 2,4-Dinitrotoluene	15.074	165	72463	8.566	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.339	232	72176	9.935	ng/ul#	95
59) Diethylphthalate	15.521	149	290539	9.550	ng/ul	99
61) Fluorene	15.757	166	312808	10.370	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.751	204	154368	10.735	ng/ul	97
63) 4-Nitroaniline	15.780	138	46331	6.370	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.833	198	32465	7.313	ng/ul#	98
67) N-Nitrosodiphenylamine	15.962	169	244555	10.380	ng/ul	99
68) 4-Bromophenyl-phenylether	16.645	248	91248	10.876	ng/ul	97
69) Hexachlorobenzene	16.762	284	110149	11.243	ng/ul	99
70) Atrazine	16.909	200	78071	9.622	ng/ul	97
71) Pentachlorophenol	17.104	266	52213	8.665	ng/ul	99
72) Phenanthrene	17.498	178	455559	10.281	ng/ul	99
74) Anthracene	17.586	178	465132	10.290	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.521	216	137717	12.493	ng/ul	99
76) Pentachlorobenzene	15.033	250	135259	11.279	ng/ul	98
77) Carbazole	17.856	167	395411	9.506	ng/ul	98
78) Di-n-butylphthalate	18.409	149	460950	8.814	ng/ul	99
80) Fluoranthene	19.503	202	507264	10.281	ng/ul	98
82) Pyrene	19.862	202	536485	10.530	ng/ul	99
83) Butylbenzylphthalate	20.744	149	188351	8.479	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.538	252	166756	10.207	ng/ul	95
85) Benzo(a)anthracene	21.609	228	535287	10.728	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.515	149	284626	8.480	ng/ul	99
87) Chrysene	21.662	228	509846	10.977	ng/ul	99
89) Di-n-octyl phthalate	22.468	149	495720	7.858	ng/ul	100
90) Benzo(b)fluoranthene	23.344	252	544052	9.664	ng/ul	99
91) Benzo(k)fluoranthene	23.397	252	540676	9.956	ng/ul	98
93) Benzo(a)pyrene	23.997	252	476761	9.766	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.709	276	617120	11.244	ng/ul	99
95) Dibenzo(a,h)anthracene	26.726	278	516205	10.334	ng/ul	97
96) Benzo(g,h,i)perylene	27.515	276	496372	10.363	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 08 01:01:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 08 00:59:34 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.224 ng/uL
RT: 3.463 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion: 88 Resp: 17999
Ion Ratio Lower Upper
88 100
43 40.7 32.8 49.2
58 86.9 64.6 97.0



#5
Pyridine
Concen: 9.582 ng/uL
RT: 3.887 min Scan# 85
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion: 79 Resp: 120815
Ion Ratio Lower Upper
79 100
52 100.1 78.1 117.1
51 49.3 37.2 55.8









#6
Benzaldehyde
Concen: 6.955 ng/ul
RT: 7.222 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion: 77 Resp: 56293
Ion Ratio Lower Upper
77 100
105 106.5 82.5 123.7
106 104.2 81.8 122.6



#8
Phenol
Concen: 8.632 ng/ul
RT: 7.269 min Scan# 660
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion: 94 Resp: 148257
Ion Ratio Lower Upper
94 100
65 27.0 22.2 33.4
66 34.1 27.3 40.9









#10
Bis(2-Chloroethyl)ether
Concen: 9.000 ng/ul
RT: 7.504 min Scan# 700
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion: 93 Resp: 123310
Ion Ratio Lower Upper
93 100
63 83.1 66.3 99.5
95 33.2 25.7 38.5



#12
2-Chlorophenol
Concen: 9.212 ng/ul
RT: 7.651 min Scan# 725
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion: 128 Resp: 125995
Ion Ratio Lower Upper
128 100
64 44.7 35.2 52.8
130 33.0 25.9 38.9





#13

2-Methylphenol

Concen: 8.009 ng/ul

RT: 8.533 min Scan# 81

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :





SSTD010006

Tgt Ion:108 Resp: 109698

Ion Ratio Lower Upper

108 100

107 86.0 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 8.976 ng/ul

RT: 8.622 min Scan# 890

Delta R.T. -0.006 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28




Tgt Ion: 45 Resp: 228540


Ion Ratio Lower Upper

45 100

77 10.7 9.0 13.4

79 8.5 6.2 9.4









#16
Acetophenone
Concen: 8.466 ng/ul
RT: 8.922 min Scan# 941
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:105 Resp: 178494
Ion Ratio Lower Upper
105 100
77 71.7 59.7 89.5
51 33.4 27.2 40.8



#17
N-Nitroso-di-n-propylamine
Concen: 7.910 ng/ul
RT: 8.904 min Scan# 938
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion: 70 Resp: 83998
Ion Ratio Lower Upper
70 100
42 74.9 62.3 93.5
101 10.5 8.0 12.0
130 23.5 17.6 26.4





#18

4-Methylphenol

Concen: 7.994 ng/ul

RT: 8.863 min Scan# 91

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :





SSTD010006

Tgt Ion:108 Resp: 118142

Ion Ratio Lower Upper

108 100

107 109.3 88.1 132.1



#19

Hexachloroethane

Concen: 9.768 ng/ul

RT: 9.180 min Scan# 985

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28




Tgt Ion:117 Resp: 51026


Ion Ratio Lower Upper

117 100

201 106.1 87.3 130.9

199 67.1 50.9 76.3









#22
Nitrobenzene
Concen: 10.308 ng/ul
RT: 9.304 min Scan# 1006
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion: 77 Resp: 127911
Ion Ratio Lower Upper
77 100
123 48.8 37.8 56.6
65 13.4 11.1 16.7



#23
Isophorone
Concen: 8.565 ng/ul
RT: 9.833 min Scan# 1096
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion: 82 Resp: 228664
Ion Ratio Lower Upper
82 100
95 6.8 5.2 7.8
138 18.9 14.9 22.3









#25
2-Nitrophenol
Concen: 9.688 ng/ul
RT: 10.022 min Scan# 1128
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:139 Resp: 64322
Ion Ratio Lower Upper
139 100
65 45.9 36.2 54.4
109 23.0 19.7 29.5



#26
2,4-Dimethylphenol
Concen: 9.745 ng/ul
RT: 10.075 min Scan# 1137
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:107 Resp: 124381
Ion Ratio Lower Upper
107 100
121 54.1 42.0 63.0
122 92.1 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 9.542 ng/ul
 RT: 10.310 min Scan# 1177
 Delta R.T. -0.000 min
 Lab File: BM037869.D
 Acq: 07 Dec 2022 13:28




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010006


Tgt Ion: 93 Resp: 153819
 Ion Ratio Lower Upper
 93 100
 95 32.1 25.5 38.3
 123 12.3 9.4 14.2



#29
 2,4-Dichlorophenol
 Concen: 9.882 ng/ul
 RT: 10.551 min Scan# 1218
 Delta R.T. -0.000 min
 Lab File: BM037869.D
 Acq: 07 Dec 2022 13:28

Tgt Ion: 162 Resp: 113571
 Ion Ratio Lower Upper
 162 100
 164 64.6 51.4 77.2
 98 30.9 26.1 39.1









#30
Naphthalene
Concen: 10.013 ng/ul
RT: 10.963 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:128 Resp: 405065
Ion Ratio Lower Upper
128 100
129 10.4 9.0 13.4
127 12.0 10.4 15.6



#32
4-Chloroaniline
Concen: 8.219 ng/ul
RT: 11.074 min Scan# 1307
Delta R.T. 0.006 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:127 Resp: 144954
Ion Ratio Lower Upper
127 100
129 33.1 25.3 37.9









#33
Hexachlorobutadiene
Concen: 10.905 ng/ul
RT: 11.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:225 Resp: 77634
Ion Ratio Lower Upper
225 100
223 62.9 50.8 76.2
227 60.0 50.5 75.7



#34
Caprolactam
Concen: 5.461 ng/ul
RT: 11.851 min Scan# 1439
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:113 Resp: 22793
Ion Ratio Lower Upper
113 100
55 197.5 172.2 258.2
56 143.7 120.7 181.1









#35
4-Chloro-3-methylphenol
Concen: 8.298 ng/ul
RT: 12.180 min Scan# 1495
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:107 Resp: 99522
Ion Ratio Lower Upper
107 100
144 28.3 23.4 35.0
142 88.7 71.0 106.4



#36
2-Methylnaphthalene
Concen: 9.221 ng/ul
RT: 12.557 min Scan# 1559
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:142 Resp: 258074
Ion Ratio Lower Upper
142 100
141 86.8 69.4 104.0









#37
1-Methylnaphthalene
Concen: 9.339 ng/ul
RT: 12.774 min Scan# 1596
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:142 Resp: 261589
Ion Ratio Lower Upper
142 100
141 93.8 72.4 108.6
116 3.3 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 12.048 ng/ul
RT: 12.921 min Scan# 1621
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:216 Resp: 143117
Ion Ratio Lower Upper
216 100
214 75.0 64.2 96.4
179 18.7 15.4 23.2
108 16.0 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 10.147 ng/ul
RT: 12.898 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:237 Resp: 66281
Ion Ratio Lower Upper
237 100
235 60.9 50.4 75.6
272 13.0 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 10.435 ng/ul
RT: 13.157 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28


Tgt Ion:196 Resp: 80926
Ion Ratio Lower Upper
196 100
198 100.9 75.0 112.6
200 35.0 24.5 36.7








#42
2,4,5-Trichlorophenol
Concen: 10.249 ng/ul
RT: 13.227 min Scan# 1673
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Instrument :
BNA_M
ClientSampleId :
SSTD010006







Tgt Ion:196 Resp: 84652
Ion Ratio Lower Upper
196 100
198 100.5 76.5 114.7
200 33.8 24.9 37.3



#43
1,1'-Biphenyl
Concen: 11.481 ng/ul
RT: 13.557 min Scan# 1729
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:154 Resp: 346134
Ion Ratio Lower Upper
154 100
153 40.3 31.2 46.8
76 11.5 9.3 13.9





#44

2-Chloronaphthalene

Concen: 11.034 ng/ul

RT: 13.598 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:162 Resp: 261052

Ion Ratio Lower Upper

162 100

164 33.4 27.3 40.9

127 36.1 29.4 44.0



#45

2-Nitroaniline

Concen: 8.639 ng/ul

RT: 13.804 min Scan# 1771

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28




Tgt Ion: 65 Resp: 57831


Ion Ratio Lower Upper

65 100

92 65.3 53.5 80.3

138 116.9 91.3 136.9





#47

Dimethylphthalate

Concen: 9.880 ng/ul

RT: 14.168 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:163 Resp: 293098

Ion Ratio Lower Upper

163 100

194 4.1 3.0 4.6

164 9.6 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 8.727 ng/ul

RT: 14.292 min Scan# 1854

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28




Tgt Ion:165 Resp: 52858


Ion Ratio Lower Upper

165 100

89 57.5 41.1 61.7

121 22.2 17.4 26.2









#50
Acenaphthylene
Concen: 10.355 ng/ul
RT: 14.439 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:152 Resp: 397690
Ion Ratio Lower Upper
152 100
151 18.5 15.3 22.9
153 12.7 10.2 15.4



#51
3-Nitroaniline
Concen: 6.392 ng/ul
RT: 14.621 min Scan# 1910
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:138 Resp: 46214
Ion Ratio Lower Upper
138 100
108 9.4 7.9 11.9
92 113.0 83.4 125.0









#52
Acenaphthene
Concen: 10.467 ng/ul
RT: 14.780 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:153 Resp: 278558
Ion Ratio Lower Upper
153 100
152 46.8 38.2 57.4
154 87.6 71.4 107.0



#53
2,4-Dinitrophenol
Concen: 4.940 ng/ul
RT: 14.827 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:184 Resp: 17475
Ion Ratio Lower Upper
184 100
63 56.5 46.2 69.4
154 52.4 45.5 68.3









#55
4-Nitrophenol
Concen: 7.341 ng/ul
RT: 14.921 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:109 Resp: 26751
Ion Ratio Lower Upper
109 100
139 167.1 119.2 178.8
65 153.3 108.7 163.1



#56
Dibenzofuran
Concen: 10.702 ng/ul
RT: 15.115 min Scan# 1994
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:168 Resp: 383464
Ion Ratio Lower Upper
168 100
139 34.3 28.1 42.1





#57

2,4-Dinitrotoluene

Concen: 8.566 ng/ul

RT: 15.074 min Scan# 1987

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:165 Resp: 72463

Ion Ratio Lower Upper

165 100

63 42.2 34.9 52.3

182 2.8 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 9.935 ng/ul

RT: 15.339 min Scan# 2032

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Tgt Ion:232 Resp: 72176




Ion Ratio Lower Upper


232 100

131 37.6 32.8 49.2

130 2.0 2.0 3.0#

166 26.7 23.2 34.8









#59
Diethylphthalate
Concen: 9.550 ng/ul
RT: 15.521 min Scan# 2063
Delta R.T. -0.006 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:149 Resp: 290539
Ion Ratio Lower Upper
149 100
177 22.5 18.5 27.7
150 12.3 9.5 14.3



#61
Fluorene
Concen: 10.370 ng/ul
RT: 15.757 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:166 Resp: 312808
Ion Ratio Lower Upper
166 100
165 97.5 77.2 115.8
167 12.7 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 10.735 ng/ul
RT: 15.751 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:204 Resp: 154368
Ion Ratio Lower Upper
204 100
206 30.9 25.5 38.3
141 50.9 42.8 64.2



#63
4-Nitroaniline
Concen: 6.370 ng/ul
RT: 15.780 min Scan# 2107
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:138 Resp: 46331
Ion Ratio Lower Upper
138 100
92 44.4 36.1 54.1
108 64.8 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 7.313 ng/ul
RT: 15.833 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:198 Resp: 32465
Ion Ratio Lower Upper
198 100
182 5.8 3.8 5.8#
77 24.8 19.3 28.9



#67
N-Nitrosodiphenylamine
Concen: 10.380 ng/ul
RT: 15.962 min Scan# 2138
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:169 Resp: 244555
Ion Ratio Lower Upper
169 100
168 67.8 53.8 80.8
167 37.8 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 10.876 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:248 Resp: 91248
Ion Ratio Lower Upper
248 100
250 97.2 78.7 118.1
141 62.1 52.7 79.1



#69
Hexachlorobenzene
Concen: 11.243 ng/ul
RT: 16.762 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:284 Resp: 110149
Ion Ratio Lower Upper
284 100
142 29.0 23.1 34.7
249 28.1 23.3 34.9





#70

Atrazine

Concen: 9.622 ng/ul

RT: 16.909 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:200 Resp: 78071

Ion Ratio Lower Upper

200 100

173 26.8 20.5 30.7

215 48.3 40.1 60.1



#71

Pentachlorophenol

Concen: 8.665 ng/ul

RT: 17.104 min Scan# 2332

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28




Tgt Ion:266 Resp: 52213


Ion Ratio Lower Upper

266 100

264 61.7 50.1 75.1

268 64.5 50.6 75.8









#72
Phenanthrene
Concen: 10.281 ng/ul
RT: 17.498 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:178 Resp: 455559
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 18.3 15.4 23.0



#74
Anthracene
Concen: 10.290 ng/ul
RT: 17.586 min Scan# 2414
Delta R.T. -0.006 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:178 Resp: 465132
Ion Ratio Lower Upper
178 100
179 15.7 11.8 17.8
176 18.5 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 12.493 ng/uL

RT: 13.521 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:216 Resp: 137717

Ion Ratio Lower Upper

216 100

214 79.8 62.6 93.8

218 47.7 38.3 57.5



#76

Pentachlorobenzene

Concen: 11.279 ng/uL

RT: 15.033 min Scan# 1980

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28




Tgt Ion:250 Resp: 135259


Ion Ratio Lower Upper

250 100

248 62.3 51.0 76.4

252 63.5 51.8 77.6









#77
 Carbazole
 Concen: 9.506 ng/ul
 RT: 17.856 min Scan# 2460
 Delta R.T. -0.000 min
 Lab File: BM037869.D
 Acq: 07 Dec 2022 13:28




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010006


Tgt Ion:167 Resp: 395411
 Ion Ratio Lower Upper
 167 100
 166 22.2 16.7 25.1
 139 12.2 9.8 14.6



#78
 Di-n-butylphthalate
 Concen: 8.814 ng/ul
 RT: 18.409 min Scan# 2554
 Delta R.T. -0.000 min
 Lab File: BM037869.D
 Acq: 07 Dec 2022 13:28

Tgt Ion:149 Resp: 460950
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 4.7 3.4 5.2









#80
Fluoranthene
Concen: 10.281 ng/ul
RT: 19.503 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:202 Resp: 507264
Ion Ratio Lower Upper
202 100
101 10.2 8.6 13.0
100 7.2 6.6 10.0



#82
Pyrene
Concen: 10.530 ng/ul
RT: 19.862 min Scan# 2801
Delta R.T. -0.006 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:202 Resp: 536485
Ion Ratio Lower Upper
202 100
101 12.7 10.0 15.0
100 10.8 7.9 11.9









#83
Butylbenzylphthalate
Concen: 8.479 ng/ul
RT: 20.744 min Scan# 2951
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:149 Resp: 188351
Ion Ratio Lower Upper
149 100
91 67.0 54.5 81.7
206 21.2 15.8 23.8



#84
3,3'-Dichlorobenzidine
Concen: 10.207 ng/ul
RT: 21.538 min Scan# 3086
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:252 Resp: 166756
Ion Ratio Lower Upper
252 100
254 62.2 53.2 79.8
126 10.4 7.8 11.8









#85
Benzo(a)anthracene
Concen: 10.728 ng/ul
RT: 21.609 min Scan# 3098
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:228 Resp: 535287
Ion Ratio Lower Upper
228 100
229 19.4 15.5 23.3
226 27.2 21.8 32.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 8.480 ng/ul
RT: 21.515 min Scan# 3082
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:149 Resp: 284626
Ion Ratio Lower Upper
149 100
167 27.4 21.5 32.3
279 3.8 3.2 4.8









#87
Chrysene
Concen: 10.977 ng/ul
RT: 21.662 min Scan# 3107
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:228 Resp: 509846
Ion Ratio Lower Upper
228 100
226 30.2 24.3 36.5
229 20.0 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 7.858 ng/ul
RT: 22.468 min Scan# 3244
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:149 Resp: 495720









#90
Benzo(b)fluoranthene
Concen: 9.664 ng/ul
RT: 23.344 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28




Instrument :
BNA_M
ClientSampleId :
SSTD010006


Tgt Ion:252 Resp: 544052
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.1 8.0 12.0



#91
Benzo(k)fluoranthene
Concen: 9.956 ng/ul
RT: 23.397 min Scan# 3402
Delta R.T. -0.000 min
Lab File: BM037869.D
Acq: 07 Dec 2022 13:28

Tgt Ion:252 Resp: 540676
Ion Ratio Lower Upper
252 100
253 21.0 17.4 26.2
125 10.9 8.1 12.1





#93

Benzo(a)pyrene

Concen: 9.766 ng/ul

RT: 23.997 min Scan# 3

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:252 Resp: 476761

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0

125 11.4 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 11.244 ng/ul

RT: 26.709 min Scan# 3965

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28




Tgt Ion:276 Resp: 617120


Ion Ratio Lower Upper

276 100

138 23.3 19.3 28.9

277 24.6 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 10.334 ng/ul

RT: 26.726 min Scan# 3968

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Instrument :

BNA_M

ClientSampleId :

SSTD010006





Tgt Ion:278 Resp: 516205

Ion Ratio Lower Upper

278 100

139 14.3 12.4 18.6

279 23.5 19.8 29.6



#96

Benzo(g,h,i)perylene

Concen: 10.363 ng/ul

RT: 27.515 min Scan# 4102

Delta R.T. -0.000 min

Lab File: BM037869.D

Acq: 07 Dec 2022 13:28

Tgt Ion:276 Resp: 496372

Ion Ratio Lower Upper

276 100

138 21.8 17.1 25.7

277 24.3 18.2 27.4

