


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM120816\
 Data File : BM008195.D
 Acq On : 08 Dec 2016 10:03
 Operator : UM/SJ
 Sample : DFTPP62
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP62

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM112916.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 02 03:12:28 2016



AutoFind: Scans 2537, 2538, 2539; Background Corrected with Scan 2531

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	79026	PASS
68	69	0.00	2	1.8	1620	PASS
69	198	0.00	100	37.6	89248	PASS
70	69	0.00	2	0.5	469	PASS
127	198	10	80	41.9	99664	PASS
197	198	0.00	2	0.5	1141	PASS
198	198	100	100	100.0	237632	PASS
199	198	5	9	6.8	16115	PASS
275	198	10	60	30.1	71600	PASS
365	198	1	100	4.4	10552	PASS
441	443	0.01	100	79.5	31061	PASS
442	198	50	100	85.8	203989	PASS
443	442	15	24	19.2	39082	PASS

m/z	Abundance
37.20	301.0
38.00	1048.0
39.10	6033.0
40.00	1759.0
44.00	741.0
49.00	735.0
50.10	23488.0
51.10	87552.0
52.10	4712.0
55.00	485.0
56.00	2830.0
57.00	5883.0
58.10	302.0
61.00	1774.0
62.00	1882.0
63.00	3727.0
64.00	605.0
65.00	1436.0
68.00	1606.0
69.00	96992.0
70.10	522.0
73.00	1422.0
74.00	10332.0
75.00	16000.0
76.10	5190.0
77.10	108504.0
78.00	6270.0
79.00	8027.0
80.00	5810.0
81.00	8817.0
82.00	2036.0
83.00	1710.0
85.00	1459.0
86.00	2625.0
87.10	1088.0
87.90	582.0
91.00	2208.0
92.00	2477.0
93.00	15124.0
94.00	878.0
96.00	763.0
96.90	332.0
98.00	11412.0
99.00	9608.0
100.00	898.0
101.00	4180.0
101.90	319.0
103.00	1476.0
104.00	3154.0
105.00	2588.0
105.90	1023.0
107.00	33920.0
108.00	5319.0
109.10	930.0
110.00	56376.0
111.00	9595.0
112.00	879.0
113.00	318.0
116.00	1608.0
117.00	34688.0
118.00	2792.0
118.90	320.0
120.00	402.0
121.00	331.0
122.10	2062.0
123.00	3498.0
124.00	1858.0
125.00	1194.0
126.10	486.0
127.00	103008.0
128.00	8519.0
129.00	43552.0
130.00	3517.0
131.10	752.0
134.00	1197.0
135.00	4180.0
136.00	1440.0
137.10	2076.0
138.00	315.0
140.00	571.0
141.00	6532.0
142.00	1980.0
143.00	1210.0
144.10	329.0
145.10	374.0
146.00	823.0
147.00	3643.0
148.00	7842.0
149.00	1990.0
151.00	787.0
152.90	2018.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

154.00	1110.0
155.10	2965.0
156.10	4925.0
156.90	951.0
158.00	1381.0
159.00	837.0
160.00	2385.0
161.00	2986.0
161.90	764.0
165.00	2353.0
166.00	1686.0
167.00	13116.0
168.00	8788.0
169.10	1330.0
171.00	640.0
171.90	1325.0
173.10	1807.0
174.10	2323.0
175.00	4659.0
175.90	1205.0
177.00	2702.0
178.00	978.0
179.00	9428.0
180.00	7763.0
181.00	3227.0
182.10	315.0
184.00	922.0
185.00	3977.0
186.00	33488.0
187.00	10586.0
188.00	1452.0
189.00	2345.0
191.00	1034.0
192.00	3385.0
193.00	3038.0
194.00	1070.0
194.90	559.0
196.00	7445.0
197.00	1183.0
198.00	225216.0
199.00	15243.0
199.90	1380.0
201.50	751.0
203.00	1820.0
204.10	8560.0
205.00	15672.0
206.10	65728.0
207.00	9628.0
208.00	3221.0
209.00	976.0
210.00	1242.0
211.00	2371.0
215.00	989.0
216.00	1240.0
217.00	19448.0
218.00	2085.0
221.00	8410.0
221.80	2164.0
223.00	4686.0
224.00	35144.0
225.10	9532.0
226.10	1304.0
227.00	16968.0
228.00	2453.0
229.10	3267.0
229.90	371.0
231.00	1319.0
233.10	320.0
234.00	1304.0
235.00	1056.0
236.00	861.0
237.00	2108.0
239.00	1029.0
240.00	565.0
241.00	1086.0
242.00	2291.0
243.10	1753.0
244.10	26720.0
245.10	2934.0
246.00	6977.0
247.00	1300.0
249.00	756.0
250.00	351.0
251.90	305.0
252.90	689.0
254.10	799.0
255.00	149376.0
256.00	21584.0
257.10	2256.0
258.00	9668.0
259.00	1533.0
265.00	4457.0
265.90	535.0
270.90	476.0
272.00	502.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

273.00	4532.0
274.00	11236.0
275.00	64024.0
276.00	8194.0
277.00	6292.0
278.00	1398.0
279.00	362.0
281.10	348.0
283.10	793.0
284.00	525.0
285.10	836.0
288.90	359.0
290.00	335.0
292.00	311.0
293.00	1598.0
293.90	365.0
295.20	395.0
296.00	22056.0
297.00	3044.0
300.90	451.0
301.90	368.0
303.00	2480.0
303.90	489.0
308.00	487.0
314.10	769.0
315.00	2250.0
316.00	1038.0
321.00	500.0
323.10	4792.0
324.10	1337.0
327.00	1249.0
327.90	507.0
332.00	405.0
333.00	469.0
334.10	3318.0
335.00	968.0
341.00	895.0
346.00	1509.0
347.00	330.0
352.00	1379.0
352.90	1041.0
354.00	2026.0
355.00	518.0
365.00	8230.0
366.00	1015.0
370.90	450.0
372.10	2432.0
372.90	530.0
383.00	749.0
390.00	313.0
402.00	972.0
403.00	1521.0
404.10	557.0
421.00	1035.0
422.10	914.0
423.00	7789.0
424.10	2732.0
425.10	362.0
441.10	23104.0
442.00	151360.0
443.10	31048.0
444.00	3484.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
37.00	496.0
38.10	1050.0
39.10	6844.0
40.00	1588.0
41.00	329.0
44.00	1156.0
49.10	952.0
50.10	26048.0
51.10	94984.0
52.00	5172.0
54.90	690.0
56.00	2870.0
57.00	6436.0
57.90	377.0
61.00	1331.0
62.00	1557.0
63.10	4170.0
64.00	526.0
65.10	2286.0
68.10	2074.0
69.00	106440.0
70.10	489.0
73.00	1373.0
74.10	10867.0
75.00	17048.0
76.10	5056.0
77.10	118896.0
78.10	8030.0
79.00	9105.0
80.00	6239.0
81.00	9613.0
82.00	2233.0
83.00	1967.0
85.00	1957.0
86.00	2831.0
87.00	1069.0
88.00	369.0
89.00	303.0
91.00	1858.0
92.00	2675.0
93.00	16616.0
94.00	1142.0
96.10	595.0
97.10	334.0
98.00	13535.0
99.00	9887.0
100.10	887.0
101.00	4910.0
103.00	2181.0
104.00	3031.0
105.10	2850.0
106.00	869.0
107.00	38576.0
108.00	5942.0
109.10	1178.0
110.00	64936.0
111.00	10667.0
112.00	1409.0
112.90	315.0
116.00	2419.0
117.00	39432.0
118.00	2608.0
118.90	361.0
119.90	398.0
122.00	2298.0
123.10	3995.0
124.00	1985.0
125.00	1811.0
126.10	393.0
127.00	119480.0
128.00	9223.0
129.00	52400.0
130.00	4405.0
131.10	876.0
131.90	380.0
133.00	373.0
134.00	1811.0
135.00	4298.0
136.00	1671.0
137.00	2654.0
140.00	860.0
141.00	7162.0
142.00	2240.0
143.00	1309.0
143.90	471.0
145.00	419.0
146.00	1566.0
147.00	4282.0
148.00	9558.0
149.00	2078.0
149.90	601.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

151.00	851.0
152.00	540.0
153.00	2428.0
154.00	1263.0
155.00	3935.0
156.10	5653.0
157.00	1247.0
157.90	1283.0
158.90	955.0
160.00	2761.0
161.00	3541.0
162.00	832.0
163.90	391.0
165.00	3045.0
166.00	2212.0
167.00	15548.0
168.00	8957.0
169.00	1560.0
169.90	452.0
171.00	551.0
171.90	1775.0
173.00	1707.0
174.00	3127.0
175.00	5955.0
176.00	1405.0
176.90	2803.0
177.90	1299.0
179.00	11757.0
180.00	8476.0
181.00	3895.0
182.00	605.0
183.00	415.0
184.00	439.0
185.10	5321.0
186.00	43144.0
187.00	12468.0
188.00	1467.0
189.00	3195.0
190.00	552.0
191.10	1583.0
192.00	3808.0
193.10	4914.0
193.90	907.0
196.00	7734.0
197.10	1556.0
198.00	287232.0
199.00	20328.0
200.00	1197.0
200.90	326.0
201.50	965.0
203.10	2066.0
204.00	11059.0
205.00	20552.0
206.10	83680.0
207.10	10767.0
208.00	3433.0
209.10	1312.0
209.90	1377.0
211.00	3014.0
212.00	488.0
215.00	985.0
216.00	1964.0
217.00	24112.0
218.00	2391.0
218.90	312.0
221.00	11086.0
222.00	2748.0
223.00	5167.0
224.00	45832.0
225.10	11535.0
226.10	1617.0
227.00	21032.0
228.00	3397.0
229.00	4533.0
229.90	405.0
231.00	1402.0
233.00	429.0
234.00	1419.0
235.00	1278.0
236.00	1453.0
237.00	2231.0
239.00	968.0
240.00	764.0
241.00	1091.0
242.10	2754.0
243.00	2548.0
244.10	34744.0
245.10	4000.0
246.00	8015.0
247.00	1688.0
248.90	1463.0
250.00	326.0
251.00	384.0
251.90	358.0
253.10	713.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

254.10	1679.0
255.00	198656.0
256.00	28624.0
257.00	2318.0
258.00	12607.0
259.00	1811.0
260.00	352.0
260.90	302.0
265.00	5179.0
265.90	739.0
271.10	339.0
272.00	687.0
273.00	5134.0
274.00	15209.0
275.00	85256.0
276.00	11614.0
277.00	7670.0
278.00	1514.0
279.00	511.0
281.00	614.0
282.00	309.0
283.10	842.0
283.90	496.0
285.00	1621.0
288.90	385.0
293.00	1650.0
294.00	436.0
294.90	455.0
296.00	28800.0
297.00	3762.0
302.00	394.0
303.00	3048.0
304.00	983.0
310.10	395.0
314.10	1231.0
315.00	3108.0
316.00	1533.0
321.00	744.0
321.90	468.0
323.10	7735.0
324.00	1720.0
324.90	306.0
326.90	1981.0
327.90	605.0
332.00	483.0
333.00	938.0
334.00	5180.0
335.00	1459.0
341.10	1033.0
342.10	493.0
346.00	1721.0
347.00	539.0
350.90	309.0
352.00	2621.0
353.00	1778.0
354.10	2695.0
355.10	572.0
365.00	12380.0
366.00	1406.0
369.90	303.0
371.00	495.0
372.00	3995.0
373.10	1424.0
382.90	1219.0
390.00	523.0
391.00	499.0
391.90	301.0
402.00	1643.0
403.00	2063.0
404.00	521.0
421.00	1717.0
422.00	1580.0
423.10	12642.0
424.10	3901.0
425.00	477.0
440.10	305.0
441.00	35832.0
442.00	236672.0
443.10	45440.0
444.10	4453.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
38.00	709.0
39.10	3614.0
40.00	1812.0
44.00	656.0
49.20	455.0
50.10	14472.0
51.10	54544.0
52.00	2795.0
56.00	2137.0
57.00	3700.0
60.90	876.0
62.00	795.0
63.10	2499.0
64.10	374.0
65.00	918.0
68.10	1180.0
69.00	64312.0
70.10	396.0
73.10	883.0
74.00	7851.0
75.10	10285.0
76.10	3933.0
77.10	69992.0
78.10	4766.0
79.00	5328.0
80.00	3822.0
81.00	5965.0
82.00	1004.0
83.10	1239.0
85.10	1129.0
86.00	1789.0
87.10	662.0
91.00	1564.0
92.10	1669.0
93.00	9981.0
93.90	607.0
96.20	553.0
98.00	7519.0
99.10	6010.0
100.00	533.0
101.00	3412.0
103.00	938.0
104.00	2172.0
105.00	2081.0
106.00	529.0
107.00	24344.0
108.00	3698.0
109.10	737.0
110.00	41552.0
111.00	6625.0
112.10	927.0
116.00	1591.0
117.00	24192.0
118.00	1707.0
119.00	376.0
120.10	361.0
122.00	2144.0
123.00	2775.0
124.00	1371.0
125.00	1069.0
126.10	408.0
127.00	76504.0
128.00	5801.0
129.00	33184.0
130.00	3075.0
131.10	448.0
134.00	791.0
135.00	2604.0
136.00	766.0
137.00	1616.0
138.00	306.0
139.80	435.0
141.00	5040.0
142.00	1474.0
143.00	1125.0
144.10	320.0
145.90	861.0
147.00	2635.0
148.00	6396.0
148.90	1513.0
149.90	305.0
151.00	454.0
153.00	1620.0
154.00	1378.0
155.00	2961.0
156.00	3434.0
157.00	540.0
158.00	942.0
159.00	601.0
160.10	1571.0
161.00	2302.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

161.90	695.0
164.90	1690.0
166.00	1529.0
167.00	11051.0
168.00	5322.0
169.00	820.0
169.90	385.0
171.10	551.0
172.00	859.0
173.10	1287.0
174.00	2225.0
175.00	3778.0
176.00	898.0
177.00	2078.0
177.90	493.0
179.00	8663.0
180.00	5528.0
181.00	2422.0
182.00	432.0
184.00	709.0
185.00	3479.0
186.10	29008.0
187.00	8268.0
188.10	896.0
189.00	1863.0
189.90	341.0
191.00	1068.0
192.00	2478.0
193.00	3053.0
194.10	700.0
196.00	5474.0
197.00	684.0
198.00	200448.0
199.00	12776.0
200.00	1090.0
201.30	809.0
203.00	1583.0
204.00	8145.0
205.00	15063.0
206.10	56776.0
207.00	7843.0
208.00	2647.0
209.00	840.0
210.00	889.0
211.00	2448.0
215.00	847.0
216.00	1394.0
217.00	16656.0
218.00	2221.0
221.00	8161.0
222.00	2489.0
223.00	4491.0
224.00	34272.0
225.00	8775.0
226.10	1126.0
227.00	17016.0
228.00	1921.0
229.00	3452.0
230.00	473.0
231.00	1066.0
234.00	1181.0
235.00	1227.0
235.90	644.0
237.10	1915.0
238.00	349.0
238.80	696.0
240.00	342.0
241.00	912.0
242.10	1388.0
243.00	1428.0
244.10	25448.0
245.10	3032.0
246.00	6277.0
247.10	744.0
248.90	1099.0
251.80	314.0
253.00	818.0
254.10	805.0
255.00	146048.0
256.00	20760.0
257.00	1671.0
258.00	9323.0
259.00	1729.0
260.90	344.0
265.00	3832.0
266.00	824.0
271.00	388.0
272.00	436.0
273.00	4872.0
274.00	11569.0
275.00	65520.0
276.00	8300.0
277.00	6234.0
277.90	872.0
278.90	355.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

281.00	422.0
283.00	1033.0
284.00	579.0
285.10	1022.0
289.00	394.0
292.10	332.0
292.90	1380.0
293.90	335.0
295.10	612.0
296.00	22240.0
297.00	2812.0
302.10	341.0
303.00	2448.0
304.00	584.0
308.00	309.0
310.00	347.0
314.00	1211.0
315.00	2715.0
316.00	1174.0
316.90	369.0
321.00	621.0
322.00	397.0
323.00	7434.0
324.00	1088.0
326.90	1471.0
328.10	596.0
332.00	555.0
332.90	623.0
334.00	5039.0
335.00	1246.0
341.10	725.0
346.00	1308.0
352.00	1688.0
353.00	1479.0
354.10	2315.0
355.10	715.0
365.00	11046.0
366.00	1572.0
371.00	335.0
372.10	3325.0
373.00	911.0
383.00	1001.0
384.00	317.0
402.00	1311.0
403.00	2073.0
404.00	658.0
421.10	1882.0
422.00	1216.0
423.00	11277.0
424.00	2772.0
425.00	373.0
441.10	34248.0
442.10	223936.0
443.10	40760.0
444.00	3793.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

Scan 2531 (17.474 min): BM008195.D
DFTPP62

m/z	Abundance
40.00	1641.0
44.00	787.0
73.00	357.0
207.10	1068.0
281.00	436.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62