

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120817\
 Data File : BM013266.D
 Acq On : 08 Dec 2017 13:45
 Operator : SJ/JU
 Sample : PB104812BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK12

Quant Time: Dec 08 16:57:01 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 08 16:56:30 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	162764	20.00	ng/ul	0.00
18) Naphthalene-d8	10.47	136	752070	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.34	164	504244	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.09	188	1232856	20.00	ng/ul	0.00
75) Chrysene-d12	21.27	240	1414969	20.00	ng/ul	0.00
83) Perylene-d12	23.50	264	1420635	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.23	96	14510	4.42	ng/uL	0.00
5) Phenol-d5	6.87	99	316443	22.47	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	194018	24.16	ng/ul	0.00
9) 2-Chlorophenol-d4	7.23	132	259158	23.55	ng/ul	0.00
13) 4-Methylphenol-d8	8.40	113	280570	23.32	ng/ul	0.00
19) Nitrobenzene-d5	8.85	128	140363	23.53	ng/ul	0.00
22) 2-Nitrophenol-d4	9.56	143	156323	24.49	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.10	165	277613	21.10	ng/ul	0.00
29) 4-Chloroaniline-d4	10.62	131	394934	32.67	ng/ul	0.00
43) Dimethylphthalate-d6	13.75	166	1059470	23.61	ng/ul	0.00
46) Acenaphthylene-d8	14.03	160	1234395	22.29	ng/ul	0.00
51) 4-Nitrophenol-d4	14.55	143	181097	23.23	ng/ul	0.00
57) Fluorene-d10	15.33	176	858252	22.24	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.46	200	163706	21.11	ng/ul	0.00
70) Anthracene-d10	17.19	188	1410778	22.85	ng/ul	0.00
76) Pyrene-d10	19.48	212	1630764	23.13	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.36	264	1649305	22.79	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120817\
Data File : BM013266.D
Acq On : 08 Dec 2017 13:45
Operator : SJ/JU
Sample : PB104812BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK12

Quant Time: Dec 08 16:57:01 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 08 16:56:30 2017
Response via : Initial Calibration

