


Data Path : S:\HPCHEM1\BNA_M\Data\BM120817\
 Data File : BM013263.D
 Acq On : 08 Dec 2017 11:55
 Operator : SJ/JU
 Sample : DFTPP59
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP59

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM113017.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 08 23:45:50 2017



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2449

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.1	67173	PASS
68	69	0.00	2	1.2	859	PASS
69	198	0.00	100	39.6	73658	PASS
70	69	0.00	2	1.0	717	PASS
127	198	10	80	44.6	83096	PASS
197	198	0.00	2	0.5	842	PASS
198	198	100	100	100.0	186112	PASS
199	198	5	9	6.7	12462	PASS
275	198	10	60	27.7	51541	PASS
365	198	1	100	4.1	7566	PASS
441	443	0.01	100	79.1	24501	PASS
442	198	50	100	86.4	160770	PASS
443	442	15	24	19.3	30984	PASS

m/z	Abundance
37.00	520.0
38.00	1279.0
39.10	6391.0
39.90	1118.0
41.00	1064.0
42.00	991.0
43.00	3143.0
43.90	2908.0
45.00	2414.0
49.10	851.0
50.10	21912.0
51.00	78016.0
52.10	4700.0
52.90	810.0
54.00	436.0
54.90	1534.0
56.00	2179.0
57.00	6671.0
58.10	591.0
58.90	312.0
59.90	1244.0
60.90	978.0
62.10	1341.0
63.10	3385.0
63.80	478.0
65.10	1403.0
65.80	355.0
67.20	607.0
67.90	1649.0
69.00	85768.0
70.00	1310.0
73.10	2374.0
74.00	9628.0
75.00	14684.0
76.10	4171.0
77.00	88792.0
78.00	6923.0
79.00	7305.0
80.00	5236.0
81.00	6799.0
82.00	1567.0
83.00	1235.0
84.10	616.0
84.90	1791.0
86.00	2185.0
87.00	980.0
89.00	488.0
91.10	1877.0
92.00	1760.0
93.10	13654.0
94.10	1809.0
96.10	704.0
97.10	770.0
98.00	10351.0
99.00	7428.0
99.90	430.0
101.00	3670.0
101.80	457.0
103.00	1373.0
103.90	2601.0
105.10	3092.0
106.00	980.0
107.00	29440.0
108.00	5355.0
108.90	1391.0
110.00	49992.0
111.00	7671.0
111.90	1097.0
112.90	415.0
116.20	1344.0
117.00	28920.0
118.10	1758.0
119.10	482.0
119.80	339.0
121.00	726.0
122.00	2328.0
123.00	4391.0
123.90	1463.0
124.90	1516.0
126.10	705.0
127.00	89728.0
128.00	7557.0
129.00	38936.0
129.90	3419.0
130.90	852.0
132.00	374.0
134.10	1448.0
135.00	3473.0
136.20	1581.0
137.00	1974.0
138.00	453.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

140.10	487.0
140.90	6007.0
142.00	2102.0
143.00	1077.0
144.20	334.0
144.90	319.0
145.80	763.0
147.00	3850.0
148.00	6195.0
149.10	2039.0
149.90	565.0
151.20	660.0
152.00	624.0
152.90	1492.0
154.00	833.0
155.10	3365.0
156.00	4706.0
157.10	831.0
158.10	1047.0
159.10	415.0
160.00	1346.0
161.00	1619.0
162.00	505.0
162.90	431.0
163.90	460.0
165.00	2062.0
166.00	1780.0
167.00	9962.0
168.10	4981.0
169.00	1291.0
169.90	360.0
170.90	521.0
172.00	794.0
173.00	903.0
173.90	2444.0
174.90	3622.0
175.90	1311.0
176.80	1878.0
178.20	743.0
179.00	7423.0
180.00	5300.0
181.00	1825.0
182.10	636.0
182.80	383.0
183.90	733.0
185.00	4067.0
186.00	25528.0
187.10	7562.0
187.90	1026.0
189.00	3315.0
189.90	746.0
191.00	1107.0
192.00	3119.0
193.00	3132.0
194.00	402.0
196.00	5385.0
197.90	185472.0
199.00	13070.0
200.10	1083.0
201.30	815.0
201.90	367.0
202.90	1029.0
204.00	8031.0
205.00	12388.0
206.00	50216.0
207.00	8484.0
208.10	2209.0
209.10	1365.0
210.20	451.0
211.10	1734.0
215.00	584.0
216.00	1051.0
216.90	14924.0
217.90	1648.0
219.90	425.0
220.90	6956.0
221.90	2449.0
223.00	2935.0
224.00	27712.0
225.00	6984.0
225.90	425.0
227.00	14297.0
227.90	1603.0
228.90	1784.0
229.90	374.0
231.00	568.0
234.00	451.0
235.00	1124.0
235.90	613.0
237.00	1278.0
239.10	498.0
240.20	302.0
240.90	713.0
241.90	1082.0
243.00	1773.0

Instrument :

BNA_M

ClientSampleId :

DFTPP59

244.00	22104.0
245.00	3037.0
246.00	5318.0
246.90	1064.0
248.90	1135.0
251.10	481.0
253.00	471.0
254.20	1028.0
255.00	117328.0
256.00	16097.0
256.90	1023.0
257.90	5614.0
259.10	817.0
260.90	403.0
265.00	3000.0
266.00	328.0
271.80	329.0
273.00	3395.0
274.00	7371.0
275.00	47072.0
276.00	6615.0
277.00	3879.0
278.00	500.0
279.00	333.0
280.90	366.0
282.00	350.0
283.00	505.0
284.90	1112.0
293.00	1138.0
294.80	407.0
296.00	13711.0
297.00	2859.0
298.10	355.0
301.90	366.0
303.10	1424.0
304.20	638.0
310.00	330.0
313.00	479.0
313.90	1001.0
315.00	1591.0
316.00	649.0
323.10	4144.0
324.00	569.0
326.70	726.0
327.10	689.0
328.20	759.0
333.00	325.0
334.00	3545.0
335.10	728.0
341.00	628.0
342.00	532.0
346.00	922.0
352.00	1426.0
352.90	558.0
354.10	1154.0
365.00	5699.0
365.90	812.0
371.10	483.0
372.00	2969.0
373.10	998.0
383.00	401.0
390.10	319.0
391.90	371.0
401.10	349.0
402.00	1283.0
403.00	1133.0
404.10	543.0
420.90	703.0
422.10	1472.0
423.10	8108.0
424.00	1659.0
441.10	19504.0
442.10	118920.0
443.10	24616.0
444.10	2173.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

m/z	Abundance
36.90	540.0
38.00	1293.0
39.00	6917.0
39.90	521.0
41.10	1497.0
41.90	801.0
43.00	3011.0
44.00	3059.0
45.00	2433.0
48.90	705.0
50.10	21176.0
51.10	79440.0
52.00	3608.0
54.10	331.0
54.90	1218.0
56.00	3382.0
57.00	6679.0
58.10	583.0
60.00	1510.0
61.00	1275.0
62.10	1093.0
63.10	3458.0
64.00	657.0
65.00	1666.0
67.00	523.0
68.10	1664.0
69.00	88960.0
70.00	842.0
71.00	498.0
73.10	1715.0
74.00	8898.0
75.00	13662.0
76.00	5921.0
77.00	91608.0
78.10	6601.0
79.00	7082.0
80.00	5473.0
81.00	7330.0
82.00	2150.0
83.20	1622.0
84.00	328.0
84.90	1186.0
86.10	2921.0
87.00	1748.0
87.90	474.0
88.90	479.0
91.00	2323.0
91.90	1623.0
93.00	11295.0
94.10	866.0
95.10	472.0
96.00	1035.0
96.90	580.0
98.00	10506.0
99.00	7842.0
99.90	372.0
101.00	4896.0
102.10	372.0
103.10	1221.0
104.00	2490.0
105.00	2363.0
106.20	720.0
107.00	31304.0
108.00	4853.0
109.10	1125.0
110.00	52848.0
111.00	8429.0
111.90	1360.0
113.00	549.0
114.80	501.0
115.90	2110.0
117.00	31072.0
118.10	2268.0
119.10	493.0
119.80	439.0
120.90	301.0
122.00	2615.0
122.90	3994.0
123.90	1665.0
125.00	1438.0
126.10	735.0
127.00	96568.0
128.00	7071.0
129.00	41648.0
130.10	4076.0
131.10	312.0
132.00	364.0
133.00	570.0
133.80	977.0
135.00	4343.0
136.10	2087.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

136.90	1303.0
137.80	587.0
139.00	439.0
139.90	620.0
141.00	6730.0
142.10	1618.0
143.10	1218.0
144.10	301.0
144.90	326.0
146.00	1224.0
146.90	3339.0
148.00	6651.0
149.00	1771.0
150.00	515.0
151.00	313.0
151.60	727.0
152.90	2340.0
154.00	1601.0
155.10	4130.0
156.00	5350.0
157.00	1300.0
158.00	1432.0
158.90	1233.0
159.90	1374.0
161.10	1807.0
161.70	731.0
162.10	692.0
162.90	337.0
164.00	400.0
164.90	2001.0
165.80	1348.0
166.20	1568.0
167.00	11742.0
168.00	7327.0
169.00	1311.0
169.90	757.0
170.80	466.0
172.10	1074.0
173.00	1810.0
174.00	2736.0
175.10	4120.0
176.00	1421.0
177.10	1842.0
177.90	709.0
178.90	7155.0
180.00	5833.0
181.00	3068.0
181.90	561.0
182.80	311.0
183.90	727.0
184.30	682.0
185.10	4844.0
186.10	32384.0
187.00	9178.0
188.00	1397.0
188.90	2310.0
190.00	384.0
191.10	1411.0
191.90	2816.0
192.90	3277.0
193.90	441.0
194.20	546.0
196.00	6147.0
197.00	1459.0
198.00	225600.0
199.00	13658.0
199.90	1578.0
201.40	594.0
203.00	2599.0
203.90	9693.0
205.10	14828.0
206.00	62968.0
207.10	9588.0
208.00	2323.0
209.10	1294.0
209.90	1416.0
210.90	2186.0
211.80	373.0
213.10	659.0
215.00	1250.0
216.00	1186.0
217.00	16696.0
218.00	2320.0
219.80	420.0
221.10	7228.0
222.00	2318.0
223.00	3501.0
224.00	32808.0
225.00	8215.0
225.80	1239.0
227.00	13942.0
228.10	2827.0
229.00	2887.0
229.80	736.0
230.90	1457.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

232.10	359.0
232.90	472.0
234.10	1053.0
234.80	691.0
235.10	761.0
236.00	696.0
236.90	1463.0
239.00	971.0
240.00	345.0
240.70	766.0
241.10	629.0
241.90	2119.0
243.10	2020.0
244.00	25656.0
245.00	4638.0
245.90	6907.0
247.00	1638.0
249.10	1050.0
250.20	426.0
251.10	364.0
252.90	863.0
254.10	1274.0
255.00	152256.0
256.00	20248.0
257.00	2170.0
258.00	7413.0
258.90	1312.0
265.00	4032.0
265.90	731.0
267.10	308.0
269.90	313.0
270.90	555.0
273.00	4616.0
274.00	10717.0
275.00	63984.0
276.00	8182.0
277.00	6431.0
278.00	1305.0
281.20	343.0
282.00	385.0
283.10	595.0
284.00	619.0
285.00	1281.0
288.90	438.0
291.90	310.0
293.00	1591.0
294.00	318.0
295.00	744.0
296.00	21112.0
297.00	2227.0
298.10	313.0
300.90	767.0
301.90	1022.0
302.90	2922.0
304.10	865.0
313.00	490.0
314.00	690.0
314.80	1667.0
315.10	1713.0
315.90	1325.0
320.90	1013.0
323.10	6243.0
324.00	1044.0
326.80	814.0
327.80	320.0
332.10	581.0
333.00	628.0
334.10	2769.0
335.10	1328.0
339.90	366.0
341.10	469.0
341.80	310.0
346.00	1578.0
352.10	2164.0
352.90	1590.0
354.20	1817.0
365.00	9801.0
365.90	1296.0
371.00	766.0
372.00	3155.0
373.10	765.0
382.90	574.0
384.10	331.0
392.00	408.0
402.10	1162.0
403.10	2189.0
404.10	666.0
421.10	2234.0
422.10	2168.0
423.10	9894.0
424.10	2095.0
424.90	309.0
441.10	29688.0
442.10	196544.0
443.10	36392.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

444.103824.0

444.80511.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

m/z	Abundance
37.90	776.0
39.00	4547.0
40.00	850.0
41.00	1090.0
42.10	819.0
43.10	3211.0
44.00	3903.0
44.90	2239.0
48.90	635.0
50.10	13288.0
51.10	44064.0
52.00	2414.0
53.00	1182.0
55.00	1605.0
56.10	1624.0
57.00	3940.0
58.10	342.0
60.00	1377.0
61.00	618.0
62.10	1209.0
63.10	1880.0
63.90	338.0
65.10	821.0
67.10	325.0
68.10	913.0
69.00	49320.0
70.20	376.0
70.90	617.0
73.00	2394.0
74.00	5390.0
75.00	8207.0
76.00	2765.0
77.00	58784.0
78.10	3207.0
78.90	3803.0
80.00	3804.0
81.00	5447.0
81.90	868.0
83.00	1425.0
83.90	322.0
85.00	1188.0
85.90	1143.0
86.80	323.0
87.30	455.0
91.00	2011.0
92.10	1230.0
93.00	8767.0
94.20	313.0
95.10	788.0
96.10	458.0
97.00	863.0
98.00	7061.0
99.00	4087.0
99.90	345.0
101.10	2637.0
102.90	810.0
103.90	1894.0
105.00	1619.0
106.00	815.0
107.00	16896.0
108.00	2453.0
109.00	527.0
110.00	29872.0
111.00	5609.0
112.00	892.0
115.90	991.0
117.00	21848.0
117.90	1143.0
119.00	618.0
120.70	367.0
122.10	1311.0
122.90	2124.0
123.90	1913.0
125.00	779.0
125.90	722.0
127.00	62992.0
128.00	4815.0
129.00	26584.0
130.00	2594.0
130.90	649.0
133.20	764.0
134.00	1132.0
135.10	2136.0
136.00	816.0
137.00	1093.0
137.70	414.0
140.00	372.0
141.00	3230.0
141.90	794.0
142.90	992.0
146.00	650.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

147.10	1674.0
148.00	3946.0
149.10	1268.0
149.90	449.0
151.10	863.0
152.10	598.0
153.10	823.0
154.00	940.0
154.90	2326.0
156.00	3834.0
157.10	427.0
157.90	650.0
159.00	699.0
160.00	1294.0
161.00	2973.0
162.00	749.0
164.90	1332.0
165.90	1112.0
167.00	7747.0
168.10	3161.0
169.00	1118.0
169.90	362.0
170.90	610.0
171.90	905.0
173.10	926.0
174.10	1885.0
175.00	3602.0
176.10	877.0
177.00	1810.0
178.90	7031.0
179.90	3545.0
180.90	1731.0
182.20	638.0
183.00	318.0
184.00	718.0
185.00	3196.0
186.00	19944.0
187.10	5837.0
188.00	926.0
188.90	1928.0
190.80	1078.0
192.00	2344.0
192.90	2793.0
193.90	371.0
196.00	4237.0
197.10	1068.0
197.90	147264.0
199.00	10658.0
199.90	1224.0
201.40	416.0
202.00	416.0
203.00	1175.0
204.10	5270.0
205.00	9665.0
206.00	44592.0
207.10	6367.0
207.90	1895.0
209.00	785.0
209.80	544.0
211.10	1352.0
213.10	311.0
214.90	476.0
215.90	1178.0
216.90	10664.0
218.00	1148.0
221.00	5782.0
221.90	2090.0
222.90	3083.0
224.00	23664.0
225.00	5284.0
226.00	972.0
227.00	11743.0
228.00	1797.0
229.00	2007.0
229.80	484.0
231.00	1366.0
233.90	1152.0
234.90	840.0
235.90	586.0
237.00	960.0
239.10	452.0
240.90	561.0
241.80	963.0
243.00	1571.0
244.00	18736.0
245.10	2312.0
246.10	5668.0
247.10	542.0
247.90	327.0
249.10	999.0
251.00	312.0
251.80	327.0
252.90	715.0
254.20	1331.0
255.00	99944.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

256.00	14016.0
257.00	1390.0
258.00	7444.0
258.80	1056.0
261.00	330.0
264.90	2262.0
273.00	3102.0
274.00	7570.0
275.00	43568.0
276.00	4672.0
277.00	5573.0
278.00	659.0
281.20	608.0
283.00	824.0
285.00	1079.0
286.10	346.0
292.00	428.0
293.10	968.0
295.10	312.0
296.00	13995.0
297.10	1943.0
301.00	440.0
303.10	1377.0
304.20	782.0
313.00	327.0
314.10	762.0
314.90	1696.0
316.10	1018.0
321.10	554.0
323.00	4095.0
324.00	1112.0
327.10	1408.0
332.00	565.0
333.00	533.0
333.90	3219.0
335.10	1293.0
341.00	769.0
341.90	540.0
345.90	1049.0
351.90	1206.0
353.10	1060.0
354.10	1266.0
365.00	7199.0
366.00	1229.0
370.90	375.0
372.20	2822.0
373.10	479.0
383.00	695.0
390.00	310.0
391.00	404.0
402.10	807.0
403.10	1634.0
404.00	560.0
421.10	1223.0
422.10	1063.0
423.10	7910.0
424.10	1970.0
441.10	24312.0
442.10	166848.0
443.10	31944.0
444.20	3450.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

m/z	Abundance
39.10	546.0
40.00	573.0
41.00	947.0
42.20	738.0
43.00	3719.0
44.00	3228.0
45.00	2543.0
53.10	348.0
54.10	440.0
55.00	1009.0
56.20	405.0
57.10	944.0
60.10	1267.0
61.00	371.0
67.90	461.0
68.90	1024.0
73.00	1400.0
75.30	464.0
77.10	500.0
79.90	322.0
81.10	444.0
82.90	507.0
84.00	589.0
91.10	434.0
93.00	306.0
107.10	395.0
109.10	316.0
118.80	417.0
128.00	397.0
135.00	386.0
137.00	317.0
143.20	314.0
146.90	435.0
180.90	338.0
206.90	1556.0
208.80	331.0
221.00	429.0
267.10	454.0
281.10	541.0
283.20	433.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59