

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM120915\
 Data File : BM003441.D
 Acq On : 10 Dec 2015 08:17
 Operator : UM/IZ
 Sample : G4690-03
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MW-06

Quant Time: Dec 10 11:35:23 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM120915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 09 18:52:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	66958	20.00	ng	0.00
21) Naphthalene-d8	10.51	136	275141	20.00	ng	0.00
38) Acenaphthene-d10	14.37	164	150616	20.00	ng	0.00
63) Phenanthrene-d10	17.12	188	315556	20.00	ng	0.00
75) Chrysene-d12	21.32	240	335023	20.00	ng	0.00
86) Perylene-d12	23.57	264	335241	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	225399	59.85	ng	0.00
7) Phenol-d6	6.89	99	174696	33.40	ng	0.00
23) Nitrobenzene-d5	8.88	82	434243	97.86	ng	0.00
41) 2,4,6-Tribromophenol	15.87	330	220348	146.49	ng	0.00
44) 2-Fluorobiphenyl	12.99	172	1000931	90.54	ng	0.00
78) Terphenyl-d14	19.76	244	1151754	81.07	ng	0.00
Target Compounds						
49) Dimethylphthalate	13.83	163	27156	2.19	ng	Qvalue # 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM120915\
Data File : BM003441.D
Acq On : 10 Dec 2015 08:17
Operator : UM/IZ
Sample : G4690-03
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MW-06

Quant Time: Dec 10 11:35:23 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM120915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 09 18:52:47 2015
Response via : Initial Calibration

