

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM120915\
 Data File : BM003455.D
 Acq On : 10 Dec 2015 16:45
 Operator : UM/IZ
 Sample : G4681-03DL 5X
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 20151203-MGP218-BRV108DL

Quant Time: Dec 10 17:46:35 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM120915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 09 18:52:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	68209	20.00	ng	0.00
21) Naphthalene-d8	10.52	136	298609	20.00	ng	0.00
38) Acenaphthene-d10	14.37	164	165216	20.00	ng	0.00
63) Phenanthrene-d10	17.12	188	331526	20.00	ng	0.00
75) Chrysene-d12	21.32	240	321336	20.00	ng	0.00
86) Perylene-d12	23.57	264	316180	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	46843	12.21	ng	0.00
7) Phenol-d6	6.89	99	41171	7.73	ng	0.00
23) Nitrobenzene-d5	8.88	82	88579	18.39	ng	0.00
41) 2,4,6-Tribromophenol	15.86	330	44005	26.67	ng	0.00
44) 2-Fluorobiphenyl	12.99	172	237676	19.60	ng	0.00
78) Terphenyl-d14	19.76	244	250750	18.40	ng	0.00
Target Compounds						
31) Naphthalene	10.59	128	9153933	585.83	ng	Qvalue 97
37) 2-Methylnaphthalene	12.19	142	1651671	154.10	ng	# 97
45) 1,1'-Biphenyl	13.20	154	139861	9.55	ng	99
48) Acenaphthylene	14.10	152	1059171	59.03	ng	100
51) Acenaphthene	14.44	154	55575	5.35	ng	99
54) Dibenzofuran	14.77	168	54108	3.60	ng	# 94
57) Fluorene	15.42	166	200772	16.13	ng	99
70) Phenanthrene	17.16	178	330159	17.75	ng	100
71) Anthracene	17.25	178	71731	3.88	ng	100
72) Carbazole	17.52	167	72100	4.48	ng	99
77) Pyrene	19.55	202	46287	2.05	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM120915\
Data File : BM003455.D
Acq On : 10 Dec 2015 16:45
Operator : UM/IZ
Sample : G4681-03DL 5X
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
20151203-MGP218-BRV108DL

Quant Time: Dec 10 17:46:35 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM120915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 09 18:52:47 2015
Response via : Initial Calibration

