


Data Path : Z:\HPCHEM1\BNA_M\Data\BM120915\
 Data File : BM003419.D
 Acq On : 09 Dec 2015 19:05
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM120915.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 09 18:52:47 2015



AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.2	33992	PASS
68	69	0.00	2	1.6	688	PASS
69	198	0.00	100	31.4	42368	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.7	61634	PASS
197	198	0.00	2	0.3	391	PASS
198	198	100	100	100.0	135005	PASS
199	198	5	9	6.8	9131	PASS
275	198	10	60	24.8	33442	PASS
365	198	1	100	3.0	4113	PASS
441	442	0.01	24	15.3	16557	PASS
442	198	50	100	80.1	108090	PASS
443	442	15	24	19.3	20837	PASS