

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121015\
 Data File : BM003458.D
 Acq On : 10 Dec 2015 18:35
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 11 06:44:13 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM120915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 09 18:52:47 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.467	0.452	3.2	112	0.00
3	Pyridine	1.308	1.296	0.9	114	0.00
4	n-Nitrosodimethylamine	0.419	0.435	-3.8	119	0.00
5 S	2-Fluorophenol	1.125	1.168	-3.8	119	0.00
6	Aniline	2.056	2.071	-0.7	117	0.00
7 S	Phenol-d6	1.562	1.609	-3.0	119	0.00
8	2-Chlorophenol	1.424	1.461	-2.6	119	0.00
9	Benzaldehyde	0.766	0.783	-2.2	112	0.00
10 C	Phenol	1.657	1.694	-2.2	119	0.00
11	bis(2-Chloroethyl)ether	1.341	1.323	1.3	115	0.00
12	1,3-Dichlorobenzene	1.565	1.576	-0.7	117	0.00
13 C	1,4-Dichlorobenzene	1.613	1.623	-0.6	118	0.00
14	1,2-Dichlorobenzene	1.541	1.550	-0.6	118	0.00
15	Benzyl Alcohol	1.081	1.148	-6.2	120	0.00
16	2,2'-oxybis(1-Chloropropane	1.433	1.412	1.5	115	0.00
17	2-Methylphenol	1.159	1.204	-3.9	120	0.00
18	Hexachloroethane	0.554	0.573	-3.4	119	0.00
19 P	n-Nitroso-di-n-propylamine	1.025	1.040	-1.5	116	0.00
20	3+4-Methylphenols	1.603	1.663	-3.7	120	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
22	Acetophenone	0.490	0.498	-1.6	117	0.00
23 S	Nitrobenzene-d5	0.323	0.336	-4.0	117	0.00
24	Nitrobenzene	0.349	0.362	-3.7	119	0.00
25	Isophorone	0.652	0.664	-1.8	114	0.00
26 C	2-Nitrophenol	0.176	0.186	-5.7	122	0.00
27	2,4-Dimethylphenol	0.307	0.312	-1.6	117	0.00
28	bis(2-Chloroethoxy)methane	0.392	0.387	1.3	115	0.00
29 C	2,4-Dichlorophenol	0.293	0.307	-4.8	119	0.00
30	1,2,4-Trichlorobenzene	0.327	0.330	-0.9	117	0.00
31	Naphthalene	1.047	1.044	0.3	117	0.00
32	Benzoic acid	0.196	0.209	-6.6	130	0.02
33	4-Chloroaniline	0.432	0.437	-1.2	118	0.00
34 C	Hexachlorobutadiene	0.194	0.197	-1.5	118	0.00
35	Caprolactam	0.097	0.094	3.1	117	0.01
36 C	4-Chloro-3-methylphenol	0.330	0.334	-1.2	116	0.00
37	2-Methylnaphthalene	0.718	0.702	2.2	115	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
39	1,2,4,5-Tetrachlorobenzene	0.643	0.661	-2.8	115	0.00
40 P	Hexachlorocyclopentadiene	0.359	0.339	5.6	105	0.00
41 S	2,4,6-Tribromophenol	0.200	0.202	-1.0	115	0.00
42 C	2,4,6-Trichlorophenol	0.419	0.452	-7.9	117	0.00
43	2,4,5-Trichlorophenol	0.441	0.465	-5.4	117	0.00
44 S	2-Fluorobiphenyl	1.468	1.468	0.0	113	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121015\
 Data File : BM003458.D
 Acq On : 10 Dec 2015 18:35
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 11 06:44:13 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM120915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 09 18:52:47 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.774	1.783	-0.5	114	0.00
46	2-Chloronaphthalene	1.303	1.330	-2.1	116	0.00
47	2-Nitroaniline	0.323	0.329	-1.9	116	0.00
48	Acenaphthylene	2.172	2.172	0.0	113	0.00
49	Dimethylphthalate	1.648	1.615	2.0	113	0.00
50	2,6-Dinitrotoluene	0.344	0.348	-1.2	112	0.00
51 C	Acenaphthene	1.259	1.242	1.4	114	0.00
52	3-Nitroaniline	0.352	0.361	-2.6	113	0.00
53 P	2,4-Dinitrophenol	0.165	0.130	21.2	92	0.00
54	Dibenzofuran	1.818	1.777	2.3	113	0.00
55 P	4-Nitrophenol	0.291	0.276	5.2	111	0.00
56	2,4-Dinitrotoluene	0.446	0.456	-2.2	114	0.00
57	Fluorene	1.507	1.464	2.9	113	0.00
58	2,3,4,6-Tetrachlorophenol	0.339	0.342	-0.9	114	0.00
59	Diethylphthalate	1.618	1.589	1.8	114	0.00
60	4-Chlorophenyl-phenylether	0.762	0.727	4.6	112	0.00
61	4-Nitroaniline	0.359	0.354	1.4	117	0.00
62	Azobenzene	1.269	1.256	1.0	113	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
64	4,6-Dinitro-2-methylphenol	0.121	0.112	7.4	105	0.00
65 c	n-Nitrosodiphenylamine	0.632	0.650	-2.8	113	0.00
66	4-Bromophenyl-phenylether	0.216	0.224	-3.7	114	0.00
67	Hexachlorobenzene	0.232	0.236	-1.7	114	0.00
68	Atrazine	0.197	0.208	-5.6	113	0.00
69 C	Pentachlorophenol	0.132	0.135	-2.3	114	0.00
70	Phenanthrene	1.122	1.116	0.5	114	0.00
71	Anthracene	1.115	1.130	-1.3	115	0.00
72	Carbazole	0.972	0.988	-1.6	115	0.00
73	Di-n-butylphthalate	1.136	1.247	-9.8	120	0.00
74 C	Fluoranthene	1.157	1.179	-1.9	118	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
76	Benzidine	0.641	0.645	-0.6	115	0.00
77	Pyrene	1.403	1.338	4.6	118	0.00
78 S	Terphenyl-d14	0.848	0.802	5.4	117	0.00
79	Butylbenzylphthalate	0.540	0.580	-7.4	132	0.00
80	Benzo(a)anthracene	1.199	1.201	-0.2	121	0.00
81	3,3'-Dichlorobenzidine	0.382	0.430	-12.6	127	0.00
82	Chrysene	1.154	1.156	-0.2	123	0.00
83	Bis(2-ethylhexyl)phthalate	0.741	0.803	-8.4	130	0.00
84 c	Di-n-octyl phthalate	1.224	1.403	-14.6	137	0.00
85	Indeno(1,2,3-cd)pyrene	1.213	1.356	-11.8	128	0.00
86 I	Perylene-d12	1.000	1.000	0.0	131	0.00
87	Benzo(b)fluoranthene	1.216	1.189	2.2	126	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121015\
Data File : BM003458.D
Acq On : 10 Dec 2015 18:35
Operator : UM/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Dec 11 06:44:13 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM120915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 09 18:52:47 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.187	1.190	-0.3	128	0.00
89 C	Benzo(a)pyrene	1.144	1.161	-1.5	128	0.00
90	Dibenzo(a,h)anthracene	1.117	1.136	-1.7	127	0.00
91	Benzo(a,h,i)perylene	1.128	1.157	-2.6	128	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0