

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121016\
 Data File : BM008242.D
 Acq On : 10 Dec 2016 21:05
 Operator : UM/SJ
 Sample : H5905-10
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA810

Quant Time: Dec 12 11:23:40 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM112116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 12 03:51:49 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	629	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.46	136	2000	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.33	164	1128	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	2024	0.40	ng/ul	-0.02
16) Chrysene-d12	21.27	240	2006	0.40	ng/ul	-0.01
20) Perylene-d12	23.50	264	1904	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.07	152	693	0.24	ng/ul	-0.01
14) Fluoranthene-d10	19.11	212	1684	0.29	ng/ul	-0.02
Target Compounds						
5) 2-Methylnaphthalene	12.14	142	184	0.05	ng/ul	93
12) Phenanthrene	17.11	178	367	0.06	ng/ul#	59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121016\
Data File : BM008242.D
Acq On : 10 Dec 2016 21:05
Operator : UM/SJ
Sample : H5905-10
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA810

Quant Time: Dec 12 11:23:40 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM112116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 12 03:51:49 2016
Response via : Initial Calibration

