

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121016\
 Data File : BM008238.D
 Acq On : 10 Dec 2016 18:41
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.445

Quant Time: Dec 12 03:52:41 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM112116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 12 02:45:24 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
3	Naphthalene	1.119	1.084	3.1	108	0.00
4 SURR	2-Methylnaphthalene-d10	0.566	0.558	1.4	110	0.00
5	2-Methylnaphthalene	0.721	0.683	5.3	106	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
7	Acenaphthylene	1.810	1.812	-0.1	104	0.00
8 C	Acenaphthene	1.377	1.320	4.1	102	0.00
9	Fluorene	1.583	1.406	11.2	94	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
11	Pentachlorophenol	0.140	0.133	5.0	88	0.00
12	Phenanthrene	1.246	1.206	3.2	87	0.00
13	Anthracene	1.170	1.137	2.8	87	0.00
14 SURR	Fluoranthene-d10	1.132	1.045	7.7	83	0.00
15 C	Fluoranthene	1.608	1.430	11.1	82	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	71	0.00
17	Pyrene	1.389	1.632	-17.5	84	0.00
18	Benzo(a)anthracene	1.228	1.337	-8.9	80	0.00
19	Chrysene	1.527	1.449	5.1	70	0.00
20 I	Perylene-d12	1.000	1.000	0.0	80	0.00
21	Benzo(b)fluoranthene	1.596	1.395	12.6	71	0.00
22	Benzo(k)fluoranthene	1.647	1.533	6.9	77	0.00
23 C	Benzo(a)pyrene	1.495	1.491	0.3	81	0.00
24	Indeno(1,2,3-cd)pyrene	1.848	1.793	3.0	77	0.00
25	Dibenzo(a,h)anthracene	1.481	1.418	4.3	77	0.00
26	Benzo(a,h,i)perylene	1.640	1.530	6.7	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0