


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM121020\
 Data File : BM028096.D
 Acq On : 10 Dec 2020 08:06
 Operator : JU/CG
 Sample : DFTPP06
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP006

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM120820.M
 Title : SVOA CALIBRATION
 Last Update : Thu Dec 10 10:24:35 2020



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2516

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.6	117622	PASS
68	69	0.00	2	1.9	2174	PASS
69	198	0.00	100	45.9	113296	PASS
70	69	0.00	2	0.5	566	PASS
127	198	10	80	53.5	132136	PASS
197	198	0.00	2	0.6	1503	PASS
198	198	100	100	100.0	246976	PASS
199	198	5	9	6.8	16805	PASS
275	198	10	60	27.2	67080	PASS
365	198	1	100	3.5	8723	PASS
441	198	0.01	100	12.9	31841	PASS
442	198	50	100	86.9	214666	PASS
443	442	15	24	19.8	42501	PASS

m/z	Abundance
39.90	512.0
43.10	159.0
44.00	1094.0
49.00	165.0
51.00	156.0
73.00	1011.0
77.00	156.0
95.90	201.0
96.20	174.0
132.90	298.0
147.10	324.0
166.90	277.0
191.00	277.0
192.90	265.0
207.00	2770.0
207.90	445.0
209.00	401.0
281.00	771.0
282.10	228.0
283.10	157.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

m/z	Abundance
37.00	439.0
38.00	1189.0
39.10	7487.0
40.00	744.0
41.00	302.0
44.00	1120.0
45.00	248.0
49.10	910.0
50.00	25568.0
51.10	95888.0
52.00	4952.0
53.00	192.0
55.00	489.0
56.00	2954.0
57.00	6498.0
57.90	330.0
61.00	1362.0
62.00	1400.0
63.00	3790.0
64.10	542.0
65.10	1813.0
67.10	235.0
68.10	1791.0
69.00	89792.0
70.00	500.0
73.00	1564.0
74.00	9794.0
75.00	14248.0
76.10	4852.0
77.10	94736.0
78.10	7411.0
79.00	6672.0
80.00	5135.0
81.00	6765.0
82.00	1832.0
83.00	1527.0
83.90	326.0
85.00	1149.0
86.00	1721.0
87.00	790.0
88.00	320.0
89.10	246.0
91.00	1603.0
92.00	2150.0
93.00	10899.0
93.90	709.0
95.00	268.0
96.00	608.0
96.80	176.0
98.00	9122.0
99.00	7532.0
99.90	740.0
101.00	3813.0
102.10	232.0
103.00	1482.0
103.90	2111.0
105.00	2033.0
106.10	998.0
107.00	27464.0
108.00	4480.0
108.90	771.0
110.00	47904.0
111.00	7913.0
112.00	1064.0
112.90	384.0
114.80	237.0
116.00	1502.0
117.00	24872.0
118.00	2304.0
118.80	258.0
120.00	235.0
120.90	160.0
122.00	1905.0
123.00	2997.0
123.90	1306.0
124.90	1155.0
126.20	744.0
127.00	96056.0
128.00	7727.0
129.00	37464.0
130.00	3048.0
131.00	733.0
132.00	357.0
132.90	453.0
134.00	1121.0
135.00	3510.0
136.00	1111.0
137.00	1738.0
138.00	395.0
138.80	176.0
140.00	504.0
141.00	4426.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

142.00	1470.0
142.90	1067.0
144.00	273.0
145.00	312.0
145.90	1025.0
147.00	2493.0
148.00	6036.0
149.00	1601.0
150.00	327.0
150.90	638.0
151.60	597.0
153.00	1429.0
154.00	1190.0
155.00	2142.0
156.00	4067.0
157.10	842.0
158.00	776.0
158.90	569.0
160.00	1377.0
161.10	2184.0
162.00	776.0
162.90	269.0
164.10	280.0
165.00	2012.0
166.00	1535.0
167.00	9587.0
168.00	7307.0
169.00	980.0
169.80	184.0
170.10	186.0
170.90	563.0
171.90	952.0
172.90	1026.0
174.00	2051.0
175.00	3024.0
176.00	783.0
177.00	1917.0
178.00	509.0
179.00	6946.0
180.00	4866.0
181.00	2104.0
182.00	415.0
183.00	248.0
184.00	429.0
185.00	3078.0
186.00	23648.0
187.00	7472.0
188.00	1039.0
189.00	1457.0
190.00	246.0
191.00	963.0
192.00	2256.0
193.10	2358.0
194.00	539.0
194.90	388.0
196.00	4871.0
197.10	740.0
198.00	167360.0
199.00	12351.0
200.00	889.0
201.40	1006.0
203.00	1085.0
204.00	6277.0
205.00	10057.0
206.10	44824.0
207.10	7849.0
208.00	2209.0
209.00	950.0
210.10	880.0
211.00	1815.0
211.90	500.0
214.90	314.0
215.90	881.0
217.00	12159.0
218.00	1348.0
219.00	158.0
221.00	6577.0
221.80	1780.0
222.90	2649.0
224.00	23032.0
225.00	6480.0
226.00	896.0
227.00	10216.0
228.00	1523.0
229.00	2309.0
229.90	338.0
231.10	848.0
232.00	181.0
232.90	323.0
233.90	652.0
234.90	665.0
235.90	541.0
237.00	1386.0
238.00	297.0
238.90	355.0
239.90	430.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

241.10	523.0
242.00	1280.0
243.00	1446.0
244.00	18048.0
245.10	2565.0
246.00	3382.0
247.00	762.0
249.00	969.0
251.00	204.0
251.90	214.0
253.00	506.0
254.10	867.0
255.00	89904.0
256.00	14065.0
257.00	1769.0
258.00	4827.0
259.00	714.0
260.00	244.0
260.90	208.0
264.00	163.0
265.00	1995.0
265.90	284.0
266.90	323.0
270.90	203.0
272.00	362.0
273.00	2590.0
274.00	7164.0
275.00	41720.0
276.00	6089.0
277.00	3306.0
278.00	949.0
278.90	223.0
281.00	605.0
282.00	227.0
283.00	476.0
283.90	286.0
285.00	690.0
293.00	674.0
293.90	348.0
295.00	253.0
296.00	10626.0
297.00	1722.0
300.90	163.0
302.00	231.0
303.00	1395.0
304.00	410.0
314.00	393.0
314.90	1305.0
316.00	537.0
321.00	297.0
323.10	3083.0
324.00	598.0
327.00	775.0
328.00	283.0
331.90	237.0
333.00	405.0
334.00	2182.0
334.90	716.0
341.00	611.0
342.10	200.0
345.90	638.0
346.90	254.0
352.10	706.0
352.90	708.0
354.00	1077.0
355.00	289.0
365.00	4282.0
366.00	746.0
369.90	157.0
371.10	258.0
372.00	1670.0
372.90	343.0
382.90	605.0
390.00	272.0
392.00	167.0
401.90	594.0
403.00	1028.0
404.00	439.0
404.90	181.0
421.00	865.0
421.90	1010.0
423.00	5594.0
424.00	3262.0
424.90	492.0
441.00	15396.0
442.00	103840.0
443.00	22008.0
444.10	2374.0
445.00	165.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

m/z	Abundance
36.00	197.0
37.00	611.0
38.00	1864.0
39.10	10706.0
40.00	1024.0
41.00	410.0
44.00	1156.0
45.00	452.0
49.10	1147.0
50.00	38888.0
51.10	143936.0
52.10	7365.0
53.00	272.0
55.00	809.0
56.00	4599.0
57.00	9641.0
58.00	501.0
59.00	190.0
61.00	1834.0
62.00	2235.0
63.10	5144.0
64.00	804.0
65.10	2472.0
66.00	165.0
67.00	249.0
68.10	2638.0
69.00	140288.0
70.00	618.0
73.10	1767.0
74.10	14564.0
75.00	22256.0
76.10	7056.0
77.10	150208.0
78.10	10250.0
79.00	10155.0
80.00	8061.0
81.00	11270.0
82.00	2674.0
83.00	2345.0
84.00	336.0
85.00	1819.0
86.00	3341.0
87.00	1369.0
87.90	430.0
89.00	289.0
91.00	2465.0
92.00	3071.0
93.00	18096.0
94.00	1092.0
94.90	311.0
96.00	906.0
98.00	14521.0
99.00	12331.0
100.00	863.0
101.00	5865.0
102.00	365.0
103.00	1869.0
104.00	4018.0
105.00	3794.0
106.10	1316.0
107.00	45272.0
108.00	7258.0
109.10	1200.0
110.00	80232.0
111.00	12519.0
112.00	1497.0
113.00	462.0
114.30	150.0
114.90	222.0
116.00	2429.0
117.00	38152.0
118.00	3052.0
118.90	320.0
120.00	615.0
121.00	325.0
122.00	3091.0
123.00	4878.0
124.00	2258.0
125.00	1903.0
127.00	160512.0
128.00	11915.0
129.00	60664.0
130.00	5096.0
131.00	1223.0
132.10	562.0
133.00	406.0
134.00	2005.0
135.00	4804.0
136.00	1732.0
137.00	2926.0
138.00	811.0
139.00	275.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

140.00	820.0
141.00	8358.0
142.00	2504.0
143.00	1878.0
144.00	418.0
145.00	473.0
146.00	1359.0
147.00	4629.0
148.00	9852.0
149.00	2184.0
150.00	666.0
151.00	867.0
151.20	860.0
151.80	538.0
153.00	2421.0
154.00	1607.0
155.00	4163.0
156.10	6021.0
157.00	1257.0
158.00	1648.0
159.00	1122.0
160.00	2405.0
161.00	3451.0
161.90	1064.0
162.90	412.0
164.00	524.0
165.00	3019.0
166.00	2412.0
167.00	16251.0
168.00	9991.0
169.00	1323.0
170.00	686.0
170.80	704.0
172.00	1243.0
173.00	1973.0
174.00	2604.0
175.00	5654.0
176.00	1746.0
176.90	2860.0
178.00	1014.0
179.00	11647.0
180.00	8632.0
181.00	3579.0
182.00	559.0
182.90	280.0
184.00	909.0
185.00	5303.0
186.00	40736.0
187.00	11523.0
188.00	1344.0
189.00	2701.0
190.00	516.0
191.00	1333.0
192.00	3633.0
193.00	4438.0
194.00	901.0
195.00	437.0
196.00	10042.0
196.70	1712.0
197.10	2008.0
198.00	298240.0
199.00	19384.0
199.90	1679.0
201.40	1365.0
203.00	2303.0
204.00	10137.0
205.10	17424.0
206.10	78024.0
207.10	12443.0
208.00	3000.0
209.00	1363.0
210.00	1534.0
211.10	3127.0
212.90	216.0
215.00	1035.0
216.00	1849.0
217.00	22504.0
218.00	3021.0
219.00	211.0
220.20	567.0
221.00	13468.0
223.00	4889.0
224.10	42896.0
225.10	11707.0
226.10	1701.0
227.00	17896.0
228.00	2521.0
229.00	4013.0
230.00	386.0
231.00	1418.0
232.00	201.0
233.00	274.0
234.00	1192.0
235.00	1276.0
236.00	928.0
237.00	2482.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

238.00	314.0
238.90	739.0
240.00	724.0
240.90	1080.0
242.10	2161.0
243.10	2221.0
244.10	31880.0
245.10	4774.0
246.00	6446.0
247.00	1290.0
248.00	337.0
249.00	1097.0
251.00	330.0
251.90	235.0
253.00	656.0
254.10	1403.0
255.00	171072.0
256.00	25216.0
257.00	2663.0
258.00	9986.0
259.00	1453.0
259.90	255.0
260.90	315.0
264.10	223.0
265.00	4085.0
266.00	643.0
269.90	275.0
270.90	413.0
272.00	409.0
273.00	4869.0
274.00	13332.0
275.00	78400.0
276.00	10471.0
277.00	6542.0
278.00	1494.0
278.80	263.0
281.00	886.0
282.00	436.0
283.00	858.0
284.10	545.0
285.00	1163.0
286.00	154.0
289.00	423.0
290.10	209.0
292.10	325.0
293.00	1684.0
294.00	551.0
295.00	702.0
296.00	21200.0
297.00	2755.0
297.90	351.0
301.00	251.0
302.00	518.0
303.00	2408.0
304.00	703.0
304.90	252.0
307.90	203.0
308.90	224.0
310.10	245.0
314.00	970.0
315.00	2508.0
316.00	1432.0
320.90	986.0
322.10	392.0
323.00	6589.0
324.00	1298.0
326.00	222.0
326.90	1600.0
328.00	765.0
331.90	496.0
333.00	826.0
334.00	4218.0
335.00	1211.0
340.00	167.0
341.00	980.0
342.10	251.0
346.00	1592.0
347.00	361.0
350.90	172.0
352.00	1766.0
353.00	1775.0
354.10	2016.0
355.10	556.0
365.00	10447.0
365.90	1270.0
370.00	279.0
371.10	438.0
372.00	3661.0
373.00	940.0
383.00	984.0
384.00	456.0
384.90	165.0
390.10	471.0
391.00	396.0
392.10	184.0
401.00	220.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

402.00	1524.0
403.00	1957.0
404.00	831.0
405.00	291.0
414.80	186.0
421.00	1944.0
422.00	2031.0
423.10	13760.0
424.00	4914.0
425.00	874.0
441.00	34896.0
442.00	239168.0
443.00	47376.0
444.00	4571.0
445.00	296.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

m/z	Abundance
35.90	201.0
37.00	524.0
38.10	1575.0
39.00	8778.0
40.00	913.0
41.10	423.0
44.00	1420.0
45.00	392.0
49.00	1017.0
50.10	30064.0
51.10	113512.0
52.10	5372.0
53.10	252.0
55.00	731.0
56.00	3505.0
57.00	7554.0
58.00	382.0
61.00	1307.0
62.00	1798.0
63.00	4718.0
64.10	666.0
65.00	1949.0
66.90	177.0
68.10	2095.0
69.00	109808.0
70.00	580.0
73.10	1708.0
74.10	10807.0
75.00	17768.0
76.10	5983.0
77.10	120760.0
78.10	8709.0
79.00	8129.0
80.00	6654.0
81.00	9313.0
82.00	2193.0
83.00	1905.0
83.80	327.0
85.00	1450.0
86.00	2837.0
87.00	1117.0
87.90	507.0
88.90	274.0
91.00	1910.0
92.00	2303.0
93.00	14248.0
93.90	1062.0
94.90	294.0
96.00	657.0
96.90	336.0
98.00	10983.0
99.00	9063.0
100.00	974.0
101.00	5158.0
101.90	237.0
102.90	1584.0
104.00	3279.0
105.00	3165.0
106.10	809.0
107.00	36144.0
108.00	6108.0
109.10	1087.0
110.00	66976.0
111.00	10354.0
112.00	1460.0
113.10	512.0
114.90	156.0
116.00	2560.0
117.00	33840.0
118.00	2297.0
118.90	242.0
120.00	472.0
121.00	314.0
122.00	2478.0
123.00	4177.0
124.00	2027.0
125.00	1790.0
126.20	802.0
127.00	139840.0
128.00	9948.0
129.00	49800.0
130.00	4501.0
131.00	1017.0
132.00	534.0
133.00	419.0
134.00	1665.0
135.00	5010.0
135.90	1780.0
137.10	2755.0
137.90	479.0
138.90	300.0
140.00	783.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

141.00	6944.0
142.00	2385.0
143.00	1473.0
144.10	344.0
144.90	576.0
146.00	1585.0
147.00	3898.0
148.00	8697.0
149.00	1817.0
150.00	466.0
151.00	1031.0
153.00	2466.0
154.00	1715.0
155.10	3587.0
156.00	5198.0
157.00	1207.0
158.00	1104.0
158.90	1064.0
160.00	2123.0
161.00	2819.0
162.00	888.0
163.00	362.0
164.00	483.0
165.00	2702.0
166.00	1982.0
167.00	13349.0
168.00	7971.0
169.00	1299.0
170.00	480.0
171.00	759.0
171.90	1059.0
173.00	1726.0
174.00	2766.0
175.10	5200.0
176.00	1386.0
177.00	2596.0
178.00	983.0
179.00	10850.0
180.00	7964.0
181.00	3350.0
181.90	532.0
183.00	249.0
183.90	867.0
185.00	4629.0
186.00	37064.0
187.00	10925.0
188.00	1398.0
188.90	2543.0
189.90	582.0
191.00	1262.0
192.00	3492.0
193.00	3844.0
193.90	630.0
195.00	420.0
196.00	8349.0
196.70	1685.0
197.10	1762.0
198.00	275328.0
199.00	18680.0
200.00	1496.0
201.50	1331.0
203.00	1849.0
204.00	10324.0
205.00	16176.0
206.10	74096.0
207.10	12128.0
208.00	3041.0
209.00	1224.0
210.20	1303.0
211.00	2627.0
212.00	579.0
212.90	250.0
214.00	190.0
214.90	944.0
216.00	1629.0
217.00	19888.0
218.00	2450.0
218.90	284.0
221.00	11992.0
223.00	4075.0
224.00	39680.0
225.10	9898.0
226.00	1383.0
227.00	17104.0
228.00	2474.0
229.00	3812.0
229.90	500.0
231.00	1616.0
231.90	256.0
233.00	348.0
234.00	1047.0
235.00	1266.0
235.90	814.0
237.00	2230.0
238.00	341.0
239.00	657.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

239.90	544.0
241.00	1036.0
242.00	2160.0
243.00	2188.0
244.10	33272.0
245.10	3967.0
246.00	6073.0
247.00	1633.0
248.00	191.0
249.00	1457.0
250.10	278.0
250.90	415.0
252.00	433.0
252.90	552.0
254.10	1731.0
255.00	174912.0
256.00	24744.0
257.00	2219.0
258.00	9983.0
259.00	1532.0
259.90	307.0
260.90	287.0
264.10	240.0
265.00	4347.0
266.00	511.0
267.00	296.0
269.90	157.0
271.00	387.0
272.10	404.0
273.00	5372.0
274.00	13715.0
275.00	81120.0
276.10	10623.0
277.00	6494.0
278.10	1375.0
281.00	755.0
282.10	249.0
283.10	777.0
284.10	634.0
285.10	1133.0
286.00	338.0
289.00	294.0
290.00	391.0
292.00	368.0
293.00	1567.0
294.00	458.0
295.00	472.0
296.00	23584.0
297.00	2945.0
300.90	354.0
302.10	359.0
303.00	2434.0
304.00	838.0
307.80	170.0
308.90	152.0
309.90	459.0
312.90	253.0
314.00	1156.0
315.00	2293.0
316.00	1375.0
317.10	153.0
321.00	736.0
322.00	332.0
323.10	7049.0
324.00	1293.0
327.00	1575.0
328.00	794.0
331.90	584.0
333.00	592.0
334.00	4601.0
335.00	1229.0
341.00	703.0
341.80	286.0
346.00	1762.0
347.00	333.0
352.00	2315.0
353.00	1819.0
354.10	2448.0
355.00	602.0
359.00	178.0
365.00	11442.0
366.00	1463.0
369.80	247.0
371.10	666.0
372.10	4231.0
373.00	1294.0
383.00	1186.0
384.10	394.0
385.00	268.0
389.90	618.0
391.00	494.0
392.10	252.0
401.00	302.0
402.00	1566.0
403.00	2770.0
404.00	1069.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006

404.90	266.0
415.00	186.0
421.00	2392.0
422.10	2245.0
423.00	16371.0
424.00	5224.0
424.90	743.0
438.90	196.0
440.10	291.0
441.00	45232.0
442.00	300992.0
443.00	58120.0
444.00	4982.0
445.00	467.0

Instrument :
BNA_M
ClientSampleId :
DFTPP006