

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121022\
 Data File : BM037927.D
 Acq On : 10 Dec 2022 09:43
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020071

Quant Time: Dec 11 21:24:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 09 21:34:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.081	152	213360	20.000	ng/ul	0.00
20) Naphthalene-d8	10.898	136	871698	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.710	164	505321	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.451	188	908313	20.000	ng/ul	0.00
79) Chrysene-d12	21.621	240	750690	20.000	ng/ul	0.00
88) Perylene-d12	24.097	264	811246	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.422	96	30387	6.467	ng/uL	0.00
4) Pyridine-d5	3.857	84	241311	17.311	ng/ul	0.00
7) Phenol-d5	7.234	99	314222	18.124	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.404	67	189046	17.907	ng/ul	0.00
11) 2-Chlorophenol-d4	7.604	132	273500	19.230	ng/ul	0.00
15) 4-Methylphenol-d8	8.792	113	249763	18.455	ng/ul	0.00
21) Nitrobenzene-d5	9.251	128	134253	19.452	ng/ul	0.00
24) 2-Nitrophenol-d4	9.981	143	142781	19.741	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.516	165	272288	19.820	ng/ul	0.00
31) 4-Chloroaniline-d4	11.033	131	339551	17.906	ng/ul	0.00
46) Dimethylphthalate-d6	14.116	166	701740	19.479	ng/ul	0.00
49) Acenaphthylene-d8	14.404	160	849874	19.302	ng/ul	0.00
54) 4-Nitrophenol-d4	14.898	143	91327	14.875	ng/ul	0.00
60) Fluorene-d10	15.698	176	603204	18.793	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.816	200	76332	15.718	ng/ul	0.00
73) Anthracene-d10	17.545	188	829660	19.571	ng/ul	0.00
81) Pyrene-d10	19.827	212	853243	18.921	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.939	264	804917	19.865	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.457	88	33768	6.850	ng/uL	96
5) Pyridine	3.875	79	246894	17.624	ng/ul	96
6) Benzaldehyde	7.210	77	118851	18.328	ng/ul	98
8) Phenol	7.263	94	315610	18.150	ng/ul	95
10) Bis(2-Chloroethyl)ether	7.492	93	258340	18.093	ng/ul	99
12) 2-Chlorophenol	7.640	128	275464	18.938	ng/ul	99
13) 2-Methylphenol	8.522	108	243747	18.317	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.604	45	457237	17.734	ng/ul	99
16) Acetophenone	8.910	105	399160	18.753	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.892	70	190400	18.989	ng/ul	97
18) 4-Methylphenol	8.857	108	265534	18.446	ng/ul	98
19) Hexachloroethane	9.169	117	106935	18.389	ng/ul	95
22) Nitrobenzene	9.292	77	286929	19.321	ng/ul	95
23) Isophorone	9.822	82	515562	18.946	ng/ul	99
25) 2-Nitrophenol	10.010	139	149582	19.727	ng/ul	95
26) 2,4-Dimethylphenol	10.063	107	288618	19.683	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.304	93	342866	18.954	ng/ul	98
29) 2,4-Dichlorophenol	10.545	162	262042	19.441	ng/ul	98
30) Naphthalene	10.951	128	905152	19.242	ng/ul	99
32) 4-Chloroaniline	11.063	127	333883	17.656	ng/ul	97
33) Hexachlorobutadiene	11.228	225	181280	20.445	ng/ul	97
34) Caprolactam	11.845	113	61310	17.733	ng/ul	92
35) 4-Chloro-3-methylphenol	12.175	107	235391	19.186	ng/ul	99
36) 2-Methylnaphthalene	12.551	142	591555	19.388	ng/ul	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121022\
 Data File : BM037927.D
 Acq On : 10 Dec 2022 09:43
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020071


Quant Time: Dec 11 21:24:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 09 21:34:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.769	142	602309	19.306	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.910	216	325283	19.701	ng/ul	100
40) Hexachlorocyclopentadiene	12.886	237	94905	10.048	ng/ul	98
41) 2,4,6-Trichlorophenol	13.151	196	199576	19.844	ng/ul	100
42) 2,4,5-Trichlorophenol	13.222	196	206771	19.714	ng/ul	99
43) 1,1'-Biphenyl	13.545	154	786252	19.357	ng/ul	99
44) 2-Chloronaphthalene	13.592	162	607233	19.379	ng/ul	99
45) 2-Nitroaniline	13.798	65	141807	18.782	ng/ul	99
47) Dimethylphthalate	14.163	163	684707	19.264	ng/ul	99
48) 2,6-Dinitrotoluene	14.286	165	134260	19.446	ng/ul	94
50) Acenaphthylene	14.433	152	931280	19.235	ng/ul	99
51) 3-Nitroaniline	14.616	138	108841	15.801	ng/ul	97
52) Acenaphthene	14.774	153	645699	19.282	ng/ul	99
53) 2,4-Dinitrophenol	14.821	184	40304	11.897	ng/ul	97
55) 4-Nitrophenol	14.916	109	66394	16.398	ng/ul	99
56) Dibenzofuran	15.104	168	864049	18.899	ng/ul	99
57) 2,4-Dinitrotoluene	15.068	165	175704	18.792	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.327	232	166755	18.900	ng/ul	97
59) Diethylphthalate	15.516	149	675818	19.086	ng/ul	98
61) Fluorene	15.751	166	702662	18.846	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.739	204	359148	19.710	ng/ul	96
63) 4-Nitroaniline	15.774	138	97422	15.598	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.827	198	73489	15.632	ng/ul	99
67) N-Nitrosodiphenylamine	15.957	169	551848	20.141	ng/ul	99
68) 4-Bromophenyl-phenylether	16.633	248	206265	20.716	ng/ul	97
69) Hexachlorobenzene	16.751	284	249375	20.982	ng/ul	96
70) Atrazine	16.904	200	176633	18.865	ng/ul	100
71) Pentachlorophenol	17.098	266	114151	17.275	ng/ul	98
72) Phenanthrene	17.492	178	955803	19.269	ng/ul	100
74) Anthracene	17.580	178	973745	19.317	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.516	216	322611	21.970	ng/uL	99
76) Pentachlorobenzene	15.027	250	311049	21.384	ng/uL	99
77) Carbazole	17.851	167	798043	18.273	ng/ul	100
78) Di-n-butylphthalate	18.404	149	998680	19.238	ng/ul	100
80) Fluoranthene	19.498	202	988214	18.569	ng/ul	98
82) Pyrene	19.856	202	1044986	18.867	ng/ul	100
83) Butylbenzylphthalate	20.739	149	365638	17.753	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.533	252	276381	18.606	ng/ul	97
85) Benzo(a)anthracene	21.603	228	1003044	19.847	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.509	149	498839	16.649	ng/ul	98
87) Chrysene	21.656	228	922116	19.446	ng/ul	100
89) Di-n-octyl phthalate	22.462	149	863244	16.607	ng/ul	100
90) Benzo(b)fluoranthene	23.339	252	987998	19.644	ng/ul	100
91) Benzo(k)fluoranthene	23.391	252	1009071	20.702	ng/ul	99
93) Benzo(a)pyrene	23.991	252	891272	20.152	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.697	276	1091681	19.166	ng/ul	97
95) Dibenzo(a,h)anthracene	26.715	278	931616	19.401	ng/ul	96
96) Benzo(g,h,i)perylene	27.497	276	806559	17.735	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 11 21:24:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 09 21:34:57 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.850 ng/uL
RT: 3.457 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion: 88 Resp: 33768
Ion Ratio Lower Upper
88 100
43 35.6 32.8 49.2
58 79.4 64.6 97.0



#5
Pyridine
Concen: 17.624 ng/uL
RT: 3.875 min Scan# 83
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion: 79 Resp: 246894
Ion Ratio Lower Upper
79 100
52 101.6 78.1 117.1
51 49.6 37.2 55.8









#6
Benzaldehyde
Concen: 18.328 ng/ul
RT: 7.210 min Scan# 650
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion: 77 Resp: 118851
Ion Ratio Lower Upper
77 100
105 102.8 82.5 123.7
106 98.1 81.8 122.6



#8
Phenol
Concen: 18.150 ng/ul
RT: 7.263 min Scan# 659
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion: 94 Resp: 315610
Ion Ratio Lower Upper
94 100
65 29.6 22.2 33.4
66 37.2 27.3 40.9









#10
Bis(2-Chloroethyl)ether
Concen: 18.093 ng/ul
RT: 7.492 min Scan# 698
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion: 93 Resp: 258340
Ion Ratio Lower Upper
93 100
63 82.3 66.3 99.5
95 31.9 25.7 38.5



#12
2-Chlorophenol
Concen: 18.938 ng/ul
RT: 7.640 min Scan# 723
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion: 128 Resp: 275464
Ion Ratio Lower Upper
128 100
64 43.5 35.2 52.8
130 33.2 25.9 38.9









#13
2-Methylphenol
Concen: 18.317 ng/ul
RT: 8.522 min Scan# 873
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:108 Resp: 243747
Ion Ratio Lower Upper
108 100
107 92.1 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.734 ng/ul
RT: 8.604 min Scan# 887
Delta R.T. -0.012 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion: 45 Resp: 457237
Ion Ratio Lower Upper
45 100
77 10.8 9.0 13.4
79 7.5 6.2 9.4









#16
Acetophenone
Concen: 18.753 ng/ul
RT: 8.910 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:105 Resp: 399160
Ion Ratio Lower Upper
105 100
77 74.1 59.7 89.5
51 32.8 27.2 40.8



#17
N-Nitroso-di-n-propylamine
Concen: 18.989 ng/ul
RT: 8.892 min Scan# 936
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion: 70 Resp: 190400
Ion Ratio Lower Upper
70 100
42 75.0 62.3 93.5
101 10.0 8.0 12.0
130 20.8 17.6 26.4









#18
4-Methylphenol
Concen: 18.446 ng/ul
RT: 8.857 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:108 Resp: 265534
Ion Ratio Lower Upper
108 100
107 107.8 88.1 132.1



#19
Hexachloroethane
Concen: 18.389 ng/ul
RT: 9.169 min Scan# 983
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:117 Resp: 106935
Ion Ratio Lower Upper
117 100
201 112.5 87.3 130.9
199 68.9 50.9 76.3









#22
Nitrobenzene
Concen: 19.321 ng/ul
RT: 9.292 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion: 77 Resp: 286929
Ion Ratio Lower Upper
77 100
123 51.0 37.8 56.6
65 13.7 11.1 16.7



#23
Isophorone
Concen: 18.946 ng/ul
RT: 9.822 min Scan# 1094
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion: 82 Resp: 515562
Ion Ratio Lower Upper
82 100
95 7.0 5.2 7.8
138 18.8 14.9 22.3









#25
2-Nitrophenol
Concen: 19.727 ng/ul
RT: 10.010 min Scan# 1126
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:139 Resp: 149582
Ion Ratio Lower Upper
139 100
65 47.7 36.2 54.4
109 28.3 19.7 29.5



#26
2,4-Dimethylphenol
Concen: 19.683 ng/ul
RT: 10.063 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:107 Resp: 288618
Ion Ratio Lower Upper
107 100
121 51.7 42.0 63.0
122 89.6 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.954 ng/ul
 RT: 10.304 min Scan# 1176
 Delta R.T. -0.000 min
 Lab File: BM037927.D
 Acq: 10 Dec 2022 09:43




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020071


Tgt Ion: 93 Resp: 342866
 Ion Ratio Lower Upper
 93 100
 95 32.9 25.5 38.3
 123 12.7 9.4 14.2



#29
 2,4-Dichlorophenol
 Concen: 19.441 ng/ul
 RT: 10.545 min Scan# 1217
 Delta R.T. -0.000 min
 Lab File: BM037927.D
 Acq: 10 Dec 2022 09:43

Tgt Ion: 162 Resp: 262042
 Ion Ratio Lower Upper
 162 100
 164 66.6 51.4 77.2
 98 33.5 26.1 39.1









#30
Naphthalene
Concen: 19.242 ng/ul
RT: 10.951 min Scan# 1286
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:128 Resp: 905152
Ion Ratio Lower Upper
128 100
129 10.5 9.0 13.4
127 12.7 10.4 15.6



#32
4-Chloroaniline
Concen: 17.656 ng/ul
RT: 11.063 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:127 Resp: 333883
Ion Ratio Lower Upper
127 100
129 33.1 25.3 37.9









#33
Hexachlorobutadiene
Concen: 20.445 ng/ul
RT: 11.228 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:225 Resp: 181280
Ion Ratio Lower Upper
225 100
223 63.0 50.8 76.2
227 66.5 50.5 75.7



#34
Caprolactam
Concen: 17.733 ng/ul
RT: 11.845 min Scan# 1438
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:113 Resp: 61310
Ion Ratio Lower Upper
113 100
55 201.3 172.2 258.2
56 140.8 120.7 181.1









#35
4-Chloro-3-methylphenol
Concen: 19.186 ng/ul
RT: 12.175 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:107 Resp: 235391
Ion Ratio Lower Upper
107 100
144 28.6 23.4 35.0
142 87.2 71.0 106.4



#36
2-Methylnaphthalene
Concen: 19.388 ng/ul
RT: 12.551 min Scan# 1558
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:142 Resp: 591555
Ion Ratio Lower Upper
142 100
141 90.2 69.4 104.0





#37

1-Methylnaphthalene

Concen: 19.306 ng/ul

RT: 12.769 min Scan# 1595

Delta R.T. -0.000 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Instrument :

BNA_M

ClientSampleId :

SSTD020071





Tgt Ion:142 Resp: 602309

Ion Ratio Lower Upper

142 100

141 89.5 72.4 108.6

116 3.2 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.701 ng/ul

RT: 12.910 min Scan# 1619

Delta R.T. -0.006 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Tgt Ion:216 Resp: 325283




Ion Ratio Lower Upper


216 100

214 80.3 64.2 96.4

179 19.7 15.4 23.2

108 15.1 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 10.048 ng/ul
RT: 12.886 min Scan# 1615
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:237 Resp: 94905
Ion Ratio Lower Upper
237 100
235 60.9 50.4 75.6
272 13.3 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 19.844 ng/ul
RT: 13.151 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:196 Resp: 199576
Ion Ratio Lower Upper
196 100
198 93.9 75.0 112.6
200 31.1 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 19.714 ng/ul
RT: 13.222 min Scan# 1672
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:196 Resp: 206771
Ion Ratio Lower Upper
196 100
198 96.7 76.5 114.7
200 31.1 24.9 37.3



#43
1,1'-Biphenyl
Concen: 19.357 ng/ul
RT: 13.545 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:154 Resp: 786252
Ion Ratio Lower Upper
154 100
153 38.9 31.2 46.8
76 10.6 9.3 13.9









#44
2-Chloronaphthalene
Concen: 19.379 ng/ul
RT: 13.592 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:162 Resp: 607233
Ion Ratio Lower Upper
162 100
164 33.3 27.3 40.9
127 37.2 29.4 44.0



#45
2-Nitroaniline
Concen: 18.782 ng/ul
RT: 13.798 min Scan# 1770
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion: 65 Resp: 141807
Ion Ratio Lower Upper
65 100
92 65.8 53.5 80.3
138 115.3 91.3 136.9









#47
Dimethylphthalate
Concen: 19.264 ng/ul
RT: 14.163 min Scan# 1832
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:163 Resp: 684707
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.5 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 19.446 ng/ul
RT: 14.286 min Scan# 1853
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:165 Resp: 134260
Ion Ratio Lower Upper
165 100
89 57.6 41.1 61.7
121 21.9 17.4 26.2









#50
Acenaphthylene
Concen: 19.235 ng/ul
RT: 14.433 min Scan# 1878
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:152 Resp: 931280
Ion Ratio Lower Upper
152 100
151 19.7 15.3 22.9
153 13.1 10.2 15.4



#51
3-Nitroaniline
Concen: 15.801 ng/ul
RT: 14.616 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:138 Resp: 108841
Ion Ratio Lower Upper
138 100
108 8.1 7.9 11.9
92 101.5 83.4 125.0









#52
Acenaphthene
Concen: 19.282 ng/ul
RT: 14.774 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:153 Resp: 645699
Ion Ratio Lower Upper
153 100
152 47.1 38.2 57.4
154 88.1 71.4 107.0



#53
2,4-Dinitrophenol
Concen: 11.897 ng/ul
RT: 14.821 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:184 Resp: 40304
Ion Ratio Lower Upper
184 100
63 58.7 46.2 69.4
154 60.6 45.5 68.3









#55
4-Nitrophenol
Concen: 16.398 ng/ul
RT: 14.916 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:109 Resp: 66394
Ion Ratio Lower Upper
109 100
139 146.8 119.2 178.8
65 135.4 108.7 163.1



#56
Dibenzofuran
Concen: 18.899 ng/ul
RT: 15.104 min Scan# 1992
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:168 Resp: 864049
Ion Ratio Lower Upper
168 100
139 35.6 28.1 42.1









#57
2,4-Dinitrotoluene
Concen: 18.792 ng/ul
RT: 15.068 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:165 Resp: 175704
Ion Ratio Lower Upper
165 100
63 41.6 34.9 52.3
182 2.5 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.900 ng/ul
RT: 15.327 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:232 Resp: 166755
Ion Ratio Lower Upper
232 100
131 44.0 32.8 49.2
130 2.1 2.0 3.0
166 29.5 23.2 34.8









#59
Diethylphthalate
Concen: 19.086 ng/ul
RT: 15.516 min Scan# 2062
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:149 Resp: 675818
Ion Ratio Lower Upper
149 100
177 22.1 18.5 27.7
150 12.1 9.5 14.3



#61
Fluorene
Concen: 18.846 ng/ul
RT: 15.751 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:166 Resp: 702662
Ion Ratio Lower Upper
166 100
165 95.4 77.2 115.8
167 13.3 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 19.710 ng/ul
RT: 15.739 min Scan# 2100
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:204 Resp: 359148
Ion Ratio Lower Upper
204 100
206 33.2 25.5 38.3
141 57.0 42.8 64.2



#63
4-Nitroaniline
Concen: 15.598 ng/ul
RT: 15.774 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:138 Resp: 97422
Ion Ratio Lower Upper
138 100
92 45.3 36.1 54.1
108 72.2 51.6 77.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 15.632 ng/ul
 RT: 15.827 min Scan# 2115
 Delta R.T. -0.006 min
 Lab File: BM037927.D
 Acq: 10 Dec 2022 09:43




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020071


Tgt Ion:198 Resp: 73489
 Ion Ratio Lower Upper
 198 100
 182 4.0 3.8 5.8
 77 24.6 19.3 28.9



#67
 N-Nitrosodiphenylamine
 Concen: 20.141 ng/ul
 RT: 15.957 min Scan# 2137
 Delta R.T. -0.000 min
 Lab File: BM037927.D
 Acq: 10 Dec 2022 09:43

Tgt Ion:169 Resp: 551848
 Ion Ratio Lower Upper
 169 100
 168 66.8 53.8 80.8
 167 37.3 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 20.716 ng/ul
RT: 16.633 min Scan# 211
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:248 Resp: 206265
Ion Ratio Lower Upper
248 100
250 101.3 78.7 118.1
141 68.1 52.7 79.1



#69
Hexachlorobenzene
Concen: 20.982 ng/ul
RT: 16.751 min Scan# 2272
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:284 Resp: 249375
Ion Ratio Lower Upper
284 100
142 31.7 23.1 34.7
249 30.4 23.3 34.9









#70
Atrazine
Concen: 18.865 ng/ul
RT: 16.904 min Scan# 2298
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:200 Resp: 176633
Ion Ratio Lower Upper
200 100
173 25.6 20.5 30.7
215 49.7 40.1 60.1



#71
Pentachlorophenol
Concen: 17.275 ng/ul
RT: 17.098 min Scan# 2331
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:266 Resp: 114151
Ion Ratio Lower Upper
266 100
264 65.6 50.1 75.1
268 63.2 50.6 75.8









#72
Phenanthrene
Concen: 19.269 ng/ul
RT: 17.492 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:178 Resp: 955803
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.2
176 19.4 15.4 23.0



#74
Anthracene
Concen: 19.317 ng/ul
RT: 17.580 min Scan# 2413
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:178 Resp: 973745
Ion Ratio Lower Upper
178 100
179 15.2 11.8 17.8
176 18.4 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.970 ng/uL
RT: 13.516 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:216 Resp: 322611
Ion Ratio Lower Upper
216 100
214 79.2 62.6 93.8
218 48.0 38.3 57.5



#76
Pentachlorobenzene
Concen: 21.384 ng/uL
RT: 15.027 min Scan# 1979
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:250 Resp: 311049
Ion Ratio Lower Upper
250 100
248 63.4 51.0 76.4
252 62.7 51.8 77.6









#77
 Carbazole
 Concen: 18.273 ng/ul
 RT: 17.851 min Scan# 2459
 Delta R.T. -0.006 min
 Lab File: BM037927.D
 Acq: 10 Dec 2022 09:43




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020071


Tgt Ion:167 Resp: 798043
 Ion Ratio Lower Upper
 167 100
 166 21.2 16.7 25.1
 139 12.2 9.8 14.6



#78
 Di-n-butylphthalate
 Concen: 19.238 ng/ul
 RT: 18.404 min Scan# 2553
 Delta R.T. -0.000 min
 Lab File: BM037927.D
 Acq: 10 Dec 2022 09:43

Tgt Ion:149 Resp: 998680
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 4.4 3.4 5.2









#80
Fluoranthene
Concen: 18.569 ng/ul
RT: 19.498 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43




Instrument :
BNA_M
ClientSampleId :
SSTD020071


Tgt Ion:202 Resp: 988214
Ion Ratio Lower Upper
202 100
101 9.8 8.6 13.0
100 7.7 6.6 10.0



#82
Pyrene
Concen: 18.867 ng/ul
RT: 19.856 min Scan# 2800
Delta R.T. -0.006 min
Lab File: BM037927.D
Acq: 10 Dec 2022 09:43

Tgt Ion:202 Resp: 1044986
Ion Ratio Lower Upper
202 100
101 12.3 10.0 15.0
100 9.7 7.9 11.9





#83

Butylbenzylphthalate

Concen: 17.753 ng/ul

RT: 20.739 min Scan# 2950

Delta R.T. -0.006 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Instrument :

BNA_M

ClientSampleId :

SSTD020071





Tgt Ion:149 Resp: 365638

Ion Ratio Lower Upper

149 100

91 67.4 54.5 81.7

206 21.2 15.8 23.8



#84

3,3'-Dichlorobenzidine

Concen: 18.606 ng/ul

RT: 21.533 min Scan# 3085

Delta R.T. -0.000 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43




Tgt Ion:252 Resp: 276381


Ion Ratio Lower Upper

252 100

254 64.3 53.2 79.8

126 9.4 7.8 11.8





#85

Benzo(a)anthracene

Concen: 19.847 ng/ul

RT: 21.603 min Scan# 3097

Delta R.T. -0.000 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Instrument :

BNA_M

ClientSampleId :

SSTD020071





Tgt Ion:228 Resp: 1003044

Ion Ratio Lower Upper

228 100

229 19.3 15.5 23.3

226 26.8 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.649 ng/ul

RT: 21.509 min Scan# 3081

Delta R.T. -0.000 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43




Tgt Ion:149 Resp: 498839


Ion Ratio Lower Upper

149 100

167 27.8 21.5 32.3

279 4.3 3.2 4.8





#87

Chrysene

Concen: 19.446 ng/ul

RT: 21.656 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Instrument :

BNA_M

ClientSampleId :

SSTD020071





Tgt Ion:228 Resp: 922116

Ion Ratio Lower Upper

228 100

226 30.2 24.3 36.5

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 16.607 ng/ul




RT: 22.462 min Scan# 3243


Delta R.T. -0.000 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Tgt Ion:149 Resp: 863244





#90

Benzo(b)fluoranthene

Concen: 19.644 ng/ul

RT: 23.339 min Scan# 3392

Delta R.T. -0.006 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Instrument :

BNA_M

ClientSampleId :

SSTD020071





Tgt Ion:252 Resp: 987998

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 10.0 8.0 12.0



#91

Benzo(k)fluoranthene

Concen: 20.702 ng/ul

RT: 23.391 min Scan# 3401

Delta R.T. -0.000 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43




Tgt Ion:252 Resp: 1009071


Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.2

125 9.1 8.1 12.1





#93

Benzo(a)pyrene

Concen: 20.152 ng/ul

RT: 23.991 min Scan# 3503

Delta R.T. -0.000 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Instrument :

BNA_M

ClientSampleId :

SSTD020071





Tgt Ion:252 Resp: 891272

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0

125 9.7 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.166 ng/ul

RT: 26.697 min Scan# 3963

Delta R.T. -0.006 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43




Tgt Ion:276 Resp: 1091681


Ion Ratio Lower Upper

276 100

138 21.4 19.3 28.9

277 25.5 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 19.401 ng/ul

RT: 26.715 min Scan# 3966

Delta R.T. -0.000 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Instrument :

BNA_M

ClientSampleId :

SSTD020071





Tgt Ion:278 Resp: 931616

Ion Ratio Lower Upper

278 100

139 13.7 12.4 18.6

279 22.9 19.8 29.6



#96

Benzo(g,h,i)perylene

Concen: 17.735 ng/ul

RT: 27.497 min Scan# 4099

Delta R.T. -0.006 min

Lab File: BM037927.D

Acq: 10 Dec 2022 09:43

Tgt Ion:276 Resp: 806559

Ion Ratio Lower Upper

276 100

138 18.7 17.1 25.7

277 22.7 18.2 27.4

