

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121024\
 Data File : BM048932.D
 Acq On : 10 Dec 2024 20:57
 Operator : RC/JU
 Sample : PB165472BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK472

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/11/2024
 Supervised By :mohammad ahmed 12/12/2024

Quant Time: Dec 10 22:50:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM120924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 09 17:52:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.581	152	3032	0.400	ng/ul	0.00
4) Naphthalene-d8	10.347	136	7926	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.214	164	4101	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.963	188	9147m	0.400	ng/ul	0.00
17) Chrysene-d12	21.158	240	6449m	0.400	ng/ul	0.00
23) Perylene-d12	23.320	264	5708m	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.159	96	25299	7.139	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.958	152	3479	0.344	ng/ul	0.01
18) Fluoranthene-d10	18.996	212	7090	0.368	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121024\
Data File : BM048932.D
Acq On : 10 Dec 2024 20:57
Operator : RC/JU
Sample : PB165472BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK472

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/11/2024
Supervised By :mohammad ahmed 12/12/2024

Quant Time: Dec 10 22:50:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM120924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 09 17:52:26 2024
Response via : Initial Calibration

