

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM121117\
 Data File : BM013304.D
 Acq On : 11 Dec 2017 14:41
 Operator : SJ/JU
 Sample : I6737-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 JHPK5

Manual Integrations
 APPROVED

Sohil
 12/12/2017 3:56:45 PM

Quant Time: Dec 12 01:19:16 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 12 01:04:42 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	130303	20.00	ng/ul	0.00
18) Naphthalene-d8	10.47	136	557391	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.33	164	358699	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.08	188	919168	20.00	ng/ul	0.00
75) Chrysene-d12	21.27	240	1143213	20.00	ng/ul	0.00
83) Perylene-d12	23.49	264	1078757	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.23	96	17409m	6.62	ng/uL	0.00
5) Phenol-d5	6.87	99	58004	5.14	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	181304	28.20	ng/ul	0.00
9) 2-Chlorophenol-d4	7.23	132	196506	22.30	ng/ul	0.00
13) 4-Methylphenol-d8	8.39	113	123634	12.84	ng/ul	0.00
19) Nitrobenzene-d5	8.85	128	134696	30.47	ng/ul	0.00
22) 2-Nitrophenol-d4	9.56	143	145945	30.85	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.09	165	248260	25.47	ng/ul	0.00
29) 4-Chloroaniline-d4	10.61	131	1197	0.13	ng/ul	0.00
43) Dimethylphthalate-d6	13.75	166	948827	29.73	ng/ul	0.00
46) Acenaphthylene-d8	14.02	160	1183955	30.06	ng/ul	0.00
51) 4-Nitrophenol-d4	14.55	143	20208m	3.64	ng/ul	0.00
57) Fluorene-d10	15.33	176	829936	30.23	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.45	200	132483	22.91	ng/ul	0.00
70) Anthracene-d10	17.17	188	1374128	29.85	ng/ul	0.00
76) Pyrene-d10	19.47	212	1658161	29.11	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.36	264	1673026	30.45	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA M\DATA\BM121117\
Data File : BM013304.D
Acq On : 11 Dec 2017 14:41
Operator : SJ/JU
Sample : I6737-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
JHPK5

Manual Integrations
APPROVED

Sohil
12/12/2017 3:56:45 PM

Quant Time: Dec 12 01:19:16 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 12 01:04:42 2017
Response via : Initial Calibration

