


Data Path : Z:\HPCHEM1\BNA_M\Data\BM121117\
 Data File : BM013297.D
 Acq On : 11 Dec 2017 09:38
 Operator : SJ/JU
 Sample : DFTPP60
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP60

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM113017.M
 Title : SVOA CALIBRATION
 Last Update : Tue Dec 12 11:20:58 2017



AutoFind: Scans 2454, 2455, 2456; Background Corrected with Scan 2449

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.3	78242	PASS
68	69	0.00	2	1.7	1471	PASS
69	198	0.00	100	41.7	87362	PASS
70	69	0.00	2	0.9	778	PASS
127	198	10	80	46.0	96410	PASS
197	198	0.00	2	0.8	1648	PASS
198	198	100	100	100.0	209621	PASS
199	198	5	9	7.0	14726	PASS
275	198	10	60	27.6	57880	PASS
365	198	1	100	4.1	8649	PASS
441	443	0.01	100	78.5	26281	PASS
442	198	50	100	82.4	172818	PASS
443	442	15	24	19.4	33480	PASS

m/z	Abundance
38.10	959.0
39.10	4834.0
39.90	1023.0
40.90	757.0
44.00	1327.0
44.90	371.0
49.10	837.0
50.10	22040.0
51.10	74912.0
52.00	3676.0
52.90	509.0
55.10	724.0
56.00	2242.0
57.00	4694.0
61.00	960.0
62.10	944.0
63.10	2809.0
63.90	489.0
65.10	1463.0
67.20	581.0
68.00	1365.0
69.00	80752.0
69.90	691.0
72.90	1566.0
74.00	9920.0
75.00	14495.0
76.00	4366.0
77.10	83400.0
78.00	6741.0
79.00	5703.0
80.00	5814.0
81.00	7181.0
82.00	1686.0
82.90	1602.0
85.00	1295.0
86.00	2326.0
87.00	1115.0
87.90	323.0
91.00	1414.0
92.10	2095.0
93.00	13564.0
93.90	447.0
95.00	467.0
96.00	584.0
98.00	9965.0
99.00	6940.0
100.10	693.0
101.00	3752.0
103.00	1453.0
104.00	2167.0
105.00	2628.0
106.00	870.0
107.00	28184.0
108.00	4061.0
108.80	635.0
109.20	406.0
110.00	43168.0
111.00	6869.0
112.00	1000.0
112.80	549.0
115.00	447.0
116.00	1783.0
117.00	26024.0
118.00	1655.0
119.00	491.0
119.80	722.0
122.00	2170.0
123.00	2842.0
123.90	1954.0
124.80	918.0
126.20	845.0
127.00	82880.0
128.10	6850.0
129.00	39536.0
129.90	3691.0
131.00	1100.0
132.00	372.0
133.10	332.0
134.00	1438.0
135.00	3449.0
136.00	1303.0
137.00	1583.0
137.70	435.0
140.10	948.0
141.00	6129.0
142.00	1651.0
143.00	990.0
144.90	356.0
146.10	1260.0
147.00	2730.0
148.00	7077.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

149.10	929.0
151.10	973.0
151.80	303.0
153.00	1346.0
154.00	1026.0
155.00	2658.0
156.00	2902.0
156.90	822.0
157.30	547.0
157.80	681.0
159.10	824.0
160.00	1303.0
160.90	2292.0
161.90	539.0
164.90	2091.0
166.00	1242.0
167.00	9648.0
168.00	5523.0
169.10	1091.0
169.90	495.0
170.90	464.0
171.80	700.0
173.10	1249.0
174.10	1517.0
174.90	2972.0
175.90	1496.0
177.00	2079.0
178.10	849.0
179.00	8396.0
179.90	4888.0
181.00	2058.0
182.00	1368.0
183.90	584.0
185.00	3256.0
186.00	26360.0
187.00	7689.0
188.10	727.0
189.00	3296.0
191.20	719.0
192.00	2334.0
193.10	1585.0
193.80	482.0
195.20	349.0
195.90	3198.0
197.00	1344.0
198.00	173952.0
199.00	12262.0
199.90	1016.0
201.20	844.0
201.50	721.0
203.10	1634.0
204.00	7660.0
205.00	11061.0
206.00	45960.0
207.00	8874.0
207.90	2188.0
209.10	1132.0
210.20	879.0
211.00	1850.0
215.00	670.0
215.80	778.0
217.00	14578.0
218.00	1391.0
221.00	5817.0
222.00	1520.0
223.00	2576.0
224.00	28000.0
225.10	6767.0
226.10	1119.0
227.00	12563.0
228.10	1931.0
228.90	2496.0
229.80	455.0
230.90	977.0
232.00	322.0
234.00	537.0
234.90	970.0
236.00	505.0
237.00	907.0
239.00	772.0
240.10	329.0
240.90	871.0
241.90	1386.0
243.10	962.0
244.10	20000.0
245.00	2672.0
246.00	5049.0
247.10	1166.0
249.10	444.0
252.90	704.0
253.90	1021.0
255.00	102448.0
256.00	17240.0
256.90	1075.0
258.00	6454.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

258.90	1624.0
265.00	2465.0
266.00	756.0
272.90	2616.0
274.00	5829.0
275.00	41272.0
276.10	5633.0
276.90	4107.0
277.90	1224.0
280.80	598.0
283.10	759.0
284.10	418.0
285.00	536.0
293.00	1474.0
294.90	331.0
296.00	15609.0
296.90	2709.0
303.10	1343.0
303.90	527.0
313.90	740.0
315.00	1016.0
315.90	469.0
321.00	430.0
322.00	339.0
323.00	3875.0
324.10	919.0
326.80	940.0
327.90	796.0
332.10	610.0
332.90	406.0
334.10	1797.0
335.00	956.0
340.90	553.0
345.90	807.0
352.00	595.0
353.00	1042.0
353.90	565.0
355.10	468.0
365.00	5049.0
366.00	901.0
371.00	493.0
372.10	2244.0
383.00	475.0
401.90	841.0
403.00	1116.0
404.20	325.0
404.80	345.0
421.20	594.0
422.10	1055.0
423.10	4817.0
424.10	1440.0
441.00	14644.0
442.10	103736.0
443.10	24040.0
444.10	2056.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

m/z	Abundance
37.10	482.0
38.00	1987.0
39.00	8008.0
39.90	484.0
43.90	1613.0
44.70	328.0
49.30	934.0
50.10	27696.0
51.10	91832.0
52.00	4881.0
52.90	356.0
55.00	673.0
56.00	2746.0
57.00	7559.0
57.90	311.0
61.00	1240.0
62.00	1209.0
63.10	4037.0
64.00	742.0
65.00	1993.0
67.30	354.0
68.00	1975.0
69.00	105000.0
69.90	1132.0
73.10	1727.0
74.00	11912.0
75.00	16912.0
76.10	5906.0
77.00	107848.0
78.10	6443.0
79.00	8153.0
80.00	6622.0
81.00	8800.0
82.10	2647.0
83.00	2100.0
84.00	614.0
84.80	1915.0
86.00	2763.0
87.10	891.0
88.10	585.0
89.00	386.0
91.10	1507.0
92.10	1899.0
93.00	15577.0
94.10	1224.0
95.10	313.0
96.10	944.0
96.90	810.0
98.00	12495.0
99.00	11002.0
100.00	392.0
100.90	4486.0
102.90	1191.0
104.10	2837.0
105.00	2866.0
106.10	637.0
107.00	32728.0
108.00	5442.0
108.90	1323.0
110.00	57208.0
111.00	8614.0
112.10	1311.0
116.10	1972.0
117.00	36600.0
117.90	3448.0
118.80	626.0
121.90	2724.0
122.90	3316.0
124.00	1944.0
124.90	1376.0
126.10	390.0
127.00	116008.0
128.00	8246.0
129.00	52032.0
130.00	4474.0
131.00	774.0
131.80	324.0
133.90	1822.0
135.00	3762.0
136.10	1730.0
137.10	1929.0
137.90	311.0
140.00	398.0
141.00	7711.0
142.10	1120.0
142.80	1641.0
146.00	1301.0
147.00	4495.0
148.00	8528.0
148.90	1579.0
150.00	633.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

150.90	858.0
151.30	1130.0
152.00	450.0
153.00	3244.0
154.10	2028.0
155.00	3366.0
156.00	5551.0
157.20	1056.0
157.80	1342.0
159.10	1304.0
159.90	2276.0
161.00	3628.0
162.10	595.0
163.00	407.0
165.00	3307.0
166.10	2546.0
167.00	14248.0
168.00	5634.0
169.10	1220.0
170.80	675.0
172.10	1004.0
172.90	1734.0
174.10	3338.0
175.00	6017.0
176.00	1422.0
177.00	3421.0
178.00	954.0
179.00	11230.0
180.00	8390.0
180.90	3411.0
182.00	471.0
183.90	933.0
185.00	4106.0
186.00	37776.0
187.00	9811.0
188.10	670.0
189.00	2575.0
190.00	507.0
191.10	1340.0
192.00	4310.0
193.00	3991.0
194.00	576.0
194.90	761.0
196.00	6188.0
197.10	1981.0
198.00	247680.0
199.00	17536.0
200.00	1260.0
201.30	875.0
201.90	636.0
202.90	1520.0
204.00	10664.0
205.10	16456.0
206.00	73352.0
207.00	11187.0
208.00	3043.0
209.10	1400.0
210.20	1257.0
210.90	2990.0
211.20	3125.0
211.80	933.0
214.90	980.0
216.00	1463.0
216.90	20080.0
218.00	2564.0
221.00	8610.0
222.10	1764.0
223.00	3685.0
224.00	37912.0
225.00	9897.0
226.00	1148.0
227.00	19392.0
227.90	2499.0
228.90	3294.0
230.00	537.0
230.90	1109.0
231.90	572.0
234.00	667.0
235.10	1001.0
236.10	858.0
236.90	2053.0
239.00	771.0
239.80	583.0
241.10	631.0
242.00	2704.0
243.00	2262.0
244.00	30520.0
245.10	4039.0
246.00	6712.0
247.10	1444.0
249.00	1408.0
252.00	515.0
253.10	605.0
254.10	850.0
255.00	153792.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

256.00	26536.0
257.00	1487.0
258.00	10899.0
259.00	1615.0
259.90	614.0
265.10	3595.0
265.90	586.0
271.10	507.0
272.10	632.0
273.00	4073.0
274.00	14995.0
275.00	65720.0
276.00	9436.0
277.00	7406.0
278.00	849.0
279.00	686.0
280.90	988.0
282.10	428.0
283.00	860.0
284.10	561.0
285.00	970.0
293.00	1228.0
294.90	714.0
296.00	22952.0
296.90	3306.0
301.00	355.0
303.00	2412.0
304.00	717.0
309.90	549.0
313.20	366.0
314.00	1408.0
315.10	1886.0
316.10	1123.0
320.90	802.0
322.00	403.0
323.10	6479.0
324.10	1397.0
327.10	1562.0
328.10	573.0
331.90	419.0
332.90	812.0
334.00	4200.0
334.90	621.0
340.00	306.0
341.00	1419.0
341.90	447.0
345.90	883.0
352.10	1685.0
353.00	1051.0
354.10	2913.0
355.20	625.0
365.00	12242.0
365.80	962.0
370.00	444.0
371.00	641.0
372.10	2590.0
373.00	1328.0
382.80	901.0
383.90	476.0
390.10	753.0
391.10	388.0
392.00	517.0
401.10	410.0
402.10	1332.0
403.00	2062.0
403.90	858.0
421.10	1384.0
422.00	1298.0
423.10	10783.0
424.10	3203.0
424.90	514.0
441.10	30912.0
442.10	202048.0
443.10	35784.0
444.00	3477.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

m/z	Abundance
36.80	390.0
38.10	1059.0
39.00	4600.0
40.00	388.0
41.00	532.0
44.00	1239.0
45.00	340.0
48.80	401.0
49.10	454.0
50.00	18544.0
51.00	67984.0
52.10	3726.0
55.00	540.0
55.90	1580.0
56.10	1575.0
57.00	4170.0
61.00	1791.0
61.90	1418.0
63.00	3596.0
64.10	536.0
65.00	1876.0
67.90	1075.0
69.00	76336.0
70.00	511.0
73.00	1346.0
74.00	9254.0
75.10	11145.0
75.90	5147.0
77.00	78608.0
78.00	5163.0
79.10	6595.0
80.00	4232.0
81.00	6424.0
82.10	1513.0
83.00	1185.0
85.00	1373.0
85.90	1491.0
87.10	958.0
88.00	634.0
88.90	328.0
91.00	1538.0
91.90	1683.0
93.00	11399.0
94.00	1291.0
95.00	344.0
96.00	1052.0
98.00	8902.0
99.00	6764.0
100.00	588.0
101.00	2926.0
102.00	322.0
103.10	1661.0
104.10	2118.0
105.10	2408.0
106.00	1045.0
107.00	29592.0
108.00	4662.0
109.20	1235.0
110.00	47032.0
111.00	5820.0
111.90	769.0
113.00	341.0
116.10	1470.0
117.00	30448.0
118.00	2392.0
118.90	561.0
122.00	1679.0
123.00	3242.0
124.00	1278.0
125.10	1027.0
127.00	90344.0
128.00	6415.0
129.00	38544.0
130.00	3145.0
130.80	483.0
132.90	311.0
134.00	1286.0
134.90	2946.0
136.00	1618.0
137.00	1272.0
138.00	473.0
140.10	432.0
141.00	6241.0
142.00	1887.0
143.00	1396.0
145.90	883.0
147.00	3660.0
148.10	5334.0
149.00	1456.0
150.10	585.0
151.00	576.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

152.30	339.0
153.00	1383.0
154.10	942.0
155.00	4213.0
156.00	5047.0
157.10	627.0
157.90	777.0
159.10	1074.0
160.00	2410.0
161.00	3066.0
162.00	568.0
165.00	2541.0
166.00	1917.0
167.00	11521.0
168.00	6444.0
168.90	827.0
170.00	556.0
170.90	745.0
172.10	976.0
173.10	1636.0
174.00	2908.0
175.10	4033.0
176.10	1277.0
177.00	2719.0
178.00	562.0
179.00	8815.0
179.90	7284.0
181.10	2547.0
182.90	305.0
183.90	664.0
184.80	3281.0
186.00	34040.0
187.00	8609.0
187.80	813.0
189.00	1377.0
191.00	1549.0
192.00	1708.0
193.00	3737.0
193.90	471.0
196.10	5917.0
197.10	1621.0
198.00	207232.0
199.00	14382.0
199.80	1239.0
201.20	773.0
203.00	1458.0
204.00	7419.0
205.00	13688.0
206.00	59144.0
207.00	10735.0
208.00	2881.0
208.90	1007.0
209.80	1129.0
210.10	1030.0
210.80	1294.0
211.10	1737.0
213.20	395.0
214.90	934.0
216.20	1334.0
217.00	17128.0
218.00	2238.0
220.90	6850.0
222.00	1779.0
223.00	3998.0
224.00	34224.0
225.10	9101.0
226.00	1023.0
227.00	18584.0
228.10	2165.0
229.00	4279.0
230.90	1656.0
231.90	329.0
232.90	670.0
234.20	540.0
235.10	776.0
235.90	1030.0
237.10	1257.0
237.90	364.0
239.00	434.0
239.80	378.0
240.90	587.0
242.10	1481.0
243.10	1892.0
244.00	26000.0
245.00	3887.0
246.00	4975.0
247.00	1195.0
248.00	387.0
248.90	804.0
250.10	358.0
251.90	466.0
253.10	852.0
254.10	619.0
255.00	143552.0
256.00	19600.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

257.00	2004.0
258.00	10074.0
258.80	1328.0
264.90	3502.0
266.00	604.0
271.00	379.0
271.90	607.0
273.10	5066.0
274.00	11662.0
275.00	66648.0
276.00	9391.0
277.00	6331.0
277.90	1109.0
281.00	671.0
281.80	401.0
283.10	937.0
285.10	1038.0
293.00	1591.0
295.10	614.0
296.00	23000.0
297.00	2269.0
298.00	507.0
301.90	394.0
303.00	2370.0
303.90	1007.0
308.90	334.0
310.00	611.0
314.10	1884.0
315.00	2857.0
316.10	1613.0
321.10	1332.0
322.00	562.0
323.10	7379.0
324.10	1048.0
326.90	1577.0
327.90	942.0
332.10	598.0
333.00	751.0
334.00	4611.0
335.00	1117.0
340.90	679.0
345.90	1240.0
352.00	1820.0
352.90	1271.0
354.10	2415.0
355.20	980.0
365.00	8656.0
366.00	1140.0
371.00	507.0
372.00	3411.0
373.20	1016.0
383.00	789.0
401.00	437.0
402.00	1588.0
403.10	1775.0
404.20	806.0
421.10	1888.0
421.90	1174.0
423.10	10499.0
424.10	2209.0
425.00	342.0
441.10	33288.0
442.10	212672.0
443.10	40616.0
444.10	4365.0
445.00	380.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

Scan 2449 (17.492 min): BM013297.D
DFTPP60

m/z	Abundance
39.90	593.0
44.00	2249.0
73.10	1812.0
95.30	396.0
133.00	351.0
147.00	692.0
178.70	366.0
190.80	358.0
207.10	2033.0
208.10	585.0
281.10	813.0
282.20	356.0
283.00	433.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60