

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121120\
 Data File : BM028105.D
 Acq On : 11 Dec 2020 18:34
 Operator : JU/CG
 Sample : SST0.406
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST0.4061

Quant Time: Dec 12 02:54:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM121120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 12 02:45:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.22	152	2669	0.40	ng/ul	0.00
4) Naphthalene-d8	11.05	136	10270	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.84	164	5863	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.56	188	12270	0.40	ng/ul	0.00
17) Chrysene-d12	21.67	240	10757	0.40	ng/ul	0.00
23) Perylene-d12	24.04	264	9038	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.43	96	1290	0.47	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.62	152	6654	0.39	ng/ul	0.00
18) Fluoranthene-d10	19.56	212	13213	0.40	ng/ul	0.00
Target Compounds					Ovalue	
2) 1,4-Dioxane	3.47	88	1478	0.502	ng/ul	100
5) Naphthalene	11.10	128	12170	0.418	ng/ul	100
7) 2-Methylnaphthalene	12.70	142	8106	0.404	ng/ul	100
8) 1-Methylnaphthalene	12.91	142	7738	0.408	ng/ul	100
10) Acenaphthylene	14.57	152	9824	0.375	ng/ul	100
11) Acenaphthene	14.91	153	8737	0.439	ng/ul	100
12) Fluorene	15.88	166	10013	0.426	ng/ul	100
14) Pentachlorophenol	17.21	266	1239	0.306	ng/ul	100
15) Phenanthrene	17.60	178	15904	0.444	ng/ul	100
16) Anthracene	17.69	178	13894	0.395	ng/ul	100
19) Fluoranthene	19.59	202	17425	0.414	ng/ul	100
20) Pyrene	19.95	202	17332	0.413	ng/ul	100
21) Benzo(a)anthracene	21.66	228	14966	0.380	ng/ul	100
22) Chrysene	21.71	228	16251	0.402	ng/ul	100
24) Benzo(b)fluoranthene	23.32	252	15500	0.427	ng/ul	100
25) Benzo(k)fluoranthene	23.37	252	14976	0.414	ng/ul	100
26) Benzo(a)pyrene	23.94	252	12771	0.388	ng/ul	100
27) Indeno(1,2,3-cd)pyrene	26.48	276	14612	0.378	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.48	278	12107	0.376	ng/ul	100
29) Benzo(a,h,i)perylene	27.22	276	12654	0.389	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121120\
Data File : BM028105.D
Acq On : 11 Dec 2020 18:34
Operator : JU/CG
Sample : SSTD0.406
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4061

Quant Time: Dec 12 02:54:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM121120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 12 02:45:43 2020
Response via : Initial Calibration

