

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121123\
 Data File : BM043227.D
 Acq On : 11 Dec 2023 15:48
 Operator : MA/JU
 Sample : SSTD3.283
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2034

Quant Time: Dec 11 16:23:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 11 14:52:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.009	152	10601	0.400	ng/ul	0.00
4) Naphthalene-d8	10.812	136	31955	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.629	164	17937	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.367	188	36301	0.400	ng/ul	0.00
17) Chrysene-d12	21.536	240	24381	0.400	ng/ul	0.00
23) Perylene-d12	23.950	264	22053	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.402	96	40247	2.560	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.440	88	43620	2.611	ng/ul#	82
14) Pentachlorophenol	17.012	266	32006	3.372	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121123\
Data File : BM043227.D
Acq On : 11 Dec 2023 15:48
Operator : MA/JU
Sample : SST3.283
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2034

Quant Time: Dec 11 16:23:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 11 14:52:13 2023
Response via : Initial Calibration

