

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121124\
 Data File : BM048957.D
 Acq On : 11 Dec 2024 17:02
 Operator : RC/JU
 Sample : SST01061
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST010049

Quant Time: Dec 12 00:27:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 12 00:21:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	150902	20.000	ng/u1	0.00
20) Naphthalene-d8	10.339	136	586563	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.210	164	335890	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.957	188	619417	20.000	ng/u1	0.00
79) Chrysene-d12	21.180	240	546908	20.000	ng/u1	0.00
88) Perylene-d12	23.997	264	549859	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.152	96	17402	4.597	ng/uL	0.00
4) Pyridine-d5	3.546	84	111423	10.681	ng/u1	0.00
7) Phenol-d5	6.757	99	111115	9.673	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.916	67	82158	11.907	ng/u1	0.00
11) 2-Chlorophenol-d4	7.110	132	88193	9.268	ng/u1	0.00
15) 4-Methylphenol-d8	8.281	113	82071	9.086	ng/u1	0.00
21) Nitrobenzene-d5	8.716	128	36738	8.641	ng/u1	0.00
24) 2-Nitrophenol-d4	9.433	143	31906	7.056	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.963	165	76014	9.049	ng/u1	0.00
31) 4-Chloroaniline-d4	10.480	131	110765	9.283	ng/u1	0.00
46) Dimethylphthalate-d6	13.627	166	222000	10.354	ng/u1	0.00
49) Acenaphthylene-d8	13.898	160	261734	10.354	ng/u1	0.00
54) 4-Nitrophenol-d4	14.416	143	30840	7.744	ng/u1	0.00
60) Fluorene-d10	15.204	176	187129	9.998	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.321	200	19242	6.175	ng/u1	0.00
73) Anthracene-d10	17.051	188	271688	10.365	ng/u1	0.00
81) Pyrene-d10	19.356	212	279661	10.189	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.803	264	246860	9.609	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.187	88	18462	4.444	ng/uL#	84
5) Pyridine	3.563	79	118701	11.085	ng/u1	96
6) Benzaldehyde	6.722	77	80107	14.112	ng/u1	99
8) Phenol	6.781	94	123406	10.239	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.010	93	102571	10.829	ng/u1	98
12) 2-Chlorophenol	7.145	128	92014	9.334	ng/u1	96
13) 2-Methylphenol	8.016	108	81388	9.122	ng/u1	92
14) 2,2'-oxybis(1-Chloropr...	8.092	45	156792	11.257	ng/u1	99
16) Acetophenone	8.386	105	144473	10.261	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.375	70	70677	10.682	ng/u1	97
18) 4-Methylphenol	8.339	108	91293	9.345	ng/u1	93
19) Hexachloroethane	8.639	117	41211	9.892	ng/u1	91
22) Nitrobenzene	8.757	77	105149	10.908	ng/u1	99
23) Isophorone	9.286	82	186012	10.432	ng/u1	100
25) 2-Nitrophenol	9.463	139	39305	7.976	ng/u1	95
26) 2,4-Dimethylphenol	9.533	107	86940	9.445	ng/u1	96
27) Bis(2-Chloroethoxy)met...	9.769	93	121941	10.956	ng/u1	99
29) 2,4-Dichlorophenol	9.992	162	77094	9.047	ng/u1	96
30) Naphthalene	10.386	128	297887	10.034	ng/u1	98
32) 4-Chloroaniline	10.498	127	108779	9.669	ng/u1	99
33) Hexachlorobutadiene	10.680	225	52753	9.148	ng/u1	100
34) Caprolactam	11.275	113	19645	7.315	ng/u1	79
35) 4-Chloro-3-methylphenol	11.633	107	70630	8.566	ng/u1	97
36) 2-Methylnaphthalene	12.010	142	185370	9.758	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121124\
 Data File : BM048957.D
 Acq On : 11 Dec 2024 17:02
 Operator : RC/JU
 Sample : SSTD01061
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD010049


Quant Time: Dec 12 00:27:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 12 00:21:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.227	142	186546	9.569	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.380	216	93898	9.755	ng/ul	98
40) Hexachlorocyclopentadiene	12.369	237	49440	10.049	ng/ul	95
41) 2,4,6-Trichlorophenol	12.627	196	51211	8.423	ng/ul	95
42) 2,4,5-Trichlorophenol	12.698	196	55622	8.591	ng/ul	95
43) 1,1'-Biphenyl	13.039	154	242739	10.569	ng/ul	99
44) 2-Chloronaphthalene	13.074	162	193845	10.696	ng/ul	99
45) 2-Nitroaniline	13.280	65	41394	9.095	ng/ul	97
47) Dimethylphthalate	13.674	163	221739	10.295	ng/ul	99
48) 2,6-Dinitrotoluene	13.786	165	32485	7.912	ng/ul	97
50) Acenaphthylene	13.927	152	291871	10.474	ng/ul	100
51) 3-Nitroaniline	14.116	138	35928	8.058	ng/ul#	98
52) Acenaphthene	14.274	153	207620	10.584	ng/ul	99
53) 2,4-Dinitrophenol	14.321	184	11524	5.220	ng/ul	94
55) 4-Nitrophenol	14.433	109	21756	7.556	ng/ul	96
56) Dibenzofuran	14.610	168	283306	10.469	ng/ul	98
57) 2,4-Dinitrotoluene	14.574	165	49110	8.228	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.839	232	39487	7.332	ng/ul	94
59) Diethylphthalate	15.051	149	208518	9.826	ng/ul	97
61) Fluorene	15.262	166	224732	10.489	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.268	204	105903	10.033	ng/ul	98
63) 4-Nitroaniline	15.280	138	32007	7.290	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.339	198	21674	6.392	ng/ul#	89
67) N-Nitrosodiphenylamine	15.480	169	173307	10.616	ng/ul	99
68) 4-Bromophenyl-phenylether	16.162	248	53002	9.073	ng/ul	93
69) Hexachlorobenzene	16.262	284	66787	9.473	ng/ul	99
70) Atrazine	16.439	200	53795	9.054	ng/ul	97
71) Pentachlorophenol	16.609	266	30400	6.784	ng/ul	96
72) Phenanthrene	16.998	178	331928	10.805	ng/ul	99
74) Anthracene	17.086	178	331278	10.745	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.998	216	91898	10.400	ng/ul	98
76) Pentachlorobenzene	14.527	250	88471	10.276	ng/ul	97
77) Carbazole	17.362	167	293860	10.582	ng/ul	99
78) Di-n-butylphthalate	17.951	149	291665	8.933	ng/ul	99
80) Fluoranthene	19.021	202	333930	10.486	ng/ul	99
82) Pyrene	19.386	202	375536	10.875	ng/ul	98
83) Butylbenzylphthalate	20.315	149	99203	7.479	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.103	252	69326	6.977	ng/ul	97
85) Benzo(a)anthracene	21.162	228	330897	9.598	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.127	149	146899	7.592	ng/ul	99
87) Chrysene	21.221	228	316385	9.550	ng/ul	100
89) Di-n-octyl phthalate	22.203	149	203609	6.569	ng/ul	100
90) Benzo(b)fluoranthene	23.115	252	304527	10.061	ng/ul	100
91) Benzo(k)fluoranthene	23.174	252	299410	9.921	ng/ul	98
93) Benzo(a)pyrene	23.862	252	278629	9.791	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.115	276	352752	9.538	ng/ul	98
95) Dibenzo(a,h)anthracene	27.179	278	285633	9.456	ng/ul	97
96) Benzo(g,h,i)perylene	28.079	276	277801	9.580	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 12 00:27:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 12 00:21:01 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.444 ng/uL
RT: 3.187 min Scan# 34
Delta R.T. 0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion: 88 Resp: 18462
Ion Ratio Lower Upper
88 100
43 23.3 24.7 37.1#
58 60.9 60.0 90.0



#5
Pyridine
Concen: 11.085 ng/uL
RT: 3.563 min Scan# 98
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion: 79 Resp: 118701
Ion Ratio Lower Upper
79 100
52 73.4 56.2 84.4
51 37.5 28.6 42.8









#6
Benzaldehyde
Concen: 14.112 ng/ul
RT: 6.722 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion: 77 Resp: 80107
Ion Ratio Lower Upper
77 100
105 91.1 73.3 109.9
106 91.0 71.9 107.9



#8
Phenol
Concen: 10.239 ng/ul
RT: 6.781 min Scan# 645
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion: 94 Resp: 123406
Ion Ratio Lower Upper
94 100
65 27.7 22.1 33.1
66 30.6 24.9 37.3









#10
Bis(2-Chloroethyl)ether
Concen: 10.829 ng/ul
RT: 7.010 min Scan# 684
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion: 93 Resp: 102571
Ion Ratio Lower Upper
93 100
63 80.5 62.9 94.3
95 33.6 26.2 39.2



#12
2-Chlorophenol
Concen: 9.334 ng/ul
RT: 7.145 min Scan# 707
Delta R.T. 0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:128 Resp: 92014
Ion Ratio Lower Upper
128 100
64 46.8 40.8 61.2
130 32.4 26.1 39.1









#13
2-Methylphenol
Concen: 9.122 ng/ul
RT: 8.016 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:108 Resp: 81388
Ion Ratio Lower Upper
108 100
107 96.2 70.7 106.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 11.257 ng/ul
RT: 8.092 min Scan# 868
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion: 45 Resp: 156792
Ion Ratio Lower Upper
45 100
77 13.6 10.6 16.0
79 10.0 8.5 12.7









#16
Acetophenone
Concen: 10.261 ng/ul
RT: 8.386 min Scan# 918
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:105 Resp: 144473
Ion Ratio Lower Upper
105 100
77 77.9 64.2 96.2
51 27.3 23.7 35.5



#17
N-Nitroso-di-n-propylamine
Concen: 10.682 ng/ul
RT: 8.375 min Scan# 916
Delta R.T. -0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion: 70 Resp: 70677
Ion Ratio Lower Upper
70 100
42 64.3 49.6 74.4
101 9.3 6.9 10.3
130 17.9 15.5 23.3





#18

4-Methylphenol

Concen: 9.345 ng/ul

RT: 8.339 min Scan# 910

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :





SSTD010049

Tgt Ion:108 Resp: 91293

Ion Ratio Lower Upper

108 100

107 119.2 89.4 134.0



#19

Hexachloroethane

Concen: 9.892 ng/ul

RT: 8.639 min Scan# 961

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:117 Resp: 41211


Ion Ratio Lower Upper

117 100

201 97.9 72.2 108.4

199 64.5 45.6 68.4









#22
Nitrobenzene
Concen: 10.908 ng/ul
RT: 8.757 min Scan# 981
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion: 77 Resp: 105149
Ion Ratio Lower Upper
77 100
123 42.7 34.3 51.5
65 11.2 10.0 15.0



#23
Isophorone
Concen: 10.432 ng/ul
RT: 9.286 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion: 82 Resp: 186012
Ion Ratio Lower Upper
82 100
95 5.4 4.7 7.1
138 17.0 13.7 20.5









#25
2-Nitrophenol
Concen: 7.976 ng/ul
RT: 9.463 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:139 Resp: 39305
Ion Ratio Lower Upper
139 100
65 46.7 41.4 62.0
109 22.2 17.8 26.6



#26
2,4-Dimethylphenol
Concen: 9.445 ng/ul
RT: 9.533 min Scan# 1113
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:107 Resp: 86940
Ion Ratio Lower Upper
107 100
121 51.6 40.2 60.4
122 90.4 68.6 102.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.956 ng/ul
 RT: 9.769 min Scan# 1153
 Delta R.T. 0.000 min
 Lab File: BM048957.D
 Acq: 11 Dec 2024 17:02




Instrument :
 BNA_M
 ClientSampleId :
 SST010049


Tgt Ion:	93	Resp:	121941
Ion	Ratio	Lower	Upper
93	100		
95	33.7	26.6	40.0
123	11.2	8.8	13.2



#29
 2,4-Dichlorophenol
 Concen: 9.047 ng/ul
 RT: 9.992 min Scan# 1191
 Delta R.T. 0.000 min
 Lab File: BM048957.D
 Acq: 11 Dec 2024 17:02

Tgt Ion:	162	Resp:	77094
Ion	Ratio	Lower	Upper
162	100		
164	67.7	50.5	75.7
98	34.9	28.3	42.5









#30
Naphthalene
Concen: 10.034 ng/ul
RT: 10.386 min Scan# 1258
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:128 Resp: 297887
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.8 10.1 15.1



#32
4-Chloroaniline
Concen: 9.669 ng/ul
RT: 10.498 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:127 Resp: 108779
Ion Ratio Lower Upper
127 100
129 33.6 26.5 39.7





#33

Hexachlorobutadiene

Concen: 9.148 ng/ul

RT: 10.680 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:225 Resp: 52753

Ion Ratio Lower Upper

225 100

223 61.4 48.9 73.3

227 63.7 50.9 76.3



#34

Caprolactam

Concen: 7.315 ng/ul

RT: 11.275 min Scan# 1409

Delta R.T. 0.006 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:113 Resp: 19645


Ion Ratio Lower Upper

113 100

55 215.1 143.5 215.3

56 136.7 96.9 145.3









#35
4-Chloro-3-methylphenol
Concen: 8.566 ng/ul
RT: 11.633 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:	107	Resp:	70630
Ion	Ratio	Lower	Upper
107	100		
144	29.7	22.2	33.4
142	85.7	66.6	99.8



#36
2-Methylnaphthalene
Concen: 9.758 ng/ul
RT: 12.010 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:	142	Resp:	185370
Ion	Ratio	Lower	Upper
142	100		
141	88.5	70.9	106.3





#37

1-Methylnaphthalene

Concen: 9.569 ng/ul

RT: 12.227 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:142 Resp: 186546

Ion Ratio Lower Upper

142 100

141 88.7 75.1 112.7

116 3.5 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.755 ng/ul

RT: 12.380 min Scan# 1597

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Tgt Ion:216 Resp: 93898




Ion Ratio Lower Upper


216 100

214 81.9 63.7 95.5

179 22.1 17.8 26.8

108 19.5 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 10.049 ng/ul
RT: 12.369 min Scan# 1595
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:237 Resp: 49440
Ion Ratio Lower Upper
237 100
235 67.3 50.5 75.7
272 12.5 9.8 14.6



#41
2,4,6-Trichlorophenol
Concen: 8.423 ng/ul
RT: 12.627 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:196 Resp: 51211
Ion Ratio Lower Upper
196 100
198 90.8 77.9 116.9
200 31.0 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 8.591 ng/ul

RT: 12.698 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:196 Resp: 55622

Ion Ratio Lower Upper

196 100

198 100.1 75.4 113.0

200 32.2 24.4 36.6



#43

1,1'-Biphenyl

Concen: 10.569 ng/ul

RT: 13.039 min Scan# 1709

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:154 Resp: 242739


Ion Ratio Lower Upper

154 100

153 39.7 31.4 47.0

76 14.9 12.2 18.4









#44
2-Chloronaphthalene
Concen: 10.696 ng/ul
RT: 13.074 min Scan# 1715
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:162 Resp: 193845
Ion Ratio Lower Upper
162 100
164 31.3 26.2 39.2
127 35.5 28.6 42.8



#45
2-Nitroaniline
Concen: 9.095 ng/ul
RT: 13.280 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion: 65 Resp: 41394
Ion Ratio Lower Upper
65 100
92 62.9 54.6 81.8
138 95.4 76.1 114.1









#47
Dimethylphthalate
Concen: 10.295 ng/ul
RT: 13.674 min Scan# 1817
Delta R.T. -0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:163 Resp: 221739
Ion Ratio Lower Upper
163 100
194 4.0 2.7 4.1
164 9.9 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 7.912 ng/ul
RT: 13.786 min Scan# 1836
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:165 Resp: 32485
Ion Ratio Lower Upper
165 100
89 64.4 52.4 78.6
121 25.5 18.0 27.0









#50
Acenaphthylene
Concen: 10.474 ng/ul
RT: 13.927 min Scan# 1860
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:152 Resp: 291871
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4
153 12.7 10.3 15.5



#51
3-Nitroaniline
Concen: 8.058 ng/ul
RT: 14.116 min Scan# 1892
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:138 Resp: 35928
Ion Ratio Lower Upper
138 100
108 13.1 8.6 13.0#
92 115.8 93.8 140.6









#52
Acenaphthene
Concen: 10.584 ng/ul
RT: 14.274 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:153 Resp: 207620
Ion Ratio Lower Upper
153 100
152 48.9 39.1 58.7
154 84.5 69.1 103.7



#53
2,4-Dinitrophenol
Concen: 5.220 ng/ul
RT: 14.321 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:184 Resp: 11524
Ion Ratio Lower Upper
184 100
63 74.5 61.2 91.8
154 52.1 48.6 73.0









#55
4-Nitrophenol
Concen: 7.556 ng/ul
RT: 14.433 min Scan# 1945
Delta R.T. 0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:109 Resp: 21756
Ion Ratio Lower Upper
109 100
139 167.0 133.8 200.6
65 154.0 132.3 198.5



#56
Dibenzofuran
Concen: 10.469 ng/ul
RT: 14.610 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:168 Resp: 283306
Ion Ratio Lower Upper
168 100
139 33.6 27.6 41.4









#57
2,4-Dinitrotoluene
Concen: 8.228 ng/ul
RT: 14.574 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:165 Resp: 49110
Ion Ratio Lower Upper
165 100
63 51.0 40.9 61.3
182 2.5 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 7.332 ng/ul
RT: 14.839 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:232 Resp: 39487
Ion Ratio Lower Upper
232 100
131 44.7 39.3 58.9
130 2.8 2.3 3.5
166 35.7 26.2 39.2





#59

Diethylphthalate

Concen: 9.826 ng/ul

RT: 15.051 min Scan# 2051

Delta R.T. -0.006 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:149 Resp: 208518

Ion Ratio Lower Upper

149 100

177 20.2 17.5 26.3

150 12.3 9.4 14.2



#61

Fluorene

Concen: 10.489 ng/ul

RT: 15.262 min Scan# 2087

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:166 Resp: 224732


Ion Ratio Lower Upper

166 100

165 100.8 79.8 119.6

167 13.3 10.6 15.8









#62
 4-Chlorophenyl-phenylether
 Concen: 10.033 ng/ul
 RT: 15.268 min Scan# 2088
 Delta R.T. 0.000 min
 Lab File: BM048957.D
 Acq: 11 Dec 2024 17:02




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010049


Tgt Ion:204 Resp: 105903
 Ion Ratio Lower Upper
 204 100
 206 33.1 26.4 39.6
 141 50.5 42.2 63.2



#63
 4-Nitroaniline
 Concen: 7.290 ng/ul
 RT: 15.280 min Scan# 2090
 Delta R.T. 0.000 min
 Lab File: BM048957.D
 Acq: 11 Dec 2024 17:02

Tgt Ion:138 Resp: 32007
 Ion Ratio Lower Upper
 138 100
 92 58.6 40.6 60.8
 108 65.6 50.3 75.5









#66
4,6-Dinitro-2-methylphenol
Concen: 6.392 ng/ul
RT: 15.339 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:198 Resp: 21674
Ion Ratio Lower Upper
198 100
182 9.7 5.3 7.9#
77 36.7 24.3 36.5#



#67
N-Nitrosodiphenylamine
Concen: 10.616 ng/ul
RT: 15.480 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:169 Resp: 173307
Ion Ratio Lower Upper
169 100
168 70.2 56.9 85.3
167 39.1 31.9 47.9









#68
4-Bromophenyl-phenylether
Concen: 9.073 ng/ul
RT: 16.162 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:248 Resp: 53002
Ion Ratio Lower Upper
248 100
250 100.6 81.4 122.0
141 61.3 58.8 88.2



#69
Hexachlorobenzene
Concen: 9.473 ng/ul
RT: 16.262 min Scan# 2257
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:284 Resp: 66787
Ion Ratio Lower Upper
284 100
142 40.8 32.0 48.0
249 30.0 24.6 37.0





#70

Atrazine

Concen: 9.054 ng/ul

RT: 16.439 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:200 Resp: 53795

Ion Ratio Lower Upper

200 100

173 26.5 20.7 31.1

215 44.4 37.7 56.5



#71

Pentachlorophenol

Concen: 6.784 ng/ul

RT: 16.609 min Scan# 2316

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:266 Resp: 30400


Ion Ratio Lower Upper

266 100

264 62.4 48.4 72.6

268 68.9 52.0 78.0





#72

Phenanthrene

Concen: 10.805 ng/ul

RT: 16.998 min Scan# 2382

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:178 Resp: 331928

Ion Ratio Lower Upper

178 100

179 15.1 12.3 18.5

176 19.5 16.2 24.2



#74

Anthracene

Concen: 10.745 ng/ul

RT: 17.086 min Scan# 2397

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:178 Resp: 331278


Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.4

176 19.9 15.4 23.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 10.400 ng/uL
RT: 12.998 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:216 Resp: 91898
Ion Ratio Lower Upper
216 100
214 76.7 63.7 95.5
218 46.7 37.9 56.9



#76
Pentachlorobenzene
Concen: 10.276 ng/uL
RT: 14.527 min Scan# 1962
Delta R.T. -0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:250 Resp: 88471
Ion Ratio Lower Upper
250 100
248 63.4 49.0 73.4
252 62.8 52.2 78.4





#77

Carbazole

Concen: 10.582 ng/ul

RT: 17.362 min Scan# 2444

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:167 Resp: 293860

Ion Ratio Lower Upper

167 100

166 22.3 17.4 26.2

139 11.7 9.7 14.5



#78

Di-n-butylphthalate

Concen: 8.933 ng/ul

RT: 17.951 min Scan# 2544

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:149 Resp: 291665


Ion Ratio Lower Upper

149 100

150 9.2 7.1 10.7

104 4.7 3.9 5.9





#80

Fluoranthene

Concen: 10.486 ng/ul

RT: 19.021 min Scan# 2726

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:202 Resp: 333930

Ion Ratio Lower Upper

202 100

101 13.0 10.2 15.4

100 10.0 8.3 12.5



#82

Pyrene

Concen: 10.875 ng/ul

RT: 19.386 min Scan# 2788

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:202 Resp: 375536


Ion Ratio Lower Upper

202 100

101 14.4 11.7 17.5

100 12.9 9.4 14.2





#83

Butylbenzylphthalate

Concen: 7.479 ng/ul

RT: 20.315 min Scan# 2946

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:149 Resp: 99203

Ion Ratio Lower Upper

149 100

91 76.4 56.2 84.4

206 17.9 14.2 21.2



#84

3,3'-Dichlorobenzidine

Concen: 6.977 ng/ul

RT: 21.103 min Scan# 3080

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:252 Resp: 69326


Ion Ratio Lower Upper

252 100

254 62.1 51.9 77.9

126 12.4 9.4 14.2









#85
Benzo(a)anthracene
Concen: 9.598 ng/ul
RT: 21.162 min Scan# 3091
Delta R.T. -0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:228 Resp: 330897
Ion Ratio Lower Upper
228 100
229 19.2 15.7 23.5
226 27.6 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 7.592 ng/ul
RT: 21.127 min Scan# 3084
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:149 Resp: 146899
Ion Ratio Lower Upper
149 100
167 25.3 20.6 31.0
279 3.6 2.5 3.7









#87
Chrysene
Concen: 9.550 ng/ul
RT: 21.221 min Scan# 3100
Delta R.T. -0.006 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:228 Resp: 316385
Ion Ratio Lower Upper
228 100
226 30.5 24.6 37.0
229 19.1 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 6.569 ng/ul
RT: 22.203 min Scan# 3267
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:149 Resp: 203609









#90
Benzo(b)fluoranthene
Concen: 10.061 ng/ul
RT: 23.115 min Scan# 3422
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02




Instrument :
BNA_M
ClientSampleId :
SSTD010049


Tgt Ion:252 Resp: 304527
Ion Ratio Lower Upper
252 100
253 21.5 17.3 25.9
125 12.7 9.8 14.8



#91
Benzo(k)fluoranthene
Concen: 9.921 ng/ul
RT: 23.174 min Scan# 3432
Delta R.T. 0.000 min
Lab File: BM048957.D
Acq: 11 Dec 2024 17:02

Tgt Ion:252 Resp: 299410
Ion Ratio Lower Upper
252 100
253 21.6 17.2 25.8
125 14.1 10.1 15.1





#93

Benzo(a)pyrene

Concen: 9.791 ng/ul

RT: 23.862 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:252 Resp: 278629

Ion Ratio Lower Upper

252 100

253 20.5 17.8 26.8

125 13.3 11.0 16.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.538 ng/ul

RT: 27.115 min Scan# 4102

Delta R.T. -0.012 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02




Tgt Ion:276 Resp: 352752


Ion Ratio Lower Upper

276 100

138 26.9 20.0 30.0

277 23.5 18.8 28.2





#95

Dibenzo(a,h)anthracene

Concen: 9.456 ng/ul

RT: 27.179 min Scan# 4114

Delta R.T. -0.006 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Instrument :

BNA_M

ClientSampleId :

SSTD010049





Tgt Ion:278 Resp: 285633

Ion Ratio Lower Upper

278 100

139 20.0 15.2 22.8

279 25.1 18.6 28.0



#96

Benzo(g,h,i)perylene

Concen: 9.580 ng/ul

RT: 28.079 min Scan# 4266

Delta R.T. -0.006 min

Lab File: BM048957.D

Acq: 11 Dec 2024 17:02

Tgt Ion:276 Resp: 277801

Ion Ratio Lower Upper

276 100

138 27.0 22.2 33.2

277 23.4 18.9 28.3

