


Data Path : Z:\HPCHEM1\BNA\_M\Data\BM121217\  
Data File : BM013333.D  
Acq On : 12 Dec 2017 09:56  
Operator : SJ/JU  
Sample : DFTPP61  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM113017.M  
Title : SVOA CALIBRATION  
Last Update : Wed Dec 13 04:59:25 2017



AutoFind: Scans 2454, 2455, 2456; Background Corrected with Scan 2448

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 38.7      | 59640   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.6       | 1063    | PASS             |
| 69          | 198          | 0.00         | 100          | 42.1      | 64773   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 281     | PASS             |
| 127         | 198          | 10           | 80           | 47.2      | 72717   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.2       | 368     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 153925  | PASS             |
| 199         | 198          | 5            | 9            | 6.6       | 10091   | PASS             |
| 275         | 198          | 10           | 60           | 31.6      | 48650   | PASS             |
| 365         | 198          | 1            | 100          | 4.7       | 7203    | PASS             |
| 441         | 443          | 0.01         | 100          | 77.8      | 21928   | PASS             |
| 442         | 198          | 50           | 100          | 89.7      | 138125  | PASS             |
| 443         | 442          | 15           | 24           | 20.4      | 28178   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 389.0     |
| 38.00  | 1179.0    |
| 39.10  | 5871.0    |
| 40.00  | 818.0     |
| 42.90  | 332.0     |
| 44.00  | 1564.0    |
| 45.00  | 341.0     |
| 49.10  | 1240.0    |
| 50.10  | 17936.0   |
| 51.10  | 73104.0   |
| 52.00  | 3523.0    |
| 55.00  | 631.0     |
| 56.00  | 3427.0    |
| 57.00  | 4217.0    |
| 61.10  | 1099.0    |
| 61.80  | 1087.0    |
| 63.00  | 3383.0    |
| 63.90  | 389.0     |
| 65.00  | 2295.0    |
| 68.00  | 1686.0    |
| 69.00  | 80512.0   |
| 69.90  | 364.0     |
| 73.10  | 2156.0    |
| 74.10  | 9028.0    |
| 75.00  | 13845.0   |
| 76.00  | 4600.0    |
| 77.00  | 80800.0   |
| 78.10  | 4543.0    |
| 79.00  | 6587.0    |
| 80.00  | 4890.0    |
| 81.00  | 6872.0    |
| 82.00  | 1093.0    |
| 83.00  | 827.0     |
| 85.00  | 1311.0    |
| 85.90  | 2013.0    |
| 86.90  | 1163.0    |
| 91.00  | 1579.0    |
| 91.90  | 2515.0    |
| 93.00  | 12140.0   |
| 94.00  | 815.0     |
| 96.10  | 908.0     |
| 96.90  | 686.0     |
| 98.00  | 9922.0    |
| 99.00  | 6735.0    |
| 99.90  | 382.0     |
| 101.00 | 3192.0    |
| 103.00 | 799.0     |
| 104.10 | 1990.0    |
| 105.00 | 2010.0    |
| 106.00 | 798.0     |
| 107.00 | 24720.0   |
| 108.00 | 4190.0    |
| 109.00 | 713.0     |
| 110.00 | 43904.0   |
| 110.90 | 7060.0    |
| 112.00 | 1356.0    |
| 113.00 | 387.0     |
| 116.00 | 1985.0    |
| 117.00 | 33248.0   |
| 118.00 | 2203.0    |
| 120.00 | 429.0     |
| 121.90 | 1332.0    |
| 123.00 | 3579.0    |
| 123.90 | 916.0     |
| 125.00 | 706.0     |
| 125.90 | 486.0     |
| 127.00 | 80112.0   |
| 128.00 | 6076.0    |
| 129.00 | 33688.0   |
| 130.00 | 3102.0    |
| 130.90 | 616.0     |
| 132.00 | 330.0     |
| 132.80 | 742.0     |
| 133.90 | 1179.0    |
| 135.00 | 3044.0    |
| 136.10 | 1476.0    |
| 137.00 | 1445.0    |
| 138.20 | 571.0     |
| 139.90 | 478.0     |
| 141.00 | 5653.0    |
| 142.10 | 2426.0    |
| 142.80 | 1073.0    |
| 144.00 | 395.0     |
| 146.00 | 1111.0    |
| 147.10 | 2806.0    |
| 148.00 | 6314.0    |
| 148.90 | 2295.0    |
| 150.90 | 785.0     |
| 152.10 | 463.0     |
| 152.90 | 2641.0    |
| 154.00 | 916.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61

|        |          |
|--------|----------|
| 154.90 | 1772.0   |
| 156.00 | 3398.0   |
| 156.90 | 1142.0   |
| 157.90 | 1101.0   |
| 158.90 | 1384.0   |
| 160.00 | 1740.0   |
| 161.10 | 1824.0   |
| 162.00 | 731.0    |
| 163.10 | 567.0    |
| 164.90 | 2415.0   |
| 166.00 | 1791.0   |
| 167.00 | 10457.0  |
| 168.00 | 6821.0   |
| 168.80 | 838.0    |
| 169.80 | 392.0    |
| 170.90 | 727.0    |
| 172.00 | 1422.0   |
| 172.90 | 1412.0   |
| 174.00 | 2255.0   |
| 175.00 | 4544.0   |
| 176.20 | 691.0    |
| 177.00 | 2087.0   |
| 177.90 | 1222.0   |
| 179.00 | 6943.0   |
| 180.00 | 4446.0   |
| 181.10 | 1681.0   |
| 182.00 | 344.0    |
| 184.10 | 584.0    |
| 185.00 | 3372.0   |
| 186.00 | 27976.0  |
| 187.00 | 8335.0   |
| 188.10 | 1162.0   |
| 189.10 | 2209.0   |
| 190.00 | 544.0    |
| 191.10 | 1313.0   |
| 192.10 | 1897.0   |
| 193.00 | 3203.0   |
| 194.00 | 521.0    |
| 195.10 | 562.0    |
| 196.00 | 5247.0   |
| 196.70 | 1106.0   |
| 198.00 | 171008.0 |
| 198.90 | 10488.0  |
| 199.80 | 477.0    |
| 200.10 | 390.0    |
| 201.70 | 980.0    |
| 203.00 | 1506.0   |
| 203.90 | 6779.0   |
| 205.00 | 11927.0  |
| 206.00 | 51768.0  |
| 207.10 | 8089.0   |
| 208.10 | 2378.0   |
| 208.90 | 873.0    |
| 210.10 | 1215.0   |
| 211.10 | 2028.0   |
| 215.00 | 780.0    |
| 216.00 | 2082.0   |
| 217.00 | 14381.0  |
| 218.10 | 1953.0   |
| 219.70 | 357.0    |
| 221.00 | 4016.0   |
| 221.60 | 2077.0   |
| 222.10 | 1450.0   |
| 223.00 | 3726.0   |
| 224.00 | 30584.0  |
| 225.00 | 6931.0   |
| 226.10 | 901.0    |
| 227.00 | 12128.0  |
| 228.00 | 1919.0   |
| 229.00 | 3399.0   |
| 231.20 | 976.0    |
| 233.00 | 557.0    |
| 234.10 | 857.0    |
| 235.10 | 962.0    |
| 236.20 | 532.0    |
| 237.00 | 1892.0   |
| 239.00 | 670.0    |
| 240.10 | 393.0    |
| 241.10 | 612.0    |
| 242.00 | 1582.0   |
| 243.00 | 1265.0   |
| 244.00 | 19912.0  |
| 245.00 | 2469.0   |
| 246.00 | 5241.0   |
| 246.80 | 972.0    |
| 247.90 | 359.0    |
| 249.00 | 1008.0   |
| 251.10 | 334.0    |
| 254.00 | 885.0    |
| 255.00 | 109304.0 |
| 256.00 | 18784.0  |
| 257.00 | 1418.0   |
| 257.90 | 6819.0   |
| 258.90 | 1303.0   |
| 264.90 | 2950.0   |

Instrument :

BNA\_M

ClientSampleId :

DFTPP61

|        |          |
|--------|----------|
| 266.10 | 542.0    |
| 271.00 | 708.0    |
| 271.80 | 328.0    |
| 273.00 | 2622.0   |
| 274.00 | 8516.0   |
| 275.00 | 52544.0  |
| 276.00 | 6534.0   |
| 277.00 | 5125.0   |
| 278.10 | 1202.0   |
| 281.10 | 360.0    |
| 282.90 | 586.0    |
| 284.00 | 833.0    |
| 285.20 | 332.0    |
| 291.80 | 439.0    |
| 292.80 | 1175.0   |
| 294.00 | 338.0    |
| 296.00 | 14042.0  |
| 297.00 | 2351.0   |
| 302.90 | 1214.0   |
| 303.90 | 794.0    |
| 313.80 | 837.0    |
| 315.00 | 1566.0   |
| 320.80 | 635.0    |
| 323.10 | 4341.0   |
| 323.90 | 647.0    |
| 326.90 | 489.0    |
| 327.90 | 587.0    |
| 332.00 | 625.0    |
| 332.90 | 727.0    |
| 334.10 | 3516.0   |
| 334.90 | 623.0    |
| 335.20 | 639.0    |
| 341.10 | 650.0    |
| 345.90 | 1139.0   |
| 352.00 | 1497.0   |
| 352.90 | 1070.0   |
| 354.00 | 1173.0   |
| 354.90 | 300.0    |
| 365.00 | 6839.0   |
| 366.00 | 1170.0   |
| 371.90 | 1583.0   |
| 373.20 | 556.0    |
| 383.00 | 324.0    |
| 401.90 | 732.0    |
| 403.10 | 1195.0   |
| 421.00 | 781.0    |
| 422.10 | 710.0    |
| 423.10 | 6340.0   |
| 424.00 | 1813.0   |
| 441.10 | 17632.0  |
| 442.10 | 124056.0 |
| 443.00 | 23176.0  |
| 444.20 | 2369.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 727.0     |
| 39.00  | 5363.0    |
| 39.90  | 384.0     |
| 44.00  | 1767.0    |
| 49.10  | 625.0     |
| 50.00  | 20016.0   |
| 51.10  | 67920.0   |
| 52.00  | 3241.0    |
| 55.00  | 483.0     |
| 56.00  | 1872.0    |
| 57.00  | 4273.0    |
| 61.10  | 508.0     |
| 61.90  | 1197.0    |
| 63.00  | 2947.0    |
| 65.20  | 1180.0    |
| 68.10  | 794.0     |
| 69.00  | 71040.0   |
| 70.00  | 480.0     |
| 73.00  | 1676.0    |
| 74.00  | 8793.0    |
| 75.00  | 13754.0   |
| 76.00  | 3878.0    |
| 77.00  | 77248.0   |
| 78.00  | 6096.0    |
| 79.00  | 5358.0    |
| 80.00  | 5013.0    |
| 81.00  | 5788.0    |
| 82.10  | 1665.0    |
| 83.00  | 1040.0    |
| 85.00  | 1147.0    |
| 86.00  | 2239.0    |
| 86.80  | 739.0     |
| 87.10  | 654.0     |
| 91.00  | 1925.0    |
| 91.90  | 1821.0    |
| 93.00  | 11694.0   |
| 94.00  | 743.0     |
| 96.00  | 1224.0    |
| 98.00  | 8899.0    |
| 99.00  | 6042.0    |
| 99.90  | 810.0     |
| 101.00 | 3617.0    |
| 101.90 | 391.0     |
| 102.90 | 1466.0    |
| 104.00 | 2769.0    |
| 104.80 | 1774.0    |
| 105.80 | 1203.0    |
| 107.00 | 24728.0   |
| 108.00 | 4169.0    |
| 108.90 | 660.0     |
| 110.00 | 41840.0   |
| 111.00 | 6870.0    |
| 112.00 | 792.0     |
| 114.80 | 343.0     |
| 115.80 | 1770.0    |
| 117.00 | 27352.0   |
| 118.00 | 1964.0    |
| 120.00 | 347.0     |
| 120.90 | 636.0     |
| 122.00 | 2245.0    |
| 122.90 | 2971.0    |
| 123.90 | 1383.0    |
| 124.90 | 523.0     |
| 127.00 | 88240.0   |
| 128.00 | 5472.0    |
| 129.00 | 37176.0   |
| 130.00 | 3351.0    |
| 130.90 | 644.0     |
| 132.00 | 512.0     |
| 132.70 | 500.0     |
| 134.00 | 1501.0    |
| 135.00 | 2502.0    |
| 136.00 | 1255.0    |
| 136.90 | 1434.0    |
| 138.20 | 445.0     |
| 139.20 | 933.0     |
| 140.20 | 586.0     |
| 141.00 | 5413.0    |
| 142.00 | 1383.0    |
| 143.00 | 1621.0    |
| 144.10 | 422.0     |
| 145.90 | 748.0     |
| 147.00 | 3255.0    |
| 148.00 | 6729.0    |
| 148.90 | 1266.0    |
| 149.90 | 367.0     |
| 151.10 | 843.0     |
| 151.90 | 412.0     |
| 153.10 | 1534.0    |
| 154.00 | 940.0     |
| 155.00 | 2723.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61

|        |          |
|--------|----------|
| 156.00 | 3818.0   |
| 156.90 | 880.0    |
| 158.00 | 760.0    |
| 159.00 | 835.0    |
| 159.80 | 1548.0   |
| 160.90 | 1830.0   |
| 162.10 | 994.0    |
| 162.80 | 415.0    |
| 163.20 | 406.0    |
| 165.00 | 2622.0   |
| 165.90 | 2175.0   |
| 167.00 | 9557.0   |
| 168.00 | 4922.0   |
| 169.00 | 1095.0   |
| 171.00 | 478.0    |
| 172.00 | 826.0    |
| 172.90 | 1624.0   |
| 173.90 | 2429.0   |
| 175.00 | 3748.0   |
| 176.10 | 871.0    |
| 176.80 | 1439.0   |
| 177.80 | 680.0    |
| 178.90 | 7000.0   |
| 180.00 | 5814.0   |
| 181.00 | 2082.0   |
| 183.90 | 418.0    |
| 184.90 | 3020.0   |
| 186.10 | 28848.0  |
| 187.00 | 8346.0   |
| 187.90 | 1007.0   |
| 188.90 | 1643.0   |
| 189.90 | 350.0    |
| 191.00 | 940.0    |
| 192.10 | 2351.0   |
| 193.00 | 3208.0   |
| 194.10 | 823.0    |
| 195.00 | 489.0    |
| 196.00 | 4835.0   |
| 197.10 | 826.0    |
| 197.90 | 184128.0 |
| 199.00 | 12755.0  |
| 199.90 | 1004.0   |
| 201.50 | 1096.0   |
| 203.10 | 1633.0   |
| 204.00 | 11164.0  |
| 205.00 | 14155.0  |
| 206.10 | 50240.0  |
| 207.10 | 9215.0   |
| 207.90 | 2326.0   |
| 208.90 | 1390.0   |
| 209.90 | 1153.0   |
| 211.10 | 1835.0   |
| 211.70 | 567.0    |
| 215.10 | 712.0    |
| 216.10 | 934.0    |
| 217.00 | 15053.0  |
| 218.00 | 2186.0   |
| 220.20 | 987.0    |
| 221.00 | 7678.0   |
| 223.10 | 2733.0   |
| 224.00 | 30248.0  |
| 225.00 | 8574.0   |
| 226.00 | 734.0    |
| 227.00 | 16220.0  |
| 227.90 | 2510.0   |
| 229.10 | 3269.0   |
| 230.00 | 845.0    |
| 231.10 | 1566.0   |
| 234.20 | 657.0    |
| 234.90 | 480.0    |
| 235.10 | 479.0    |
| 236.10 | 903.0    |
| 236.90 | 632.0    |
| 239.00 | 475.0    |
| 240.00 | 338.0    |
| 240.80 | 1012.0   |
| 242.00 | 985.0    |
| 243.00 | 2358.0   |
| 244.00 | 24616.0  |
| 244.90 | 3381.0   |
| 245.90 | 6212.0   |
| 246.90 | 1454.0   |
| 249.00 | 1192.0   |
| 251.90 | 373.0    |
| 253.10 | 684.0    |
| 254.10 | 1182.0   |
| 255.00 | 119664.0 |
| 256.00 | 17848.0  |
| 256.90 | 1496.0   |
| 257.90 | 8026.0   |
| 259.00 | 2112.0   |
| 265.00 | 3563.0   |
| 266.00 | 403.0    |
| 267.10 | 436.0    |
| 271.10 | 465.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61

|        |          |
|--------|----------|
| 272.00 | 307.0    |
| 273.00 | 3983.0   |
| 274.00 | 10052.0  |
| 275.00 | 56400.0  |
| 276.00 | 6499.0   |
| 277.00 | 5070.0   |
| 277.90 | 796.0    |
| 281.00 | 676.0    |
| 282.10 | 446.0    |
| 282.90 | 823.0    |
| 283.90 | 651.0    |
| 285.00 | 1230.0   |
| 291.90 | 566.0    |
| 292.80 | 828.0    |
| 293.90 | 357.0    |
| 294.80 | 344.0    |
| 296.00 | 17632.0  |
| 296.90 | 2282.0   |
| 302.90 | 2130.0   |
| 303.90 | 662.0    |
| 313.80 | 647.0    |
| 314.20 | 754.0    |
| 315.00 | 1634.0   |
| 316.10 | 793.0    |
| 321.10 | 541.0    |
| 322.00 | 311.0    |
| 323.00 | 5845.0   |
| 324.10 | 915.0    |
| 325.80 | 313.0    |
| 326.80 | 872.0    |
| 328.10 | 471.0    |
| 332.00 | 500.0    |
| 334.10 | 2988.0   |
| 335.10 | 1132.0   |
| 341.20 | 946.0    |
| 345.90 | 980.0    |
| 352.00 | 1433.0   |
| 352.80 | 932.0    |
| 353.90 | 1864.0   |
| 365.00 | 7285.0   |
| 365.90 | 1481.0   |
| 370.80 | 341.0    |
| 372.00 | 3893.0   |
| 373.00 | 814.0    |
| 382.90 | 722.0    |
| 390.20 | 377.0    |
| 391.20 | 388.0    |
| 392.00 | 364.0    |
| 402.10 | 1478.0   |
| 403.00 | 2095.0   |
| 421.00 | 1115.0   |
| 422.10 | 900.0    |
| 423.00 | 8807.0   |
| 423.90 | 2540.0   |
| 425.00 | 311.0    |
| 441.00 | 26328.0  |
| 442.10 | 164224.0 |
| 443.10 | 34680.0  |
| 444.10 | 2581.0   |
| 445.10 | 460.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 567.0     |
| 39.00  | 2188.0    |
| 40.00  | 848.0     |
| 42.90  | 322.0     |
| 43.90  | 1821.0    |
| 49.20  | 747.0     |
| 50.10  | 10598.0   |
| 51.00  | 37896.0   |
| 52.10  | 1816.0    |
| 56.10  | 1350.0    |
| 57.00  | 1943.0    |
| 60.90  | 417.0     |
| 61.90  | 983.0     |
| 63.00  | 1817.0    |
| 64.00  | 472.0     |
| 65.10  | 983.0     |
| 67.90  | 709.0     |
| 69.00  | 42768.0   |
| 73.10  | 576.0     |
| 74.10  | 4117.0    |
| 75.00  | 7622.0    |
| 76.00  | 2351.0    |
| 77.00  | 42688.0   |
| 78.00  | 2409.0    |
| 79.00  | 2983.0    |
| 80.10  | 2263.0    |
| 81.00  | 4476.0    |
| 82.00  | 873.0     |
| 83.10  | 970.0     |
| 84.00  | 514.0     |
| 84.80  | 644.0     |
| 85.90  | 1428.0    |
| 87.00  | 719.0     |
| 91.00  | 845.0     |
| 92.10  | 1422.0    |
| 93.00  | 7125.0    |
| 96.00  | 496.0     |
| 98.10  | 4483.0    |
| 99.00  | 3904.0    |
| 99.90  | 505.0     |
| 101.00 | 2153.0    |
| 102.10 | 438.0     |
| 102.80 | 671.0     |
| 104.00 | 1502.0    |
| 105.10 | 1448.0    |
| 107.00 | 14956.0   |
| 107.90 | 1943.0    |
| 108.90 | 582.0     |
| 110.00 | 24424.0   |
| 111.00 | 2781.0    |
| 112.00 | 1032.0    |
| 116.20 | 802.0     |
| 117.00 | 17248.0   |
| 118.10 | 892.0     |
| 119.90 | 341.0     |
| 122.00 | 961.0     |
| 123.00 | 1345.0    |
| 124.00 | 832.0     |
| 125.10 | 851.0     |
| 127.00 | 49800.0   |
| 128.00 | 4215.0    |
| 129.00 | 21800.0   |
| 130.00 | 1341.0    |
| 130.80 | 624.0     |
| 132.90 | 660.0     |
| 134.00 | 981.0     |
| 134.90 | 2562.0    |
| 136.10 | 569.0     |
| 137.10 | 1330.0    |
| 138.10 | 417.0     |
| 140.20 | 405.0     |
| 141.00 | 3396.0    |
| 142.00 | 1312.0    |
| 142.70 | 398.0     |
| 146.00 | 657.0     |
| 147.00 | 1870.0    |
| 148.00 | 3313.0    |
| 148.90 | 1076.0    |
| 150.00 | 429.0     |
| 151.00 | 454.0     |
| 153.00 | 1026.0    |
| 154.10 | 686.0     |
| 154.90 | 1679.0    |
| 156.10 | 2408.0    |
| 157.90 | 367.0     |
| 158.90 | 467.0     |
| 159.90 | 900.0     |
| 161.00 | 1113.0    |
| 161.80 | 496.0     |
| 164.90 | 1444.0    |
| 165.80 | 646.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61



|        |          |
|--------|----------|
| 167.00 | 6439.0   |
| 168.00 | 3896.0   |
| 170.80 | 413.0    |
| 173.10 | 630.0    |
| 173.90 | 1215.0   |
| 175.10 | 1931.0   |
| 175.80 | 785.0    |
| 177.00 | 1857.0   |
| 177.90 | 615.0    |
| 179.00 | 5898.0   |
| 180.00 | 3735.0   |
| 181.10 | 1372.0   |
| 183.80 | 343.0    |
| 185.10 | 2229.0   |
| 186.00 | 16240.0  |
| 187.10 | 4915.0   |
| 187.80 | 608.0    |
| 189.10 | 1223.0   |
| 191.10 | 785.0    |
| 191.90 | 1927.0   |
| 193.00 | 1628.0   |
| 194.00 | 714.0    |
| 195.10 | 599.0    |
| 196.00 | 3941.0   |
| 197.90 | 106640.0 |
| 199.00 | 7031.0   |
| 200.10 | 1067.0   |
| 201.60 | 587.0    |
| 202.70 | 539.0    |
| 203.10 | 739.0    |
| 204.00 | 4296.0   |
| 205.00 | 8199.0   |
| 206.10 | 31792.0  |
| 207.00 | 7327.0   |
| 208.10 | 1765.0   |
| 209.00 | 1040.0   |
| 209.90 | 910.0    |
| 211.00 | 1567.0   |
| 214.90 | 864.0    |
| 215.90 | 955.0    |
| 217.00 | 11435.0  |
| 218.10 | 1127.0   |
| 220.00 | 300.0    |
| 221.00 | 4368.0   |
| 222.10 | 808.0    |
| 222.90 | 2091.0   |
| 224.00 | 20312.0  |
| 225.00 | 3649.0   |
| 226.00 | 760.0    |
| 227.00 | 10844.0  |
| 227.80 | 1314.0   |
| 229.00 | 2246.0   |
| 231.10 | 570.0    |
| 234.00 | 550.0    |
| 235.10 | 843.0    |
| 237.10 | 934.0    |
| 240.00 | 362.0    |
| 241.00 | 470.0    |
| 242.00 | 1039.0   |
| 243.00 | 1026.0   |
| 244.10 | 13804.0  |
| 245.00 | 2229.0   |
| 245.90 | 3487.0   |
| 246.90 | 468.0    |
| 248.80 | 484.0    |
| 252.80 | 441.0    |
| 253.10 | 392.0    |
| 254.10 | 566.0    |
| 255.00 | 82480.0  |
| 256.00 | 12029.0  |
| 257.10 | 1011.0   |
| 258.00 | 5433.0   |
| 259.10 | 736.0    |
| 264.90 | 1824.0   |
| 265.90 | 337.0    |
| 267.00 | 411.0    |
| 272.10 | 331.0    |
| 272.90 | 2849.0   |
| 274.00 | 6654.0   |
| 275.00 | 37008.0  |
| 276.00 | 5787.0   |
| 276.90 | 3449.0   |
| 278.00 | 832.0    |
| 281.00 | 499.0    |
| 282.00 | 309.0    |
| 283.00 | 739.0    |
| 284.10 | 498.0    |
| 285.00 | 1237.0   |
| 289.10 | 320.0    |
| 293.00 | 1229.0   |
| 296.00 | 13417.0  |
| 296.90 | 1665.0   |
| 303.20 | 1233.0   |
| 304.00 | 433.0    |
| 314.00 | 462.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61

|        |          |
|--------|----------|
| 314.90 | 966.0    |
| 315.90 | 870.0    |
| 321.90 | 701.0    |
| 323.00 | 3974.0   |
| 324.10 | 476.0    |
| 326.70 | 468.0    |
| 327.10 | 497.0    |
| 327.80 | 326.0    |
| 332.00 | 357.0    |
| 333.90 | 2377.0   |
| 335.00 | 684.0    |
| 339.80 | 389.0    |
| 341.10 | 676.0    |
| 345.90 | 597.0    |
| 346.20 | 538.0    |
| 352.10 | 1108.0   |
| 352.70 | 445.0    |
| 353.20 | 1158.0   |
| 354.00 | 1567.0   |
| 355.10 | 515.0    |
| 365.00 | 7485.0   |
| 365.90 | 639.0    |
| 371.10 | 420.0    |
| 372.00 | 2057.0   |
| 373.20 | 539.0    |
| 383.00 | 501.0    |
| 389.90 | 311.0    |
| 391.00 | 366.0    |
| 401.90 | 1405.0   |
| 403.00 | 1060.0   |
| 420.90 | 938.0    |
| 422.00 | 1270.0   |
| 423.00 | 7593.0   |
| 424.00 | 1961.0   |
| 441.00 | 21824.0  |
| 442.10 | 126096.0 |
| 443.10 | 26680.0  |
| 444.10 | 1866.0   |
| 445.00 | 307.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61

Scan 2448 (17.486 min): BM013333.D  
DFTPP61

| m/z    | Abundance |
|--------|-----------|
| 40.10  | 596.0     |
| 44.00  | 1478.0    |
| 48.80  | 318.0     |
| 73.10  | 835.0     |
| 85.90  | 358.0     |
| 132.80 | 392.0     |
| 147.00 | 486.0     |
| 166.90 | 551.0     |
| 191.00 | 562.0     |
| 207.00 | 2853.0    |
| 208.10 | 343.0     |
| 209.10 | 456.0     |
| 220.90 | 374.0     |
| 280.90 | 709.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP61