

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121222\
 Data File : BM037962.D
 Acq On : 12 Dec 2022 12:06
 Operator : CG/JU
 Sample : N149539BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS539

Quant Time: Dec 13 00:55:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 09 21:34:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	227468	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.893	136	967893	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.704	164	568252	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.445	188	1032201	20.000	ng/u1	0.00
79) Chrysene-d12	21.621	240	747967	20.000	ng/u1	0.00
88) Perylene-d12	24.098	264	859394	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	26304	5.251	ng/uL	0.00
4) Pyridine-d5	3.852	84	387396	26.067	ng/u1	0.00
7) Phenol-d5	7.228	99	546225	29.551	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.393	67	314205	27.916	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.599	132	461828	30.457	ng/u1	-0.01
15) 4-Methylphenol-d8	8.787	113	436203	30.232	ng/u1	0.00
21) Nitrobenzene-d5	9.246	128	233844	30.515	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.969	143	251135	31.271	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.510	165	462803	30.340	ng/u1	-0.01
31) 4-Chloroaniline-d4	11.028	131	559758	26.584	ng/u1	-0.01
46) Dimethylphthalate-d6	14.110	166	1222464	30.175	ng/u1	-0.01
49) Acenaphthylene-d8	14.398	160	1459560	29.478	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.898	143	178781	25.894	ng/u1	0.00
60) Fluorene-d10	15.692	176	1032574	28.608	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.816	200	165528	29.995	ng/u1	0.00
73) Anthracene-d10	17.545	188	1427921	29.641	ng/u1	0.00
81) Pyrene-d10	19.827	212	1415232	31.498	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.939	264	1297849	30.235	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	54147	10.303	ng/uL	97
5) Pyridine	3.869	79	373491	25.008	ng/u1	97
6) Benzaldehyde	7.205	77	258351	37.369	ng/u1	99
8) Phenol	7.252	94	560371	30.228	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.487	93	441216	28.984	ng/u1	98
12) 2-Chlorophenol	7.628	128	484885	31.267	ng/u1	100
13) 2-Methylphenol	8.516	108	440986	31.084	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.604	45	760047	27.650	ng/u1	98
16) Acetophenone	8.904	105	690342	30.421	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.893	70	335172	31.354	ng/u1	98
18) 4-Methylphenol	8.851	108	476696	31.061	ng/u1	100
19) Hexachloroethane	9.157	117	190627	30.748	ng/u1	96
22) Nitrobenzene	9.287	77	504897	30.620	ng/u1	98
23) Isophorone	9.816	82	945111	31.279	ng/u1	99
25) 2-Nitrophenol	10.004	139	274123	32.558	ng/u1	97
26) 2,4-Dimethylphenol	10.057	107	500986	30.770	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.298	93	589414	29.345	ng/u1	99
29) 2,4-Dichlorophenol	10.534	162	459794	30.723	ng/u1	98
30) Naphthalene	10.945	128	1542484	29.531	ng/u1	100
32) 4-Chloroaniline	11.057	127	461659	21.987	ng/u1	100
33) Hexachlorobutadiene	11.222	225	300480	30.520	ng/u1	98
34) Caprolactam	11.845	113	124217	32.356	ng/u1	92
35) 4-Chloro-3-methylphenol	12.169	107	433457	31.819	ng/u1	96
36) 2-Methylnaphthalene	12.539	142	1026648	30.304	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121222\
 Data File : BM037962.D
 Acq On : 12 Dec 2022 12:06
 Operator : CG/JU
 Sample : N149539BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS539

Quant Time: Dec 13 00:55:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 09 21:34:57 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.763	142	1032698	29.811	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.904	216	550931	29.672	ng/ul	98
40) Hexachlorocyclopentadiene	12.881	237	176650	16.632	ng/ul	99
41) 2,4,6-Trichlorophenol	13.145	196	348032	30.773	ng/ul	99
42) 2,4,5-Trichlorophenol	13.216	196	362739	30.754	ng/ul	99
43) 1,1'-Biphenyl	13.545	154	1323939	28.985	ng/ul	99
44) 2-Chloronaphthalene	13.586	162	1042570	29.588	ng/ul	97
45) 2-Nitroaniline	13.792	65	272562	32.103	ng/ul	99
47) Dimethylphthalate	14.163	163	1248390	31.233	ng/ul	99
48) 2,6-Dinitrotoluene	14.286	165	252332	32.501	ng/ul	98
50) Acenaphthylene	14.428	152	1625141	29.850	ng/ul	98
51) 3-Nitroaniline	14.610	138	242380	31.290	ng/ul	97
52) Acenaphthene	14.769	153	1114207	29.588	ng/ul	99
53) 2,4-Dinitrophenol	14.816	184	104235	27.360	ng/ul	96
55) 4-Nitrophenol	14.916	109	131045	28.780	ng/ul	98
56) Dibenzofuran	15.104	168	1468925	28.571	ng/ul	99
57) 2,4-Dinitrotoluene	15.069	165	333775	31.745	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.328	232	297418	29.977	ng/ul	99
59) Diethylphthalate	15.516	149	1240608	31.156	ng/ul	99
61) Fluorene	15.751	166	1221008	29.121	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.739	204	603491	29.452	ng/ul	96
63) 4-Nitroaniline	15.775	138	235589	33.542	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.828	198	164762	30.840	ng/ul#	98
67) N-Nitrosodiphenylamine	15.951	169	985359	31.647	ng/ul	99
68) 4-Bromophenyl-phenylether	16.633	248	361710	31.968	ng/ul	98
69) Hexachlorobenzene	16.751	284	425167	31.479	ng/ul	97
70) Atrazine	16.898	200	189233	17.785	ng/ul	99
71) Pentachlorophenol	17.092	266	220869	29.413	ng/ul	99
72) Phenanthrene	17.486	178	1704990	30.247	ng/ul	99
74) Anthracene	17.580	178	1730089	30.202	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.510	216	539243	32.315	ng/uL	99
76) Pentachlorobenzene	15.022	250	515163	31.165	ng/uL	100
77) Carbazole	17.851	167	1445882	29.133	ng/ul	99
78) Di-n-butylphthalate	18.398	149	1980837	33.578	ng/ul	99
80) Fluoranthene	19.498	202	1715009	32.342	ng/ul	98
82) Pyrene	19.857	202	1749586	31.703	ng/ul	98
83) Butylbenzylphthalate	20.739	149	694022	33.820	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.533	252	495910	33.506	ng/ul	97
85) Benzo(a)anthracene	21.604	228	1589833	31.572	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.509	149	1000496	33.514	ng/ul	98
87) Chrysene	21.657	228	1464681	31.001	ng/ul	100
89) Di-n-octyl phthalate	22.462	149	1645592	29.884	ng/ul	100
90) Benzo(b)fluoranthene	23.339	252	1616398	30.338	ng/ul	99
91) Benzo(k)fluoranthene	23.392	252	1634123	31.647	ng/ul	99
93) Benzo(a)pyrene	23.992	252	1456563	31.089	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.697	276	1778831	29.480	ng/ul	99
95) Dibenzo(a,h)anthracene	26.709	278	1564856	30.763	ng/ul	99
96) Benzo(g,h,i)perylene	27.503	276	1469215	30.497	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121222\
 Data File : BM037962.D
 Acq On : 12 Dec 2022 12:06
 Operator : CG/JU
 Sample : N149539BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS539

Quant Time: Dec 13 00:55:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 09 21:34:57 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.303 ng/uL
RT: 3.452 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion: 88 Resp: 54147
Ion Ratio Lower Upper
88 100
43 39.0 32.8 49.2
58 82.9 64.6 97.0



#5
Pyridine
Concen: 25.008 ng/uL
RT: 3.869 min Scan# 116
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion: 79 Resp: 373491
Ion Ratio Lower Upper
79 100
52 93.4 78.1 117.1
51 45.8 37.2 55.8









#6
Benzaldehyde
Concen: 37.369 ng/ul
RT: 7.205 min Scan# 684
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion: 77 Resp: 258351
Ion Ratio Lower Upper
77 100
105 104.0 82.5 123.7
106 101.7 81.8 122.6



#8
Phenol
Concen: 30.228 ng/ul
RT: 7.252 min Scan# 691
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion: 94 Resp: 560371
Ion Ratio Lower Upper
94 100
65 28.3 22.2 33.4
66 35.5 27.3 40.9









#10
Bis(2-Chloroethyl)ether
Concen: 28.984 ng/ul
RT: 7.487 min Scan# 71
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion: 93 Resp: 441216
Ion Ratio Lower Upper
93 100
63 81.1 66.3 99.5
95 32.7 25.7 38.5



#12
2-Chlorophenol
Concen: 31.267 ng/ul
RT: 7.628 min Scan# 755
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:128 Resp: 484885
Ion Ratio Lower Upper
128 100
64 44.5 35.2 52.8
130 32.3 25.9 38.9





#13

2-Methylphenol

Concen: 31.084 ng/ul

RT: 8.516 min Scan# 906

Delta R.T. -0.012 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :





SLCS539

Tgt Ion:108 Resp: 440986

Ion Ratio Lower Upper

108 100

107 87.3 71.9 107.9



#14

2,2'-oxybis(1-chloropropane)

Concen: 27.650 ng/ul

RT: 8.604 min Scan# 921

Delta R.T. -0.012 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06




Tgt Ion: 45 Resp: 760047


Ion Ratio Lower Upper

45 100

77 10.6 9.0 13.4

79 9.0 6.2 9.4









#16
Acetophenone
Concen: 30.421 ng/ul
RT: 8.904 min Scan# 91
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:105 Resp: 690342
Ion Ratio Lower Upper
105 100
77 73.7 59.7 89.5
51 32.7 27.2 40.8



#17
N-Nitroso-di-n-propylamine
Concen: 31.354 ng/ul
RT: 8.893 min Scan# 970
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion: 70 Resp: 335172
Ion Ratio Lower Upper
70 100
42 76.0 62.3 93.5
101 10.2 8.0 12.0
130 23.8 17.6 26.4









#18
4-Methylphenol
Concen: 31.061 ng/ul
RT: 8.851 min Scan# 90
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:108 Resp: 476696
Ion Ratio Lower Upper
108 100
107 110.3 88.1 132.1



#19
Hexachloroethane
Concen: 30.748 ng/ul
RT: 9.157 min Scan# 1015
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:117 Resp: 190627
Ion Ratio Lower Upper
117 100
201 103.7 87.3 130.9
199 64.2 50.9 76.3









#22
Nitrobenzene
Concen: 30.620 ng/ul
RT: 9.287 min Scan# 1037
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion: 77 Resp: 504897
Ion Ratio Lower Upper
77 100
123 48.2 37.8 56.6
65 13.2 11.1 16.7



#23
Isophorone
Concen: 31.279 ng/ul
RT: 9.816 min Scan# 1127
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion: 82 Resp: 945111
Ion Ratio Lower Upper
82 100
95 6.8 5.2 7.8
138 19.0 14.9 22.3









#25
2-Nitrophenol
Concen: 32.558 ng/ul
RT: 10.004 min Scan# 1159
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:139 Resp: 274123
Ion Ratio Lower Upper
139 100
65 47.5 36.2 54.4
109 26.1 19.7 29.5



#26
2,4-Dimethylphenol
Concen: 30.770 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:107 Resp: 500986
Ion Ratio Lower Upper
107 100
121 52.2 42.0 63.0
122 89.4 71.4 107.2





#27

Bis(2-Chloroethoxy)methane

Concen: 29.345 ng/ul

RT: 10.298 min Scan# 1209

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :

SLCS539





Tgt Ion: 93 Resp: 589414

Ion Ratio Lower Upper

93 100

95 32.8 25.5 38.3

123 11.4 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 30.723 ng/ul

RT: 10.534 min Scan# 1249

Delta R.T. -0.012 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06




Tgt Ion:162 Resp: 459794


Ion Ratio Lower Upper

162 100

164 66.0 51.4 77.2

98 33.5 26.1 39.1









#30
Naphthalene
Concen: 29.531 ng/ul
RT: 10.945 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:128 Resp: 1542484
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 13.1 10.4 15.6



#32
4-Chloroaniline
Concen: 21.987 ng/ul
RT: 11.057 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:127 Resp: 461659
Ion Ratio Lower Upper
127 100
129 31.8 25.3 37.9









#33
Hexachlorobutadiene
Concen: 30.520 ng/ul
RT: 11.222 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:225 Resp: 300480
Ion Ratio Lower Upper
225 100
223 61.1 50.8 76.2
227 62.7 50.5 75.7



#34
Caprolactam
Concen: 32.356 ng/ul
RT: 11.845 min Scan# 1472
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:113 Resp: 124217
Ion Ratio Lower Upper
113 100
55 196.3 172.2 258.2
56 147.9 120.7 181.1









#35
4-Chloro-3-methylphenol
Concen: 31.819 ng/ul
RT: 12.169 min Scan# 1527
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:107 Resp: 433457
Ion Ratio Lower Upper
107 100
144 26.7 23.4 35.0
142 84.9 71.0 106.4



#36
2-Methylnaphthalene
Concen: 30.304 ng/ul
RT: 12.539 min Scan# 1590
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:142 Resp: 1026648
Ion Ratio Lower Upper
142 100
141 88.6 69.4 104.0





#37

1-Methylnaphthalene

Concen: 29.811 ng/ul

RT: 12.763 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :

SLCS539





Tgt Ion:142 Resp: 1032698

Ion Ratio Lower Upper

142 100

141 90.6 72.4 108.6

116 3.7 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 29.672 ng/ul

RT: 12.904 min Scan# 1652

Delta R.T. -0.012 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Tgt Ion:216 Resp: 550931




Ion Ratio Lower Upper


216 100

214 79.1 64.2 96.4

179 20.1 15.4 23.2

108 17.0 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 16.632 ng/ul
RT: 12.881 min Scan# 1648
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:237 Resp: 176650
Ion Ratio Lower Upper
237 100
235 62.9 50.4 75.6
272 14.7 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 30.773 ng/ul
RT: 13.145 min Scan# 1693
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:196 Resp: 348032
Ion Ratio Lower Upper
196 100
198 95.1 75.0 112.6
200 29.7 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 30.754 ng/ul
RT: 13.216 min Scan# 1705
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:196 Resp: 362739
Ion Ratio Lower Upper
196 100
198 96.7 76.5 114.7
200 31.3 24.9 37.3



#43
1,1'-Biphenyl
Concen: 28.985 ng/ul
RT: 13.545 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:154 Resp: 1323939
Ion Ratio Lower Upper
154 100
153 39.6 31.2 46.8
76 11.6 9.3 13.9









#44
2-Chloronaphthalene
Concen: 29.588 ng/ul
RT: 13.586 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:162 Resp: 1042570
Ion Ratio Lower Upper
162 100
164 32.3 27.3 40.9
127 37.9 29.4 44.0



#45
2-Nitroaniline
Concen: 32.103 ng/ul
RT: 13.792 min Scan# 1803
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion: 65 Resp: 272562
Ion Ratio Lower Upper
65 100
92 68.3 53.5 80.3
138 114.0 91.3 136.9





#47

Dimethylphthalate

Concen: 31.233 ng/ul

RT: 14.163 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :

SLCS539





Tgt Ion:163 Resp: 1248390

Ion Ratio Lower Upper

163 100

194 4.1 3.0 4.6

164 9.7 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 32.501 ng/ul

RT: 14.286 min Scan# 1887

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06




Tgt Ion:165 Resp: 252332


Ion Ratio Lower Upper

165 100

89 53.3 41.1 61.7

121 21.0 17.4 26.2









#50
Acenaphthylene
Concen: 29.850 ng/ul
RT: 14.428 min Scan# 1911
Delta R.T. -0.012 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:152 Resp: 1625141
Ion Ratio Lower Upper
152 100
151 19.7 15.3 22.9
153 13.7 10.2 15.4



#51
3-Nitroaniline
Concen: 31.290 ng/ul
RT: 14.610 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:138 Resp: 242380
Ion Ratio Lower Upper
138 100
108 9.3 7.9 11.9
92 107.9 83.4 125.0









#52
Acenaphthene
Concen: 29.588 ng/ul
RT: 14.769 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:153 Resp: 1114207
Ion Ratio Lower Upper
153 100
152 47.8 38.2 57.4
154 88.1 71.4 107.0



#53
2,4-Dinitrophenol
Concen: 27.360 ng/ul
RT: 14.816 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:184 Resp: 104235
Ion Ratio Lower Upper
184 100
63 63.5 46.2 69.4
154 57.5 45.5 68.3









#55
4-Nitrophenol
Concen: 28.780 ng/ul
RT: 14.916 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:109 Resp: 131045
Ion Ratio Lower Upper
109 100
139 152.1 119.2 178.8
65 138.1 108.7 163.1



#56
Dibenzofuran
Concen: 28.571 ng/ul
RT: 15.104 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:168 Resp: 1468925
Ion Ratio Lower Upper
168 100
139 35.9 28.1 42.1









#57
2,4-Dinitrotoluene
Concen: 31.745 ng/ul
RT: 15.069 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:165 Resp: 333775
Ion Ratio Lower Upper
165 100
63 42.1 34.9 52.3
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 29.977 ng/ul
RT: 15.328 min Scan# 2064
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:232 Resp: 297418
Ion Ratio Lower Upper
232 100
131 42.1 32.8 49.2
130 2.5 2.0 3.0
166 29.6 23.2 34.8









#59
Diethylphthalate
Concen: 31.156 ng/ul
RT: 15.516 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:149 Resp: 1240608
Ion Ratio Lower Upper
149 100
177 22.3 18.5 27.7
150 11.9 9.5 14.3



#61
Fluorene
Concen: 29.121 ng/ul
RT: 15.751 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:166 Resp: 1221008
Ion Ratio Lower Upper
166 100
165 97.8 77.2 115.8
167 12.9 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 29.452 ng/ul
RT: 15.739 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:204 Resp: 603491
Ion Ratio Lower Upper
204 100
206 33.0 25.5 38.3
141 56.9 42.8 64.2



#63
4-Nitroaniline
Concen: 33.542 ng/ul
RT: 15.775 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:138 Resp: 235589
Ion Ratio Lower Upper
138 100
92 45.8 36.1 54.1
108 72.3 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 30.840 ng/ul
RT: 15.828 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:198 Resp: 164762
Ion Ratio Lower Upper
198 100
182 3.6 3.8 5.8#
77 23.1 19.3 28.9



#67
N-Nitrosodiphenylamine
Concen: 31.647 ng/ul
RT: 15.951 min Scan# 2170
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:169 Resp: 985359
Ion Ratio Lower Upper
169 100
168 66.4 53.8 80.8
167 36.0 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 31.968 ng/ul
RT: 16.633 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:248 Resp: 361710
Ion Ratio Lower Upper
248 100
250 97.0 78.7 118.1
141 64.4 52.7 79.1



#69
Hexachlorobenzene
Concen: 31.479 ng/ul
RT: 16.751 min Scan# 2306
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:284 Resp: 425167
Ion Ratio Lower Upper
284 100
142 31.2 23.1 34.7
249 30.3 23.3 34.9









#70
Atrazine
Concen: 17.785 ng/ul
RT: 16.898 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:200 Resp: 189233
Ion Ratio Lower Upper
200 100
173 26.6 20.5 30.7
215 50.6 40.1 60.1



#71
Pentachlorophenol
Concen: 29.413 ng/ul
RT: 17.092 min Scan# 2364
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:266 Resp: 220869
Ion Ratio Lower Upper
266 100
264 62.1 50.1 75.1
268 64.6 50.6 75.8









#72
Phenanthrene
Concen: 30.247 ng/ul
RT: 17.486 min Scan# 2431
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:178 Resp: 1704990
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 18.9 15.4 23.0



#74
Anthracene
Concen: 30.202 ng/ul
RT: 17.580 min Scan# 2447
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:178 Resp: 1730089
Ion Ratio Lower Upper
178 100
179 15.2 11.8 17.8
176 18.8 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.315 ng/uL
RT: 13.510 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:216 Resp: 539243
Ion Ratio Lower Upper
216 100
214 79.0 62.6 93.8
218 49.5 38.3 57.5



#76
Pentachlorobenzene
Concen: 31.165 ng/uL
RT: 15.022 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:250 Resp: 515163
Ion Ratio Lower Upper
250 100
248 63.8 51.0 76.4
252 64.2 51.8 77.6





#77

Carbazole

Concen: 29.133 ng/ul

RT: 17.851 min Scan# 2493

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :

SLCS539





Tgt Ion:167 Resp: 1445882

Ion Ratio Lower Upper

167 100

166 21.3 16.7 25.1

139 12.6 9.8 14.6



#78

Di-n-butylphthalate

Concen: 33.578 ng/ul

RT: 18.398 min Scan# 2586

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06




Tgt Ion:149 Resp: 1980837


Ion Ratio Lower Upper

149 100

150 9.2 7.3 10.9

104 4.6 3.4 5.2









#80
Fluoranthene
Concen: 32.342 ng/ul
RT: 19.498 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:202 Resp: 1715009
Ion Ratio Lower Upper
202 100
101 9.9 8.6 13.0
100 7.7 6.6 10.0



#82
Pyrene
Concen: 31.703 ng/ul
RT: 19.857 min Scan# 2834
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:202 Resp: 1749586
Ion Ratio Lower Upper
202 100
101 13.2 10.0 15.0
100 10.6 7.9 11.9









#83
Butylbenzylphthalate
Concen: 33.820 ng/ul
RT: 20.739 min Scan# 2984
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:149 Resp: 694022
Ion Ratio Lower Upper
149 100
91 67.6 54.5 81.7
206 20.1 15.8 23.8



#84
3,3'-Dichlorobenzidine
Concen: 33.506 ng/ul
RT: 21.533 min Scan# 3119
Delta R.T. 0.000 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:252 Resp: 495910
Ion Ratio Lower Upper
252 100
254 63.6 53.2 79.8
126 10.3 7.8 11.8





#85

Benzo(a)anthracene

Concen: 31.572 ng/ul

RT: 21.604 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :

SLCS539





Tgt Ion:228 Resp: 1589833

Ion Ratio Lower Upper

228 100

229 19.5 15.5 23.3

226 27.2 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 33.514 ng/ul

RT: 21.509 min Scan# 3115

Delta R.T. 0.000 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06




Tgt Ion:149 Resp: 1000496


Ion Ratio Lower Upper

149 100

167 27.8 21.5 32.3

279 4.1 3.2 4.8





#87

Chrysene

Concen: 31.001 ng/ul

RT: 21.657 min Scan# 3140

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :

SLCS539





Tgt Ion:228 Resp: 1464681

Ion Ratio Lower Upper

228 100

226 30.3 24.3 36.5

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 29.884 ng/ul




RT: 22.462 min Scan# 3277


Delta R.T. 0.000 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Tgt Ion:149 Resp: 1645592









#90
Benzo(b)fluoranthene
Concen: 30.338 ng/ul
RT: 23.339 min Scan# 3426
Delta R.T. -0.006 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06




Instrument :
BNA_M
ClientSampleId :
SLCS539


Tgt Ion:252 Resp: 1616398
Ion Ratio Lower Upper
252 100
253 22.5 17.6 26.4
125 10.1 8.0 12.0



#91
Benzo(k)fluoranthene
Concen: 31.647 ng/ul
RT: 23.392 min Scan# 3435
Delta R.T. 0.000 min
Lab File: BM037962.D
Acq: 12 Dec 2022 12:06

Tgt Ion:252 Resp: 1634123
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.2
125 9.7 8.1 12.1





#93

Benzo(a)pyrene

Concen: 31.089 ng/ul

RT: 23.992 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :

SLCS539





Tgt Ion:252 Resp: 1456563

Ion Ratio Lower Upper

252 100

253 21.3 17.4 26.0

125 10.2 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 29.480 ng/ul

RT: 26.697 min Scan# 3997

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06




Tgt Ion:276 Resp: 1778831


Ion Ratio Lower Upper

276 100

138 23.3 19.3 28.9

277 25.1 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 30.763 ng/ul

RT: 26.709 min Scan# 3999

Delta R.T. -0.006 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Instrument :

BNA_M

ClientSampleId :

SLCS539





Tgt Ion:278 Resp: 1564856

Ion Ratio Lower Upper

278 100

139 15.0 12.4 18.6

279 23.9 19.8 29.6



#96

Benzo(g,h,i)perylene

Concen: 30.497 ng/ul

RT: 27.503 min Scan# 4134

Delta R.T. 0.000 min

Lab File: BM037962.D

Acq: 12 Dec 2022 12:06

Tgt Ion:276 Resp: 1469215

Ion Ratio Lower Upper

276 100

138 20.1 17.1 25.7

277 24.4 18.2 27.4

