

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121222\
 Data File : BM037979.D
 Acq On : 12 Dec 2022 23:08
 Operator : CG/JU
 Sample : PB149541BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS541

Quant Time: Dec 13 00:59:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 09 21:34:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.075	152	203368	20.000	ng/u1	0.00
20) Naphthalene-d8	10.892	136	876957	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.704	164	535816	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.445	188	992258	20.000	ng/u1	0.00
79) Chrysene-d12	21.615	240	727536	20.000	ng/u1	0.00
88) Perylene-d12	24.091	264	758422	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.416	96	21463	4.792	ng/uL	0.00
4) Pyridine-d5	3.852	84	341686	25.716	ng/u1	0.00
7) Phenol-d5	7.228	99	449787	27.217	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.398	67	266573	26.491	ng/u1	0.00
11) 2-Chlorophenol-d4	7.598	132	387679	28.597	ng/u1	-0.01
15) 4-Methylphenol-d8	8.787	113	361088	27.992	ng/u1	0.00
21) Nitrobenzene-d5	9.245	128	195561	28.165	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.975	143	209020	28.726	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.510	165	405337	29.328	ng/u1	-0.01
31) 4-Chloroaniline-d4	11.033	131	487696	25.564	ng/u1	0.00
46) Dimethylphthalate-d6	14.110	166	1099544	28.784	ng/u1	-0.01
49) Acenaphthylene-d8	14.398	160	1308315	28.023	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.898	143	150423	23.105	ng/u1	0.00
60) Fluorene-d10	15.692	176	955932	28.088	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.810	200	136172	25.668	ng/u1	0.00
73) Anthracene-d10	17.545	188	1303486	28.147	ng/u1	0.00
81) Pyrene-d10	19.827	212	1267232	28.996	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.927	264	1124554	29.686	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	44398	9.449	ng/uL	94
5) Pyridine	3.869	79	321764	24.097	ng/u1	94
6) Benzaldehyde	7.204	77	218841	35.406	ng/u1	97
8) Phenol	7.257	94	462272	27.891	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.492	93	368489	27.075	ng/u1	99
12) 2-Chlorophenol	7.634	128	409811	29.558	ng/u1	98
13) 2-Methylphenol	8.516	108	360204	28.399	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.610	45	671233	27.313	ng/u1	99
16) Acetophenone	8.904	105	582753	28.724	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.892	70	282285	29.536	ng/u1	98
18) 4-Methylphenol	8.851	108	390008	28.424	ng/u1	97
19) Hexachloroethane	9.163	117	158780	28.646	ng/u1	96
22) Nitrobenzene	9.292	77	433046	28.986	ng/u1	100
23) Isophorone	9.822	82	780293	28.502	ng/u1	99
25) 2-Nitrophenol	10.004	139	226360	29.673	ng/u1	93
26) 2,4-Dimethylphenol	10.057	107	409896	27.786	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.298	93	497554	27.340	ng/u1	97
29) 2,4-Dichlorophenol	10.539	162	403795	29.779	ng/u1	99
30) Naphthalene	10.945	128	1349042	28.506	ng/u1	99
32) 4-Chloroaniline	11.057	127	410638	21.585	ng/u1	99
33) Hexachlorobutadiene	11.222	225	273451	30.655	ng/u1	100
34) Caprolactam	11.851	113	103172	29.661	ng/u1	92
35) 4-Chloro-3-methylphenol	12.169	107	371544	30.102	ng/u1	95
36) 2-Methylnaphthalene	12.545	142	907409	29.562	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121222\
 Data File : BM037979.D
 Acq On : 12 Dec 2022 23:08
 Operator : CG/JU
 Sample : PB149541BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS541

Quant Time: Dec 13 00:59:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 09 21:34:57 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.763	142	916569	29.203	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.904	216	506021	28.903	ng/ul	98
40) Hexachlorocyclopentadiene	12.880	237	132738	13.254	ng/ul	98
41) 2,4,6-Trichlorophenol	13.145	196	308737	28.951	ng/ul	98
42) 2,4,5-Trichlorophenol	13.216	196	328090	29.500	ng/ul	99
43) 1,1'-Biphenyl	13.545	154	1199685	27.855	ng/ul	99
44) 2-Chloronaphthalene	13.586	162	945167	28.447	ng/ul	99
45) 2-Nitroaniline	13.792	65	238931	29.845	ng/ul	99
47) Dimethylphthalate	14.157	163	1114174	29.562	ng/ul	99
48) 2,6-Dinitrotoluene	14.286	165	222264	30.361	ng/ul	99
50) Acenaphthylene	14.427	152	1459017	28.420	ng/ul	99
51) 3-Nitroaniline	14.610	138	209736	28.715	ng/ul	97
52) Acenaphthene	14.768	153	1020884	28.751	ng/ul	99
53) 2,4-Dinitrophenol	14.816	184	79666	22.177	ng/ul	94
55) 4-Nitrophenol	14.910	109	114777	26.734	ng/ul	93
56) Dibenzofuran	15.098	168	1360807	28.070	ng/ul	98
57) 2,4-Dinitrotoluene	15.068	165	294031	29.658	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.321	232	266376	28.473	ng/ul	97
59) Diethylphthalate	15.515	149	1125653	29.981	ng/ul	98
61) Fluorene	15.745	166	1148962	29.062	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.739	204	575840	29.804	ng/ul	97
63) 4-Nitroaniline	15.768	138	204843	30.930	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.827	198	134433	26.176	ng/ul#	98
67) N-Nitrosodiphenylamine	15.951	169	913103	30.507	ng/ul	98
68) 4-Bromophenyl-phenylether	16.627	248	336569	30.943	ng/ul	98
69) Hexachlorobenzene	16.751	284	400961	30.882	ng/ul	98
70) Atrazine	16.898	200	160917	15.732	ng/ul	98
71) Pentachlorophenol	17.092	266	182179	25.237	ng/ul	97
72) Phenanthrene	17.486	178	1574923	29.065	ng/ul	100
74) Anthracene	17.580	178	1595809	28.979	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.510	216	496894	30.976	ng/ul	99
76) Pentachlorobenzene	15.021	250	481683	30.313	ng/ul	98
77) Carbazole	17.845	167	1285385	26.942	ng/ul	99
78) Di-n-butylphthalate	18.398	149	1791223	31.586	ng/ul	100
80) Fluoranthene	19.492	202	1552667	30.103	ng/ul	99
82) Pyrene	19.856	202	1623371	30.242	ng/ul	99
83) Butylbenzylphthalate	20.733	149	587042	29.410	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.527	252	433805	30.133	ng/ul	99
85) Benzo(a)anthracene	21.597	228	1464075	29.891	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.503	149	875499	30.150	ng/ul	99
87) Chrysene	21.650	228	1360934	29.614	ng/ul	99
89) Di-n-octyl phthalate	22.456	149	1399966	28.808	ng/ul	100
90) Benzo(b)fluoranthene	23.333	252	1422592	30.256	ng/ul	99
91) Benzo(k)fluoranthene	23.380	252	1403443	30.798	ng/ul	99
93) Benzo(a)pyrene	23.980	252	1270087	30.718	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.685	276	1412316	26.522	ng/ul	99
95) Dibenzo(a,h)anthracene	26.703	278	1270773	28.307	ng/ul	98
96) Benzo(g,h,i)perylene	27.485	276	1199553	28.214	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121222\
 Data File : BM037979.D
 Acq On : 12 Dec 2022 23:08
 Operator : CG/JU
 Sample : PB149541BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS541

Quant Time: Dec 13 00:59:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 09 21:34:57 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.449 ng/uL
RT: 3.452 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion: 88 Resp: 44398
Ion Ratio Lower Upper
88 100
43 42.6 32.8 49.2
58 87.7 64.6 97.0



#5
Pyridine
Concen: 24.097 ng/uL
RT: 3.869 min Scan# 116
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion: 79 Resp: 321764
Ion Ratio Lower Upper
79 100
52 102.5 78.1 117.1
51 51.1 37.2 55.8









#6
Benzaldehyde
Concen: 35.406 ng/ul
RT: 7.204 min Scan# 684
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion: 77 Resp: 218841
Ion Ratio Lower Upper
77 100
105 100.9 82.5 123.7
106 98.5 81.8 122.6



#8
Phenol
Concen: 27.891 ng/ul
RT: 7.257 min Scan# 692
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion: 94 Resp: 462272
Ion Ratio Lower Upper
94 100
65 30.0 22.2 33.4
66 37.6 27.3 40.9





#10

Bis(2-Chloroethyl)ether

Concen: 27.075 ng/ul

RT: 7.492 min Scan# 732

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :

SLCS541





Tgt Ion: 93 Resp: 368489

Ion Ratio Lower Upper

93 100

63 82.6 66.3 99.5

95 32.8 25.7 38.5



#12

2-Chlorophenol

Concen: 29.558 ng/ul

RT: 7.634 min Scan# 756

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08




Tgt Ion:128 Resp: 409811


Ion Ratio Lower Upper

128 100

64 44.7 35.2 52.8

130 34.3 25.9 38.9





#13

2-Methylphenol

Concen: 28.399 ng/ul

RT: 8.516 min Scan# 906

Delta R.T. -0.012 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :





SLCS541

Tgt Ion:108 Resp: 360204

Ion Ratio Lower Upper

108 100

107 87.2 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 27.313 ng/ul

RT: 8.610 min Scan# 922

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08




Tgt Ion: 45 Resp: 671233


Ion Ratio Lower Upper

45 100

77 10.7 9.0 13.4

79 8.2 6.2 9.4









#16
Acetophenone
Concen: 28.724 ng/ul
RT: 8.904 min Scan# 91
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:105 Resp: 582753
Ion Ratio Lower Upper
105 100
77 74.3 59.7 89.5
51 34.3 27.2 40.8



#17
N-Nitroso-di-n-propylamine
Concen: 29.536 ng/ul
RT: 8.892 min Scan# 970
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion: 70 Resp: 282285
Ion Ratio Lower Upper
70 100
42 79.1 62.3 93.5
101 10.5 8.0 12.0
130 23.4 17.6 26.4









#18
4-Methylphenol
Concen: 28.424 ng/ul
RT: 8.851 min Scan# 90
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:108 Resp: 390008
Ion Ratio Lower Upper
108 100
107 113.0 88.1 132.1



#19
Hexachloroethane
Concen: 28.646 ng/ul
RT: 9.163 min Scan# 1016
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:117 Resp: 158780
Ion Ratio Lower Upper
117 100
201 112.5 87.3 130.9
199 68.5 50.9 76.3









#22
Nitrobenzene
Concen: 28.986 ng/ul
RT: 9.292 min Scan# 1038
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion: 77 Resp: 433046
Ion Ratio Lower Upper
77 100
123 47.2 37.8 56.6
65 14.0 11.1 16.7



#23
Isophorone
Concen: 28.502 ng/ul
RT: 9.822 min Scan# 1128
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion: 82 Resp: 780293
Ion Ratio Lower Upper
82 100
95 7.2 5.2 7.8
138 19.1 14.9 22.3









#25
2-Nitrophenol
Concen: 29.673 ng/ul
RT: 10.004 min Scan# 1159
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:139 Resp: 226360
Ion Ratio Lower Upper
139 100
65 49.9 36.2 54.4
109 28.5 19.7 29.5



#26
2,4-Dimethylphenol
Concen: 27.786 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:107 Resp: 409896
Ion Ratio Lower Upper
107 100
121 51.2 42.0 63.0
122 90.1 71.4 107.2





#27

Bis(2-Chloroethoxy)methane

Concen: 27.340 ng/ul

RT: 10.298 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :

SLCS541





Tgt Ion: 93 Resp: 497554

Ion Ratio Lower Upper

93 100

95 33.5 25.5 38.3

123 11.1 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 29.779 ng/ul

RT: 10.539 min Scan# 1250

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08




Tgt Ion:162 Resp: 403795


Ion Ratio Lower Upper

162 100

164 64.8 51.4 77.2

98 31.8 26.1 39.1









#30
Naphthalene
Concen: 28.506 ng/ul
RT: 10.945 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:128 Resp: 1349042
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.4
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 21.585 ng/ul
RT: 11.057 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:127 Resp: 410638
Ion Ratio Lower Upper
127 100
129 32.2 25.3 37.9









#33
Hexachlorobutadiene
Concen: 30.655 ng/ul
RT: 11.222 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:225 Resp: 273451
Ion Ratio Lower Upper
225 100
223 63.0 50.8 76.2
227 63.3 50.5 75.7



#34
Caprolactam
Concen: 29.661 ng/ul
RT: 11.851 min Scan# 1473
Delta R.T. -0.000 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:113 Resp: 103172
Ion Ratio Lower Upper
113 100
55 198.6 172.2 258.2
56 146.1 120.7 181.1









#35
 4-Chloro-3-methylphenol
 Concen: 30.102 ng/ul
 RT: 12.169 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BM037979.D
 Acq: 12 Dec 2022 23:08




Instrument :
 BNA_M
 ClientSampleId :
 SLCS541


Tgt Ion:107 Resp: 371544
 Ion Ratio Lower Upper
 107 100
 144 27.8 23.4 35.0
 142 83.8 71.0 106.4



#36
 2-Methylnaphthalene
 Concen: 29.562 ng/ul
 RT: 12.545 min Scan# 1591
 Delta R.T. -0.006 min
 Lab File: BM037979.D
 Acq: 12 Dec 2022 23:08

Tgt Ion:142 Resp: 907409
 Ion Ratio Lower Upper
 142 100
 141 90.2 69.4 104.0





#37

1-Methylnaphthalene

Concen: 29.203 ng/ul

RT: 12.763 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :

SLCS541





Tgt Ion:142 Resp: 916569

Ion Ratio Lower Upper

142 100

141 90.7 72.4 108.6

116 3.3 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 28.903 ng/ul

RT: 12.904 min Scan# 1652

Delta R.T. -0.012 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Tgt Ion:216 Resp: 506021




Ion Ratio Lower Upper


216 100

214 78.2 64.2 96.4

179 19.4 15.4 23.2

108 16.4 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 13.254 ng/ul
RT: 12.880 min Scan# 1
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:237 Resp: 132738
Ion Ratio Lower Upper
237 100
235 64.2 50.4 75.6
272 13.9 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 28.951 ng/ul
RT: 13.145 min Scan# 1693
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:196 Resp: 308737
Ion Ratio Lower Upper
196 100
198 95.5 75.0 112.6
200 30.0 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 29.500 ng/ul
RT: 13.216 min Scan# 1705
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:196 Resp: 328090
Ion Ratio Lower Upper
196 100
198 96.6 76.5 114.7
200 31.3 24.9 37.3



#43
1,1'-Biphenyl
Concen: 27.855 ng/ul
RT: 13.545 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:154 Resp: 1199685
Ion Ratio Lower Upper
154 100
153 39.6 31.2 46.8
76 11.6 9.3 13.9









#44
2-Chloronaphthalene
Concen: 28.447 ng/ul
RT: 13.586 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:162 Resp: 945167
Ion Ratio Lower Upper
162 100
164 32.4 27.3 40.9
127 36.8 29.4 44.0



#45
2-Nitroaniline
Concen: 29.845 ng/ul
RT: 13.792 min Scan# 1803
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion: 65 Resp: 238931
Ion Ratio Lower Upper
65 100
92 67.8 53.5 80.3
138 113.6 91.3 136.9









#47
Dimethylphthalate
Concen: 29.562 ng/ul
RT: 14.157 min Scan# 1865
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:163 Resp: 1114174
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.0 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 30.361 ng/ul
RT: 14.286 min Scan# 1887
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:165 Resp: 222264
Ion Ratio Lower Upper
165 100
89 51.7 41.1 61.7
121 20.1 17.4 26.2









#50
Acenaphthylene
Concen: 28.420 ng/ul
RT: 14.427 min Scan# 1911
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:152 Resp: 1459017
Ion Ratio Lower Upper
152 100
151 19.6 15.3 22.9
153 13.2 10.2 15.4



#51
3-Nitroaniline
Concen: 28.715 ng/ul
RT: 14.610 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:138 Resp: 209736
Ion Ratio Lower Upper
138 100
108 9.1 7.9 11.9
92 107.0 83.4 125.0









#52
Acenaphthene
Concen: 28.751 ng/ul
RT: 14.768 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:153 Resp: 1020884
Ion Ratio Lower Upper
153 100
152 47.3 38.2 57.4
154 88.5 71.4 107.0



#53
2,4-Dinitrophenol
Concen: 22.177 ng/ul
RT: 14.816 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:184 Resp: 79666
Ion Ratio Lower Upper
184 100
63 64.5 46.2 69.4
154 59.0 45.5 68.3









#55
4-Nitrophenol
Concen: 26.734 ng/ul
RT: 14.910 min Scan# 1993
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:109 Resp: 114777
Ion Ratio Lower Upper
109 100
139 135.8 119.2 178.8
65 132.2 108.7 163.1



#56
Dibenzofuran
Concen: 28.070 ng/ul
RT: 15.098 min Scan# 2025
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:168 Resp: 1360807
Ion Ratio Lower Upper
168 100
139 36.4 28.1 42.1





#57

2,4-Dinitrotoluene

Concen: 29.658 ng/ul

RT: 15.068 min Scan# 2019

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :

SLCS541





Tgt Ion:165 Resp: 294031

Ion Ratio Lower Upper

165 100

63 43.4 34.9 52.3

182 2.4 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 28.473 ng/ul

RT: 15.321 min Scan# 2063

Delta R.T. -0.012 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Tgt Ion:232 Resp: 266376




Ion Ratio Lower Upper


232 100

131 43.6 32.8 49.2

130 2.0 2.0 3.0

166 29.9 23.2 34.8









#59
Diethylphthalate
Concen: 29.981 ng/ul
RT: 15.515 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:149 Resp: 1125653
Ion Ratio Lower Upper
149 100
177 22.2 18.5 27.7
150 12.4 9.5 14.3



#61
Fluorene
Concen: 29.062 ng/ul
RT: 15.745 min Scan# 2135
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:166 Resp: 1148962
Ion Ratio Lower Upper
166 100
165 95.3 77.2 115.8
167 13.3 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 29.804 ng/ul
RT: 15.739 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:204 Resp: 575840
Ion Ratio Lower Upper
204 100
206 32.2 25.5 38.3
141 56.4 42.8 64.2



#63
4-Nitroaniline
Concen: 30.930 ng/ul
RT: 15.768 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:138 Resp: 204843
Ion Ratio Lower Upper
138 100
92 47.9 36.1 54.1
108 73.8 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 26.176 ng/ul
RT: 15.827 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:198 Resp: 134433
Ion Ratio Lower Upper
198 100
182 3.4 3.8 5.8#
77 23.2 19.3 28.9



#67
N-Nitrosodiphenylamine
Concen: 30.507 ng/ul
RT: 15.951 min Scan# 2170
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:169 Resp: 913103
Ion Ratio Lower Upper
169 100
168 66.1 53.8 80.8
167 37.5 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 30.943 ng/ul
RT: 16.627 min Scan# 21
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:248 Resp: 336569
Ion Ratio Lower Upper
248 100
250 96.6 78.7 118.1
141 68.6 52.7 79.1



#69
Hexachlorobenzene
Concen: 30.882 ng/ul
RT: 16.751 min Scan# 2306
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:284 Resp: 400961
Ion Ratio Lower Upper
284 100
142 30.8 23.1 34.7
249 28.5 23.3 34.9





#70

Atrazine

Concen: 15.732 ng/ul

RT: 16.898 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :

SLCS541





Tgt Ion:200 Resp: 160917

Ion Ratio Lower Upper

200 100

173 26.0 20.5 30.7

215 48.6 40.1 60.1



#71

Pentachlorophenol

Concen: 25.237 ng/ul

RT: 17.092 min Scan# 2364

Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08




Tgt Ion:266 Resp: 182179


Ion Ratio Lower Upper

266 100

264 63.6 50.1 75.1

268 66.1 50.6 75.8









#72
Phenanthrene
Concen: 29.065 ng/ul
RT: 17.486 min Scan# 2431
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:178 Resp: 1574923
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 19.1 15.4 23.0



#74
Anthracene
Concen: 28.979 ng/ul
RT: 17.580 min Scan# 2447
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:178 Resp: 1595809
Ion Ratio Lower Upper
178 100
179 15.1 11.8 17.8
176 18.3 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 30.976 ng/uL
RT: 13.510 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:216 Resp: 496894
Ion Ratio Lower Upper
216 100
214 78.8 62.6 93.8
218 47.7 38.3 57.5



#76
Pentachlorobenzene
Concen: 30.313 ng/uL
RT: 15.021 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:250 Resp: 481683
Ion Ratio Lower Upper
250 100
248 61.3 51.0 76.4
252 63.2 51.8 77.6









#77
 Carbazole
 Concen: 26.942 ng/ul
 RT: 17.845 min Scan# 2492
 Delta R.T. -0.012 min
 Lab File: BM037979.D
 Acq: 12 Dec 2022 23:08




Instrument :
 BNA_M
 ClientSampleId :
 SLCS541


Tgt Ion:167 Resp: 1285385
 Ion Ratio Lower Upper
 167 100
 166 21.1 16.7 25.1
 139 12.6 9.8 14.6



#78
 Di-n-butylphthalate
 Concen: 31.586 ng/ul
 RT: 18.398 min Scan# 2586
 Delta R.T. -0.006 min
 Lab File: BM037979.D
 Acq: 12 Dec 2022 23:08

Tgt Ion:149 Resp: 1791223
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 4.8 3.4 5.2









#80
Fluoranthene
Concen: 30.103 ng/ul
RT: 19.492 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:202 Resp: 1552667
Ion Ratio Lower Upper
202 100
101 10.4 8.6 13.0
100 8.1 6.6 10.0



#82
Pyrene
Concen: 30.242 ng/ul
RT: 19.856 min Scan# 2834
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:202 Resp: 1623371
Ion Ratio Lower Upper
202 100
101 11.8 10.0 15.0
100 9.9 7.9 11.9









#83
Butylbenzylphthalate
Concen: 29.410 ng/ul
RT: 20.733 min Scan# 2984
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:149 Resp: 587042
Ion Ratio Lower Upper
149 100
91 69.1 54.5 81.7
206 19.9 15.8 23.8



#84
3,3'-Dichlorobenzidine
Concen: 30.133 ng/ul
RT: 21.527 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:252 Resp: 433805
Ion Ratio Lower Upper
252 100
254 65.6 53.2 79.8
126 9.0 7.8 11.8









#85
Benzo(a)anthracene
Concen: 29.891 ng/ul
RT: 21.597 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:228 Resp: 1464075
Ion Ratio Lower Upper
228 100
229 19.5 15.5 23.3
226 26.5 21.8 32.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 30.150 ng/ul
RT: 21.503 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:149 Resp: 875499
Ion Ratio Lower Upper
149 100
167 27.5 21.5 32.3
279 4.0 3.2 4.8





#87

Chrysene

Concen: 29.614 ng/ul

RT: 21.650 min Scan# 3140

Delta R.T. -0.012 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :

SLCS541





Tgt Ion:228 Resp: 1360934

Ion Ratio Lower Upper

228 100

226 30.0 24.3 36.5

229 19.3 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 28.808 ng/ul




RT: 22.456 min Scan# 3276


Delta R.T. -0.006 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Tgt Ion:149 Resp: 1399966









#90
Benzo(b)fluoranthene
Concen: 30.256 ng/ul
RT: 23.333 min Scan# 3425
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08




Instrument :
BNA_M
ClientSampleId :
SLCS541


Tgt Ion:252 Resp: 1422592
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 9.3 8.0 12.0



#91
Benzo(k)fluoranthene
Concen: 30.798 ng/ul
RT: 23.380 min Scan# 3433
Delta R.T. -0.012 min
Lab File: BM037979.D
Acq: 12 Dec 2022 23:08

Tgt Ion:252 Resp: 1403443
Ion Ratio Lower Upper
252 100
253 22.4 17.4 26.2
125 9.6 8.1 12.1





#93

Benzo(a)pyrene

Concen: 30.718 ng/ul

RT: 23.980 min Scan# 3

Delta R.T. -0.012 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :

SLCS541





Tgt Ion:252 Resp: 1270087

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 11.2 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 26.522 ng/ul

RT: 26.685 min Scan# 3995

Delta R.T. -0.018 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08




Tgt Ion:276 Resp: 1412316


Ion Ratio Lower Upper

276 100

138 22.7 19.3 28.9

277 24.8 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 28.307 ng/ul

RT: 26.703 min Scan# 3999

Delta R.T. -0.012 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Instrument :

BNA_M

ClientSampleId :

SLCS541





Tgt Ion:278 Resp: 1270773

Ion Ratio Lower Upper

278 100

139 14.0 12.4 18.6

279 23.9 19.8 29.6



#96

Benzo(g,h,i)perylene

Concen: 28.214 ng/ul

RT: 27.485 min Scan# 4131

Delta R.T. -0.018 min

Lab File: BM037979.D

Acq: 12 Dec 2022 23:08

Tgt Ion:276 Resp: 1199553

Ion Ratio Lower Upper

276 100

138 19.2 17.1 25.7

277 24.9 18.2 27.4

