


Data Path : Z:\HPCHEM1\BNA M\Data\BM121315\
Data File : BM003501.D
Acq On : 13 Dec 2015 09:05
Operator : UM/IZ
Sample : DFTPP56
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP56

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM120115.M
Title : SVOA CALIBRATION
Last Update : Tue Dec 15 00:18:50 2015



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2455

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.2	13497	PASS
68	69	0.00	2	0.8	129	PASS
69	198	0.00	100	33.5	16605	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.7	23629	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	49581	PASS
199	198	5	9	6.7	3339	PASS
275	198	10	60	24.6	12217	PASS
365	198	1	100	2.9	1457	PASS
441	443	0.01	100	76.5	5439	PASS
442	198	50	100	73.0	36213	PASS
443	442	15	24	19.6	7114	PASS

Scan 2455 (17.527 min): BM003501.D
DFTPP56

m/z	Abundance
73.00	872.0
147.00	414.0
207.00	672.0

Instrument :

BNA_M

ClientSampleId :

DFTPP56

m/z	Abundance
39.00	977.0
50.10	3240.0
51.10	13213.0
52.00	725.0
56.00	334.0
57.00	1063.0
63.00	628.0
65.00	348.0
67.90	338.0
69.00	15482.0
73.00	527.0
74.00	1716.0
75.00	2855.0
76.10	939.0
77.10	19280.0
78.00	1441.0
79.00	1410.0
80.00	949.0
81.00	1471.0
82.00	385.0
85.90	425.0
91.00	433.0
92.00	405.0
93.00	2165.0
98.00	1617.0
99.00	1786.0
101.00	732.0
104.00	466.0
105.00	536.0
107.00	5848.0
108.00	1004.0
110.00	11003.0
111.00	1824.0
116.00	302.0
117.00	4950.0
118.00	363.0
122.00	409.0
123.00	567.0
125.00	320.0
127.00	21168.0
128.00	1730.0
129.00	8183.0
130.00	718.0
135.00	657.0
137.00	431.0
141.00	982.0
142.00	427.0
147.00	674.0
148.00	1218.0
149.00	392.0
153.00	377.0
155.00	568.0
156.00	873.0
160.00	346.0
161.00	485.0
164.90	394.0
166.10	346.0
167.00	2017.0
168.00	1936.0
169.00	316.0
173.90	406.0
175.00	807.0
177.00	390.0
179.00	1483.0
180.00	1244.0
181.00	582.0
185.00	795.0
186.10	5620.0
187.00	1683.0
188.00	314.0
189.00	342.0
192.00	562.0
193.00	671.0
196.00	1217.0
198.00	41672.0
199.00	2920.0
204.00	1530.0
205.00	2533.0
206.10	10550.0
207.10	1883.0
208.00	420.0
210.90	423.0
216.00	326.0
217.00	2957.0
218.00	443.0
221.00	1478.0
221.80	431.0
223.00	600.0
224.10	5997.0
225.10	1446.0
227.00	2344.0
228.00	352.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

228.90	532.0
237.00	343.0
244.10	4268.0
245.10	557.0
246.00	873.0
255.10	22144.0
256.00	3078.0
257.00	370.0
258.00	1332.0
265.00	448.0
273.10	701.0
274.00	1759.0
275.10	9288.0
276.10	1250.0
277.00	813.0
296.10	2489.0
297.10	395.0
323.10	807.0
334.00	384.0
365.10	1014.0
372.10	414.0
423.10	1120.0
424.10	590.0
441.10	3367.0
442.10	23384.0
443.10	4792.0
444.00	456.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
39.00	1146.0
50.10	3979.0
51.10	16432.0
52.00	838.0
56.00	544.0
57.00	1249.0
63.00	840.0
65.00	356.0
68.10	388.0
69.00	19752.0
73.00	510.0
74.00	2237.0
75.00	3444.0
76.00	1197.0
77.10	24120.0
78.10	1624.0
79.00	1621.0
80.00	1227.0
81.00	1739.0
82.00	489.0
82.90	416.0
85.00	350.0
86.00	454.0
91.00	451.0
92.00	602.0
93.00	2695.0
98.00	2178.0
99.00	2091.0
101.00	972.0
103.00	324.0
103.90	646.0
105.00	718.0
107.00	7867.0
108.00	1240.0
110.00	13717.0
111.00	2229.0
116.00	415.0
117.00	6510.0
118.00	524.0
122.00	506.0
123.00	719.0
124.00	360.0
127.00	27728.0
128.00	2156.0
129.00	10662.0
130.00	999.0
133.90	326.0
134.90	875.0
136.00	416.0
137.00	540.0
141.00	1460.0
142.00	377.0
143.00	347.0
147.00	786.0
148.00	1805.0
149.10	430.0
153.00	461.0
155.00	755.0
156.00	1225.0
160.00	403.0
161.00	688.0
164.90	617.0
166.00	457.0
167.10	2951.0
168.00	2294.0
169.00	334.0
173.00	345.0
174.00	648.0
175.00	1171.0
175.90	314.0
177.00	558.0
179.00	2025.0
180.00	1652.0
181.00	734.0
185.00	1044.0
186.00	7874.0
187.00	2359.0
188.00	305.0
189.00	543.0
192.00	660.0
193.10	814.0
196.00	1780.0
198.00	58608.0
199.00	3940.0
200.00	309.0
202.90	418.0
204.10	2256.0
205.10	3672.0
206.10	15788.0
207.10	2667.0
208.00	658.0
211.00	653.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

216.00	398.0
217.00	4023.0
218.10	504.0
221.10	2252.0
221.80	571.0
223.00	934.0
224.10	8629.0
225.10	2158.0
227.10	3695.0
228.00	641.0
229.00	791.0
237.00	455.0
242.00	504.0
243.10	413.0
244.10	6512.0
245.00	822.0
246.00	1277.0
255.10	32544.0
256.10	4570.0
257.00	379.0
258.00	1981.0
265.00	828.0
273.00	933.0
274.10	2514.0
275.10	14428.0
276.10	1817.0
277.00	1223.0
296.10	3979.0
297.10	542.0
303.10	430.0
315.00	455.0
323.10	1184.0
327.00	312.0
334.00	770.0
352.10	336.0
354.10	326.0
365.00	1760.0
372.00	598.0
423.10	2156.0
424.10	711.0
441.10	6258.0
442.10	41904.0
443.10	8291.0
444.10	801.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
39.00	814.0
50.00	2657.0
51.10	10847.0
52.00	707.0
56.00	344.0
57.00	936.0
63.00	586.0
69.00	14582.0
73.10	489.0
74.10	1524.0
75.00	2320.0
76.00	869.0
77.10	17744.0
78.00	1283.0
79.00	1169.0
80.00	1057.0
81.00	1310.0
82.00	337.0
83.00	322.0
85.00	321.0
85.90	438.0
91.00	364.0
92.00	373.0
93.00	2077.0
98.00	1659.0
99.00	1409.0
101.00	857.0
104.00	389.0
105.00	577.0
107.00	6146.0
108.00	977.0
110.00	10962.0
111.00	1683.0
116.00	374.0
117.00	5139.0
118.00	413.0
122.10	453.0
123.00	553.0
127.00	21992.0
128.00	1669.0
129.00	8268.0
130.00	598.0
135.00	687.0
136.00	300.0
137.00	373.0
141.00	1127.0
142.00	389.0
147.00	720.0
148.00	1404.0
149.00	324.0
153.00	397.0
155.00	681.0
156.00	973.0
159.90	363.0
161.00	554.0
165.00	501.0
166.00	359.0
167.00	2443.0
168.00	1519.0
169.00	302.0
174.00	492.0
175.00	956.0
176.00	344.0
177.00	412.0
179.00	1824.0
180.00	1437.0
181.00	567.0
185.00	935.0
186.00	6678.0
187.00	1926.0
189.00	515.0
192.00	581.0
193.00	681.0
196.10	1389.0
198.00	48464.0
199.00	3157.0
203.00	391.0
204.10	1829.0
205.00	3115.0
206.10	12523.0
207.10	2091.0
208.00	553.0
211.00	531.0
217.00	3484.0
217.90	480.0
221.10	2137.0
221.90	441.0
223.00	760.0
224.10	7381.0
225.00	1793.0
227.00	3016.0
228.10	433.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

229.00	524.0
237.00	367.0
242.10	409.0
243.00	394.0
244.10	5571.0
245.10	700.0
246.00	1132.0
255.10	28752.0
256.10	4076.0
257.00	314.0
258.00	1579.0
265.00	604.0
273.00	889.0
274.00	2262.0
275.10	12936.0
276.10	1524.0
277.00	1108.0
296.10	3563.0
297.00	471.0
303.10	410.0
315.00	450.0
323.10	1124.0
324.10	322.0
334.10	695.0
354.10	308.0
365.10	1598.0
372.10	584.0
403.10	337.0
421.00	342.0
423.10	2090.0
424.10	843.0
441.10	6692.0
442.10	43352.0
443.10	8259.0
444.10	868.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56