

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM121317\
 Data File : BM013401.D
 Acq On : 14 Dec 2017 07:27
 Operator : SJ/JU
 Sample : I6775-08DL2 30X
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F9A11DL2

Quant Time: Dec 14 10:44:30 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 14 06:00:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	149914	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	626934	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	388421	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.07	188	974859	20.00	ng/ul	0.00
75) Chrysene-d12	21.26	240	1097617	20.00	ng/ul	0.00
83) Perylene-d12	23.49	264	987609	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.21	96	126	0.04	ng/uL	-0.01
5) Phenol-d5	6.87	99	13700	1.06	ng/ul	0.01
7) Bis-(2-Chloroethyl)ether-d	7.02	67	8843	1.20	ng/ul	0.00
9) 2-Chlorophenol-d4	7.22	132	11006	1.09	ng/ul	0.00
13) 4-Methylphenol-d8	8.39	113	11380	1.03	ng/ul	0.00
19) Nitrobenzene-d5	8.84	128	5792	1.16	ng/ul	0.00
22) 2-Nitrophenol-d4	9.56	143	5903	1.11	ng/ul	0.01
26) 2,4-Dichlorophenol-d3	10.10	165	11304	1.03	ng/ul	0.01
29) 4-Chloroaniline-d4	10.62	131	4768	0.47	ng/ul	0.01
43) Dimethylphthalate-d6	13.74	166	43915	1.27	ng/ul	0.00
46) Acenaphthylene-d8	14.02	160	52978	1.24	ng/ul	0.00
51) 4-Nitrophenol-d4	14.56	143	3060	0.51	ng/ul	0.02
57) Fluorene-d10	15.32	176	40158	1.35	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.45	200	1785	0.29	ng/ul	0.00
70) Anthracene-d10	17.17	188	66830	1.37	ng/ul	0.00
76) Pyrene-d10	19.47	212	78310	1.43	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.34	264	62589	1.24	ng/ul	0.00

Target Compounds

					Ovalue	
28) Naphthalene	10.51	128	930229	28.840	ng/ul	99
49) Acenaphthene	14.39	153	136300	5.069	ng/ul	97
53) Dibenzofuran	14.72	168	91460	2.304	ng/ul	94
58) Fluorene	15.37	166	103718	3.158	ng/ul	96
69) Phenanthrene	17.12	178	433539	7.828	ng/ul	98
72) Carbazole	17.48	167	226476	4.651	ng/ul	96
74) Fluoranthene	19.13	202	73346	1.081	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA M\DATA\BM121317\
Data File : BM013401.D
Acq On : 14 Dec 2017 07:27
Operator : SJ/JU
Sample : I6775-08DL2 30X
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
F9A11DL2

Quant Time: Dec 14 10:44:30 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM113017.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 14 06:00:20 2017
Response via : Initial Calibration

