


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM121416\
 Data File : BM008286.D
 Acq On : 13 Dec 2016 23:58
 Operator : UM/SJ
 Sample : DFTPP35
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP35

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-2016\SOM02.2-EPA-BM080316.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 12 01:07:13 2016



AutoFind: Scans 2535, 2536, 2537; Background Corrected with Scan 2529

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.7	70242	PASS
68	69	0.00	2	1.8	1384	PASS
69	198	0.00	100	34.2	78189	PASS
70	69	0.00	2	0.7	586	PASS
127	198	10	80	40.4	92426	PASS
197	198	0.00	2	0.6	1365	PASS
198	198	100	100	100.0	228778	PASS
199	198	5	9	6.4	14609	PASS
275	198	10	60	31.3	71634	PASS
365	198	1	100	4.7	10866	PASS
441	443	0.01	100	76.7	31600	PASS
442	198	50	100	93.1	213056	PASS
443	442	15	24	19.3	41205	PASS

m/z	Abundance
37.10	375.0
38.00	1176.0
39.10	5693.0
40.00	2064.0
44.00	1138.0
49.20	747.0
50.10	22600.0
51.10	86984.0
52.10	4097.0
55.10	520.0
56.00	2602.0
57.00	5297.0
61.00	1582.0
62.10	1557.0
63.10	3946.0
64.00	626.0
65.10	1349.0
66.10	318.0
68.10	1985.0
69.00	94184.0
70.10	646.0
73.00	1390.0
74.10	10524.0
75.10	16536.0
76.10	5606.0
77.10	106792.0
78.10	7653.0
79.10	8047.0
80.00	5734.0
81.00	7872.0
82.00	1944.0
83.00	1584.0
84.90	1380.0
86.00	2711.0
87.00	1300.0
88.00	320.0
91.00	2062.0
92.10	1819.0
93.00	13499.0
94.00	925.0
96.00	849.0
97.20	426.0
98.10	11229.0
99.00	8817.0
100.00	790.0
101.00	4382.0
103.10	1412.0
104.00	2672.0
105.00	2619.0
106.00	983.0
107.00	33504.0
108.10	5657.0
109.00	893.0
110.00	61576.0
111.00	9144.0
112.00	1180.0
112.90	394.0
116.10	2225.0
117.00	33344.0
118.00	2634.0
119.10	423.0
120.00	412.0
122.00	2020.0
123.00	3919.0
124.00	1708.0
125.00	1374.0
125.90	370.0
126.20	305.0
127.00	105016.0
128.10	8496.0
129.00	45880.0
130.00	3748.0
131.10	866.0
131.90	367.0
132.90	319.0
134.00	1475.0
135.00	3637.0
136.00	1738.0
137.10	2474.0
138.00	555.0
139.00	401.0
140.10	540.0
141.00	6030.0
142.00	2055.0
142.90	1356.0
144.10	401.0
145.00	363.0
146.00	997.0
147.00	3878.0
148.00	7994.0
149.00	1939.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

150.00	480.0
151.10	858.0
152.10	611.0
153.00	2004.0
154.00	1411.0
155.00	3987.0
156.00	5223.0
157.20	1140.0
158.00	1161.0
159.00	912.0
160.00	1717.0
161.00	3045.0
161.90	1163.0
164.00	320.0
165.00	2648.0
166.00	2325.0
167.10	13484.0
168.00	7368.0
169.00	1237.0
170.00	681.0
170.80	634.0
171.90	822.0
173.00	1853.0
174.00	2924.0
175.00	5253.0
175.90	1073.0
177.00	2759.0
178.00	1053.0
179.00	10250.0
180.00	6980.0
181.00	2570.0
182.00	618.0
182.70	335.0
183.90	807.0
185.00	4943.0
186.00	36480.0
187.00	10135.0
188.00	1317.0
189.00	2647.0
191.00	1535.0
192.00	3474.0
193.10	3561.0
194.10	1062.0
195.10	384.0
196.00	6306.0
197.10	1634.0
198.00	245568.0
199.00	15762.0
200.00	1639.0
201.50	1126.0
203.00	1745.0
204.00	10631.0
205.10	14758.0
206.10	68664.0
207.00	10335.0
208.00	2570.0
209.00	1077.0
210.30	956.0
211.10	2300.0
211.70	451.0
212.10	449.0
215.00	733.0
216.00	1796.0
217.00	17696.0
217.90	2599.0
221.00	7604.0
221.80	2506.0
223.00	4214.0
224.10	38872.0
225.00	10301.0
226.00	1225.0
227.00	17608.0
228.10	2352.0
229.00	3970.0
229.90	503.0
231.00	1087.0
232.10	310.0
232.90	659.0
234.10	1467.0
235.10	1131.0
236.00	817.0
237.00	1898.0
238.00	305.0
239.00	625.0
239.90	574.0
240.90	1013.0
242.00	1802.0
243.10	2612.0
244.10	30408.0
245.10	4353.0
246.00	7212.0
246.90	1182.0
248.10	396.0
249.00	1123.0
252.90	567.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

254.10	1302.0
255.00	164096.0
256.00	24640.0
257.00	1811.0
258.00	9967.0
259.00	1631.0
263.90	419.0
265.00	4161.0
266.00	815.0
267.00	328.0
270.90	464.0
271.90	359.0
273.00	4054.0
274.00	12683.0
275.00	72248.0
276.10	9615.0
277.00	6086.0
278.00	1337.0
281.00	622.0
282.10	330.0
283.10	628.0
284.00	740.0
285.10	822.0
293.00	1553.0
293.90	433.0
296.00	21616.0
297.00	3120.0
300.90	333.0
302.10	383.0
303.00	2402.0
304.00	616.0
310.10	433.0
314.00	980.0
315.00	2581.0
316.00	953.0
321.00	584.0
322.00	493.0
323.10	6787.0
324.00	1252.0
327.00	1182.0
328.00	604.0
332.00	557.0
333.00	558.0
334.00	4247.0
335.10	877.0
341.00	1125.0
346.00	1322.0
352.00	1906.0
353.00	1700.0
354.00	2342.0
354.90	794.0
365.00	10569.0
365.90	1365.0
370.90	497.0
372.10	3594.0
373.00	784.0
383.10	950.0
390.00	316.0
402.00	1094.0
403.00	1854.0
404.00	575.0
421.00	1451.0
422.10	1412.0
423.00	9121.0
424.10	2378.0
425.10	616.0
441.10	26496.0
442.10	176320.0
443.10	34920.0
444.00	2957.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
36.90	373.0
38.20	925.0
39.10	5066.0
40.00	2151.0
44.00	1542.0
49.00	699.0
50.10	20680.0
51.10	82752.0
52.00	3833.0
54.90	535.0
56.00	2244.0
57.00	5524.0
60.00	345.0
61.10	1057.0
62.00	1249.0
63.10	3588.0
64.00	526.0
65.00	1986.0
68.10	1423.0
69.00	90864.0
70.00	644.0
73.10	1301.0
74.00	9831.0
75.10	15568.0
76.10	5385.0
77.10	102968.0
78.10	7583.0
79.00	8386.0
80.00	6003.0
81.00	7827.0
82.00	2535.0
83.10	1832.0
85.00	1121.0
86.00	2550.0
87.10	1234.0
88.10	363.0
91.00	1815.0
92.00	1666.0
93.00	13811.0
94.00	801.0
96.00	444.0
98.00	10949.0
99.00	8903.0
100.00	801.0
101.00	4559.0
102.00	375.0
103.00	1542.0
104.00	2919.0
105.00	2725.0
106.10	944.0
107.00	35712.0
108.10	5455.0
109.10	1057.0
110.00	58848.0
111.00	9102.0
112.00	1152.0
113.10	372.0
116.00	2053.0
117.00	34512.0
118.00	2379.0
120.10	396.0
122.00	2834.0
123.00	3362.0
124.00	1753.0
125.00	1405.0
127.00	108512.0
128.00	8471.0
129.00	44800.0
130.00	3699.0
131.10	792.0
133.90	1335.0
135.10	3609.0
136.00	1290.0
137.00	2201.0
137.80	629.0
140.00	370.0
141.00	6963.0
142.00	1988.0
143.10	1815.0
143.90	371.0
145.00	428.0
146.00	881.0
147.00	4231.0
148.00	7900.0
149.00	1898.0
150.00	402.0
151.00	760.0
152.00	503.0
153.00	2333.0
154.00	1676.0
155.00	4218.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

156.10	5663.0
157.10	1083.0
157.90	1202.0
159.00	1015.0
160.00	2110.0
161.10	3447.0
162.00	636.0
163.90	451.0
165.00	2414.0
166.00	2254.0
167.00	14117.0
168.00	7495.0
169.00	1182.0
170.00	560.0
171.00	552.0
172.00	1527.0
173.00	1623.0
174.00	2733.0
175.00	5631.0
176.00	1534.0
177.00	3052.0
178.00	1106.0
179.00	10201.0
180.00	7597.0
181.00	3655.0
182.10	737.0
184.00	1052.0
185.10	4976.0
186.00	41232.0
187.00	12980.0
188.00	1228.0
189.00	2998.0
190.00	491.0
191.00	1581.0
192.00	3810.0
193.00	3959.0
194.00	977.0
195.10	502.0
196.00	7733.0
197.10	1571.0
198.00	274112.0
199.00	17344.0
200.10	1509.0
201.60	1073.0
202.10	373.0
203.00	2462.0
204.10	11193.0
205.10	18752.0
206.10	78032.0
207.10	12010.0
208.00	2998.0
209.00	1029.0
210.00	1204.0
211.10	3018.0
214.90	998.0
216.00	1830.0
217.00	22240.0
218.10	2897.0
221.00	11183.0
222.00	3164.0
223.00	4206.0
224.00	44816.0
225.00	11408.0
226.10	1150.0
227.00	20736.0
228.00	3091.0
229.00	4216.0
230.00	630.0
231.00	1500.0
232.00	335.0
233.00	521.0
234.00	1185.0
235.00	1446.0
236.00	955.0
237.00	1703.0
238.90	622.0
239.90	680.0
241.00	1269.0
242.00	2410.0
243.10	2695.0
244.10	33560.0
245.10	4638.0
246.00	7961.0
247.00	1501.0
249.00	1328.0
249.80	356.0
250.90	353.0
251.90	358.0
252.90	932.0
254.00	1335.0
255.00	188736.0
256.00	27192.0
257.00	2490.0
258.00	12878.0
259.00	1721.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

261.00	345.0
265.00	4451.0
266.00	959.0
271.00	458.0
272.10	493.0
273.00	5653.0
274.00	15997.0
275.00	85216.0
276.10	11402.0
277.00	8906.0
278.00	1470.0
278.90	340.0
281.00	949.0
282.00	381.0
283.00	993.0
284.10	583.0
285.00	1208.0
290.10	466.0
292.10	448.0
293.00	1686.0
293.90	540.0
295.10	685.0
296.00	27240.0
297.00	3947.0
302.20	463.0
303.10	3113.0
304.10	722.0
308.00	420.0
313.00	330.0
314.00	1159.0
315.00	3637.0
316.00	1721.0
321.00	785.0
322.00	489.0
323.00	8502.0
324.00	1753.0
326.90	1471.0
328.00	945.0
331.90	449.0
333.00	899.0
334.00	4999.0
335.00	1714.0
341.00	1220.0
346.00	1406.0
347.00	312.0
352.00	2465.0
353.10	1614.0
354.10	2559.0
355.10	501.0
365.00	12331.0
366.10	1495.0
371.00	689.0
372.10	5218.0
373.00	811.0
383.00	818.0
389.90	349.0
390.20	367.0
390.90	308.0
392.00	380.0
401.10	391.0
402.10	2076.0
403.00	2512.0
404.00	784.0
421.00	2191.0
422.10	2156.0
423.10	13053.0
424.10	3242.0
425.00	866.0
441.10	38312.0
442.10	256512.0
443.10	49680.0
444.00	4286.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
38.10	513.0
39.10	3158.0
40.10	1902.0
44.00	1093.0
49.00	355.0
50.10	11780.0
51.10	40992.0
52.10	2008.0
55.00	353.0
56.00	1132.0
57.00	2597.0
61.00	521.0
62.00	964.0
63.00	1884.0
65.10	973.0
68.10	746.0
69.00	49520.0
70.10	469.0
73.10	1048.0
74.10	5574.0
75.00	7987.0
76.00	2463.0
77.10	56120.0
78.10	4030.0
79.00	4130.0
80.00	2960.0
81.00	4398.0
82.00	1329.0
83.10	819.0
84.80	631.0
86.00	1210.0
87.00	638.0
91.00	1101.0
92.10	1040.0
93.00	7978.0
94.00	511.0
96.00	343.0
98.00	6660.0
99.00	4494.0
100.10	397.0
101.00	2518.0
103.00	800.0
104.00	1382.0
105.10	1337.0
106.00	529.0
107.00	18320.0
108.10	2915.0
109.20	531.0
110.00	32688.0
111.10	4898.0
112.00	690.0
116.00	1266.0
117.00	18872.0
118.10	1131.0
120.00	319.0
122.00	1667.0
123.00	1820.0
124.00	847.0
124.90	916.0
127.00	63752.0
128.00	4686.0
129.00	25824.0
129.90	2311.0
131.10	386.0
133.90	772.0
135.10	2040.0
136.00	711.0
137.00	1183.0
137.90	328.0
140.00	491.0
141.00	3679.0
141.90	1187.0
143.00	550.0
145.10	388.0
146.10	653.0
147.00	1985.0
148.00	4889.0
149.00	1000.0
150.00	349.0
151.10	490.0
153.00	1321.0
154.00	869.0
155.00	2122.0
156.00	3143.0
157.00	609.0
158.10	489.0
158.90	407.0
159.90	1020.0
161.00	1809.0
165.00	1447.0
166.00	1366.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

167.00	8421.0
168.10	5251.0
169.10	907.0
170.00	327.0
171.00	525.0
172.00	863.0
173.00	1100.0
174.10	1824.0
175.10	3319.0
176.00	725.0
177.00	1499.0
178.00	443.0
179.00	6808.0
180.00	4734.0
181.00	2463.0
184.00	542.0
185.00	3356.0
186.00	24552.0
187.00	7278.0
188.00	700.0
189.00	1590.0
191.00	1024.0
192.00	2368.0
193.00	2881.0
194.00	889.0
196.00	4301.0
197.10	892.0
198.00	166656.0
199.00	10722.0
200.10	794.0
201.50	641.0
203.00	1467.0
204.10	6655.0
205.10	11382.0
206.10	47592.0
207.10	7092.0
208.00	1721.0
209.00	725.0
210.10	949.0
211.00	1706.0
212.00	302.0
215.00	713.0
215.90	1291.0
217.00	14693.0
218.00	2146.0
220.30	448.0
221.00	6323.0
222.00	1574.0
223.00	3023.0
224.10	28880.0
225.10	7011.0
226.00	695.0
227.00	14730.0
228.00	1557.0
229.00	2975.0
231.00	1014.0
234.00	999.0
235.00	818.0
236.00	957.0
237.00	1010.0
239.10	483.0
239.90	430.0
241.00	637.0
242.00	1882.0
243.00	1496.0
244.00	21720.0
245.00	2656.0
246.00	5153.0
247.00	1059.0
249.00	613.0
250.90	363.0
252.10	327.0
253.00	502.0
254.00	975.0
255.00	121248.0
256.00	18072.0
257.00	1545.0
258.00	8108.0
259.00	805.0
265.00	3689.0
266.00	458.0
270.90	370.0
273.00	3894.0
274.00	9501.0
275.00	57440.0
276.00	6433.0
277.00	4462.0
278.00	1055.0
281.10	665.0
282.00	357.0
283.00	591.0
283.90	532.0
285.10	973.0
288.80	308.0
293.00	1298.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

293.90	539.0
295.10	373.0
296.00	18032.0
296.90	2588.0
302.20	305.0
303.10	2069.0
304.00	577.0
307.80	305.0
313.90	577.0
315.00	2325.0
316.00	1252.0
321.00	887.0
323.10	6415.0
324.00	992.0
327.00	1181.0
328.10	456.0
332.80	462.0
334.10	3578.0
335.00	1556.0
341.10	888.0
345.90	1185.0
346.80	308.0
352.10	1769.0
353.10	1356.0
354.00	2039.0
355.00	625.0
365.00	9700.0
366.00	1343.0
371.00	474.0
372.10	2926.0
373.00	708.0
383.00	910.0
390.00	384.0
391.00	326.0
402.00	1017.0
403.00	1595.0
404.10	726.0
421.00	1707.0
422.10	1424.0
423.00	10812.0
424.10	2770.0
441.10	29992.0
442.10	206336.0
443.10	39016.0
444.10	3278.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

Scan 2529 (17.462 min): BM008286.D
DFTPP35

m/z	Abundance
40.00	1987.0
44.10	1277.0
73.10	702.0
207.10	1057.0
208.00	330.0
209.00	477.0
281.10	391.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35