


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM121420\
 Data File : BM028110.D
 Acq On : 14 Dec 2020 10:14
 Operator : JU/CG
 Sample : DFTPP08
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP080

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121120.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Dec 12 03:21:54 2020



AutoFind: Scans 2512, 2513, 2514; Background Corrected with Scan 2506

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.8	91000	PASS
68	69	0.00	2	2.0	1674	PASS
69	198	0.00	100	47.7	83842	PASS
70	69	0.00	2	0.5	446	PASS
127	198	10	80	54.7	96160	PASS
197	198	0.00	2	0.8	1339	PASS
198	198	100	100	100.0	175786	PASS
199	198	5	9	7.0	12389	PASS
275	198	10	60	27.7	48672	PASS
365	198	1	100	3.7	6513	PASS
441	198	0.01	100	13.0	22808	PASS
442	198	50	100	85.8	150781	PASS
443	442	15	24	19.2	28949	PASS

Scan 2506 (17.927 min): BM028110.D
DFTPP08

m/z	Abundance
40.00	420.0
44.00	2150.0
48.80	178.0
73.10	460.0
95.90	254.0
132.80	150.0
147.00	194.0
190.90	235.0
207.00	1761.0
207.90	281.0
221.10	206.0
281.10	312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

m/z	Abundance
37.00	409.0
38.00	1007.0
39.10	6769.0
40.00	820.0
41.10	254.0
44.00	2152.0
44.90	228.0
49.00	913.0
50.10	22632.0
51.10	83584.0
52.00	4473.0
52.90	228.0
55.10	480.0
56.00	2659.0
57.00	5001.0
57.90	164.0
61.00	1273.0
62.00	1143.0
63.00	3370.0
64.00	364.0
65.00	1376.0
66.00	171.0
68.10	1475.0
69.00	74328.0
70.00	409.0
73.10	1288.0
74.00	7336.0
75.00	11951.0
76.00	4285.0
77.10	79896.0
78.10	5291.0
79.00	5457.0
80.00	4939.0
81.00	5764.0
82.00	1291.0
83.00	1390.0
84.00	173.0
85.00	886.0
86.00	2026.0
87.00	709.0
87.90	361.0
88.90	227.0
91.00	1206.0
92.00	1796.0
93.00	10330.0
94.00	705.0
95.00	169.0
96.00	500.0
97.10	158.0
98.00	7625.0
99.00	7652.0
99.90	606.0
101.00	2856.0
101.90	229.0
103.10	1020.0
104.00	1809.0
105.00	2223.0
106.00	606.0
107.00	23960.0
108.00	3382.0
109.10	695.0
110.00	41552.0
111.00	7113.0
112.00	846.0
112.90	318.0
116.00	1198.0
117.00	22800.0
118.00	1886.0
118.90	199.0
119.90	294.0
122.00	1752.0
123.00	2759.0
123.90	1250.0
125.00	1045.0
126.20	744.0
127.00	79952.0
128.00	6923.0
129.00	30864.0
130.00	2467.0
131.00	513.0
132.10	269.0
132.90	169.0
134.00	840.0
135.00	2908.0
136.00	1039.0
137.00	1589.0
137.80	236.0
139.00	181.0
140.00	371.0
141.00	4227.0
142.00	1161.0
142.90	1024.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

143.80	199.0
145.00	194.0
146.00	652.0
147.00	2324.0
148.00	5299.0
149.00	1410.0
150.00	322.0
151.10	524.0
152.00	353.0
153.00	1230.0
153.90	882.0
155.00	2191.0
156.00	3379.0
157.00	373.0
158.00	459.0
158.90	592.0
159.90	1154.0
161.00	1984.0
161.90	544.0
165.00	1405.0
165.90	1142.0
167.00	8567.0
168.00	6597.0
169.00	983.0
170.00	247.0
170.90	305.0
172.00	700.0
172.90	876.0
174.00	1727.0
175.00	2975.0
175.90	925.0
176.90	1435.0
177.90	596.0
179.00	5584.0
180.00	4699.0
181.00	1669.0
182.00	389.0
183.10	190.0
184.00	353.0
185.00	2872.0
186.00	19728.0
187.00	5738.0
188.00	768.0
189.00	1102.0
189.90	294.0
191.00	1025.0
191.90	1907.0
193.00	2363.0
194.10	422.0
194.80	198.0
196.00	4608.0
197.10	1058.0
197.90	136384.0
199.00	9342.0
199.90	798.0
201.50	766.0
203.00	981.0
204.00	5849.0
205.00	8788.0
206.00	36192.0
207.10	6190.0
208.00	1458.0
209.00	806.0
210.00	691.0
211.00	1383.0
214.90	429.0
216.00	826.0
217.00	10509.0
218.00	1210.0
221.00	6677.0
221.90	1272.0
223.00	2065.0
224.00	20552.0
225.00	5238.0
226.10	776.0
227.00	8105.0
228.00	1259.0
229.00	1651.0
230.00	217.0
231.00	897.0
234.00	566.0
235.00	667.0
235.90	347.0
237.00	1504.0
237.90	252.0
239.00	394.0
240.00	315.0
240.90	476.0
242.10	1024.0
243.00	1178.0
244.00	14935.0
245.00	2251.0
246.00	2835.0
247.00	567.0
247.90	243.0
249.00	399.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

249.90	260.0
251.00	233.0
251.90	154.0
253.10	430.0
253.90	777.0
255.00	78704.0
256.00	13130.0
257.00	1201.0
258.00	4583.0
259.00	677.0
264.90	2078.0
265.90	336.0
271.10	200.0
272.00	271.0
273.00	2399.0
274.00	5906.0
275.00	33776.0
276.00	4235.0
277.00	3051.0
278.00	620.0
281.00	529.0
283.00	355.0
284.00	242.0
285.00	508.0
293.00	622.0
294.00	402.0
294.90	316.0
296.00	9112.0
297.00	1328.0
302.00	223.0
303.00	1017.0
303.90	308.0
304.90	176.0
309.80	169.0
314.00	309.0
314.90	882.0
316.00	592.0
321.00	366.0
322.00	191.0
323.00	2739.0
324.10	526.0
327.00	584.0
328.00	343.0
331.90	174.0
332.90	271.0
334.00	1892.0
335.00	474.0
340.80	374.0
346.00	634.0
347.00	321.0
352.00	913.0
352.90	668.0
354.10	782.0
355.00	240.0
365.00	4297.0
365.90	569.0
370.90	222.0
372.00	1533.0
373.00	458.0
383.00	487.0
383.90	170.0
389.90	188.0
400.90	165.0
401.90	656.0
402.90	711.0
403.90	281.0
421.00	669.0
421.90	599.0
423.00	4961.0
423.90	2582.0
425.00	411.0
441.00	12546.0
442.00	84920.0
443.00	17072.0
444.10	1842.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

m/z	Abundance
37.00	646.0
38.00	1607.0
39.10	8741.0
40.10	1146.0
40.90	265.0
44.00	2343.0
44.90	313.0
49.10	871.0
50.10	30008.0
51.10	109368.0
52.00	5620.0
53.00	272.0
55.00	612.0
56.00	3378.0
57.00	7641.0
57.90	241.0
59.00	170.0
61.00	1911.0
62.00	1716.0
63.00	3763.0
64.00	582.0
65.00	2099.0
66.10	153.0
68.10	2180.0
69.00	101584.0
70.00	519.0
73.00	1536.0
74.00	11086.0
75.00	16496.0
76.10	5359.0
77.10	108968.0
78.10	7398.0
79.00	7382.0
80.00	6000.0
81.00	8145.0
81.90	1857.0
83.00	1612.0
84.00	205.0
85.00	1225.0
86.00	2624.0
87.00	1013.0
88.00	447.0
88.90	259.0
91.00	1435.0
92.00	2703.0
93.00	14112.0
94.10	1098.0
94.90	168.0
96.00	752.0
97.00	278.0
98.00	10506.0
99.00	9028.0
100.00	608.0
101.00	4669.0
102.10	298.0
103.00	1509.0
104.00	3038.0
105.00	2812.0
106.10	817.0
107.00	33400.0
108.00	4877.0
109.10	918.0
110.00	57920.0
111.00	9840.0
112.00	1095.0
112.90	375.0
114.90	185.0
116.00	2059.0
117.00	32344.0
118.00	2463.0
119.00	365.0
120.10	527.0
122.00	2687.0
123.00	3852.0
124.00	1970.0
125.00	1268.0
126.20	647.0
127.00	118144.0
128.00	8996.0
129.00	46072.0
130.00	3772.0
130.90	954.0
131.90	440.0
133.00	370.0
133.90	1312.0
135.00	3526.0
136.00	1763.0
137.00	2237.0
137.90	465.0
138.90	260.0
140.10	395.0
141.00	6067.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

141.90	1818.0
143.00	1216.0
144.00	312.0
145.00	345.0
146.00	1209.0
147.00	3381.0
148.00	7230.0
149.00	1782.0
150.00	395.0
151.10	726.0
152.00	445.0
153.00	1614.0
154.00	1282.0
155.00	3425.0
156.00	4950.0
156.90	915.0
157.90	1228.0
158.90	826.0
160.00	1838.0
161.00	2524.0
162.00	755.0
163.00	328.0
164.00	307.0
165.00	2360.0
166.00	2081.0
167.00	12587.0
168.00	7983.0
169.00	1209.0
169.90	508.0
171.00	561.0
171.90	1067.0
173.00	1171.0
174.00	2343.0
175.00	3934.0
176.00	1126.0
177.00	2244.0
178.00	851.0
179.00	9034.0
180.00	6810.0
181.00	2929.0
182.00	492.0
183.00	160.0
183.90	774.0
185.00	3902.0
186.00	30592.0
187.00	8881.0
188.00	1279.0
189.00	1873.0
190.00	443.0
190.90	1108.0
191.90	2852.0
193.00	3106.0
194.00	877.0
194.90	238.0
196.00	5976.0
197.10	1589.0
198.00	210048.0
199.00	14763.0
200.00	1075.0
201.60	975.0
203.00	1921.0
204.00	8076.0
205.00	13549.0
206.00	57720.0
207.10	9533.0
208.00	2516.0
209.00	989.0
210.10	906.0
211.00	2311.0
211.70	447.0
213.00	165.0
214.90	602.0
216.00	1148.0
217.00	15698.0
218.00	2014.0
219.10	258.0
221.00	9389.0
221.70	2306.0
223.00	3197.0
224.00	31416.0
225.00	7748.0
226.00	999.0
227.00	12930.0
228.00	2020.0
229.00	2587.0
230.10	409.0
231.00	1134.0
232.00	214.0
233.00	189.0
233.90	744.0
235.00	839.0
235.90	766.0
237.00	1652.0
238.90	468.0
239.90	698.0
240.90	720.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

242.00	1741.0
243.00	1893.0
244.00	24880.0
245.00	3197.0
246.00	5245.0
247.00	1228.0
248.10	271.0
249.00	803.0
250.90	313.0
252.00	334.0
253.10	549.0
254.10	1400.0
255.00	129880.0
256.00	19680.0
257.00	1717.0
258.00	7239.0
259.00	1287.0
260.10	164.0
261.00	238.0
264.00	275.0
265.00	2735.0
265.90	464.0
271.00	431.0
272.00	333.0
273.00	3922.0
274.00	9766.0
275.00	58640.0
276.00	7693.0
277.00	5040.0
278.00	962.0
281.00	565.0
281.90	246.0
283.00	773.0
284.00	539.0
285.10	790.0
289.00	174.0
292.00	165.0
292.90	987.0
293.90	338.0
295.00	381.0
296.00	15259.0
297.00	2375.0
301.00	170.0
301.90	287.0
303.00	1758.0
304.10	623.0
305.00	212.0
308.00	208.0
308.90	174.0
313.00	167.0
313.90	819.0
315.00	1791.0
315.90	831.0
316.90	164.0
321.10	552.0
322.00	247.0
323.00	4603.0
324.00	979.0
327.00	932.0
327.90	308.0
331.90	319.0
333.00	487.0
334.00	3097.0
335.00	621.0
341.00	622.0
341.90	168.0
346.00	1222.0
346.90	280.0
352.00	1674.0
352.90	1099.0
354.00	1510.0
354.90	237.0
365.00	7690.0
366.00	1190.0
369.90	163.0
371.00	342.0
372.00	2609.0
373.10	478.0
383.00	730.0
384.00	242.0
390.10	312.0
391.00	243.0
402.00	1072.0
403.00	1479.0
404.00	763.0
404.90	224.0
421.00	1347.0
422.00	1653.0
423.00	9535.0
424.00	3674.0
425.00	800.0
438.90	156.0
440.10	254.0
441.00	26856.0
442.00	172928.0
443.00	33696.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

443.902955.0

444.90210.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

m/z	Abundance
37.00	434.0
38.00	1068.0
39.10	6582.0
39.90	774.0
41.00	323.0
44.00	2172.0
45.00	245.0
49.10	742.0
50.00	20904.0
51.10	80048.0
52.00	3955.0
53.00	188.0
54.90	445.0
56.00	2436.0
57.00	5312.0
58.00	266.0
61.00	1179.0
62.00	1142.0
63.00	2975.0
64.10	509.0
65.00	1444.0
68.10	1367.0
69.00	75616.0
70.00	412.0
73.10	1313.0
74.10	7814.0
75.00	12147.0
76.10	3701.0
77.10	80480.0
78.10	6055.0
79.00	5341.0
80.00	4407.0
81.00	5815.0
82.00	1495.0
83.00	1244.0
85.00	1022.0
86.00	1909.0
87.00	846.0
88.00	300.0
89.00	150.0
90.90	1285.0
92.10	1638.0
93.00	10792.0
93.90	636.0
95.10	171.0
96.00	633.0
96.90	198.0
98.00	8277.0
99.00	6523.0
99.90	639.0
101.00	3701.0
101.80	225.0
103.00	978.0
104.00	2358.0
105.00	2192.0
106.20	659.0
107.00	24912.0
108.00	3737.0
109.10	697.0
110.00	43184.0
111.00	6909.0
112.00	828.0
112.90	320.0
114.90	167.0
116.00	1539.0
117.00	25768.0
118.00	1392.0
119.00	317.0
120.00	477.0
122.00	1824.0
123.00	2942.0
123.90	1180.0
125.00	1085.0
126.20	770.0
127.00	90384.0
128.00	6679.0
129.00	35848.0
130.00	3268.0
131.00	514.0
131.90	204.0
133.00	302.0
134.00	1075.0
135.00	2984.0
136.00	1332.0
137.00	1696.0
138.00	363.0
138.90	166.0
140.00	644.0
141.00	4882.0
142.00	1116.0
142.90	1205.0
144.00	254.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

145.00	333.0
146.00	840.0
147.00	2709.0
148.00	6347.0
148.90	1459.0
149.90	437.0
151.10	668.0
152.00	389.0
152.90	1405.0
154.00	1180.0
155.00	2898.0
156.00	3750.0
157.00	874.0
158.00	1011.0
159.00	838.0
160.00	1383.0
161.00	2219.0
162.10	624.0
163.00	185.0
164.00	244.0
165.00	1930.0
166.00	1495.0
167.00	9838.0
168.00	6454.0
169.00	756.0
170.00	288.0
171.00	470.0
172.00	791.0
173.00	1114.0
174.00	1995.0
175.10	3443.0
176.00	860.0
177.00	2202.0
178.00	619.0
178.90	7052.0
180.00	5270.0
181.00	2268.0
182.00	432.0
183.00	317.0
184.00	485.0
185.00	3354.0
186.00	24648.0
187.00	7082.0
187.90	914.0
189.00	1659.0
189.90	343.0
190.90	1184.0
192.00	2122.0
193.00	2612.0
194.00	586.0
194.80	372.0
196.00	5271.0
197.10	1370.0
198.00	180928.0
199.00	13063.0
199.90	902.0
201.40	1287.0
202.10	205.0
203.00	1511.0
204.00	6839.0
205.00	11348.0
206.00	49264.0
207.00	7566.0
208.00	2178.0
208.90	700.0
210.10	1134.0
211.00	2135.0
211.80	437.0
214.90	409.0
216.00	1038.0
217.00	12586.0
217.90	1775.0
221.00	8030.0
221.80	1800.0
223.00	2498.0
224.00	27944.0
225.00	6836.0
226.00	899.0
227.00	12669.0
228.00	1554.0
229.00	2873.0
230.00	254.0
231.00	965.0
233.00	163.0
234.00	784.0
235.00	908.0
236.00	617.0
237.00	1270.0
238.00	178.0
239.00	470.0
239.80	285.0
241.00	625.0
242.00	1359.0
243.00	1510.0
244.00	23160.0
245.00	2912.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

246.00	4225.0
246.90	905.0
248.00	222.0
248.90	852.0
249.90	259.0
252.00	184.0
252.90	421.0
254.10	958.0
255.00	113024.0
256.00	16768.0
257.00	1329.0
258.00	6959.0
259.00	1058.0
260.90	188.0
263.90	204.0
265.00	2935.0
265.90	442.0
270.90	370.0
271.90	324.0
273.00	3227.0
274.00	9192.0
275.00	53600.0
276.00	6868.0
277.00	4681.0
277.90	737.0
281.10	399.0
281.90	295.0
283.00	658.0
284.10	297.0
285.00	924.0
289.00	202.0
289.90	252.0
291.90	206.0
293.00	1115.0
293.90	491.0
294.80	274.0
296.00	15019.0
297.00	1883.0
301.00	284.0
302.00	321.0
303.00	1918.0
303.90	360.0
305.00	158.0
307.90	235.0
309.10	217.0
309.90	267.0
313.00	175.0
314.00	776.0
315.00	1458.0
316.00	859.0
316.90	268.0
321.10	351.0
322.10	244.0
323.00	4512.0
324.00	994.0
326.90	1050.0
328.00	496.0
331.90	536.0
333.00	519.0
334.00	3326.0
335.00	789.0
341.00	516.0
346.00	969.0
347.00	178.0
352.00	1214.0
353.00	858.0
354.00	1749.0
355.10	403.0
365.00	7552.0
365.90	1038.0
371.00	404.0
372.00	2484.0
373.10	838.0
383.00	820.0
383.80	223.0
390.00	329.0
390.90	261.0
391.90	213.0
401.90	879.0
403.00	1559.0
404.00	743.0
404.90	183.0
421.00	1648.0
421.90	1342.0
423.00	9844.0
424.00	3208.0
424.80	485.0
439.00	199.0
440.20	413.0
441.00	29024.0
442.00	194496.0
443.00	36080.0
444.00	3857.0
445.00	340.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080