

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121522\
 Data File : BM038031.D
 Acq On : 15 Dec 2022 14:36
 Operator : CG/JU
 Sample : N5983-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COLB6

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/16/2022
 Supervised By :mohammad ahmed 12/17/2022

Quant Time: Dec 15 23:43:28 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 15 21:53:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	4231	0.400	ng/u1	0.00
4) Naphthalene-d8	10.878	136	13179	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.684	164	8137	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.421	188	18780	0.400	ng/u1	# 0.00
17) Chrysene-d12	21.582	240	13568	0.400	ng/u1	0.00
23) Perylene-d12	24.038	264	12227	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.410	96	19812	3.426	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.456	152	4527	0.235	ng/u1	0.00
18) Fluoranthene-d10	19.436	212	12999	0.264	ng/u1	0.00
Target Compounds						
					Qvalue	
7) 2-Methylnaphthalene	12.528	142	3670	0.150	ng/u1	100
8) 1-Methylnaphthalene	12.742	142	3845	0.153	ng/u1	99
14) Pentachlorophenol	17.071	266	261523	33.633	ng/u1	99
15) Phenanthrene	17.459	178	1331	0.021	ng/u1#	51
19) Fluoranthene	19.464	202	2201	0.031	ng/u1#	67
20) Pyrene	19.831	202	6684	0.093	ng/u1#	78
22) Chrysene	21.620	228	1331m	0.023	ng/u1	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121522\
Data File : BM038031.D
Acq On : 15 Dec 2022 14:36
Operator : CG/JU
Sample : N5983-05
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
COLB6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/16/2022
Supervised By :mohammad ahmed 12/17/2022


Quant Time: Dec 15 23:43:28 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 15 21:53:17 2022
Response via : Initial Calibration





#7
2-Methylnaphthalene
Concen: 0.150 ng/ul
RT: 12.528 min Scan# 2079
Delta R.T. -0.000 min
Lab File: BM038031.D
Acq: 15 Dec 2022 14:36




Instrument :
BNA_M
ClientSampleId :
COLB6




Tgt Ion:142 Resp: 3670
Ion Ratio Lower Upper
142 100
141 88.3 70.4 105.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/16/2022
Supervised By :mohammad ahmed 12/17/2022




#8
1-Methylnaphthalene
Concen: 0.153 ng/ul
RT: 12.742 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BM038031.D
Acq: 15 Dec 2022 14:36




Tgt Ion:142 Resp: 3845
Ion Ratio Lower Upper
142 100
141 92.6 73.6 110.4





#14
Pentachlorophenol
Concen: 33.633 ng/ul
RT: 17.071 min Scan# 3053
Delta R.T. 0.000 min
Lab File: BM038031.D
Acq: 15 Dec 2022 14:36




Instrument :
BNA_M
ClientSampleId :
COLB6




Tgt Ion:266 Resp: 261523
Ion Ratio Lower Upper
266 100
264 62.5 49.9 74.9
268 64.2 52.0 78.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/16/2022
Supervised By :mohammad ahmed 12/17/2022




#15
Phenanthrene
Concen: 0.021 ng/ul
RT: 17.459 min Scan# 3145
Delta R.T. -0.004 min
Lab File: BM038031.D
Acq: 15 Dec 2022 14:36




Tgt Ion:178 Resp: 1331
Ion Ratio Lower Upper
178 100
179 48.0 12.6 19.0#
176 32.5 15.8 23.6#





#19
Fluoranthene
Concen: 0.031 ng/u1
RT: 19.464 min Scan# 3599
Delta R.T. -0.005 min
Lab File: BM038031.D
Acq: 15 Dec 2022 14:36




Instrument :
BNA_M
ClientSampleId :
COLB6



Tgt Ion:202 Resp: 2201
Ion Ratio Lower Upper
202 100
101 29.5 10.6 15.8#
100 18.3 7.9 11.9#




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/16/2022
Supervised By :mohammad ahmed 12/17/2022



#20
Pyrene
Concen: 0.093 ng/u1
RT: 19.831 min Scan# 3678
Delta R.T. -0.000 min
Lab File: BM038031.D
Acq: 15 Dec 2022 14:36

Tgt Ion:202 Resp: 6684
Ion Ratio Lower Upper
202 100
101 28.6 12.1 18.1#
100 15.4 9.8 14.6#





#22

Chrysene

Concen: 0.023 ng/ul m

RT: 21.620 min Scan# 4205

Delta R.T. -0.003 min

Lab File: BM038031.D


Acq: 15 Dec 2022 14:36

Instrument :

BNA_M

ClientSampleId :

C0LB6



Tgt Ion:228 Resp: 133

Ion Ratio Lower Upper

228 100

226 37.7 24.1 36.1

229 31.6 15.7 23.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 12/16/2022

Supervised By :mohammad ahmed 12/17/2022

