

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121522\
 Data File : BM038034.D
 Acq On : 15 Dec 2022 16:27
 Operator : CG/JU
 Sample : N5983-18
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COLF4

Quant Time: Dec 15 23:44:10 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 15 21:53:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.064	152	5131	0.400	ng/u1	0.00
4) Naphthalene-d8	10.878	136	15894	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.684	164	9016	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.421	188	19042	0.400	ng/u1	0.00
17) Chrysene-d12	21.585	240	14679	0.400	ng/u1	0.00
23) Perylene-d12	24.041	264	14017	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.406	96	24547	3.500	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.456	152	6660	0.287	ng/u1	0.00
18) Fluoranthene-d10	19.441	212	16698	0.314	ng/u1	0.00
Target Compounds						
14) Pentachlorophenol	17.071	266	864	0.110	ng/u1	Qvalue 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121522\
Data File : BM038034.D
Acq On : 15 Dec 2022 16:27
Operator : CG/JU
Sample : N5983-18
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COLF4


Quant Time: Dec 15 23:44:10 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 15 21:53:17 2022
Response via : Initial Calibration





#14
 Pentachlorophenol
 Concen: 0.110 ng/ul
 RT: 17.071 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BM038034.D
 Acq: 15 Dec 2022 16:27

Instrument :
 BNA_M
 ClientSampleId :
 COLF4



Tgt Ion:	266	Resp:	864
Ion	Ratio	Lower	Upper
266	100		
264	60.8	49.9	74.9
268	58.3	52.0	78.0

