

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121522\
 Data File : BM038041.D
 Acq On : 15 Dec 2022 20:47
 Operator : CG/JU
 Sample : N5984-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0LC3

Quant Time: Dec 15 23:45:57 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 15 21:53:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.059	152	5080	0.400	ng/u1	0.00
4) Naphthalene-d8	10.873	136	15759	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.685	164	8555	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.422	188	17176	0.400	ng/u1	0.00
17) Chrysene-d12	21.583	240	11421	0.400	ng/u1	0.00
23) Perylene-d12	24.038	264	11312	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.410	96	31066	4.474	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.451	152	6580	0.286	ng/u1	0.00
18) Fluoranthene-d10	19.436	212	12110	0.293	ng/u1	0.00
Target Compounds						
14) Pentachlorophenol	17.071	266	312	0.044	ng/u1	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121522\
Data File : BM038041.D
Acq On : 15 Dec 2022 20:47
Operator : CG/JU
Sample : N5984-01
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COLC3

Quant Time: Dec 15 23:45:57 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 15 21:53:17 2022
Response via : Initial Calibration





#14
 Pentachlorophenol
 Concen: 0.044 ng/ul
 RT: 17.071 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BM038041.D
 Acq: 15 Dec 2022 20:47

Instrument :
 BNA_M
 ClientSampleId :
 C0LC3

Tgt Ion:	266	Resp:	312
Ion	Ratio	Lower	Upper
266	100		
264	61.9	49.9	74.9
268	66.0	52.0	78.0

