

Data Path : Z:\HPCHEM1\BNA M\Data\BM121615\
 Data File : BM003574.D
 Acq On : 16 Dec 2015 11:32
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.455

Quant Time: Dec 17 01:08:42 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 16 05:49:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2 SURR	1,4-Dioxane-d8	0.171	0.177	-3.5	86	0.00
3	1,4-Dioxane	0.218	0.214	1.8	82	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
5	Naphthalene	0.901	0.911	-1.1	85	0.00
6 SURR	2-Methylnaphthalene-d10	0.479	0.480	-0.2	84	0.00
7	2-Methylnaphthalene	0.605	0.626	-3.5	86	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
9	Acenaphthylene	1.528	1.445	5.4	82	0.00
10 C	Acenaphthene	1.232	1.280	-3.9	88	0.00
11	Fluorene	1.365	1.472	-7.8	90	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
13	Pentachlorophenol	0.057	0.034	40.4#	56	0.00
14	Phenanthrene	1.032	1.065	-3.2	91	0.00
15	Anthracene	0.893	0.882	1.2	88	0.00
16 SURR	Fluoranthene-d10	0.849	0.881	-3.8	91	0.00
17 C	Fluoranthene	1.120	1.179	-5.3	91	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
19	Pyrene	1.080	1.150	-6.5	93	0.00
20	Benzo(a)anthracene	0.977	0.970	0.7	89	0.00
21	Chrysene	1.119	1.166	-4.2	99	0.00
22 I	Perylene-d12	1.000	1.000	0.0	105	0.00
23	Benzo(b)fluoranthene	1.365	1.332	2.4	98	0.00
24	Benzo(k)fluoranthene	1.237	1.318	-6.5	112	-0.01
25 C	Benzo(a)pyrene	1.152	1.168	-1.4	105	0.00
26	Indeno(1,2,3-cd)pyrene	1.335	1.251	6.3	104	0.00
27	Dibenzo(a,h)anthracene	1.075	0.994	7.5	102	0.00
28	Benzo(q,h,i)perylene	1.162	1.098	5.5	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0