

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM121616\
 Data File : BM008352.D
 Acq On : 15 Dec 2016 20:37
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02056

Quant Time: Dec 16 02:38:13 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-2016\SOM02.2-EPA-BM121416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 16 02:33:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2	1,4-Dioxane	0.497	0.504	-1.4	119	0.00
3 S	1,4-Dioxane-d8	0.425	0.445	-4.7	121	0.00
4	Benzaldehyde	1.027	1.214	-18.2	129	0.00
5 S	Phenol-d5	1.551	1.613	-4.0	127	0.00
6	Phenol	1.611	1.684	-4.5	127	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.892	0.976	-9.4	131	0.00
8	Bis(2-Chloroethyl)ether	1.203	1.309	-8.8	130	0.00
9 S	2-Chlorophenol-d4	1.167	1.273	-9.1	129	0.00
10	2-Chlorophenol	1.179	1.293	-9.7	128	0.00
11	2-Methylphenol	1.179	1.253	-6.3	128	0.00
12	2,2'-oxybis(1-Chloropropane	1.367	1.545	-13.0	136	0.00
13 S	4-Methylphenol-d8	1.202	1.271	-5.7	127	0.00
14	Acetophenone	1.965	2.117	-7.7	131	0.00
15 P	N-Nitroso-di-n-propylamine	1.010	1.135	-12.4	130	0.00
16	4-Methylphenol	1.279	1.330	-4.0	125	0.00
17	Hexachloroethane	0.538	0.586	-8.9	127	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	129	0.00
19 S	Nitrobenzene-d5	0.145	0.154	-6.2	127	0.00
20	Nitrobenzene	0.373	0.413	-10.7	131	0.00
21	Isophorone	0.670	0.716	-6.9	128	0.00
22 S	2-Nitrophenol-d4	0.160	0.169	-5.6	124	0.00
23 C	2-Nitrophenol	0.164	0.179	-9.1	130	0.00
24	2,4-Dimethylphenol	0.366	0.393	-7.4	130	0.00
25	Bis(2-Chloroethoxy)methane	0.380	0.421	-10.8	133	0.00
26 S	2,4-Dichlorophenol-d3	0.294	0.319	-8.5	128	0.00
27 C	2,4-Dichlorophenol	0.296	0.313	-5.7	125	0.00
28	Naphthalene	0.947	1.015	-7.2	128	0.00
29 S	4-Chloroaniline-d4	0.336	0.365	-8.6	122	0.00
30	4-Chloroaniline	0.339	0.371	-9.4	126	0.00
31 C	Hexachlorobutadiene	0.240	0.264	-10.0	133	0.00
32	Caprolactam	0.093	0.082	11.8	105	0.01
33 C	4-Chloro-3-methylphenol	0.320	0.323	-0.9	118	0.00
34	2-Methylnaphthalene	0.684	0.719	-5.1	126	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
36	1,2,4,5-Tetrachlorobenzene	0.626	0.703	-12.3	127	0.00
37	Hexachlorocyclopentadiene	0.253	0.228	9.9	127	0.00
38 C	2,4,6-Trichlorophenol	0.355	0.386	-8.7	124	0.00
39	2,4,5-Trichlorophenol	0.388	0.391	-0.8	114	0.00
40	1,1'-Biphenyl	1.334	1.480	-10.9	125	0.00
41	2-Chloronaphthalene	1.023	1.121	-9.6	124	0.00
42	2-Nitroaniline	0.303	0.316	-4.3	117	0.00
43 S	Dimethylphthalate-d6	1.329	1.367	-2.9	115	0.00
44	Dimethylphthalate	1.303	1.347	-3.4	115	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM121616\
 Data File : BM008352.D
 Acq On : 15 Dec 2016 20:37
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02056

Quant Time: Dec 16 02:38:13 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-2016\SOM02.2-EPA-BM121416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 16 02:33:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.254	0.265	-4.3	114	0.00
46 S	Acenaphthylene-d8	1.571	1.698	-8.1	120	0.00
47	Acenaphthylene	1.599	1.727	-8.0	121	0.00
48	3-Nitroaniline	0.253	0.242	4.3	109	0.00
49 C	Acenaphthene	1.099	1.178	-7.2	122	0.00
50	2,4-Dinitrophenol	0.147	0.162	-10.2	160#	0.00
51 S	4-Nitrophenol-d4	0.196	0.140	28.6#	92	0.00
52	4-Nitrophenol	0.188	0.135	28.2#	91	0.00
53	Dibenzofuran	1.602	1.696	-5.9	119	0.00
54	2,4-Dinitrotoluene	0.380	0.376	1.1	108	0.00
55	2,3,4,6-Tetrachlorophenol	0.358	0.360	-0.6	112	0.00
56	Diethylphthalate	1.296	1.301	-0.4	113	0.00
57 S	Fluorene-d10	1.156	1.205	-4.2	118	0.00
58	Fluorene	1.315	1.362	-3.6	116	0.00
59	4-Chlorophenyl-phenylether	0.699	0.746	-6.7	121	0.00
60	4-Nitroaniline	0.243	0.211	13.2	103	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.110	7.6	102	0.00
63	4,6-Dinitro-2-methylphenol	0.122	0.109	10.7	96	0.00
64	N-Nitrosodiphenylamine	0.544	0.624	-14.7	113	0.00
65	4-Bromophenyl-phenylether	0.220	0.249	-13.2	111	0.00
66	Hexachlorobenzene	0.249	0.276	-10.8	110	0.00
67	Atrazine	0.223	0.228	-2.2	105	0.00
68 C	Pentachlorophenol	0.135	0.113	16.3	92	0.00
69	Phenanthrene	1.025	1.125	-9.8	109	0.00
70 S	Anthracene-d10	0.871	0.955	-9.6	108	0.00
71	Anthracene	1.040	1.140	-9.6	108	0.00
72	Carbazole	0.916	0.924	-0.9	102	0.00
73	Di-n-butylphthalate	1.025	1.084	-5.8	104	0.00
74 C	Fluoranthene	1.259	1.258	0.1	100	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
76 S	Pyrene-d10	0.772	0.867	-12.3	100	0.00
77	Pyrene	0.982	1.095	-11.5	100	0.00
78	Butylbenzylphthalate	0.352	0.377	-7.1	95	0.00
79	3,3'-Dichlorobenzidine	0.351	0.360	-2.6	90	0.00
80	Benzo(a)anthracene	1.034	1.106	-7.0	95	0.00
81	Bis(2-ethylhexyl)phthalate	0.516	0.558	-8.1	95	0.00
82	Chrysene	0.994	1.078	-8.5	96	0.00
83 I	Perylene-d12	1.000	1.000	0.0	97	0.00
84	Di-n-octyl phthalate	0.960	0.989	-3.0	95	0.00
85	Benzo(b)fluoranthene	1.080	1.202	-11.3	99	0.00
86	Benzo(k)fluoranthene	1.033	1.116	-8.0	92	0.00
87 S	Benzo(a)pyrene-d12	0.840	0.925	-10.1	96	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM121616\
Data File : BM008352.D
Acq On : 15 Dec 2016 20:37
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02056

Quant Time: Dec 16 02:38:13 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-2016\SOM02.2-EPA-BM121416.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 16 02:33:30 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.040	1.146	-10.2	97	0.00
89	Indeno(1,2,3-cd)pyrene	1.263	1.413	-11.9	98	0.00
90	Dibenzo(a,h)anthracene	1.061	1.176	-10.8	97	0.00
91	Benzo(g,h,i)perylene	1.056	1.166	-10.4	96	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0