


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121619\  
 Data File : BM024108.D  
 Acq On : 16 Dec 2019 14:56  
 Operator : JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\8270-BM121619.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Dec 16 19:58:15 2019



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2450

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 36.7      | 87706   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 2070    | PASS             |
| 69          | 198          | 0.00         | 100          | 47.3      | 112872  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 659     | PASS             |
| 127         | 198          | 10           | 80           | 50.8      | 121349  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 943     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 238656  | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 16368   | PASS             |
| 275         | 198          | 10           | 60           | 31.7      | 75661   | PASS             |
| 365         | 198          | 1            | 100          | 4.1       | 9900    | PASS             |
| 441         | 198          | 0.01         | 100          | 9.2       | 21890   | PASS             |
| 442         | 198          | 50           | 100          | 58.9      | 140584  | PASS             |
| 443         | 442          | 15           | 24           | 18.7      | 26263   | PASS             |